

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100421\
 Data File : BG050432.D
 Acq On : 5 Oct 2021 3:58
 Operator : CG/JU
 Sample : M3879-06
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 EW407

Manual Integrations
 APPROVED

mohammad
 10/5/2021 12:07:06 PM

Quant Time: Oct 05 04:59:02 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG093021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 30 16:27:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.273	152	59424	20.00	ng/ul	0.00
20) Naphthalene-d8	11.093	136	250203	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.883	164	160520	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.627	188	329525	20.00	ng/ul	0.00
79) Chrysene-d12	21.928	240	294001	20.00	ng/ul	-0.02
88) Perylene-d12	25.353	264	288093	20.00	ng/ul	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.637	96	3794	2.35	ng/uL	0.00
4) Pyridine-d5	4.066	84	30988	6.43	ng/ul	0.00
7) Phenol-d5	7.404	99	60863	11.72	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.580	67	50542	15.47	ng/ul	0.00
11) 2-Chlorophenol-d4	7.791	132	48462	13.32	ng/ul	0.00
15) 4-Methylphenol-d8	8.955	113	27940	7.11	ng/ul	0.00
21) Nitrobenzene-d5	9.437	128	28777	16.44	ng/ul	0.00
24) 2-Nitrophenol-d4	10.165	143	30283	16.12	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.700	165	48240	13.00	ng/ul	0.00
31) 4-Chloroaniline-d4	11.223	131	51351	9.69	ng/ul	0.00
46) Dimethylphthalate-d6	14.278	166	162822	14.82	ng/ul	-0.01
49) Acenaphthylene-d8	14.583	160	206324	14.96	ng/ul	0.00
54) 4-Nitrophenol-d4	15.048	143	23382m	11.05	ng/ul	0.00
60) Fluorene-d10	15.870	176	145799	14.73	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.976	200	15727	9.57	ng/ul	0.00
73) Anthracene-d10	17.727	188	210649	15.13	ng/ul	0.00
81) Pyrene-d10	20.001	212	245969	15.73	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.112	264	211391	14.59	ng/ul	-0.03

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100421\
Data File : BG050432.D
Acq On : 5 Oct 2021 3:58
Operator : CG/JU
Sample : M3879-06
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
EW407

Manual Integrations
APPROVED

mohammad
10/5/2021 12:07:06 PM

Quant Time: Oct 05 04:59:02 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG093021.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 30 16:27:24 2021
Response via : Initial Calibration

