


Data Path : Z:\svoasrv\HPCHEM1\BNA\_G\Data\BG100421\  
 Data File : BG050406.D  
 Acq On : 4 Oct 2021 9:04  
 Operator : CG/JU  
 Sample : DFTPP424  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP424

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA\_G\Methods\SFAM-EPA-BG093021.M  
 Title : SVOA CALIBRATION  
 Last Update : Thu Sep 30 16:27:24 2021



AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2470

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 46.5         | 41112      | PASS                |
| 68             | 69              | 0.00            | 2               | 0.8          | 357        | PASS                |
| 69             | 198             | 0.00            | 100             | 47.7         | 42155      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.8          | 353        | PASS                |
| 127            | 198             | 10              | 80              | 51.3         | 45392      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.0          | 0          | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 88421      | PASS                |
| 199            | 198             | 5               | 9               | 6.5          | 5774       | PASS                |
| 275            | 198             | 10              | 60              | 26.4         | 23323      | PASS                |
| 365            | 198             | 1               | 100             | 3.6          | 3204       | PASS                |
| 441            | 198             | 0.01            | 100             | 8.2          | 7288       | PASS                |
| 442            | 198             | 50              | 100             | 54.5         | 48211      | PASS                |
| 443            | 442             | 15              | 24              | 19.1         | 9193       | PASS                |

| m/z    | Abundance |
|--------|-----------|
| 39.90  | 539.0     |
| 43.20  | 178.0     |
| 44.00  | 914.0     |
| 55.00  | 224.0     |
| 57.00  | 152.0     |
| 73.00  | 190.0     |
| 133.00 | 242.0     |
| 190.90 | 220.0     |
| 192.90 | 228.0     |
| 207.00 | 502.0     |
| 281.00 | 205.0     |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP424

| m/z    | Abundance |
|--------|-----------|
| 36.90  | 241.0     |
| 38.10  | 589.0     |
| 39.00  | 3239.0    |
| 40.00  | 635.0     |
| 40.90  | 328.0     |
| 44.00  | 619.0     |
| 49.00  | 232.0     |
| 50.00  | 12452.0   |
| 51.00  | 47352.0   |
| 52.00  | 2411.0    |
| 55.00  | 252.0     |
| 56.00  | 1420.0    |
| 57.00  | 3564.0    |
| 58.10  | 332.0     |
| 61.00  | 858.0     |
| 61.90  | 497.0     |
| 63.00  | 1809.0    |
| 64.00  | 237.0     |
| 65.00  | 1238.0    |
| 69.00  | 46360.0   |
| 70.00  | 426.0     |
| 73.10  | 547.0     |
| 74.00  | 4960.0    |
| 75.00  | 8088.0    |
| 76.00  | 2771.0    |
| 77.00  | 53248.0   |
| 78.00  | 3670.0    |
| 79.00  | 3223.0    |
| 80.00  | 2425.0    |
| 81.00  | 4229.0    |
| 81.90  | 878.0     |
| 82.90  | 1130.0    |
| 85.00  | 765.0     |
| 86.00  | 1221.0    |
| 87.00  | 571.0     |
| 87.90  | 222.0     |
| 91.00  | 781.0     |
| 92.00  | 887.0     |
| 93.00  | 6527.0    |
| 94.00  | 467.0     |
| 95.00  | 175.0     |
| 95.90  | 423.0     |
| 98.00  | 4638.0    |
| 99.00  | 3136.0    |
| 99.90  | 441.0     |
| 101.00 | 2232.0    |
| 102.90 | 653.0     |
| 104.00 | 1321.0    |
| 105.00 | 1525.0    |
| 106.00 | 475.0     |
| 107.00 | 14887.0   |
| 108.00 | 2465.0    |
| 109.00 | 625.0     |
| 110.00 | 26048.0   |
| 111.00 | 4290.0    |
| 112.00 | 498.0     |
| 112.90 | 156.0     |
| 114.80 | 159.0     |
| 116.10 | 689.0     |
| 117.00 | 14298.0   |
| 118.00 | 977.0     |
| 120.00 | 233.0     |
| 121.90 | 1307.0    |
| 122.90 | 1772.0    |
| 124.00 | 738.0     |
| 125.00 | 749.0     |
| 127.00 | 47008.0   |
| 128.00 | 3869.0    |
| 129.00 | 20704.0   |
| 129.90 | 1597.0    |
| 130.90 | 222.0     |
| 134.00 | 631.0     |
| 134.90 | 1772.0    |
| 136.00 | 634.0     |
| 137.00 | 1082.0    |
| 139.00 | 203.0     |
| 140.00 | 368.0     |
| 140.90 | 2956.0    |
| 141.90 | 770.0     |
| 143.00 | 535.0     |
| 144.00 | 163.0     |
| 146.00 | 660.0     |
| 147.00 | 1490.0    |
| 148.00 | 3277.0    |
| 149.00 | 831.0     |
| 150.90 | 381.0     |
| 151.50 | 161.0     |
| 152.10 | 169.0     |
| 152.90 | 809.0     |
| 154.00 | 534.0     |
| 155.00 | 1207.0    |
| 156.00 | 1825.0    |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP424

|        |         |
|--------|---------|
| 157.10 | 398.0   |
| 157.90 | 644.0   |
| 159.00 | 256.0   |
| 160.00 | 848.0   |
| 160.90 | 1238.0  |
| 162.00 | 326.0   |
| 165.00 | 1023.0  |
| 166.00 | 990.0   |
| 167.00 | 4965.0  |
| 168.00 | 2473.0  |
| 169.00 | 315.0   |
| 170.00 | 239.0   |
| 170.90 | 304.0   |
| 171.90 | 438.0   |
| 173.00 | 592.0   |
| 174.00 | 938.0   |
| 175.00 | 1779.0  |
| 176.00 | 454.0   |
| 176.90 | 1046.0  |
| 177.90 | 443.0   |
| 178.90 | 4042.0  |
| 180.00 | 2586.0  |
| 180.90 | 1181.0  |
| 182.10 | 222.0   |
| 183.90 | 527.0   |
| 185.00 | 1995.0  |
| 186.00 | 14740.0 |
| 187.00 | 4300.0  |
| 188.00 | 247.0   |
| 189.00 | 998.0   |
| 189.90 | 208.0   |
| 190.90 | 488.0   |
| 192.00 | 1293.0  |
| 193.00 | 1203.0  |
| 194.00 | 448.0   |
| 196.00 | 2256.0  |
| 197.90 | 85696.0 |
| 198.90 | 5844.0  |
| 200.00 | 579.0   |
| 201.50 | 412.0   |
| 203.00 | 971.0   |
| 204.00 | 4157.0  |
| 205.00 | 6010.0  |
| 206.00 | 25720.0 |
| 207.00 | 4159.0  |
| 207.90 | 1320.0  |
| 209.00 | 347.0   |
| 210.00 | 378.0   |
| 210.30 | 341.0   |
| 211.00 | 835.0   |
| 214.90 | 443.0   |
| 215.90 | 409.0   |
| 216.90 | 8005.0  |
| 217.90 | 920.0   |
| 219.80 | 241.0   |
| 221.00 | 3151.0  |
| 221.80 | 847.0   |
| 222.90 | 1557.0  |
| 224.00 | 14835.0 |
| 225.00 | 3952.0  |
| 226.00 | 268.0   |
| 227.00 | 6668.0  |
| 228.00 | 961.0   |
| 229.00 | 1303.0  |
| 230.90 | 518.0   |
| 234.00 | 393.0   |
| 235.00 | 651.0   |
| 235.90 | 305.0   |
| 236.90 | 452.0   |
| 237.90 | 187.0   |
| 239.10 | 290.0   |
| 240.00 | 303.0   |
| 241.00 | 442.0   |
| 242.00 | 944.0   |
| 243.10 | 1084.0  |
| 244.00 | 10238.0 |
| 245.00 | 1444.0  |
| 246.00 | 2357.0  |
| 246.90 | 470.0   |
| 248.10 | 160.0   |
| 248.90 | 372.0   |
| 251.90 | 215.0   |
| 253.10 | 180.0   |
| 255.00 | 52616.0 |
| 256.00 | 8877.0  |
| 257.00 | 519.0   |
| 258.00 | 3353.0  |
| 258.90 | 650.0   |
| 264.90 | 1593.0  |
| 266.00 | 285.0   |
| 268.20 | 188.0   |
| 271.90 | 151.0   |
| 273.00 | 1444.0  |
| 274.00 | 3334.0  |
| 275.00 | 20688.0 |
| 276.00 | 3220.0  |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP424

|        |         |
|--------|---------|
| 276.90 | 2309.0  |
| 277.80 | 385.0   |
| 278.90 | 154.0   |
| 281.00 | 199.0   |
| 282.10 | 157.0   |
| 283.00 | 260.0   |
| 284.90 | 534.0   |
| 293.00 | 425.0   |
| 296.00 | 6890.0  |
| 297.00 | 947.0   |
| 303.00 | 692.0   |
| 304.10 | 241.0   |
| 313.00 | 165.0   |
| 314.00 | 240.0   |
| 315.00 | 674.0   |
| 316.00 | 462.0   |
| 321.10 | 248.0   |
| 321.80 | 177.0   |
| 323.00 | 1937.0  |
| 324.00 | 413.0   |
| 326.90 | 333.0   |
| 327.90 | 198.0   |
| 332.00 | 198.0   |
| 333.00 | 214.0   |
| 334.00 | 1245.0  |
| 335.00 | 427.0   |
| 341.00 | 222.0   |
| 346.00 | 469.0   |
| 352.10 | 463.0   |
| 353.00 | 486.0   |
| 353.90 | 386.0   |
| 365.00 | 2984.0  |
| 366.00 | 452.0   |
| 371.90 | 792.0   |
| 373.00 | 230.0   |
| 382.80 | 153.0   |
| 401.80 | 183.0   |
| 402.20 | 195.0   |
| 403.00 | 394.0   |
| 421.00 | 385.0   |
| 422.00 | 455.0   |
| 423.00 | 1878.0  |
| 424.00 | 354.0   |
| 441.00 | 6029.0  |
| 442.10 | 39096.0 |
| 443.00 | 7180.0  |
| 444.10 | 942.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP424

| m/z    | Abundance |
|--------|-----------|
| 37.20  | 196.0     |
| 38.00  | 724.0     |
| 39.00  | 3920.0    |
| 40.00  | 669.0     |
| 41.10  | 261.0     |
| 43.00  | 208.0     |
| 44.00  | 569.0     |
| 49.00  | 312.0     |
| 50.00  | 12612.0   |
| 51.00  | 46440.0   |
| 52.00  | 2665.0    |
| 55.00  | 287.0     |
| 56.00  | 1348.0    |
| 57.00  | 3788.0    |
| 57.90  | 260.0     |
| 61.00  | 586.0     |
| 61.90  | 889.0     |
| 63.00  | 1847.0    |
| 63.90  | 386.0     |
| 65.00  | 1150.0    |
| 68.00  | 1071.0    |
| 69.00  | 49712.0   |
| 70.10  | 302.0     |
| 73.00  | 579.0     |
| 74.00  | 5028.0    |
| 75.00  | 8360.0    |
| 76.00  | 2950.0    |
| 77.00  | 56904.0   |
| 78.00  | 3887.0    |
| 79.00  | 3581.0    |
| 80.00  | 2746.0    |
| 81.00  | 4281.0    |
| 82.00  | 1275.0    |
| 83.00  | 839.0     |
| 84.90  | 594.0     |
| 86.00  | 1194.0    |
| 87.00  | 429.0     |
| 91.00  | 996.0     |
| 92.00  | 1009.0    |
| 93.00  | 7514.0    |
| 93.90  | 495.0     |
| 95.10  | 165.0     |
| 96.10  | 461.0     |
| 98.00  | 5133.0    |
| 99.00  | 3741.0    |
| 99.90  | 433.0     |
| 101.00 | 2223.0    |
| 101.80 | 165.0     |
| 102.90 | 720.0     |
| 104.00 | 1544.0    |
| 105.00 | 1136.0    |
| 105.90 | 553.0     |
| 107.00 | 16224.0   |
| 107.90 | 2470.0    |
| 110.00 | 27080.0   |
| 111.00 | 4141.0    |
| 112.00 | 519.0     |
| 116.00 | 928.0     |
| 117.00 | 16116.0   |
| 118.00 | 1120.0    |
| 118.80 | 267.0     |
| 120.10 | 245.0     |
| 122.00 | 1029.0    |
| 122.90 | 1716.0    |
| 124.00 | 827.0     |
| 124.90 | 738.0     |
| 127.00 | 54096.0   |
| 128.00 | 3856.0    |
| 129.00 | 21832.0   |
| 130.00 | 1897.0    |
| 131.00 | 456.0     |
| 132.00 | 174.0     |
| 134.00 | 891.0     |
| 135.00 | 1511.0    |
| 136.00 | 713.0     |
| 136.90 | 879.0     |
| 137.90 | 174.0     |
| 140.00 | 291.0     |
| 140.90 | 3324.0    |
| 142.00 | 846.0     |
| 143.00 | 615.0     |
| 144.90 | 209.0     |
| 146.00 | 848.0     |
| 146.90 | 1476.0    |
| 148.00 | 4039.0    |
| 149.00 | 661.0     |
| 150.90 | 469.0     |
| 151.90 | 333.0     |
| 152.90 | 955.0     |
| 154.00 | 817.0     |
| 155.00 | 1738.0    |
| 156.00 | 2736.0    |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP424

|        |          |
|--------|----------|
| 157.00 | 429.0    |
| 158.00 | 408.0    |
| 158.80 | 563.0    |
| 159.90 | 1116.0   |
| 161.00 | 1836.0   |
| 161.90 | 171.0    |
| 162.90 | 178.0    |
| 164.90 | 1273.0   |
| 166.00 | 912.0    |
| 167.00 | 6191.0   |
| 168.00 | 3056.0   |
| 168.90 | 400.0    |
| 169.90 | 202.0    |
| 170.90 | 228.0    |
| 172.00 | 634.0    |
| 172.90 | 903.0    |
| 173.90 | 1484.0   |
| 175.00 | 2372.0   |
| 175.90 | 645.0    |
| 176.90 | 1170.0   |
| 178.00 | 322.0    |
| 178.90 | 5127.0   |
| 180.00 | 3186.0   |
| 181.00 | 1193.0   |
| 182.10 | 361.0    |
| 182.80 | 165.0    |
| 184.00 | 571.0    |
| 185.00 | 3046.0   |
| 186.00 | 14857.0  |
| 187.00 | 5137.0   |
| 187.90 | 550.0    |
| 189.00 | 1192.0   |
| 189.90 | 282.0    |
| 191.00 | 669.0    |
| 192.00 | 1519.0   |
| 193.00 | 1980.0   |
| 194.00 | 418.0    |
| 194.90 | 288.0    |
| 196.00 | 3015.0   |
| 197.90 | 106064.0 |
| 198.90 | 6456.0   |
| 199.90 | 618.0    |
| 201.20 | 273.0    |
| 203.00 | 1095.0   |
| 204.00 | 4257.0   |
| 205.00 | 8103.0   |
| 206.00 | 31144.0  |
| 207.00 | 5458.0   |
| 207.90 | 1619.0   |
| 208.90 | 255.0    |
| 210.00 | 484.0    |
| 211.00 | 1023.0   |
| 211.90 | 427.0    |
| 215.00 | 335.0    |
| 216.10 | 709.0    |
| 217.00 | 9166.0   |
| 217.90 | 1442.0   |
| 219.90 | 226.0    |
| 221.00 | 4464.0   |
| 221.80 | 1384.0   |
| 223.00 | 2039.0   |
| 224.00 | 17160.0  |
| 225.00 | 4279.0   |
| 226.00 | 326.0    |
| 227.00 | 8625.0   |
| 227.90 | 1153.0   |
| 229.00 | 1372.0   |
| 229.90 | 246.0    |
| 231.00 | 550.0    |
| 233.00 | 178.0    |
| 234.00 | 544.0    |
| 235.00 | 717.0    |
| 235.90 | 281.0    |
| 237.00 | 742.0    |
| 239.00 | 313.0    |
| 239.90 | 229.0    |
| 241.00 | 458.0    |
| 242.00 | 1083.0   |
| 243.00 | 1019.0   |
| 244.00 | 12428.0  |
| 245.00 | 1706.0   |
| 246.00 | 3295.0   |
| 246.90 | 633.0    |
| 248.80 | 334.0    |
| 250.00 | 162.0    |
| 250.90 | 213.0    |
| 252.10 | 179.0    |
| 252.90 | 353.0    |
| 255.00 | 67672.0  |
| 256.00 | 10102.0  |
| 257.10 | 797.0    |
| 258.00 | 4699.0   |
| 259.00 | 635.0    |
| 260.90 | 180.0    |
| 265.00 | 1660.0   |
| 266.10 | 272.0    |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP424

|        |         |
|--------|---------|
| 270.10 | 156.0   |
| 271.80 | 237.0   |
| 273.00 | 1811.0  |
| 274.00 | 5180.0  |
| 275.00 | 27808.0 |
| 276.00 | 3469.0  |
| 277.00 | 2557.0  |
| 278.00 | 358.0   |
| 280.80 | 180.0   |
| 282.00 | 159.0   |
| 283.00 | 368.0   |
| 283.90 | 279.0   |
| 285.10 | 538.0   |
| 293.00 | 682.0   |
| 293.90 | 235.0   |
| 296.00 | 8744.0  |
| 296.90 | 1127.0  |
| 297.80 | 204.0   |
| 301.90 | 265.0   |
| 303.00 | 1111.0  |
| 304.10 | 252.0   |
| 314.00 | 356.0   |
| 315.00 | 952.0   |
| 315.90 | 393.0   |
| 321.00 | 186.0   |
| 322.10 | 208.0   |
| 323.00 | 1954.0  |
| 324.00 | 535.0   |
| 327.00 | 573.0   |
| 328.00 | 288.0   |
| 332.00 | 167.0   |
| 333.00 | 276.0   |
| 334.00 | 1677.0  |
| 335.00 | 364.0   |
| 340.90 | 339.0   |
| 346.00 | 653.0   |
| 352.10 | 840.0   |
| 353.00 | 531.0   |
| 354.00 | 758.0   |
| 355.00 | 327.0   |
| 365.00 | 3596.0  |
| 366.00 | 472.0   |
| 370.90 | 156.0   |
| 372.00 | 1138.0  |
| 373.00 | 297.0   |
| 382.70 | 246.0   |
| 383.00 | 250.0   |
| 383.80 | 158.0   |
| 402.00 | 559.0   |
| 403.00 | 722.0   |
| 404.10 | 220.0   |
| 421.00 | 566.0   |
| 422.10 | 293.0   |
| 423.10 | 3133.0  |
| 424.00 | 793.0   |
| 441.10 | 8021.0  |
| 442.10 | 55752.0 |
| 443.10 | 10822.0 |
| 444.00 | 1434.0  |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP424

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 233.0     |
| 38.00  | 354.0     |
| 39.00  | 2201.0    |
| 40.00  | 695.0     |
| 44.00  | 522.0     |
| 49.00  | 262.0     |
| 50.00  | 7238.0    |
| 51.00  | 29544.0   |
| 52.00  | 1576.0    |
| 54.90  | 327.0     |
| 56.00  | 879.0     |
| 57.00  | 1921.0    |
| 61.00  | 230.0     |
| 62.00  | 417.0     |
| 63.00  | 1134.0    |
| 64.00  | 195.0     |
| 65.00  | 655.0     |
| 69.00  | 30392.0   |
| 70.00  | 332.0     |
| 73.00  | 268.0     |
| 74.00  | 3084.0    |
| 75.00  | 5194.0    |
| 76.00  | 1562.0    |
| 77.00  | 35256.0   |
| 78.00  | 2572.0    |
| 78.90  | 2510.0    |
| 80.00  | 1694.0    |
| 81.00  | 2867.0    |
| 82.00  | 613.0     |
| 83.00  | 620.0     |
| 85.00  | 528.0     |
| 86.00  | 822.0     |
| 86.90  | 386.0     |
| 91.00  | 664.0     |
| 92.00  | 917.0     |
| 93.00  | 4254.0    |
| 93.90  | 342.0     |
| 95.00  | 237.0     |
| 97.10  | 245.0     |
| 97.90  | 2906.0    |
| 98.90  | 2123.0    |
| 101.00 | 1759.0    |
| 103.00 | 567.0     |
| 103.90 | 1118.0    |
| 104.90 | 716.0     |
| 105.90 | 371.0     |
| 107.00 | 10513.0   |
| 107.90 | 1968.0    |
| 108.80 | 347.0     |
| 110.00 | 17040.0   |
| 110.90 | 2788.0    |
| 111.90 | 395.0     |
| 116.00 | 579.0     |
| 117.00 | 10351.0   |
| 118.00 | 645.0     |
| 119.90 | 237.0     |
| 121.90 | 808.0     |
| 123.00 | 1127.0    |
| 123.90 | 641.0     |
| 125.00 | 553.0     |
| 127.00 | 35072.0   |
| 128.00 | 2888.0    |
| 129.00 | 14556.0   |
| 129.90 | 1304.0    |
| 130.90 | 327.0     |
| 131.80 | 239.0     |
| 132.90 | 193.0     |
| 133.90 | 544.0     |
| 134.90 | 1274.0    |
| 136.00 | 425.0     |
| 137.00 | 630.0     |
| 139.90 | 225.0     |
| 141.00 | 2156.0    |
| 142.00 | 629.0     |
| 143.00 | 472.0     |
| 145.90 | 347.0     |
| 146.90 | 1193.0    |
| 148.00 | 2291.0    |
| 149.00 | 260.0     |
| 150.00 | 220.0     |
| 151.10 | 329.0     |
| 152.90 | 675.0     |
| 153.90 | 394.0     |
| 155.00 | 1293.0    |
| 156.00 | 1434.0    |
| 157.00 | 305.0     |
| 157.90 | 552.0     |
| 158.90 | 335.0     |
| 159.90 | 560.0     |
| 160.90 | 929.0     |
| 162.10 | 273.0     |
| 164.90 | 856.0     |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP424

|        |         |
|--------|---------|
| 166.00 | 855.0   |
| 166.90 | 3601.0  |
| 168.00 | 2082.0  |
| 168.90 | 483.0   |
| 169.80 | 161.0   |
| 171.20 | 163.0   |
| 171.90 | 393.0   |
| 173.00 | 458.0   |
| 173.90 | 796.0   |
| 175.00 | 1650.0  |
| 176.10 | 398.0   |
| 176.90 | 743.0   |
| 178.90 | 3204.0  |
| 180.00 | 2166.0  |
| 181.00 | 962.0   |
| 182.00 | 164.0   |
| 183.90 | 311.0   |
| 185.00 | 1921.0  |
| 186.00 | 11132.0 |
| 187.00 | 3188.0  |
| 188.00 | 441.0   |
| 189.00 | 770.0   |
| 189.90 | 175.0   |
| 191.00 | 609.0   |
| 191.90 | 1274.0  |
| 193.00 | 1290.0  |
| 194.00 | 343.0   |
| 196.00 | 2485.0  |
| 197.90 | 73504.0 |
| 198.90 | 5022.0  |
| 199.80 | 433.0   |
| 201.40 | 304.0   |
| 203.10 | 881.0   |
| 204.00 | 3239.0  |
| 205.00 | 5906.0  |
| 206.00 | 22168.0 |
| 207.00 | 3865.0  |
| 208.00 | 1282.0  |
| 209.00 | 378.0   |
| 210.00 | 718.0   |
| 211.00 | 1167.0  |
| 214.90 | 412.0   |
| 216.10 | 621.0   |
| 216.90 | 6097.0  |
| 217.90 | 819.0   |
| 220.90 | 2576.0  |
| 222.00 | 1005.0  |
| 223.00 | 1490.0  |
| 224.00 | 11662.0 |
| 225.00 | 3257.0  |
| 226.10 | 392.0   |
| 227.00 | 5546.0  |
| 228.00 | 680.0   |
| 229.00 | 1247.0  |
| 229.80 | 252.0   |
| 231.00 | 353.0   |
| 234.00 | 450.0   |
| 235.00 | 340.0   |
| 236.10 | 322.0   |
| 237.00 | 341.0   |
| 239.00 | 295.0   |
| 240.00 | 191.0   |
| 241.00 | 375.0   |
| 242.00 | 594.0   |
| 243.00 | 952.0   |
| 244.00 | 9361.0  |
| 245.00 | 1005.0  |
| 246.00 | 2068.0  |
| 246.80 | 428.0   |
| 248.00 | 172.0   |
| 248.90 | 365.0   |
| 252.70 | 234.0   |
| 255.00 | 53576.0 |
| 256.00 | 6942.0  |
| 256.90 | 503.0   |
| 258.00 | 3028.0  |
| 259.00 | 647.0   |
| 265.00 | 1392.0  |
| 266.10 | 289.0   |
| 271.90 | 278.0   |
| 273.00 | 1628.0  |
| 274.00 | 3821.0  |
| 275.00 | 21472.0 |
| 276.00 | 2621.0  |
| 276.90 | 2061.0  |
| 277.90 | 423.0   |
| 283.00 | 303.0   |
| 285.00 | 346.0   |
| 293.00 | 345.0   |
| 294.00 | 205.0   |
| 296.00 | 7015.0  |
| 297.00 | 1013.0  |
| 300.90 | 159.0   |
| 303.00 | 742.0   |
| 304.10 | 276.0   |
| 307.90 | 163.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP424

|        |         |
|--------|---------|
| 314.00 | 321.0   |
| 315.00 | 898.0   |
| 316.00 | 459.0   |
| 320.90 | 214.0   |
| 322.00 | 190.0   |
| 323.00 | 2160.0  |
| 324.00 | 519.0   |
| 326.80 | 470.0   |
| 327.90 | 330.0   |
| 332.10 | 168.0   |
| 333.20 | 162.0   |
| 334.00 | 1178.0  |
| 335.10 | 293.0   |
| 341.10 | 176.0   |
| 345.90 | 429.0   |
| 352.00 | 472.0   |
| 352.90 | 395.0   |
| 354.00 | 632.0   |
| 355.00 | 188.0   |
| 365.00 | 3032.0  |
| 365.90 | 371.0   |
| 370.00 | 152.0   |
| 371.00 | 192.0   |
| 372.00 | 917.0   |
| 372.90 | 291.0   |
| 383.00 | 180.0   |
| 401.10 | 164.0   |
| 402.00 | 518.0   |
| 403.00 | 561.0   |
| 404.00 | 185.0   |
| 421.10 | 425.0   |
| 422.10 | 365.0   |
| 423.00 | 2492.0  |
| 424.10 | 447.0   |
| 441.00 | 7813.0  |
| 442.00 | 49784.0 |
| 443.00 | 9577.0  |
| 444.10 | 729.0   |

Instrument :  
BNA\_G  
ClientSampleId :  
DFTPP424