

Data Path : Z:\HPCHEM1\BNA G\DATA\BG100615\
 Data File : BG019057.D
 Acq On : 6 Oct 2015 22:29
 Operator : UM/NP
 Sample : G3915-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C03G3


Quant Time: Oct 07 01:09:00 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG100415.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 07 01:07:34 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.04	152	18353	20.00	ng/ul	0.00
18) Naphthalene-d8	10.85	136	80553	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.67	164	56576	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.42	188	151398	20.00	ng/ul	0.00
78) Chrysene-d12	21.69	240	191852	20.00	ng/ul	0.00
86) Perylene-d12	24.91	264	188833	20.00	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL	
5) Phenol-d5	7.20	99	3740	2.36	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.37	67	6780	6.62	ng/ul	0.00
9) 2-Chlorophenol-d4	7.57	132	7557	6.37	ng/ul	0.00
13) 4-Methylphenol-d8	8.74	113	5910	4.75	ng/ul	0.00
19) Nitrobenzene-d5	9.20	128	3942	6.47	ng/ul	0.00
22) 2-Nitrophenol-d4	9.93	143	4361	6.96	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.47	165	8693	6.93	ng/ul	0.00
29) 4-Chloroaniline-d4	0.00	131	0	0.00	ng/ul	
44) Dimethylphthalate-d6	14.07	166	30598	6.86	ng/ul	0.00
47) Acenaphthylene-d8	14.37	160	32609	5.91	ng/ul	0.00
52) 4-Nitrophenol-d4	14.86	143	2147	2.38	ng/ul	0.00
58) Fluorene-d10	15.66	176	25643	6.35	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.76	200	4288	5.22	ng/ul	-0.01
71) Anthracene-d10	17.52	188	47113	6.61	ng/ul	0.00
79) Pyrene-d10	19.79	212	57485	6.47	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.69	264	60777	6.29	ng/ul	-0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
45) Dimethylphthalate	14.12	163	24617	5.40	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#45

Dimethylphthalate

Concen: 5.40 ng/ul

RT: 14.12 min Scan# 1920

Delta R.T. -0.01 min

Lab File: BG019057.D

Acq: 6 Oct 2015 22:29

Instrument :

BNA_G

ClientSampleId :

C03G3

Tot Ion:163 Resp: 24617

Ion Ratio Lower Upper

163 100

194 4.7 3.3 4.9

164 10.0 8.1 12.1

