

Data Path : Z:\HPCHEM1\BNA G\Data\BG100716\RAW\
 Data File : BG024210.D
 Acq On : 6 Oct 2016 22:11
 Operator : UM/SJ
 Sample : H5082-06
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BDPH9

Quant Time: Oct 07 09:26:59 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 07 06:34:32 2016
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.303	152	179503	20.00	ng/ul	0.00
18) Naphthalene-d8	11.129	136	762456	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.913	164	617520	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.657	188	1312778	20.00	ng/ul	0.00
75) Chrysene-d12	21.958	240	1403379	20.00	ng/ul	0.00
83) Perylene-d12	25.413	264	1441680	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.656	96	13271	3.51	ng/uL	0.00
5) Phenol-d5	7.428	99	341649	20.40	ng/ul	0.00
7) Bis-(2-Chloroethyl)eth...	7.616	67	241421	21.32	ng/ul	0.00
9) 2-Chlorophenol-d4	7.822	132	247805	21.49	ng/ul	0.00
13) 4-Methylphenol-d8	8.985	113	259278	19.15	ng/ul	0.00
19) Nitrobenzene-d5	9.473	128	126936	23.46	ng/ul	0.00
22) 2-Nitrophenol-d4	10.201	143	139817	23.10	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.730	165	275091	20.86	ng/ul	0.00
29) 4-Chloroaniline-d4	11.259	131	346868	25.04	ng/ul	0.00
43) Dimethylphthalate-d6	14.308	166	1017492	23.60	ng/ul	0.00
46) Acenaphthylene-d8	14.614	160	1152736	23.34	ng/ul	0.00
51) 4-Nitrophenol-d4	15.072	143	174117	22.77	ng/ul	0.00
57) Fluorene-d10	15.900	176	947999	23.25	ng/ul	0.00
62) 4,6-Dinitro-2-methylph...	16.000	200	177559	22.51	ng/ul	0.00
70) Anthracene-d10	17.751	188	1495838	25.23	ng/ul	0.00
76) Pyrene-d10	20.019	212	1695519	25.89	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.166	264	1553381	23.95	ng/ul	0.00
Target Compounds						
44) Dimethylphthalate	14.355	163	959505	22.70	ng/ul	Ovalue 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\Data\BG100716\RAW\
Data File : BG024210.D
Acq On : 6 Oct 2016 22:11
Operator : UM/SJ
Sample : H5082-06
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BDPH9

Quant Time: Oct 07 09:26:59 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Oct 07 06:34:32 2016
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 22.70 ng/ul

RT: 14.355 min Scan# 19

Delta R.T. 0.001 min

Lab File: BG024210.D

Acq: 6 Oct 2016 22:11

Instrument :

BNA_G

ClientSampleId :

BDPH9

Tot Ion:163 Resp: 959505

Ion Ratio Lower Upper

163 100

194 4.3 3.4 5.2

164 10.3 8.8 13.2

