


Data Path : Z:\HPCHEM1\BNA G\DATA\BG100717\
Data File : BG029081.D
Acq On : 8 Oct 2017 14:15
Operator : SJ/JU
Sample : PB102938EL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampled :
SBLK38

Manual Integrations
APPROVED

Sohil
10/9/2017 6:50:18 PM

Quant Time: Oct 09 07:39:12 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG100717MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 09 07:29:39 2017
Response via : Initial Calibration






Data Path : Z:\HPCHEM1\BNA G\DATA\BG100717\
Data File : BG029081.D
Acq On : 8 Oct 2017 14:15
Operator : SJ/JU
Sample : PB102938BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampled :
SBLK38

Manual Integrations
APPROVED

Sohil
10/9/2017 6:50:18 PM

Quant Time: Oct 09 07:34:46 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG100717MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 09 07:29:39 2017
Response via : Initial Calibration



TIC: BG029081.D

(14) Anthracene-d10 (S)

17.559min (-0.097) 20.65ng/ul

response 521873

Ion	Exp%	Act%
188.00	100	100
94.00	11.10	9.27
80.00	11.80	9.77
0.00	0.00	0.00




Data Path : Z:\HPCHEM1\BNA G\DATA\BG100717\
Data File : BG029081.D
Acq On : 8 Oct 2017 14:15
Operator : SJ/JU
Sample : PB102938BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK38

Manual Integrations
APPROVED

Sohil
10/9/2017 6:50:18 PM

Quant Time: Oct 09 07:34:46 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG100717MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 09 07:29:39 2017
Response via : Initial Calibration



TIC: BG029081.D

(14) Anthracene-d10 (S)

17.659min (+0.003) 29.67ng/ul mgus/9/17

response 750066

Ion	Exp%	Act%
188.00	100	100
94.00	11.10	10.66
80.00	11.80	9.40#
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG100717\
 Data File : BG029081.D
 Acq On : 8 Oct 2017 14:15
 Operator : SJ/JU
 Sample : PB102938BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK38

Manual Integrations
 APPROVED

Sohil
 10/9/2017 6:50:18 PM

Quant Time: Oct 09 07:39:12 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG100717MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 09 07:29:39 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.21	152	68444	20.00	ng/ul	0.00
2) Naphthalene-d8	11.03	136	314261	20.00	ng/ul	0.00
6) Acenaphthene-d10	14.82	164	212506	20.00	ng/ul	0.00
12) Phenanthrene-d10	17.56	188	521873	20.00	ng/ul	0.00
17) Chrysene-d12	21.84	240	544877	20.00	ng/ul	0.00
22) Perylene-d12	25.18	264	538735	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
7) Acenaphthylene-d8	14.52	160	601539	27.22	ng/ul	0.00
10) Fluorene-d10	15.81	176	456162	28.46	ng/ul	0.00
14) Anthracene-d10	17.66	188	750066m	29.67	ng/ul	0.00
18) Pyrene-d10	19.93	212	842796	30.23	ng/ul	0.00
25) Benzo(a)pyrene-d12	24.95	264	769129	29.80	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

2017/10/9