

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG100819\
 Data File : BG043090.D
 Acq On : 8 Oct 2019 18:00
 Operator : JU
 Sample : K5141-08MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Oct 09 01:09:21 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG092519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 25 15:35:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.05	152	55230	20.00	ng	-0.03
21) Naphthalene-d8	10.86	136	207235	20.00	ng	-0.02
39) Acenaphthene-d10	14.67	164	155194	20.00	ng	-0.02
64) Phenanthrene-d10	17.42	188	392462	20.00	ng	-0.02
76) Chrysene-d12	21.69	240	432956	20.00	ng	-0.03
87) Perylene-d12	24.90	264	501363	20.00	ng	-0.05

System Monitoring Compounds

5) 2-Fluorophenol	5.64	112	402997	130.22	ng	-0.01
7) Phenol-d6	7.24	99	584603	131.18	ng	0.00
23) Nitrobenzene-d5	9.22	82	352704	82.72	ng	-0.02
42) 2,4,6-Tribromophenol	16.17	330	391034	146.53	ng	-0.01
45) 2-Fluorobiphenyl	13.31	172	912923	85.28	ng	-0.02
79) Terphenyl-d14	20.03	244	1912903	94.15	ng	-0.02

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.55	88	45228	35.052	ng	# 60
3) Pyridine	3.96	79	78268	21.567	ng	88
4) n-Nitrosodimethylamine	3.85	42	66022	37.831	ng	# 86
6) Aniline	7.39	93	125616	23.350	ng	99
8) 2-Chlorophenol	7.63	128	153572	47.583	ng	96
9) Benzaldehyde	7.20	77	83387	30.620	ng	# 77
10) Phenol	7.26	94	213036	52.102	ng	94
11) bis(2-Chloroethyl)ether	7.48	93	139596	44.125	ng	92
12) 1,3-Dichlorobenzene	7.95	146	163280	39.658	ng	97
13) 1,4-Dichlorobenzene	8.09	146	167950	40.920	ng	95
14) 1,2-Dichlorobenzene	8.41	146	156097	39.948	ng	94
15) Benzyl Alcohol	8.30	79	148758	44.397	ng	95
16) 2,2'-oxybis(1-Chloropropan	8.58	45	192859	31.743	ng	96
17) 2-Methylphenol	8.51	107	133018	46.493	ng	96
18) Hexachloroethane	9.14	117	59346	38.393	ng	91
19) n-Nitroso-di-n-propylamine	8.86	70	118209	40.394	ng	92
20) 3+4-Methylphenols	8.83	107	188363	48.043	ng	96
22) Acetophenone	8.88	105	227647	42.616	ng	# 95
24) Nitrobenzene	9.26	77	182566	44.891	ng	98
25) Isophorone	9.78	82	344532	46.004	ng	98
26) 2-Nitrophenol	9.97	139	90445	52.242	ng	94
27) 2,4-Dimethylphenol	10.03	122	154518	58.117	ng	96
28) bis(2-Chloroethoxy)methane	10.26	93	186051	43.786	ng	97
29) 2,4-Dichlorophenol	10.52	162	171701	51.926	ng	100
30) 1,2,4-Trichlorobenzene	10.72	180	180915	42.368	ng	99
31) Naphthalene	10.91	128	459430	45.036	ng	99
32) Benzoic acid	10.14	122	33736	19.633	ng	95
33) 4-Chloroaniline	11.03	127	42214	9.536	ng	95
34) Hexachlorobutadiene	11.20	225	126692	39.625	ng	98
35) Caprolactam	11.79	113	30087	25.802	ng	# 83
36) 4-Chloro-3-methylphenol	12.15	107	191383	53.232	ng	95
37) 2-Methylnaphthalene	12.51	142	353618	45.438	ng	96
38) 1-Methylnaphthalene	12.73	142	339600	46.117	ng	98
40) 1,2,4,5-Tetrachlorobenzene	12.88	216	236250	40.487	ng	100


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG100819\
 Data File : BG043090.D
 Acq On : 8 Oct 2019 18:00
 Operator : JU
 Sample : K5141-08MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Oct 09 01:09:21 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG092519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 25 15:35:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	12.86	237	309407	83.221	ng	97
43) 2,4,6-Trichlorophenol	13.12	196	165748	48.530	ng	100
44) 2,4,5-Trichlorophenol	13.19	196	185040	52.593	ng	97
46) 1,1'-Biphenyl	13.51	154	493201	41.907	ng	98
47) 2-Chloronaphthalene	13.56	162	387042	43.863	ng	99
48) 2-Nitroaniline	13.76	65	133267	45.416	ng	90
49) Acenaphthylene	14.40	152	638375	47.281	ng	99
50) Dimethylphthalate	14.13	163	729580	62.743	ng	98
51) 2,6-Dinitrotoluene	14.25	165	126332	51.017	ng	93
52) Acenaphthene	14.75	154	384478	42.543	ng	98
53) 3-Nitroaniline	14.59	138	85720	33.496	ng	95
54) 2,4-Dinitrophenol	14.78	184	99234	64.470	ng	96
55) Dibenzofuran	15.07	168	629053	47.053	ng	99
56) 4-Nitrophenol	14.90	139	208718	107.041	ng	96
57) 2,4-Dinitrotoluene	15.03	165	184823	50.968	ng	95
58) Fluorene	15.73	166	537735	47.113	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.30	232	184353	49.212	ng	99
60) Diethylphthalate	15.49	149	566672	47.885	ng	99
61) 4-Chlorophenyl-phenylether	15.71	204	309339	45.190	ng	97
62) 4-Nitroaniline	15.75	138	141611	50.738	ng	98
63) Azobenzene	16.01	77	484550	44.974	ng	98
65) 4,6-Dinitro-2-methylphenol	15.80	198	97694	43.990	ng	92
66) n-Nitrosodiphenylamine	15.93	169	496527	47.925	ng	99
67) 4-Bromophenyl-phenylether	16.61	248	225180	45.715	ng	98
68) Hexachlorobenzene	16.73	284	248348	45.282	ng	96
69) Atrazine	16.88	200	193752	44.411	ng	95
70) Pentachlorophenol	17.07	266	282252	89.188	ng	99
71) Phenanthrene	17.47	178	908820	47.433	ng	97
72) Anthracene	17.55	178	931158	48.681	ng	99
73) Carbazole	17.82	167	876396	49.409	ng	99
74) Di-n-butylphthalate	18.38	149	1006226	47.546	ng	99
75) Fluoranthene	19.47	202	1214966	47.941	ng	99
77) Benzidine	19.65	184	322861	29.782	ng	99
78) Pyrene	19.83	202	1227268	47.492	ng	99
80) Butylbenzylphthalate	20.71	149	469875	47.903	ng	100
81) Benzo(a)anthracene	21.67	228	1244374	46.127	ng	99
82) 3,3'-Dichlorobenzidine	21.58	252	383481	36.243	ng	99
83) Chrysene	21.74	228	1218971	46.562	ng	99
84) Bis(2-ethylhexyl)phthalate	21.58	149	662325	47.454	ng	99
85) Di-n-octyl phthalate	22.79	149	1127558	49.404	ng	96
86) Indeno(1,2,3-cd)pyrene	28.57	276	1613206	49.499	ng	100
88) Benzo(b)fluoranthene	23.89	252	1357816	47.864	ng	99
89) Benzo(k)fluoranthene	23.96	252	1312663	46.774	ng	100
90) Benzo(a)pyrene	24.76	252	1311820	49.013	ng	100
91) Dibenzo(a,h)anthracene	28.64	278	1354555	49.355	ng	97
92) Benzo(g,h,i)perylene	29.73	276	1339334	50.366	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.05 min Scan# 845
Delta R.T. -0.03 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Instrument :
BNA_G
ClientSampleId :

Tot Ion:152 Resp: 55230
Ion Ratio Lower Upper
152 100
150 150.0 115.0 172.4
115 59.2 47.0 70.6




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 35.052 ng
RT: 3.55 min Scan# 78
Delta R.T. -0.01 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion: 88 Resp: 45228
Ion Ratio Lower Upper
88 100
58 79.7 74.3 111.5
43 100.2 29.8 44.8#



Abundance Ion 88.00 (87.70 to 88.70): BGC
Ion 58.00 (57.70 to 58.70): BGC
Ion 43.00 (42.70 to 43.70): BGC





#3

Pvridine

Concen: 21.567 ng

RT: 3.96 min Scan# 148

Delta R.T. -0.00 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :





Tot Ion: 79 Resp: 78268

Ion Ratio Lower Upper

79 100

52 79.3 73.0 109.4

51 36.2 35.1 52.7



#4

n-Nitrosodimethylamine

Concen: 37.831 ng

RT: 3.85 min Scan# 130

Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00




Tgt Ion: 42 Resp: 66022


Ion Ratio Lower Upper

42 100

74 115.3 80.5 120.7

44 5.6 6.9 10.3#





#5

2-Fluorophenol

Concen: 130.219 ng

RT: 5.64 min Scan# 435

Delta R.T. -0.01 min

Lab File: BG043090.D





Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :

Tot Ion:112	Resp:	402997
Ion Ratio	Lower	Upper
112	100	
64	64.3	56.6 84.8
63	36.1	35.1 52.7



#6

Aniline

Concen: 23.350 ng




RT: 7.39 min Scan# 732


Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Tgt Ion: 93	Resp:	125616
Ion Ratio	Lower	Upper
93	100	
66	43.0	35.0 52.6
65	21.9	17.4 26.0





#7

Phenol-d6

Concen: 131.175 ng

RT: 7.24 min Scan# 706

Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :





Tot Ion: 99 Resp: 584603

Ion Ratio Lower Upper

99 100

42 18.6 17.2 25.8

71 40.2 33.1 49.7



#8

2-Chlorophenol

Concen: 47.583 ng

RT: 7.63 min Scan# 773

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00




Tgt Ion: 128 Resp: 153572


Ion Ratio Lower Upper

128 100

130 31.4 12.5 52.5

64 48.5 32.8 72.8





#9

Benzaldehyde

Concen: 30.620 ng

RT: 7.20 min Scan# 699

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :





Tot Ion: 77 Resp: 83387

Ion Ratio Lower Upper

77 100

105 104.5 64.6 104.6

106 101.5 59.4 99.4#



#10

Phenol

Concen: 52.102 ng

RT: 7.26 min Scan# 710

Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00




Tgt Ion: 94 Resp: 213036


Ion Ratio Lower Upper

94 100

65 32.8 16.6 56.6

66 48.0 32.4 72.4





#11

bis(2-Chloroethyl)ether

Concen: 44.125 ng

RT: 7.48 min Scan# 747

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :





Tot Ion: 93 Resp: 139596

Ion Ratio Lower Upper

93 100

63 78.4 67.4 107.4

95 31.0 11.9 51.9



#12

1,3-Dichlorobenzene

Concen: 39.658 ng

RT: 7.95 min Scan# 827

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00




Tgt Ion: 146 Resp: 163280


Ion Ratio Lower Upper

146 100

148 65.4 50.7 76.1

75 32.2 27.7 41.5





#13

1,4-Dichlorobenzene

Concen: 40.920 ng

RT: 8.09 min Scan# 852

Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :


Tot Ion:146 Resp: 167950

Ion Ratio Lower Upper




146 100

148 67.7 50.9 76.3

111 42.9 31.8 47.8



Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 39.948 ng

RT: 8.41 min Scan# 906

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion:146 Resp: 156097

Ion Ratio Lower Upper



146 100


148 64.9 53.4 80.2

111 41.0 38.5 57.7



Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 44.397 ng

RT: 8.30 min Scan# 887

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampled :


Tot Ion: 79 Resp: 148758

Ion Ratio Lower Upper

79 100

108 65.2 55.7 83.5




77 63.1 53.3 79.9



Abundance Ion 79.00 (78.70 to 79.70): BGC

Ion 108.00 (107.70 to 108.70): BGC

Ion 77.00 (76.70 to 77.70): BGC



#16

2,2'-oxybis(1-Chloropropane)

Concen: 31.743 ng

RT: 8.58 min Scan# 935

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion: 45 Resp: 192859

Ion Ratio Lower Upper

45 100

77 14.1 0.0 32.8



79 11.2 0.0 29.4




Abundance Ion 45.00 (44.70 to 45.70): BGC

Ion 77.00 (76.70 to 77.70): BGC

Ion 79.00 (78.70 to 79.70): BGC






#17
2-Methylphenol
Concen: 46.493 ng
RT: 8.51 min Scan# 922
Delta R.T. -0.01 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Instrument :
BNA_G
ClientSampleId :

Tot Ion:	107	Resp:	133018
Ion	Ratio	Lower	Upper
107	100		
108	116.9	89.9	134.9
77	57.0	45.4	68.2
79	55.8	41.4	62.0






Abundance

Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E


Ion 77.00 (76.70 to 77.70): BG

Ion 79.00 (78.70 to 79.70): BG



#18
Hexachloroethane
Concen: 38.393 ng
RT: 9.14 min Scan# 1030
Delta R.T. -0.03 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion:	117	Resp:	59346
Ion	Ratio	Lower	Upper
117	100		
119	90.5	74.6	111.8
201	117.7	106.0	159.0






Abundance

Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E









#19
n-Nitroso-di-n-propylamine
Concen: 40.394 ng
RT: 8.86 min Scan# 982
Delta R.T. -0.03 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Instrument :
BNA_G
ClientSampled :

Tot Ion: 70 Resp: 118209
Ion Ratio Lower Upper
70 100
42 58.1 52.6 79.0
101 11.2 8.7 13.1
130 23.3 15.8 23.8




Abundance Ion 70.00 (69.70 to 70.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 101.00 (100.70 to 101.70): BGC
Ion 130.00 (129.70 to 130.70): BGC






#20
3+4-Methylphenols
Concen: 48.043 ng
RT: 8.83 min Scan# 978
Delta R.T. -0.01 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion:107 Resp: 188363
Ion Ratio Lower Upper
107 100
108 90.6 64.9 104.9
77 37.0 19.4 59.4
79 36.2 16.2 56.2



Abundance Ion 107.00 (106.70 to 107.70): BGC
Ion 108.00 (107.70 to 108.70): BGC
Ion 77.00 (76.70 to 77.70): BGC
Ion 79.00 (78.70 to 79.70): BGC





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.86 min Scan# 1323

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :

Tot Ion:136 Resp: 207235



Ion Ratio Lower Upper

136 100

137 11.8 8.5 12.7

54 8.6 7.4 11.2

68 4.2 3.6 5.4





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BGC

Ion 68.00 (67.70 to 68.70): BGC



#22

Acetophenone

Concen: 42.616 ng

RT: 8.88 min Scan# 985

Delta R.T. -0.03 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Tgt Ion:105 Resp: 227647



Ion Ratio Lower Upper

105 100

71 2.7 1.2 1.8#

51 29.5 27.2 40.8

120 21.7 16.9 25.3





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BGC

Ion 51.00 (50.70 to 51.70): BGC

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 82.724 ng
 RT: 9.22 min Scan# 1043
 Delta R.T. -0.02 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion: 82 Resp: 352704
 Ion Ratio Lower Upper
 82 100
 128 40.9 29.5 44.3
 54 52.7 44.6 67.0



#24
 Nitrobenzene
 Concen: 44.891 ng
 RT: 9.26 min Scan# 1050
 Delta R.T. -0.02 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00

Tgt Ion: 77 Resp: 182566
 Ion Ratio Lower Upper
 77 100
 123 41.0 31.9 47.9
 65 15.0 12.4 18.6





#25

Isophorone

Concen: 46.004 ng

RT: 9.78 min Scan# 1139

Delta R.T. -0.03 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampled :


Tot Ion: 82 Resp: 344532

Ion Ratio Lower Upper




82 100

95 8.0 7.0 10.4

138 18.2 13.8 20.6



Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 95.00 (94.70 to 95.70): BGC
Ion 138.00 (137.70 to 138.70): BGC



#26

2-Nitrophenol

Concen: 52.242 ng

RT: 9.97 min Scan# 1171

Delta R.T. -0.03 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion: 139 Resp: 90445

Ion Ratio Lower Upper



139 100


109 37.7 31.0 46.6

65 53.3 48.4 72.6



Abundance Ion 139.00 (138.70 to 139.70): BGC
Ion 109.00 (108.70 to 109.70): BGC
Ion 65.00 (64.70 to 65.70): BGC






#27
2,4-Dimethylphenol
Concen: 58.117 ng
RT: 10.03 min Scan# 1182
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Instrument :
BNA_G
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
122	100		
107	125.8	102.2	153.2
121	54.1	48.5	72.7






Abundance

Ion 122.00 (121.70 to 122.70): E


Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28
bis(2-Chloroethoxy)methane
Concen: 43.786 ng
RT: 10.26 min Scan# 1221
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion	Ratio	Lower	Upper
93	100		
95	33.2	25.4	38.0
123	12.9	9.3	13.9






Abundance

Ion 93.00 (92.70 to 93.70): BGC

Ion 95.00 (94.70 to 95.70): BGC

Ion 123.00 (122.70 to 123.70): E









#29
 2,4-Dichlorophenol
 Concen: 51.926 ng
 RT: 10.52 min Scan# 1264
 Delta R.T. -0.01 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00




Instrument :
 BNA_G
 ClientSampled :


Tot Ion	Ratio	Lower	Upper
162	100		
164	64.7	44.7	84.7
98	36.3	16.0	56.0



#30
 1,2,4-Trichlorobenzene
 Concen: 42.368 ng
 RT: 10.72 min Scan# 1299
 Delta R.T. -0.02 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00

Tgt Ion	Ratio	Lower	Upper
180	100		
182	99.2	79.3	118.9
145	26.4	22.0	33.0









#31
Naphthalene
Concen: 45.036 ng
RT: 10.91 min Scan# 1331
Delta R.T. -0.03 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00




Instrument :
BNA_G
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
128	100		
129	11.6	8.8	13.2
127	14.4	11.3	16.9



#32
Benzoic acid
Concen: 19.633 ng
RT: 10.14 min Scan# 1201
Delta R.T. -0.05 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion	Ratio	Lower	Upper
122	100		
105	134.9	107.5	147.5
77	96.0	79.8	119.8





#33

4-Chloroaniline

Concen: 9.536 ng

RT: 11.03 min Scan# 1352


Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :
BNA_G
ClientSampled :

Tot Ion	Ratio	Lower	Upper
127	100		
129	33.2	26.2	39.2
65	33.8	30.6	45.8
92	16.6	16.2	24.4






Abundance

Ion 129.00 (128.70 to 129.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 65.00 (64.70 to 65.70): BG

Ion 92.00 (91.70 to 92.70): BG



#34

Hexachlorobutadiene

Concen: 39.625 ng


RT: 11.20 min Scan# 1380

Delta R.T. -0.03 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.7	49.3	73.9
227	63.4	52.8	79.2






Abundance

Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E





#35

Caprolactam

Concen: 25.802 ng

RT: 11.79 min Scan# 1481

Delta R.T. -0.03 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :


Tot Ion:113 Resp: 30087

Ion Ratio Lower Upper

113 100

55 174.3 175.5 215.5#




56 135.7 142.3 182.3#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BGC

Ion 56.00 (55.70 to 56.70): BGC



#36

4-Chloro-3-methylphenol

Concen: 53.232 ng

RT: 12.15 min Scan# 1542

Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion:107 Resp: 191383

Ion Ratio Lower Upper

107 100

144 27.3 20.7 31.1



142 77.8 58.6 87.8




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
 2-Methylnaphthalene
 Concen: 45.438 ng
 RT: 12.51 min Scan# 1604
 Delta R.T. -0.01 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00




Instrument :
 BNA_G
 ClientSampled :


Tot Ion	Ratio	Lower	Upper
142	100		
141	90.9	69.0	103.4
115	40.3	32.8	49.2



#38
 1-Methylnaphthalene
 Concen: 46.117 ng
 RT: 12.73 min Scan# 1641
 Delta R.T. -0.01 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00

Tgt Ion	Ratio	Lower	Upper
142	100		
141	92.1	74.9	112.3
116	4.5	3.7	5.5






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.67 min Scan# 1972
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Instrument :
BNA_G
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
164	100		
162	100.1	78.5	117.7
160	45.0	35.9	53.9






Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.487 ng
RT: 12.88 min Scan# 1666
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion	Ratio	Lower	Upper
216	100		
214	78.0	62.4	93.6
179	19.5	15.4	23.2
108	16.5	12.6	19.0




Abundance


Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 83.221 ng

RT: 12.86 min Scan# 1663

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampled :


Tot Ion:237 Resp: 309407

Ion Ratio Lower Upper




237 100

235 64.1 46.7 86.7

272 13.3 0.0 32.3



Abundance Ion 236.80 (236.50 to 237.50): E
Ion 234.90 (234.60 to 235.60): E
Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 146.528 ng

RT: 16.17 min Scan# 2226

Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion:330 Resp: 391034

Ion Ratio Lower Upper



330 100


332 95.9 78.2 117.2

141 28.9 26.0 39.0



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#43
 2,4,6-Trichlorophenol
 Concen: 48.530 ng
 RT: 13.12 min Scan# 1707
 Delta R.T. -0.01 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00




Instrument :
 BNA_G
 ClientSampled :


Tot Ion	Ratio	Lower	Upper
196	100		
198	94.2	75.5	113.3
200	31.1	24.9	37.3



#44
 2,4,5-Trichlorophenol
 Concen: 52.593 ng
 RT: 13.19 min Scan# 1720
 Delta R.T. -0.01 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00

Tgt Ion	Ratio	Lower	Upper
196	100		
198	94.4	75.1	112.7
97	42.6	37.0	55.4
132	24.4	17.5	26.3









#45
2-Fluorobiphenyl
Concen: 85.278 ng
RT: 13.31 min Scan# 1739
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00




Instrument :
BNA_G
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
172	100		
171	36.6	29.6	44.4
170	24.0	19.6	29.4



#46
1,1'-Biphenyl
Concen: 41.907 ng
RT: 13.51 min Scan# 1774
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion	Ratio	Lower	Upper
154	100		
153	41.9	21.4	61.4
76	9.5	0.0	31.3









#47
2-Chloronaphthalene
Concen: 43.863 ng
RT: 13.56 min Scan# 1782
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Instrument :
BNA_G
ClientSampleId :

Tot Ion:162 Resp: 387042
Ion Ratio Lower Upper
162 100
127 38.2 31.0 46.6
164 33.7 26.2 39.2




Abundance Ion 162.00 (161.70 to 162.70): E
Ion 127.00 (126.70 to 127.70): E
Ion 164.00 (163.70 to 164.70): E






#48
2-Nitroaniline
Concen: 45.416 ng
RT: 13.76 min Scan# 1816
Delta R.T. -0.01 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion: 65 Resp: 133267
Ion Ratio Lower Upper
65 100
92 66.1 52.0 78.0
138 103.7 70.4 105.6



Abundance Ion 65.00 (64.70 to 65.70): BGC
Ion 92.00 (91.70 to 92.70): BGC
Ion 138.00 (137.70 to 138.70): E





#49

Acenaphthylene

Concen: 47.281 ng

RT: 14.40 min Scan# 1926

Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :


Tot Ion:152 Resp: 638375

Ion Ratio Lower Upper




152 100

151 20.3 16.8 25.2

153 13.6 11.3 16.9



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 62.743 ng

RT: 14.13 min Scan# 1880

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion:163 Resp: 729580

Ion Ratio Lower Upper



163 100


194 4.7 4.4 6.6

164 10.2 8.6 12.8



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E









#51
 2,6-Dinitrotoluene
 Concen: 51.017 ng
 RT: 14.25 min Scan# 1899
 Delta R.T. -0.02 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion:165 Resp: 126332
 Ion Ratio Lower Upper
 165 100
 63 58.9 50.3 75.5
 89 54.2 48.9 73.3



#52
 Acenaphthene
 Concen: 42.543 ng
 RT: 14.75 min Scan# 1984
 Delta R.T. -0.01 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00

Tgt Ion:154 Resp: 384478
 Ion Ratio Lower Upper
 154 100
 153 116.2 91.4 137.0
 152 54.5 42.8 64.2





#53

3-Nitroaniline

Concen: 33.496 ng

RT: 14.59 min Scan# 1957

Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :


Tot Ion:138 Resp: 85720

Ion Ratio Lower Upper

138 100

108 10.4 7.9 11.9




92 115.7 97.0 145.6



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BGC



#54

2,4-Dinitrophenol

Concen: 64.470 ng

RT: 14.78 min Scan# 1990

Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion:184 Resp: 99234

Ion Ratio Lower Upper

184 100

63 67.1 55.0 82.6



154 65.9 56.5 84.7




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BGC

Ion 154.00 (153.70 to 154.70): E









#55
Dibenzenofuran
Concen: 47.053 ng
RT: 15.07 min Scan# 2040
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00




Instrument :
BNA_G
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
168	100		
139	39.6	32.6	48.8
169	14.2	11.2	16.8



#56
4-Nitrophenol
Concen: 107.041 ng
RT: 14.90 min Scan# 2011
Delta R.T. -0.00 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion	Ratio	Lower	Upper
139	100		
109	97.2	79.4	119.4
65	107.0	92.5	132.5





#57

2,4-Dinitrotoluene

Concen: 50.968 ng

RT: 15.03 min Scan# 2033

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :

Tot Ion:165 Resp: 184823


Ion Ratio Lower Upper

165 100

63 42.8 36.5 54.7

89 65.4 55.5 83.3

182 2.6 2.4 3.6






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BG

Ion 89.00 (88.70 to 89.70): BG

Ion 182.00 (181.70 to 182.70): E



#58

Fluorene

Concen: 47.113 ng

RT: 15.73 min Scan# 2151

Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion:166 Resp: 537735

Ion Ratio Lower Upper

166 100

165 98.1 77.1 115.7



167 14.1 11.0 16.6




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#59
2,3,4,6-Tetrachlorophenol
Concen: 49.212 ng
RT: 15.30 min Scan# 2079
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00




Instrument :
BNA_G
ClientSampleId :


Tot Ion:232	Resp:	184353
Ion Ratio	Lower	Upper
232	100	
131	35.8	29.6 44.4
130	2.4	1.9 2.9
166	27.6	21.8 32.8



#60
Diethylphthalate
Concen: 47.885 ng
RT: 15.49 min Scan# 2111
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion:149	Resp:	566672
Ion Ratio	Lower	Upper
149	100	
177	24.1	18.8 28.2
150	13.1	10.6 15.8









#61
4-Chlorophenyl-phenylether
Concen: 45.190 ng
RT: 15.71 min Scan# 2149
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00




Instrument :
BNA_G
ClientSampled :


Tot Ion	Ratio	Lower	Upper
204	100		
206	34.4	28.5	42.7
141	55.7	42.4	63.6



#62
4-Nitroaniline
Concen: 50.738 ng
RT: 15.75 min Scan# 2155
Delta R.T. -0.01 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion	Ratio	Lower	Upper
138	100		
92	52.2	32.9	72.9
108	99.4	81.3	121.3





#63

Azobenzene

Concen: 44.974 ng

RT: 16.01 min Scan# 2199

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :

Tot Ion: 77 Resp: 484550


Ion Ratio Lower Upper

77 100

182 32.0 10.2 50.2

105 20.6 0.0 39.1

51 34.5 13.9 53.9






Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 182.00 (181.70 to 182.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 51.00 (50.70 to 51.70): BGC



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.42 min Scan# 2439

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion: 188 Resp: 392462

Ion Ratio Lower Upper

188 100

94 5.9 5.7 8.5



80 7.1 6.5 9.7




Abundance Ion 188.00 (187.70 to 188.70): BGC

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC









#65
 4,6-Dinitro-2-methylphenol
 Concen: 43.990 ng
 RT: 15.80 min Scan# 2163
 Delta R.T. -0.01 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion:198 Resp: 97694
 Ion Ratio Lower Upper
 198 100
 51 38.5 25.3 65.3
 105 33.7 16.6 56.6



#66
 n-Nitrosodiphenylamine
 Concen: 47.925 ng
 RT: 15.93 min Scan# 2186
 Delta R.T. -0.01 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00

Tgt Ion:169 Resp: 496527
 Ion Ratio Lower Upper
 169 100
 168 68.7 54.8 82.2
 167 37.1 30.6 46.0









#67
4-Bromophenyl-phenylether
Concen: 45.715 ng
RT: 16.61 min Scan# 2301
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00




Instrument :
BNA_G
ClientSampleId :


Tot Ion:248 Resp: 225180
Ion Ratio Lower Upper
248 100
250 95.8 75.8 113.8
141 62.0 51.1 76.7



#68
Hexachlorobenzene
Concen: 45.282 ng
RT: 16.73 min Scan# 2322
Delta R.T. -0.02 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion:284 Resp: 248348
Ion Ratio Lower Upper
284 100
142 24.2 21.6 32.4
249 31.8 24.1 36.1





#69

Atrazine

Concen: 44.411 ng

RT: 16.88 min Scan# 2347

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :


Tot Ion:200 Resp: 193752

Ion Ratio Lower Upper

200 100

173 25.5 4.0 44.0




215 51.9 27.4 67.4



Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E



#70

Pentachlorophenol

Concen: 89.188 ng

RT: 17.07 min Scan# 2380

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion:266 Resp: 282252

Ion Ratio Lower Upper

266 100

268 63.2 52.2 78.2



264 63.1 50.7 76.1




Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E





#71

Phenanthrene

Concen: 47.433 ng

RT: 17.47 min Scan# 2447

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :


Tot Ion:178 Resp: 908820

Ion Ratio Lower Upper

178 100

176 20.3 15.1 22.7




179 16.7 14.1 21.1



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Anthracene

Concen: 48.681 ng

RT: 17.55 min Scan# 2462

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion:178 Resp: 931158

Ion Ratio Lower Upper

178 100

176 19.9 15.5 23.3



179 17.1 13.5 20.3




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#73
 Carbazole
 Concen: 49.409 ng
 RT: 17.82 min Scan# 2508
 Delta R.T. -0.02 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion:167 Resp: 876396
 Ion Ratio Lower Upper
 167 100
 166 23.3 18.6 28.0
 139 12.8 10.8 16.2



#74
 Di-n-butylphthalate
 Concen: 47.546 ng
 RT: 18.38 min Scan# 2603
 Delta R.T. -0.02 min
 Lab File: BG043090.D
 Acq: 8 Oct 2019 18:00

Tgt Ion:149 Resp: 1006226
 Ion Ratio Lower Upper
 149 100
 150 9.9 7.7 11.5
 104 6.1 5.0 7.6





#75

Fluoranthene

Concen: 47.941 ng

RT: 19.47 min Scan# 2788

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :


Tot Ion:202 Resp: 1214966

Ion Ratio Lower Upper




202 100

101 7.1 0.0 27.3

203 19.3 0.0 39.5



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.69 min Scan# 3166

Delta R.T. -0.03 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion:240 Resp: 432956

Ion Ratio Lower Upper



240 100


120 5.6 4.8 7.2

236 26.1 21.4 32.0



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#77

Benzidine

Concen: 29.782 ng

RT: 19.65 min Scan# 2819

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Instrument :

BNA_G

ClientSampleId :

Tot Ion:184 Resp: 322861

Ion	Ratio	Lower	Upper
184	100		
185	15.5	11.8	17.8
183	13.0	10.1	15.1






Abundance

Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 47.492 ng

RT: 19.83 min Scan# 2850


Delta R.T. -0.01 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Tgt Ion:202 Resp: 1227268

Ion	Ratio	Lower	Upper
202	100		
200	22.8	17.6	26.4
203	19.2	15.2	22.8






Abundance

Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 94.147 ng

RT: 20.03 min Scan# 2883

Delta R.T. -0.02 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Instrument :

BNA_G

ClientSampleId :

Tot Ion:244 Resp: 1912903




Ion	Ratio	Lower	Upper
244	100		
212	8.6	7.4	11.0
122	7.8	7.0	10.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 47.903 ng

RT: 20.71 min Scan# 3000


Delta R.T. -0.03 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Tgt Ion:149 Resp: 469875



Ion	Ratio	Lower	Upper
149	100		
91	71.5	57.7	86.5
206	28.2	22.6	33.8




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BG

Ion 206.00 (205.70 to 206.70): E





#81

Benzo(a)anthracene

Concen: 46.127 ng

RT: 21.67 min Scan# 3163

Delta R.T. -0.03 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampled :


Tot Ion:228 Resp: 1244374

Ion Ratio Lower Upper

228 100

226 28.2 23.0 34.4




229 21.0 16.3 24.5



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

3,3'-Dichlorobenzidine

Concen: 36.243 ng

RT: 21.58 min Scan# 3148

Delta R.T. -0.03 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion:252 Resp: 383481

Ion Ratio Lower Upper

252 100

254 65.4 51.9 77.9



126 7.0 6.1 9.1




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E









#83
Chrysene
Concen: 46.562 ng
RT: 21.74 min Scan# 3174
Delta R.T. -0.03 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00




Instrument :
BNA_G
ClientSampled :


Tot Ion:228 Resp: 1218971
Ion Ratio Lower Upper
228 100
226 31.8 25.6 38.4
229 20.7 16.9 25.3



#84
Bis(2-ethylhexyl)phthalate
Concen: 47.454 ng
RT: 21.58 min Scan# 3147
Delta R.T. -0.03 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion:149 Resp: 662325
Ion Ratio Lower Upper
149 100
167 31.2 25.4 38.2
279 7.0 5.1 7.7





#85

Di-n-octyl phthalate

Concen: 49.404 ng

RT: 22.79 min Scan# 3353

Delta R.T. -0.04 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampleId :


Tot Ion:149 Resp: 1127558

Ion Ratio Lower Upper

149 100

167 1.8 1.4 2.0




43 9.1 8.7 13.1



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BGC



#86

Indeno(1,2,3-cd)pyrene

Concen: 49.499 ng

RT: 28.57 min Scan# 4337

Delta R.T. -0.08 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion:276 Resp: 1613206

Ion Ratio Lower Upper

276 100

138 8.8 7.0 10.6



277 26.3 20.9 31.3




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#87

Pervlene-d12

Concen: 20.000 ng

RT: 24.90 min Scan# 3713

Delta R.T. -0.05 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00

Instrument :

BNA_G

ClientSampled :


Tot Ion:264 Resp: 501363

Ion Ratio Lower Upper

264 100

260 23.9 18.7 28.1




265 23.5 17.7 26.5



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#88

Benzo(b)fluoranthene

Concen: 47.864 ng

RT: 23.89 min Scan# 3540

Delta R.T. -0.04 min

Lab File: BG043090.D

Acq: 8 Oct 2019 18:00


Tgt Ion:252 Resp: 1357816

Ion Ratio Lower Upper

252 100

253 23.4 18.2 27.2



125 6.1 4.6 7.0




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#89
Benzo(k)fluoranthene
Concen: 46.774 ng
RT: 23.96 min Scan# 3552
Delta R.T. -0.04 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00




Instrument :
BNA_G
ClientSampleId :


Tot Ion:252 Resp: 1312663
Ion Ratio Lower Upper
252 100
253 23.8 18.9 28.3
125 5.8 4.5 6.7



#90
Benzo(a)pyrene
Concen: 49.013 ng
RT: 24.76 min Scan# 3688
Delta R.T. -0.05 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion:252 Resp: 1311820
Ion Ratio Lower Upper
252 100
253 22.9 18.2 27.4
125 6.2 5.0 7.6









#91
Dibenzo(a,h)anthracene
Concen: 49.355 ng
RT: 28.64 min Scan# 4349
Delta R.T. -0.08 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Instrument :
BNA_G
ClientSampleId :

Tot Ion:278 Resp: 1354555

Ion	Ratio	Lower	Upper
278	100		
139	9.1	6.8	10.2
279	23.3	20.1	30.1



#92
Benzo(g,h,i)perylene
Concen: 50.366 ng
RT: 29.73 min Scan# 4534
Delta R.T. -0.08 min
Lab File: BG043090.D
Acq: 8 Oct 2019 18:00

Tgt Ion:276 Resp: 1339334

Ion	Ratio	Lower	Upper
276	100		
277	24.5	18.4	27.6
138	11.8	8.6	12.8

