


Data Path : Z:\HPCHEM1\BNA_G\Data\BG101017\
 Data File : BG029133.D
 Acq On : 10 Oct 2017 9:48
 Operator : SJ/JU
 Sample : DFTPP53
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP53

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG100717MA.M
 Title : SVOA CALIBRATION
 Last Update : Wed Oct 11 04:15:46 2017



AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2470


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.6	39128	PASS
68	69	0.00	2	1.5	569	PASS
69	198	0.00	100	30.9	37072	PASS
70	69	0.00	2	0.6	217	PASS
127	198	10	80	41.0	49162	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	119949	PASS
199	198	5	9	7.0	8389	PASS
275	198	10	60	29.2	35082	PASS
365	198	1	100	4.0	4749	PASS
441	442	0.01	100	14.9	15937	PASS
442	198	50	100	88.9	106690	PASS
443	442	15	24	19.4	20722	PASS

Data Path : Z:\HPCHEM1\BNA_G\Data\BG101017\
Data File : BG029133.D
Acq On : 10 Oct 2017 9:48
Operator : SJ/JU
Sample : DFTPP53
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP53

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG100717MA.M
Title : SVOA CALIBRATION
Last Update : Wed Oct 11 04:15:46 2017



AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2470

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.6	39128	PASS
68	69	0.00	2	1.5	569	PASS
69	198	0.00	100	30.9	37072	PASS
70	69	0.00	2	0.6	217	PASS
127	198	10	80	41.0	49162	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	119949	PASS
199	198	5	9	7.0	8389	PASS
275	198	10	60	29.2	35082	PASS
365	198	1	100	4.0	4749	PASS
441	442	0.01	100	14.9	15937	PASS
442	198	50	100	88.9	106690	PASS
443	442	15	24	19.4	20722	PASS

m/z	Abundance
37.15	196.0
38.15	282.0
39.15	2618.0
40.05	394.0
41.25	229.0
44.15	546.0
49.15	576.0
50.15	8905.0
51.15	35992.0
52.15	1703.0
55.15	290.0
56.05	1459.0
57.15	2951.0
58.05	222.0
61.15	510.0
62.15	427.0
63.15	1459.0
64.15	269.0
65.15	849.0
68.05	760.0
69.15	34184.0
70.05	382.0
73.05	449.0
74.15	3774.0
75.15	5552.0
76.15	2335.0
77.15	37888.0
78.15	2623.0
79.15	2152.0
80.05	2165.0
81.15	3068.0
82.15	602.0
83.15	954.0
85.15	591.0
86.05	707.0
87.15	426.0
91.05	656.0
92.05	809.0
93.15	4267.0
94.15	410.0
98.15	2781.0
99.15	2366.0
99.95	168.0
101.05	1757.0
103.05	436.0
104.05	866.0
105.05	1339.0
107.15	12426.0
108.15	2029.0
109.05	284.0
110.05	21256.0
111.05	2963.0
112.05	331.0
113.05	220.0
116.05	513.0
117.05	8830.0
118.05	679.0
119.05	200.0
120.15	192.0
122.05	765.0
123.15	1601.0
124.15	545.0
125.05	581.0
127.15	43296.0
128.15	3439.0
129.05	16316.0
130.15	1725.0
131.15	392.0
134.05	554.0
135.15	1509.0
135.95	296.0
137.05	779.0
139.05	266.0
140.05	282.0
141.05	2169.0
141.95	434.0
143.05	647.0
144.95	169.0
146.05	423.0
147.15	892.0
148.15	2327.0
149.15	580.0
151.15	270.0
153.05	758.0
154.05	544.0
155.15	1562.0
156.05	1653.0
157.15	454.0
157.65	191.0
158.05	408.0
159.05	358.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

160.05	894.0
161.15	1053.0
162.05	427.0
162.65	182.0
164.15	180.0
165.05	837.0
166.05	759.0
167.05	4343.0
168.15	2395.0
168.95	355.0
169.15	362.0
171.15	261.0
172.05	413.0
173.05	596.0
174.05	1068.0
175.15	2155.0
175.95	376.0
177.05	756.0
178.15	325.0
179.05	2844.0
180.15	3024.0
181.05	1318.0
182.95	297.0
184.05	416.0
185.15	1892.0
186.15	13498.0
187.15	4555.0
188.25	442.0
189.05	1295.0
190.15	205.0
191.05	546.0
192.05	1406.0
193.15	1362.0
194.15	601.0
195.05	359.0
196.15	2754.0
198.05	95064.0
199.05	6989.0
200.15	520.0
201.15	249.0
203.05	658.0
204.15	3794.0
205.15	6075.0
206.15	27544.0
207.05	5079.0
208.15	1159.0
209.05	457.0
210.15	478.0
210.95	1319.0
212.15	174.0
214.95	224.0
216.05	533.0
217.05	7361.0
218.15	1116.0
221.05	3772.0
223.05	1066.0
224.15	14640.0
225.05	3448.0
226.05	445.0
227.05	5851.0
228.05	978.0
229.05	1285.0
231.05	805.0
233.05	241.0
234.05	410.0
234.95	741.0
236.05	224.0
237.05	594.0
239.15	398.0
240.05	383.0
241.05	404.0
242.05	801.0
243.05	705.0
244.05	10649.0
245.15	1936.0
246.05	2834.0
247.15	830.0
248.15	308.0
248.85	225.0
249.15	252.0
252.05	262.0
253.15	410.0
255.05	57384.0
256.05	8788.0
257.15	941.0
258.05	3837.0
258.95	543.0
265.05	1317.0
266.95	173.0
271.15	177.0
272.15	225.0
273.15	1402.0
274.15	4641.0
275.05	24496.0
276.05	3469.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

277.05	3071.0
278.05	539.0
282.15	167.0
283.15	252.0
283.95	224.0
285.05	407.0
290.05	200.0
291.25	193.0
293.05	418.0
294.25	150.0
296.15	7564.0
297.15	792.0
298.15	290.0
301.05	195.0
303.05	743.0
304.15	201.0
310.15	291.0
314.15	273.0
315.05	740.0
316.05	477.0
321.15	432.0
323.15	2429.0
324.05	442.0
327.05	392.0
332.25	259.0
333.15	240.0
334.05	1438.0
335.15	411.0
341.15	274.0
345.95	487.0
347.05	185.0
352.15	749.0
353.05	731.0
354.25	647.0
365.05	3139.0
366.05	520.0
370.95	201.0
372.15	1246.0
373.05	211.0
382.95	211.0
384.35	212.0
402.05	422.0
403.15	547.0
403.95	325.0
415.25	164.0
420.95	328.0
422.15	310.0
423.15	3359.0
424.15	553.0
441.15	10093.0
442.15	64960.0
443.15	13263.0
444.15	1620.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

m/z	Abundance
38.25	576.0
39.15	2887.0
40.15	348.0
41.15	293.0
44.15	626.0
49.15	451.0
50.15	11367.0
51.15	46888.0
52.15	2412.0
53.15	176.0
55.05	285.0
56.05	1152.0
57.15	2903.0
58.05	198.0
61.15	781.0
62.15	311.0
63.15	2442.0
64.15	298.0
65.15	759.0
66.45	182.0
68.05	947.0
69.15	45768.0
73.15	532.0
74.15	4724.0
75.15	7524.0
76.15	2148.0
77.15	50624.0
78.15	3217.0
79.15	3482.0
80.05	2670.0
81.15	3684.0
82.15	1018.0
83.15	908.0
85.15	540.0
86.05	1066.0
87.15	640.0
88.15	223.0
91.15	882.0
92.05	921.0
93.15	5725.0
94.05	188.0
95.05	202.0
96.05	294.0
98.15	5251.0
99.05	3279.0
100.15	493.0
101.05	2627.0
103.15	732.0
104.15	1336.0
105.15	1081.0
106.15	609.0
107.15	16872.0
108.15	2561.0
109.25	920.0
110.05	28584.0
111.05	3558.0
112.05	852.0
115.35	153.0
116.15	904.0
117.15	12286.0
118.05	716.0
119.05	270.0
120.05	230.0
121.05	184.0
122.05	1211.0
123.05	1975.0
124.05	586.0
125.15	628.0
127.15	57008.0
128.15	4398.0
129.05	22336.0
130.15	1831.0
131.15	295.0
131.95	154.0
134.05	609.0
135.05	2035.0
136.05	483.0
137.15	810.0
139.25	348.0
139.55	365.0
140.15	304.0
141.15	3323.0
142.05	1060.0
142.95	599.0
144.05	376.0
145.05	296.0
146.15	612.0
147.05	1329.0
148.15	3078.0
148.95	513.0
150.15	388.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

150.95	217.0
151.35	405.0
151.95	193.0
153.05	872.0
154.05	1017.0
155.15	2303.0
156.05	2894.0
157.15	734.0
158.05	614.0
159.15	515.0
160.15	1117.0
161.05	1661.0
162.15	809.0
165.05	1398.0
166.05	605.0
167.05	5463.0
168.15	2566.0
169.15	550.0
170.05	255.0
172.05	608.0
173.05	725.0
174.15	1201.0
175.15	2161.0
176.15	561.0
177.05	1552.0
178.05	498.0
179.05	5391.0
180.05	3508.0
181.05	1200.0
182.05	279.0
183.05	183.0
184.05	383.0
185.05	2448.0
186.15	18576.0
187.15	5194.0
188.05	757.0
189.05	1515.0
190.15	343.0
191.15	852.0
192.05	1403.0
193.15	1573.0
194.15	572.0
196.05	3596.0
198.05	141824.0
199.05	9081.0
199.95	799.0
201.55	890.0
203.15	1129.0
204.05	5573.0
205.15	9326.0
206.15	39544.0
207.15	5995.0
208.15	1488.0
209.05	682.0
210.05	687.0
210.35	594.0
211.15	1337.0
211.95	392.0
213.05	244.0
214.05	181.0
214.95	525.0
216.05	870.0
217.05	10862.0
218.05	1287.0
218.75	181.0
221.05	6027.0
222.15	1287.0
223.05	2380.0
224.15	23600.0
225.15	5555.0
226.05	742.0
227.05	9355.0
228.05	1672.0
229.05	2181.0
230.05	281.0
231.05	860.0
233.05	216.0
234.15	925.0
235.05	851.0
236.05	514.0
237.05	746.0
238.05	226.0
239.05	274.0
240.15	189.0
241.05	578.0
242.15	1072.0
243.15	1433.0
244.15	16568.0
245.15	2293.0
246.05	3659.0
247.05	622.0
248.95	503.0
251.05	234.0
252.25	305.0
253.15	465.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

255.05	94264.0
256.15	13593.0
257.05	858.0
258.05	5170.0
259.05	1149.0
260.05	176.0
261.05	234.0
264.15	179.0
265.05	2197.0
266.15	372.0
269.55	180.0
270.05	285.0
271.05	284.0
272.05	601.0
273.05	2610.0
274.05	6975.0
275.05	43280.0
276.15	4855.0
277.05	3228.0
278.15	679.0
282.35	225.0
283.15	419.0
284.15	237.0
285.15	949.0
289.05	238.0
291.15	265.0
292.05	159.0
293.15	883.0
293.85	219.0
296.05	11518.0
297.15	1870.0
298.15	218.0
301.25	266.0
302.15	201.0
303.15	1664.0
304.15	458.0
310.25	155.0
314.15	541.0
315.05	1238.0
316.15	766.0
321.15	481.0
321.95	161.0
323.15	3741.0
324.05	932.0
325.15	298.0
326.15	208.0
327.05	847.0
328.15	199.0
329.15	165.0
332.15	398.0
333.15	299.0
334.15	1902.0
335.15	783.0
336.25	235.0
341.15	459.0
346.05	678.0
346.95	162.0
352.15	957.0
353.15	926.0
354.15	1030.0
355.25	221.0
359.05	152.0
365.05	5739.0
366.15	669.0
371.15	460.0
372.15	2234.0
373.15	496.0
383.05	792.0
389.85	211.0
391.05	173.0
402.15	710.0
403.05	982.0
403.85	229.0
420.95	972.0
422.25	712.0
423.15	5933.0
424.15	1220.0
425.35	202.0
429.25	176.0
441.15	17664.0
442.15	121224.0
443.15	23472.0
444.15	2177.0
444.95	208.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

m/z	Abundance
38.15	306.0
39.15	2111.0
41.15	249.0
44.15	795.0
45.05	197.0
49.15	326.0
50.15	9067.0
51.15	34504.0
52.15	2074.0
55.05	359.0
56.15	1035.0
57.15	2643.0
58.15	168.0
61.15	543.0
62.05	399.0
63.15	1531.0
64.05	206.0
65.05	799.0
69.05	31264.0
70.15	269.0
73.15	236.0
74.15	3877.0
75.15	6486.0
76.15	1800.0
77.15	39712.0
78.15	3227.0
79.05	2714.0
80.05	2393.0
81.15	2749.0
82.05	692.0
83.05	834.0
85.15	791.0
86.15	728.0
87.15	389.0
88.05	184.0
90.15	208.0
91.15	700.0
92.15	841.0
93.05	3910.0
94.05	400.0
95.15	174.0
96.15	245.0
98.05	3312.0
99.05	2824.0
99.95	195.0
101.05	1644.0
102.25	270.0
103.05	803.0
104.15	1217.0
105.05	1193.0
106.05	395.0
107.15	12341.0
108.15	2049.0
110.05	22688.0
111.05	4023.0
112.05	504.0
114.95	231.0
116.05	785.0
117.15	9712.0
118.15	874.0
120.15	172.0
122.05	745.0
123.05	1763.0
123.95	606.0
125.15	447.0
127.15	47184.0
128.15	3673.0
129.15	18544.0
130.05	1379.0
131.15	305.0
134.15	616.0
135.05	1328.0
136.05	626.0
137.05	815.0
138.05	190.0
139.15	182.0
140.15	248.0
141.05	2347.0
142.05	681.0
143.05	608.0
144.15	194.0
146.05	663.0
147.15	1304.0
148.05	2471.0
149.15	492.0
150.15	190.0
151.15	381.0
152.15	223.0
153.05	1145.0
153.95	520.0
155.05	1535.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

156.05	2648.0
157.15	429.0
158.05	597.0
159.25	289.0
160.05	851.0
161.15	1243.0
162.15	397.0
164.25	268.0
165.15	1246.0
166.15	932.0
167.15	4790.0
168.05	2831.0
169.15	698.0
170.85	284.0
171.05	278.0
172.05	332.0
173.15	807.0
174.15	1263.0
175.15	2814.0
176.15	560.0
177.05	1198.0
178.15	679.0
179.05	4352.0
180.15	3210.0
181.15	1688.0
182.05	273.0
184.15	515.0
185.05	2697.0
186.15	15378.0
187.15	5219.0
188.15	684.0
189.05	1180.0
190.15	286.0
191.05	466.0
192.05	1342.0
193.15	1854.0
194.25	471.0
195.05	166.0
196.15	3163.0
198.05	122960.0
199.05	9098.0
200.05	628.0
201.15	348.0
202.25	315.0
203.15	1359.0
204.15	5012.0
205.05	8894.0
206.15	35968.0
207.15	4441.0
208.05	1563.0
208.95	216.0
210.05	681.0
210.95	1266.0
212.05	345.0
213.05	202.0
215.05	506.0
216.15	923.0
217.05	10510.0
218.15	811.0
219.05	258.0
221.05	5240.0
222.05	1364.0
223.15	2450.0
224.15	20184.0
225.05	5364.0
227.05	9644.0
228.15	1374.0
229.15	1307.0
230.15	459.0
231.05	880.0
233.25	199.0
234.15	605.0
234.95	669.0
236.05	342.0
237.15	474.0
238.95	340.0
240.15	188.0
241.05	461.0
242.05	895.0
243.15	1491.0
244.15	14723.0
245.05	1626.0
246.05	2883.0
247.05	681.0
249.15	733.0
250.05	181.0
251.25	475.0
251.85	161.0
253.05	608.0
255.05	88144.0
256.05	11198.0
257.05	991.0
258.05	4802.0
259.15	799.0
264.15	290.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53

265.05	2336.0
266.15	176.0
271.05	348.0
271.95	360.0
272.25	374.0
273.05	2589.0
274.05	6844.0
275.05	37472.0
276.15	4827.0
277.05	2973.0
278.05	540.0
282.05	154.0
283.05	569.0
284.05	269.0
285.05	581.0
289.15	244.0
290.05	244.0
291.05	294.0
293.05	734.0
294.05	305.0
296.05	11638.0
297.15	1593.0
298.15	192.0
301.15	262.0
303.15	1601.0
304.15	375.0
308.05	208.0
309.95	186.0
314.15	518.0
315.05	1396.0
316.15	1204.0
321.15	196.0
322.05	242.0
323.15	3642.0
324.05	644.0
327.05	434.0
328.05	420.0
332.15	269.0
333.15	501.0
334.15	2403.0
335.05	759.0
341.05	408.0
346.05	706.0
347.05	217.0
352.15	1281.0
353.05	638.0
354.15	934.0
355.05	246.0
365.05	5371.0
366.05	839.0
371.05	352.0
372.15	2022.0
373.05	602.0
383.05	512.0
384.25	170.0
390.15	271.0
401.35	211.0
402.25	999.0
403.15	1285.0
404.05	485.0
421.15	1144.0
422.05	830.0
423.15	7060.0
424.15	1178.0
441.15	20056.0
442.15	133888.0
443.15	25432.0
444.15	2638.0

Instrument :

BNA_G

ClientSampleId :

DFTPP53

m/z	Abundance
43.15	200.0
44.15	542.0
49.05	286.0
55.15	247.0
81.05	181.0
91.15	153.0
157.95	152.0
207.05	639.0
281.15	189.0

Instrument :
BNA_G
ClientSampleId :
DFTPP53