

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101117\
 Data File : BG029152.D
 Acq On : 11 Oct 2017 00:01
 Operator : SJ/JU
 Sample : SSTD02070
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02070

Quant Time: Oct 11 04:04:27 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG101117.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 11 04:00:54 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.22	152	43013	20.00	ng/ul	0.00
2) Naphthalene-d8	11.03	136	192389	20.00	ng/ul	0.00
6) Acenaphthene-d10	14.83	164	131104	20.00	ng/ul	0.00
12) Phenanthrene-d10	17.56	188	307251	20.00	ng/ul	0.00
18) Chrysene-d12	21.85	240	326752	20.00	ng/ul	0.00
23) Perylene-d12	25.20	264	326994	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 2,4-Dichlorophenol-d3	10.65	165	77211	21.96	ng/ul	0.00
7) Acenaphthylene-d8	14.53	160	284427	21.23	ng/ul	0.00
10) Fluorene-d10	15.81	176	223986	21.06	ng/ul	0.00
15) Anthracene-d10	17.66	188	325259	21.48	ng/ul	0.00
19) Pyrene-d10	19.94	212	379368	22.15	ng/ul	0.00
26) Benzo(a)pyrene-d12	24.97	264	352161	22.27	ng/ul	0.00
Target Compounds				Qvalue		
4) Naphthalene	11.09	128	217810	22.195	ng/ul	100
5) 2-Methylnaphthalene	12.68	142	166225	21.312	ng/ul	100
8) Acenaphthylene	14.55	152	272926	21.308	ng/ul	100
9) Acenaphthene	14.89	153	193744	22.256	ng/ul	100
11) Fluorene	15.87	166	224007	20.298	ng/ul	100
13) Pentachlorophenol	17.21	266	53291	26.483	ng/ul	100
14) Phenanthrene	17.61	178	354903	22.241	ng/ul	100
16) Anthracene	17.70	178	370545	22.564	ng/ul	100
17) Fluoranthene	19.60	202	430371	22.252	ng/ul	100
20) Pyrene	19.97	202	440209	22.047	ng/ul	100
21) Benzo(a)anthracene	21.83	228	441376	22.574	ng/ul	100
22) Chrysene	21.89	228	397898	22.638	ng/ul	100
24) Benzo(b)fluoranthene	24.13	252	443016	21.945	ng/ul	100
25) Benzo(k)fluoranthene	24.20	252	431301	22.056	ng/ul	100
27) Benzo(a)pyrene	25.04	252	423268	21.765	ng/ul	100
28) Indeno(1,2,3-cd)pyrene	29.04	276	401485	18.305	ng/ul	100
29) Dibenzo(a,h)anthracene	29.10	278	352407	19.034	ng/ul	100
30) Benzo(g,h,i)perylene	30.24	276	286531	16.327	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101117\
Data File : BG029152.D
Acq On : 11 Oct 2017 00:01
Operator : SJ/JU
Sample : SSTD02070
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTD02070

Quant Time: Oct 11 04:04:27 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG101117.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 11 04:00:54 2017
Response via : Initial Calibration

