

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG101118\
 Data File : BG037276.D
 Acq On : 12 Oct 2018 4:28
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 12 05:03:58 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 11 13:59:58 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	0.638	0.599	6.1	125	0.00
3	Pyridine	1.515	1.526	-0.7	124	0.00
4	n-Nitrosodimethylamine	0.588	0.585	0.5	128	0.00
5 S	2-Fluorophenol	1.136	1.131	0.4	125	0.00
6	Aniline	2.258	2.215	1.9	124	0.00
7 S	Phenol-d6	1.725	1.713	0.7	123	0.00
8	2-Chlorophenol	1.330	1.360	-2.3	125	0.00
9	Benzaldehyde	1.188	1.124	5.4	119	0.00
10 C	Phenol	1.817	1.811	0.3	124	0.00
11	bis(2-Chloroethyl)ether	1.484	1.458	1.8	126	0.00
12	1,3-Dichlorobenzene	1.564	1.501	4.0	124	0.00
13 C	1,4-Dichlorobenzene	1.583	1.549	2.1	126	0.00
14	1,2-Dichlorobenzene	1.534	1.474	3.9	124	0.00
15	Benzyl Alcohol	1.349	1.382	-2.4	126	0.00
16	2,2'-oxybis(1-Chloropropane	2.934	2.817	4.0	122	0.00
17	2-Methylphenol	1.217	1.225	-0.7	124	0.00
18	Hexachloroethane	0.540	0.529	2.0	123	0.00
19 P	n-Nitroso-di-n-propylamine	1.306	1.286	1.5	124	0.00
20	3+4-Methylphenols	1.701	1.748	-2.8	125	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
22	Acetophenone	0.504	0.490	2.8	123	0.00
23 S	Nitrobenzene-d5	0.304	0.336	-10.5	131	0.00
24	Nitrobenzene	0.326	0.353	-8.3	129	0.00
25	Isophorone	0.694	0.679	2.2	121	0.00
26 C	2-Nitrophenol	0.116	0.128	-10.3	141	0.00
27	2,4-Dimethylphenol	0.290	0.286	1.4	123	0.00
28	bis(2-Chloroethoxy)methane	0.434	0.429	1.2	123	0.00
29 C	2,4-Dichlorophenol	0.295	0.311	-5.4	124	0.00
30	1,2,4-Trichlorobenzene	0.359	0.351	2.2	125	0.00
31	Naphthalene	0.972	0.949	2.4	124	0.00
32	Benzoic acid	0.160	0.189	-18.1	146	0.01
33	4-Chloroaniline	0.456	0.447	2.0	120	0.00
34 C	Hexachlorobutadiene	0.223	0.210	5.8	121	0.00
35	Caprolactam	0.120	0.130	-8.3	126	0.00
36 C	4-Chloro-3-methylphenol	0.348	0.361	-3.7	123	0.00
37	2-Methylnaphthalene	0.709	0.701	1.1	125	0.00
38	1-Methylnaphthalene	0.693	0.676	2.5	122	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
40	1,2,4,5-Tetrachlorobenzene	0.648	0.621	4.2	121	0.00
41 P	Hexachlorocyclopentadiene	0.267	0.272	-1.9	126	0.00
42 S	2,4,6-Tribromophenol	0.196	0.208	-6.1	124	0.00
43 C	2,4,6-Trichlorophenol	0.376	0.407	-8.2	127	0.00
44	2,4,5-Trichlorophenol	0.404	0.432	-6.9	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG101118\
 Data File : BG037276.D
 Acq On : 12 Oct 2018 4:28
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 12 05:03:58 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 11 13:59:58 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.332	1.270	4.7	121	0.00
46	1,1'-Biphenyl	1.645	1.604	2.5	122	0.00
47	2-Chloronaphthalene	1.243	1.203	3.2	122	0.00
48	2-Nitroaniline	0.300	0.350	-16.7	136	0.00
49	Acenaphthylene	1.901	1.847	2.8	121	0.00
50	Dimethylphthalate	1.572	1.543	1.8	120	0.00
51	2,6-Dinitrotoluene	0.284	0.313	-10.2	132	0.00
52 C	Acenaphthene	1.174	1.153	1.8	122	0.00
53	3-Nitroaniline	0.329	0.353	-7.3	129	0.00
54 P	2,4-Dinitrophenol	0.079	0.075	5.1	121	0.00
55	Dibenzofuran	1.905	1.814	4.8	119	0.00
56 P	4-Nitrophenol	0.256	0.271	-5.9	126	0.00
57	2,4-Dinitrotoluene	0.356	0.398	-11.8	134	0.00
58	Fluorene	1.441	1.370	4.9	120	0.00
59	2,3,4,6-Tetrachlorophenol	0.358	0.380	-6.1	121	0.00
60	Diethylphthalate	1.629	1.600	1.8	119	0.00
61	4-Chlorophenyl-phenylether	0.842	0.808	4.0	120	0.00
62	4-Nitroaniline	0.345	0.367	-6.4	128	0.00
63	Azobenzene	1.562	1.490	4.6	120	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	124	0.00
65	4,6-Dinitro-2-methylphenol	0.056	0.060	-7.1	126	0.00
66 c	n-Nitrosodiphenylamine	0.587	0.578	1.5	119	0.00
67	4-Bromophenyl-phenylether	0.216	0.210	2.8	118	0.00
68	Hexachlorobenzene	0.224	0.214	4.5	117	0.00
69	Atrazine	0.218	0.219	-0.5	117	0.00
70 C	Pentachlorophenol	0.144	0.148	-2.8	122	0.00
71	Phenanthrene	1.018	0.967	5.0	117	0.00
72	Anthracene	0.996	0.969	2.7	120	0.00
73	Carbazole	1.027	0.993	3.3	118	0.00
74	Di-n-butylphthalate	1.100	1.129	-2.6	118	0.00
75 C	Fluoranthene	1.190	1.132	4.9	116	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	121	0.00
77	Benzidine	0.765	0.676	11.6	96	0.00
78	Pyrene	1.303	1.272	2.4	117	0.00
79 S	Terphenyl-d14	0.952	0.895	6.0	113	0.00
80	Butylbenzylphthalate	0.507	0.537	-5.9	123	0.00
81	Benzo(a)anthracene	1.198	1.165	2.8	117	0.00
82	3,3'-Dichlorobenzidine	0.462	0.469	-1.5	117	0.00
83	Chrysene	1.143	1.105	3.3	117	0.00
84	Bis(2-ethylhexyl)phthalate	0.728	0.761	-4.5	121	0.00
85 c	Di-n-octyl phthalate	1.182	1.256	-6.3	123	0.00
86	Indeno(1,2,3-cd)pyrene	1.266	1.303	-2.9	121	-0.01
87 I	Perylene-d12	1.000	1.000	0.0	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG101118\
Data File : BG037276.D
Acq On : 12 Oct 2018 4:28
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 12 05:03:58 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 11 13:59:58 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.087	1.035	4.8	118	0.00
89	Benzo(k)fluoranthene	1.073	1.044	2.7	120	0.00
90 C	Benzo(a)pyrene	1.043	1.014	2.8	120	0.00
91	Dibenzo(a,h)anthracene	0.998	0.972	2.6	120	0.00
92	Benzo(g,h,i)perylene	0.996	0.973	2.3	120	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0