




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101121\
Data File : BG050530.D
Acq On : 11 Oct 2021 8:52
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 11 11:11:57 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 06 15:51:01 2021
Response via : Initial Calibration



TIC: BG050530.D\data.ms

(70) Pentachlorophenol (C)




17.266min (+ 0.007) 25289.66 ng

response 68451

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.20	61.08
264.00	61.30	63.59
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101121\
Data File : BG050530.D
Acq On : 11 Oct 2021 8:52
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 11 11:11:57 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 06 15:51:01 2021
Response via : Initial Calibration



TIC: BG050530.D\data.ms

(77) Benzidine

19.840min (+ 0.007) 280898.07 ng

response 537150

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.40	14.99
183.00	11.70	12.68
0.00	0.00	0.00