

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101215\
 Data File : BG019163.D
 Acq On : 12 Oct 2015 15:34
 Operator : UM/NP
 Sample : G3989-06
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 OW1-C8-GW

Quant Time: Oct 13 01:37:20 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 10 07:04:31 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.03	152	19452	20.00	ng	0.00
21) Naphthalene-d8	10.84	136	89853	20.00	ng	0.00
38) Acenaphthene-d10	14.66	164	62153	20.00	ng	0.00
63) Phenanthrene-d10	17.41	188	148460	20.00	ng	0.00
75) Chrysene-d12	21.68	240	173909	20.00	ng	0.00
86) Perylene-d12	24.91	264	171197	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	72195	75.57	ng	0.00
7) Phenol-d6	7.20	99	73296	49.94	ng	0.00
23) Nitrobenzene-d5	9.20	82	171106	105.22	ng	0.00
41) 2,4,6-Tribromophenol	16.15	330	135267	159.71	ng	0.00
44) 2-Fluorobiphenyl	13.28	172	415441	102.78	ng	0.00
78) Terphenyl-d14	20.02	244	586528	89.09	ng	0.00
Target Compounds						
49) Dimethylphthalate	14.11	163	21336	4.35	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101215\
Data File : BG019163.D
Acq On : 12 Oct 2015 15:34
Operator : UM/NP
Sample : G3989-06
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
OW1-C8-GW

Quant Time: Oct 13 01:37:20 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 10 07:04:31 2015
Response via : Initial Calibration

