

Data Path : Z:\HPCHEM1\BNA_G\Data\BG101217\
 Data File : BG029168.D
 Acq On : 11 Oct 2017 16:38
 Operator : SJ/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02078

Quant Time: Oct 11 18:32:09 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG100717MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 11 04:15:46 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	66	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	62	0.00
3	Naphthalene	1.041	1.131	-8.6	65	0.00
4	2-Methylnaphthalene	0.790	0.868	-9.9	66	0.00
5	1-Methylnaphthalene	0.754	0.817	-8.4	66	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	65	0.00
7 S	Acenaphthylene-d8	2.080	2.175	-4.6	65	0.00
8	Acenaphthylene	2.003	2.031	-1.4	63	0.00
9 C	Acenaphthene	1.363	1.457	-6.9	67	0.00
10 S	Fluorene-d10	1.509	1.685	-11.7	69	0.00
11	Fluorene	1.595	1.667	-4.5	65	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	61	0.00
13	Phenanthrene	1.067	1.149	-7.7	64	0.00
14 S	Anthracene-d10	0.969	1.061	-9.5	65	0.00
15	Anthracene	1.091	1.180	-8.2	64	0.00
16 C	Fluoranthene	1.251	1.416	-13.2	66	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	66	0.00
18 S	Pyrene-d10	1.023	1.159	-13.3	70	0.00
19	Pyrene	1.247	1.327	-6.4	65	0.00
20	Benzo(a)anthracene	1.178	1.345	-14.2	70	0.00
21	Chrysene	1.079	1.228	-13.8	70	0.00
22 I	Perylene-d12	1.000	1.000	0.0	67	0.00
23	Benzo(b)fluoranthene	1.196	1.271	-6.3	69	0.00
24	Benzo(k)fluoranthene	1.168	1.313	-12.4	73	0.00
25 S	Benzo(a)pyrene-d12	0.958	1.064	-11.1	73	0.00
26 C	Benzo(a)pyrene	1.162	1.278	-10.0	71	0.00
27	Indeno(1,2,3-cd)pyrene	1.319	1.174	11.0	58	0.00
28	Dibenzo(a,h)anthracene	1.092	1.061	2.8	64	0.00
29	Benzo(g,h,i)perylene	1.088	0.870	20.0#	52	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0