

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101316\
 Data File : BG024322.D
 Acq On : 13 Oct 2016 16:00
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02023

Quant Time: Oct 13 16:59:52 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 01:34:23 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2	1,4-Dioxane	0.456	0.490	-7.5	127	0.00
3 S	1,4-Dioxane-d8	0.421	0.431	-2.4	129	0.00
4	Benzaldehyde	1.313	1.442	-9.8	115	0.00
5 S	Phenol-d5	1.866	1.949	-4.4	120	0.00
6	Phenol	1.894	2.000	-5.6	120	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.261	1.265	-0.3	117	0.00
8	Bis(2-Chloroethyl)ether	1.394	1.483	-6.4	119	0.00
9 S	2-Chlorophenol-d4	1.285	1.375	-7.0	121	0.00
10	2-Chlorophenol	1.302	1.331	-2.2	118	0.00
11	2-Methylphenol	1.403	1.466	-4.5	121	0.00
12	2,2'-oxybis(1-Chloropropane	2.784	2.667	4.2	107	0.00
13 S	4-Methylphenol-d8	1.509	1.583	-4.9	120	0.00
14	Acetophenone	2.240	2.407	-7.5	119	0.00
15 P	N-Nitroso-di-n-propylamine	1.351	1.390	-2.9	115	0.00
16	4-Methylphenol	1.488	1.646	-10.6	128	0.00
17	Hexachloroethane	0.602	0.597	0.8	117	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
19 S	Nitrobenzene-d5	0.142	0.157	-10.6	126	0.00
20	Nitrobenzene	0.449	0.484	-7.8	122	0.00
21	Isophorone	0.816	0.847	-3.8	114	0.00
22 S	2-Nitrophenol-d4	0.159	0.184	-15.7	136	0.00
23 C	2-Nitrophenol	0.173	0.193	-11.6	126	0.00
24	2,4-Dimethylphenol	0.426	0.454	-6.6	120	0.00
25	Bis(2-Chloroethoxy)methane	0.448	0.465	-3.8	115	0.00
26 S	2,4-Dichlorophenol-d3	0.346	0.359	-3.8	117	0.00
27 C	2,4-Dichlorophenol	0.332	0.355	-6.9	122	0.00
28	Naphthalene	0.963	1.006	-4.5	119	0.00
29 S	4-Chloroaniline-d4	0.363	0.409	-12.7	123	0.00
30	4-Chloroaniline	0.369	0.418	-13.3	125	0.00
31 C	Hexachlorobutadiene	0.266	0.285	-7.1	121	0.00
32	Caprolactam	0.126	0.124	1.6	113	0.00
33 C	4-Chloro-3-methylphenol	0.406	0.421	-3.7	119	0.00
34	2-Methylnaphthalene	0.766	0.811	-5.9	124	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
36	1,2,4,5-Tetrachlorobenzene	0.590	0.688	-16.6	123	0.00
37	Hexachlorocyclopentadiene	0.432	0.446	-3.2	110	0.00
38 C	2,4,6-Trichlorophenol	0.382	0.418	-9.4	117	0.00
39	2,4,5-Trichlorophenol	0.420	0.462	-10.0	117	0.00
40	1,1'-Biphenyl	1.295	1.415	-9.3	114	0.00
41	2-Chloronaphthalene	1.015	1.118	-10.1	117	0.00
42	2-Nitroaniline	0.386	0.430	-11.4	116	0.00
43 S	Dimethylphthalate-d6	1.396	1.416	-1.4	108	0.00
44	Dimethylphthalate	1.369	1.418	-3.6	109	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101316\
 Data File : BG024322.D
 Acq On : 13 Oct 2016 16:00
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02023

Quant Time: Oct 13 16:59:52 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 13 01:34:23 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.266	0.305	-14.7	119	0.00
46 S	Acenaphthylene-d8	1.600	1.750	-9.4	114	0.00
47	Acenaphthylene	1.539	1.663	-8.1	114	0.00
48	3-Nitroaniline	0.251	0.265	-5.6	112	0.00
49 C	Acenaphthene	1.100	1.145	-4.1	112	0.00
50	2,4-Dinitrophenol	0.178	0.187	-5.1	119	0.00
51 S	4-Nitrophenol-d4	0.248	0.242	2.4	105	0.00
52	4-Nitrophenol	0.300	0.285	5.0	99	0.00
53	Dibenzofuran	1.619	1.711	-5.7	111	0.00
54	2,4-Dinitrotoluene	0.394	0.418	-6.1	111	0.00
55	2,3,4,6-Tetrachlorophenol	0.413	0.449	-8.7	115	0.00
56	Diethylphthalate	1.443	1.422	1.5	102	0.00
57 S	Fluorene-d10	1.320	1.393	-5.5	109	0.00
58	Fluorene	1.339	1.332	0.5	104	0.00
59	4-Chlorophenyl-phenylether	0.743	0.765	-3.0	109	0.00
60	4-Nitroaniline	0.289	0.277	4.2	102	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.120	0.138	-15.0	119	0.00
63	4,6-Dinitro-2-methylphenol	0.126	0.141	-11.9	109	0.00
64	N-Nitrosodiphenylamine	0.566	0.608	-7.4	102	0.00
65	4-Bromophenyl-phenylether	0.232	0.259	-11.6	110	0.00
66	Hexachlorobenzene	0.258	0.286	-10.9	107	0.00
67	Atrazine	0.231	0.254	-10.0	105	0.00
68 C	Pentachlorophenol	0.157	0.179	-14.0	111	0.00
69	Phenanthrene	1.010	1.060	-5.0	99	0.00
70 S	Anthracene-d10	0.903	0.972	-7.6	101	0.00
71	Anthracene	1.028	1.101	-7.1	101	0.00
72	Carbazole	0.807	0.906	-12.3	96	0.00
73	Di-n-butylphthalate	0.986	1.112	-12.8	105	0.00
74 C	Fluoranthene	1.020	1.194	-17.1	97	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
76 S	Pyrene-d10	0.933	0.991	-6.2	97	0.00
77	Pyrene	1.068	1.133	-6.1	97	0.00
78	Butylbenzylphthalate	0.429	0.481	-12.1	105	0.00
79	3,3'-Dichlorobenzidine	0.372	0.443	-19.1	108	0.00
80	Benzo(a)anthracene	1.110	1.199	-8.0	100	0.00
81	Bis(2-ethylhexyl)phthalate	0.607	0.658	-8.4	100	0.00
82	Chrysene	1.057	1.128	-6.7	98	0.00
83 I	Perylene-d12	1.000	1.000	0.0	102	0.00
84	Di-n-octyl phthalate	0.938	1.098	-17.1	103	0.00
85	Benzo(b)fluoranthene	1.110	1.183	-6.6	103	0.00
86	Benzo(k)fluoranthene	1.070	1.116	-4.3	100	0.00
87 S	Benzo(a)pyrene-d12	0.900	0.960	-6.7	102	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101316\
Data File : BG024322.D
Acq On : 13 Oct 2016 16:00
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02023

Quant Time: Oct 13 16:59:52 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 13 01:34:23 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.058	1.137	-7.5	103	0.00
89	Indeno(1,2,3-cd)pyrene	1.253	1.420	-13.3	109	0.00
90	Dibenzo(a,h)anthracene	1.059	1.176	-11.0	107	0.01
91	Benzo(a,h,i)perylene	1.053	1.141	-8.4	106	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0