


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101321\
 Data File : BG050563.D
 Acq On : 13 Oct 2021 9:01
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Oct 14 10:56:22 2021






AutoFind: Scans 2473, 2474, 2475; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.9	48296	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	49.6	48997	PASS
70	69	0.00	2	0.5	263	PASS
127	198	10	80	53.0	52344	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	98848	PASS
199	198	5	9	7.1	6969	PASS
275	198	10	60	26.7	26352	PASS
365	198	1	100	3.6	3603	PASS
441	198	0.01	100	8.3	8173	PASS
442	442	50	100	100.0	53707	PASS
443	442	15	24	19.4	10441	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101321\
Data File : BG050563.D
Acq On : 13 Oct 2021 9:01
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 14 04:17:18 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 06 15:51:01 2021
Response via : Initial Calibration



TIC: BG050563.D\data.ms

(70) Pentachlorophenol (C)

17.261min (+ 0.002) 18283.16 ng




response 101591

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.20	62.12
264.00	61.30	62.98
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101321\
Data File : BG050563.D
Acq On : 13 Oct 2021 9:01
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 14 04:17:18 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 06 15:51:01 2021
Response via : Initial Calibration



TIC: BG050563.D\data.ms

(77) Benzidine

19.834min (+ 0.002) 0.00 ng

response 743183

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.40	15.60
183.00	11.70	12.26
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/13/2021	BNA_G	BG050563.D
Compound Name	Response	Retention Time
DDT	337356	21.068
DDD	3371	20.668
DDE	175	20.116
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3546	340902	1.04