


Data Path : Z:\HPCHEM1\BNA G\Data\BG101415\
Data File : BG019180.D
Acq On : 13 Oct 2015 12:56
Operator : UM/NP
Sample : DFTPP01
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP01

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG100415.M
Title : SVOA CALIBRATION
Last Update : Wed Oct 14 04:49:14 2015



AutoFind: Scans 2552, 2553, 2554; Background Corrected with Scan 2548

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	49.0	8262	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	41.1	6926	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.9	8245	PASS
197	198	0.00	2	0.4	72	PASS
198	198	100	100	100.0	16857	PASS
199	198	5	9	7.2	1208	PASS
275	198	10	60	30.7	5177	PASS
365	198	1	100	5.1	861	PASS
441	442	0.01	24	16.2	2542	PASS
442	198	50	100	93.1	15696	PASS
443	442	15	24	20.3	3186	PASS

Scan 2548 (17.811 min): BG019180.D
DFTPP01

m/z	Abundance
40.00	197.0
44.00	517.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

m/z	Abundance
39.10	506.0
40.10	364.0
44.10	480.0
50.10	1969.0
51.10	6361.0
52.00	476.0
56.00	190.0
57.00	557.0
62.10	192.0
63.10	219.0
69.00	6023.0
74.00	624.0
75.00	990.0
76.00	280.0
77.10	6178.0
78.10	276.0
79.00	547.0
80.00	311.0
81.10	496.0
82.10	159.0
90.90	196.0
92.00	153.0
93.00	601.0
98.00	788.0
99.10	382.0
101.00	255.0
103.90	190.0
107.00	2067.0
108.10	458.0
110.00	3356.0
111.10	526.0
116.10	179.0
117.00	1770.0
123.00	240.0
127.00	5991.0
128.10	646.0
129.00	2297.0
135.00	293.0
141.00	371.0
147.00	246.0
147.90	386.0
155.00	194.0
156.00	226.0
157.10	209.0
159.00	152.0
161.10	183.0
167.00	714.0
168.00	437.0
174.10	228.0
175.00	385.0
176.00	177.0
177.00	179.0
177.20	178.0
179.00	625.0
180.00	238.0
185.10	311.0
186.00	2046.0
187.00	598.0
191.90	240.0
193.00	190.0
195.90	362.0
196.70	217.0
198.00	11829.0
198.90	1007.0
203.10	185.0
204.00	628.0
205.00	1014.0
206.00	3540.0
207.00	657.0
208.00	237.0
217.00	1299.0
217.90	192.0
221.10	412.0
222.10	164.0
223.00	328.0
224.00	2309.0
225.00	498.0
227.00	955.0
227.90	191.0
229.00	151.0
244.00	1536.0
245.00	259.0
246.00	301.0
255.00	8685.0
256.00	1078.0
258.00	608.0
265.00	195.0
272.90	293.0
274.00	574.0
275.00	3522.0
276.00	592.0
277.00	428.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

296.00	1144.0
297.00	231.0
303.00	177.0
323.00	341.0
334.00	241.0
365.00	548.0
403.00	198.0
422.10	152.0
423.00	410.0
441.00	1294.0
442.00	8354.0
443.00	1586.0
444.00	174.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

m/z	Abundance
39.10	635.0
40.00	259.0
44.10	296.0
49.00	154.0
50.10	2439.0
51.10	10604.0
52.10	556.0
56.00	379.0
57.10	715.0
63.10	486.0
65.00	223.0
69.00	8532.0
74.10	757.0
75.10	1488.0
76.00	505.0
77.10	9678.0
78.10	509.0
79.00	737.0
80.00	480.0
81.00	886.0
82.00	191.0
86.00	231.0
91.00	196.0
92.00	225.0
93.00	1133.0
98.00	911.0
99.00	629.0
100.90	384.0
103.10	198.0
104.00	304.0
105.00	264.0
107.00	3256.0
108.00	490.0
109.10	151.0
110.00	4830.0
111.00	696.0
116.00	214.0
117.00	2409.0
122.10	180.0
123.10	299.0
123.90	158.0
125.10	159.0
127.00	10273.0
128.00	915.0
129.00	4276.0
130.00	272.0
133.90	279.0
134.90	430.0
136.00	169.0
137.10	198.0
141.00	479.0
141.90	212.0
143.00	171.0
146.90	283.0
148.00	925.0
148.90	170.0
153.00	244.0
153.90	174.0
155.10	431.0
156.00	549.0
160.00	224.0
161.00	411.0
165.00	314.0
166.00	352.0
167.00	899.0
168.00	646.0
168.90	166.0
173.10	224.0
174.00	199.0
175.00	491.0
176.10	168.0
179.00	971.0
180.00	581.0
181.00	306.0
185.00	485.0
186.00	3602.0
187.10	946.0
188.00	211.0
189.00	204.0
192.00	297.0
193.00	356.0
196.00	545.0
198.00	20072.0
199.00	1393.0
202.90	237.0
204.00	941.0
205.00	1620.0
206.00	5921.0
207.00	924.0
207.90	300.0
211.00	226.0
216.00	180.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

217.00	1484.0
221.00	528.0
221.90	291.0
223.00	555.0
224.00	3393.0
225.00	901.0
227.00	1370.0
228.10	265.0
229.00	224.0
235.00	176.0
236.90	172.0
242.00	230.0
242.90	245.0
244.10	3058.0
245.10	384.0
246.00	794.0
254.10	306.0
255.00	14554.0
256.00	2039.0
256.90	181.0
258.00	1043.0
259.00	194.0
265.00	489.0
273.10	538.0
274.00	1096.0
275.00	5921.0
276.00	824.0
277.00	645.0
296.00	2303.0
296.90	352.0
303.00	290.0
314.10	160.0
314.90	220.0
323.00	652.0
324.00	174.0
334.00	474.0
335.10	153.0
352.00	166.0
354.00	192.0
365.00	915.0
366.00	247.0
372.00	247.0
401.90	158.0
403.00	196.0
421.10	224.0
422.00	311.0
423.00	1002.0
424.00	268.0
441.00	2887.0
442.00	16520.0
443.10	3386.0
443.90	322.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

m/z	Abundance
39.10	385.0
40.00	293.0
44.00	407.0
50.10	2093.0
51.10	7821.0
52.10	495.0
56.10	193.0
57.10	426.0
63.10	364.0
65.00	187.0
69.00	6225.0
74.10	753.0
75.10	1025.0
76.00	343.0
77.10	7634.0
78.10	547.0
79.10	495.0
80.00	485.0
81.00	420.0
82.10	171.0
86.10	187.0
91.90	191.0
93.00	907.0
98.00	789.0
99.00	649.0
101.10	250.0
103.00	181.0
104.00	352.0
105.00	254.0
107.00	2481.0
108.00	407.0
110.00	4093.0
111.10	639.0
117.00	2515.0
118.10	271.0
121.80	162.0
123.00	432.0
127.00	8471.0
128.00	531.0
129.00	3748.0
130.10	200.0
135.10	205.0
137.10	173.0
140.90	440.0
147.00	213.0
148.00	752.0
149.00	151.0
153.00	267.0
153.90	180.0
155.00	316.0
156.00	355.0
158.00	203.0
160.00	415.0
160.90	305.0
165.00	156.0
166.10	166.0
167.00	828.0
168.00	414.0
174.00	195.0
175.00	461.0
176.10	197.0
179.00	919.0
180.10	547.0
180.90	277.0
185.00	331.0
186.10	3265.0
187.00	1070.0
189.00	297.0
192.00	384.0
192.90	298.0
196.00	351.0
198.00	18672.0
198.90	1224.0
203.00	161.0
204.00	943.0
205.00	1471.0
206.00	5321.0
207.00	851.0
208.00	304.0
211.00	311.0
215.90	167.0
217.00	1785.0
218.10	198.0
221.10	738.0
222.20	166.0
223.00	394.0
224.00	3785.0
225.00	1154.0
227.00	1792.0
227.90	329.0
229.00	366.0
242.10	187.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01

244.00	2477.0
245.10	346.0
246.00	647.0
247.00	211.0
253.00	203.0
253.90	182.0
255.00	14420.0
256.00	2260.0
258.00	925.0
259.00	178.0
265.00	382.0
273.00	429.0
274.00	1168.0
275.00	6090.0
276.10	788.0
276.90	791.0
285.00	222.0
296.00	2106.0
297.10	370.0
303.00	266.0
314.90	271.0
323.00	780.0
324.10	164.0
334.00	451.0
346.00	163.0
352.00	202.0
354.10	192.0
365.00	1120.0
366.00	165.0
371.90	402.0
382.90	190.0
402.00	201.0
403.00	195.0
420.90	207.0
422.00	198.0
423.00	991.0
424.00	267.0
441.00	3446.0
442.00	22216.0
443.00	4588.0
444.00	499.0

Instrument :
BNA_G
ClientSampleId :
DFTPP01