

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101416\
 Data File : BG024369.D
 Acq On : 15 Oct 2016 2:07
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02028

Quant Time: Oct 15 05:07:46 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 14 23:22:45 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2	1,4-Dioxane	0.456	0.393	13.8	96	0.00
3 S	1,4-Dioxane-d8	0.421	0.389	7.6	110	0.00
4	Benzaldehyde	1.313	1.418	-8.0	107	0.00
5 S	Phenol-d5	1.866	1.778	4.7	103	0.00
6	Phenol	1.894	1.873	1.1	106	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.261	1.151	8.7	100	0.00
8	Bis(2-Chloroethyl)ether	1.394	1.388	0.4	105	0.00
9 S	2-Chlorophenol-d4	1.285	1.276	0.7	105	0.00
10	2-Chlorophenol	1.302	1.223	6.1	102	0.00
11	2-Methylphenol	1.403	1.417	-1.0	110	0.00
12	2,2'-oxybis(1-Chloropropane	2.784	2.513	9.7	95	0.00
13 S	4-Methylphenol-d8	1.509	1.561	-3.4	111	0.00
14	Acetophenone	2.240	2.290	-2.2	107	0.00
15 P	N-Nitroso-di-n-propylamine	1.351	1.338	1.0	104	0.00
16	4-Methylphenol	1.488	1.519	-2.1	111	0.00
17	Hexachloroethane	0.602	0.581	3.5	108	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
19 S	Nitrobenzene-d5	0.142	0.157	-10.6	122	0.00
20	Nitrobenzene	0.449	0.443	1.3	108	0.00
21	Isophorone	0.816	0.833	-2.1	108	0.00
22 S	2-Nitrophenol-d4	0.159	0.171	-7.5	122	0.00
23 C	2-Nitrophenol	0.173	0.183	-5.8	116	0.00
24	2,4-Dimethylphenol	0.426	0.421	1.2	107	0.00
25	Bis(2-Chloroethoxy)methane	0.448	0.441	1.6	105	0.00
26 S	2,4-Dichlorophenol-d3	0.346	0.356	-2.9	112	0.00
27 C	2,4-Dichlorophenol	0.332	0.346	-4.2	114	0.00
28	Naphthalene	0.963	0.969	-0.6	111	0.00
29 S	4-Chloroaniline-d4	0.363	0.424	-16.8	124	0.00
30	4-Chloroaniline	0.369	0.419	-13.6	121	0.00
31 C	Hexachlorobutadiene	0.266	0.285	-7.1	117	0.00
32	Caprolactam	0.126	0.126	0.0	110	0.00
33 C	4-Chloro-3-methylphenol	0.406	0.426	-4.9	117	0.00
34	2-Methylnaphthalene	0.766	0.788	-2.9	116	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	122	0.00
36	1,2,4,5-Tetrachlorobenzene	0.590	0.641	-8.6	123	0.00
37	Hexachlorocyclopentadiene	0.432	0.413	4.4	110	0.00
38 C	2,4,6-Trichlorophenol	0.382	0.397	-3.9	120	0.00
39	2,4,5-Trichlorophenol	0.420	0.426	-1.4	117	0.00
40	1,1'-Biphenyl	1.295	1.282	1.0	111	0.00
41	2-Chloronaphthalene	1.015	1.016	-0.1	115	0.00
42	2-Nitroaniline	0.386	0.422	-9.3	123	0.00
43 S	Dimethylphthalate-d6	1.396	1.416	-1.4	117	0.00
44	Dimethylphthalate	1.369	1.350	1.4	112	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101416\
 Data File : BG024369.D
 Acq On : 15 Oct 2016 2:07
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02028

Quant Time: Oct 15 05:07:46 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 14 23:22:45 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.266	0.294	-10.5	123	0.00
46 S	Acenaphthylene-d8	1.600	1.639	-2.4	115	0.00
47	Acenaphthylene	1.539	1.572	-2.1	116	0.00
48	3-Nitroaniline	0.251	0.284	-13.1	129	0.00
49 C	Acenaphthene	1.100	1.106	-0.5	116	0.00
50	2,4-Dinitrophenol	0.178	0.198	-11.2	136	0.00
51 S	4-Nitrophenol-d4	0.248	0.256	-3.2	120	0.00
52	4-Nitrophenol	0.300	0.295	1.7	111	0.00
53	Dibenzofuran	1.619	1.647	-1.7	115	0.00
54	2,4-Dinitrotoluene	0.394	0.435	-10.4	124	0.00
55	2,3,4,6-Tetrachlorophenol	0.413	0.436	-5.6	120	0.00
56	Diethylphthalate	1.443	1.416	1.9	109	0.00
57 S	Fluorene-d10	1.320	1.353	-2.5	114	0.00
58	Fluorene	1.339	1.350	-0.8	113	0.00
59	4-Chlorophenyl-phenylether	0.743	0.733	1.3	112	0.00
60	4-Nitroaniline	0.289	0.286	1.0	114	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.120	0.139	-15.8	136	0.00
63	4,6-Dinitro-2-methylphenol	0.126	0.144	-14.3	126	0.00
64	N-Nitrosodiphenylamine	0.566	0.585	-3.4	111	0.00
65	4-Bromophenyl-phenylether	0.232	0.252	-8.6	121	0.00
66	Hexachlorobenzene	0.258	0.276	-7.0	117	0.00
67	Atrazine	0.231	0.245	-6.1	115	0.00
68 C	Pentachlorophenol	0.157	0.179	-14.0	125	0.00
69	Phenanthrene	1.010	1.044	-3.4	110	0.00
70 S	Anthracene-d10	0.903	0.929	-2.9	109	0.00
71	Anthracene	1.028	1.075	-4.6	112	0.00
72	Carbazole	0.807	0.917	-13.6	110	0.00
73	Di-n-butylphthalate	0.986	1.087	-10.2	116	0.00
74 C	Fluoranthene	1.020	1.235	-21.1	113	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
76 S	Pyrene-d10	0.933	0.952	-2.0	111	0.00
77	Pyrene	1.068	1.089	-2.0	111	0.00
78	Butylbenzylphthalate	0.429	0.438	-2.1	114	0.00
79	3,3'-Dichlorobenzidine	0.372	0.437	-17.5	128	0.00
80	Benzo(a)anthracene	1.110	1.151	-3.7	115	0.00
81	Bis(2-ethylhexyl)phthalate	0.607	0.620	-2.1	113	0.00
82	Chrysene	1.057	1.091	-3.2	113	0.00
83 I	Perylene-d12	1.000	1.000	0.0	126	0.00
84	Di-n-octyl phthalate	0.938	0.988	-5.3	115	0.00
85	Benzo(b)fluoranthene	1.110	1.118	-0.7	120	0.00
86	Benzo(k)fluoranthene	1.070	1.036	3.2	115	0.00
87 S	Benzo(a)pyrene-d12	0.900	0.925	-2.8	122	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101416\
Data File : BG024369.D
Acq On : 15 Oct 2016 2:07
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02028

Quant Time: Oct 15 05:07:46 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Oct 14 23:22:45 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.058	1.076	-1.7	120	0.00
89	Indeno(1,2,3-cd)pyrene	1.253	1.305	-4.2	124	0.00
90	Dibenzo(a,h)anthracene	1.059	1.112	-5.0	126	0.00
91	Benzo(a,h,i)perylene	1.053	1.066	-1.2	122	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0