


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG101417\
Data File : BG029189.D
Acq On : 13 Oct 2017 10:39
Operator : SJ/JU
Sample : DFTPP57
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP57

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG101417.M
Title : SVOA CALIBRATION
Last Update : Mon Oct 16 02:34:29 2017



AutoFind: Scans 2469, 2470, 2471; Background Corrected with Scan 2462

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.3	55000	PASS
68	69	0.00	2	0.9	469	PASS
69	198	0.00	100	39.2	54960	PASS
70	69	0.00	2	0.4	238	PASS
127	198	10	80	44.6	62501	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	140125	PASS
199	198	5	9	7.1	9969	PASS
275	198	10	60	30.2	42346	PASS
365	198	1	100	3.9	5521	PASS
441	442	0.01	100	15.6	15937	PASS
442	198	50	100	72.8	101970	PASS
443	442	15	24	19.6	19939	PASS

Scan 2462 (17.903 min): BG029189.D
DFTPP57

m/z	Abundance
39.90	194.0
44.00	872.0
207.00	155.0
445.00	178.0

Instrument :
BNA_G
ClientSampleId :
DFTPP57

m/z	Abundance
38.00	458.0
39.00	3934.0
39.90	320.0
43.90	645.0
49.10	589.0
50.00	12673.0
51.00	48704.0
52.00	2353.0
52.90	192.0
54.90	209.0
56.00	1372.0
57.00	3262.0
58.00	168.0
61.00	736.0
62.00	694.0
63.00	2263.0
64.00	321.0
65.00	1339.0
66.90	213.0
69.00	48512.0
70.00	175.0
73.00	446.0
74.00	6422.0
75.00	7797.0
76.10	2775.0
77.00	58552.0
78.00	3059.0
79.00	4026.0
80.00	2659.0
81.00	4818.0
82.00	960.0
83.00	721.0
85.00	750.0
85.90	1386.0
86.90	611.0
87.80	164.0
90.90	1116.0
92.00	1230.0
93.00	7055.0
93.80	457.0
96.00	352.0
97.90	4715.0
99.00	4070.0
100.00	319.0
101.00	2297.0
101.80	220.0
103.00	1063.0
103.90	1887.0
105.00	1491.0
106.20	921.0
107.00	18120.0
108.00	2485.0
110.00	31000.0
111.00	4548.0
111.90	740.0
112.90	178.0
116.00	1312.0
117.00	12004.0
118.00	1032.0
121.00	306.0
121.90	1533.0
123.00	2174.0
123.90	853.0
125.00	578.0
127.00	60016.0
128.00	4209.0
129.00	23024.0
130.00	1754.0
130.90	399.0
134.00	571.0
134.90	2129.0
136.00	643.0
137.00	946.0
138.00	212.0
138.90	250.0
140.00	461.0
141.00	3554.0
141.90	920.0
142.90	570.0
144.10	243.0
144.90	178.0
146.00	848.0
146.90	1720.0
147.90	3690.0
149.00	663.0
150.00	191.0
151.20	409.0
151.80	483.0
152.90	1054.0
153.90	854.0
155.00	1729.0

Instrument :
BNA_G
ClientSampleId :
DFTPP57

156.00	2923.0
157.00	491.0
157.90	792.0
159.00	637.0
159.90	1002.0
161.00	1222.0
162.00	522.0
163.90	228.0
164.90	1627.0
166.00	1223.0
167.00	7220.0
168.00	3206.0
169.10	362.0
170.00	216.0
171.00	533.0
172.00	765.0
173.00	983.0
174.00	1895.0
175.00	2858.0
176.00	925.0
177.00	1372.0
178.90	5693.0
180.00	3867.0
180.90	1629.0
181.90	459.0
183.00	166.0
184.00	606.0
185.00	2600.0
186.00	22768.0
187.00	6435.0
188.00	777.0
189.00	1714.0
190.90	927.0
192.00	1575.0
192.90	2161.0
194.00	719.0
195.00	295.0
196.00	3386.0
197.90	143040.0
198.90	9066.0
199.90	1209.0
201.00	400.0
201.30	389.0
202.10	304.0
203.00	1269.0
204.00	6996.0
205.00	10230.0
206.00	42704.0
207.00	5118.0
207.90	1212.0
209.00	546.0
210.00	710.0
211.00	1822.0
211.90	303.0
214.90	689.0
216.10	958.0
217.00	12127.0
218.00	1507.0
218.80	205.0
221.00	5153.0
222.00	1372.0
223.00	3231.0
224.00	25328.0
225.00	5647.0
226.00	514.0
227.00	11257.0
228.00	1823.0
229.00	2806.0
229.80	194.0
231.00	935.0
232.00	241.0
232.90	219.0
234.00	1000.0
235.00	1029.0
236.00	311.0
236.90	772.0
239.00	696.0
240.00	324.0
241.00	555.0
242.00	1531.0
243.00	1576.0
244.00	19968.0
245.00	2196.0
246.00	3691.0
247.00	1032.0
248.00	345.0
249.00	760.0
250.00	442.0
252.00	180.0
253.00	764.0
255.00	102800.0
256.00	13515.0
257.00	1279.0
258.00	6237.0
259.00	1142.0

Instrument :
BNA_G
ClientSampleId :
DFTPP57

260.00	279.0
260.90	263.0
265.00	2713.0
265.90	269.0
271.00	322.0
272.00	266.0
273.00	3202.0
274.00	8374.0
275.00	44648.0
276.00	4950.0
277.00	4758.0
278.00	759.0
281.90	314.0
283.10	318.0
283.90	198.0
285.00	539.0
289.10	170.0
293.00	838.0
293.90	272.0
296.00	13916.0
297.00	1890.0
297.90	216.0
301.90	351.0
303.00	1716.0
304.00	463.0
305.00	165.0
307.90	253.0
314.00	867.0
315.00	1486.0
316.10	1056.0
316.80	265.0
321.00	460.0
322.20	264.0
323.00	3830.0
324.00	586.0
326.90	1095.0
328.00	336.0
331.90	428.0
333.00	457.0
334.00	3275.0
335.00	813.0
335.80	151.0
341.00	681.0
345.90	1200.0
351.10	205.0
352.00	1089.0
353.00	711.0
354.00	866.0
355.10	402.0
365.00	6611.0
365.90	589.0
371.00	533.0
372.00	1963.0
373.00	574.0
383.00	675.0
390.00	314.0
391.10	310.0
401.00	205.0
402.10	1072.0
403.00	1219.0
404.00	280.0
421.00	859.0
422.00	1162.0
423.00	7265.0
424.00	1397.0
441.00	18504.0
442.00	125144.0
443.00	22944.0
444.10	2026.0

Instrument :
BNA_G
ClientSampleId :
DFTPP57

m/z	Abundance
35.70	197.0
37.00	277.0
37.90	647.0
39.00	5289.0
40.00	281.0
44.00	850.0
49.00	838.0
50.00	16111.0
51.00	64592.0
52.00	2958.0
53.10	239.0
55.00	507.0
55.90	1952.0
57.00	4681.0
58.00	304.0
61.00	754.0
62.00	939.0
63.00	3556.0
64.00	412.0
65.00	1647.0
68.00	1409.0
69.00	64288.0
69.90	274.0
73.10	382.0
74.00	6158.0
75.00	10455.0
76.00	4150.0
77.00	74400.0
78.00	4634.0
79.00	4984.0
80.00	4130.0
81.00	5295.0
81.90	1370.0
82.90	1037.0
85.00	809.0
86.00	1625.0
86.90	721.0
87.90	189.0
91.00	1402.0
92.00	1134.0
93.00	8159.0
94.00	623.0
96.00	459.0
98.00	6916.0
99.00	4633.0
100.00	658.0
101.00	2780.0
102.90	757.0
104.00	2365.0
105.00	1962.0
106.00	1034.0
107.00	23568.0
108.00	3490.0
110.00	37056.0
111.00	6584.0
111.90	766.0
113.00	160.0
114.90	177.0
116.00	1630.0
117.00	18424.0
117.90	1193.0
118.90	245.0
120.00	328.0
122.00	1434.0
123.00	2005.0
123.90	942.0
124.90	1144.0
127.00	73752.0
128.00	6423.0
129.00	31072.0
129.90	2590.0
131.00	566.0
131.90	240.0
134.00	1003.0
135.00	2641.0
136.00	810.0
137.00	1382.0
137.80	334.0
138.80	159.0
139.90	400.0
141.00	4235.0
142.00	981.0
142.90	1101.0
144.00	427.0
145.00	320.0
145.90	899.0
146.90	2213.0
147.90	4830.0
149.00	973.0
150.00	227.0
151.10	565.0

Instrument :
BNA_G
ClientSampleId :
DFTPP57

151.90	332.0
153.00	1652.0
154.00	993.0
155.00	2034.0
156.00	3163.0
156.90	580.0
157.90	651.0
159.00	629.0
159.90	1032.0
160.90	1645.0
161.90	468.0
163.00	341.0
164.00	181.0
164.90	1517.0
166.00	1095.0
167.00	7663.0
168.00	3779.0
169.00	982.0
170.00	172.0
170.90	602.0
171.90	993.0
172.90	1346.0
174.00	1730.0
175.00	4087.0
175.80	1067.0
177.00	1657.0
178.10	480.0
178.90	6234.0
180.00	5621.0
181.00	2107.0
182.00	366.0
184.00	686.0
185.10	3698.0
186.00	25248.0
187.00	7397.0
187.90	649.0
188.90	1948.0
190.00	332.0
191.00	635.0
192.00	2418.0
193.00	2376.0
193.90	589.0
195.00	321.0
195.90	4926.0
197.90	160256.0
198.90	12008.0
199.90	979.0
201.70	735.0
203.00	1406.0
204.00	7616.0
205.00	11594.0
206.00	52080.0
207.00	6608.0
208.00	1534.0
209.00	898.0
209.90	519.0
210.10	510.0
211.00	1850.0
214.90	543.0
216.90	13400.0
218.00	1817.0
220.00	308.0
221.00	6547.0
221.60	1925.0
221.90	1838.0
223.00	3459.0
224.00	27736.0
225.00	6984.0
226.00	789.0
227.00	13663.0
228.00	1967.0
228.90	2137.0
229.90	335.0
231.00	964.0
231.90	290.0
232.90	164.0
234.00	954.0
235.00	922.0
236.00	479.0
237.00	817.0
239.00	237.0
239.80	541.0
241.00	561.0
242.00	1701.0
243.00	1876.0
244.00	20032.0
245.00	2745.0
246.00	4964.0
246.90	1116.0
247.90	273.0
248.90	677.0
250.00	228.0
251.30	337.0
251.90	255.0
252.90	683.0

Instrument :
BNA_G
ClientSampleId :
DFTPP57

255.00	110944.0
256.00	16632.0
257.00	1072.0
258.00	7164.0
259.00	1061.0
260.00	462.0
261.20	220.0
263.90	216.0
265.00	1948.0
265.90	593.0
269.90	182.0
270.60	209.0
271.00	179.0
272.00	490.0
273.00	3688.0
274.00	8328.0
275.00	50096.0
276.00	6653.0
276.90	3486.0
278.00	727.0
278.90	179.0
283.00	542.0
284.00	301.0
285.00	873.0
286.10	177.0
289.10	268.0
292.00	294.0
293.00	1020.0
295.00	414.0
296.00	13807.0
296.90	1947.0
299.00	157.0
302.00	229.0
303.00	1774.0
304.00	609.0
308.00	213.0
309.00	267.0
312.90	200.0
313.90	661.0
315.00	1808.0
316.00	941.0
321.00	630.0
321.90	191.0
323.00	4973.0
324.10	807.0
327.00	653.0
328.00	306.0
329.10	192.0
332.10	281.0
333.10	612.0
334.00	2398.0
335.00	658.0
336.10	161.0
341.00	341.0
345.90	953.0
347.00	209.0
352.10	1192.0
353.10	859.0
354.00	1250.0
355.00	466.0
365.00	6308.0
365.90	605.0
370.00	230.0
371.00	309.0
372.00	1671.0
373.10	530.0
383.00	363.0
384.00	240.0
389.90	270.0
391.00	269.0
392.10	201.0
402.00	870.0
403.00	1072.0
404.00	190.0
421.10	956.0
422.00	1157.0
423.00	6418.0
424.00	1021.0
425.00	192.0
441.00	19032.0
442.00	115992.0
443.00	24336.0
444.00	2204.0

Instrument :
BNA_G
ClientSampleId :
DFTPP57

m/z	Abundance
38.00	494.0
39.00	4035.0
40.00	182.0
41.00	155.0
44.00	629.0
49.00	693.0
50.00	11965.0
51.00	51704.0
52.00	2943.0
55.00	222.0
56.00	1563.0
57.00	3532.0
58.00	220.0
60.90	346.0
61.90	662.0
63.00	2586.0
64.00	284.0
65.00	1090.0
69.00	52080.0
69.90	265.0
73.00	558.0
74.00	5166.0
75.00	7686.0
76.10	2762.0
77.00	60160.0
78.00	4078.0
79.00	3552.0
80.00	3260.0
80.90	5102.0
82.00	1184.0
83.00	833.0
85.00	1102.0
86.00	1297.0
86.90	645.0
87.90	252.0
91.00	1125.0
92.00	927.0
93.00	6395.0
94.00	437.0
97.00	156.0
98.00	4664.0
99.00	4049.0
100.00	431.0
101.00	2591.0
101.90	305.0
102.90	673.0
103.90	1212.0
105.00	1118.0
106.10	727.0
107.00	18488.0
108.00	2703.0
109.00	795.0
110.00	28264.0
111.00	4510.0
111.90	567.0
112.90	169.0
115.90	873.0
117.00	12149.0
118.00	1046.0
120.00	220.0
122.00	1134.0
123.00	1732.0
124.00	905.0
125.00	736.0
127.00	53736.0
128.00	4980.0
129.00	23408.0
129.90	2018.0
130.90	726.0
134.00	626.0
135.00	2109.0
136.00	685.0
136.90	1015.0
140.90	3406.0
141.90	538.0
142.90	606.0
144.00	197.0
144.90	185.0
146.00	601.0
146.90	1240.0
148.00	3420.0
149.00	807.0
150.00	198.0
150.90	359.0
153.00	1040.0
154.00	1005.0
155.00	1602.0
156.00	2708.0
156.90	312.0
158.00	421.0
159.00	871.0

Instrument :
BNA_G
ClientSampleId :
DFTPP57

160.00	1190.0
161.00	1643.0
161.90	279.0
165.00	1493.0
165.90	1053.0
167.00	6040.0
168.00	2788.0
168.90	699.0
170.90	334.0
171.90	620.0
173.00	1032.0
174.00	1339.0
175.00	2003.0
175.90	558.0
176.90	1326.0
177.80	434.0
178.90	4864.0
180.00	3429.0
181.00	1576.0
181.90	158.0
182.10	168.0
183.00	208.0
184.00	502.0
185.00	1998.0
186.00	18640.0
187.00	4472.0
187.90	519.0
189.00	1416.0
190.90	824.0
191.90	1211.0
193.00	1425.0
193.90	598.0
196.00	3587.0
197.90	117080.0
198.90	8834.0
199.90	665.0
201.40	676.0
202.90	1015.0
204.00	5359.0
205.00	7821.0
206.00	34784.0
207.00	4562.0
207.90	1327.0
208.90	498.0
210.90	1427.0
212.00	260.0
215.10	596.0
216.00	917.0
216.90	10231.0
217.90	1010.0
221.00	3660.0
221.90	1217.0
223.00	1959.0
224.00	20280.0
225.00	5460.0
225.80	416.0
227.00	9011.0
228.00	1477.0
229.00	1924.0
229.90	207.0
230.90	394.0
231.80	188.0
233.00	173.0
234.00	698.0
234.90	713.0
235.90	288.0
237.00	633.0
238.90	370.0
240.90	430.0
242.00	977.0
243.00	1277.0
244.00	13416.0
245.00	2052.0
246.00	2650.0
246.90	378.0
248.90	527.0
251.90	230.0
253.00	665.0
255.00	75264.0
256.00	11156.0
256.90	873.0
258.00	4905.0
258.90	700.0
264.00	186.0
265.00	1677.0
265.90	312.0
271.00	162.0
273.00	2354.0
274.00	5280.0
275.00	32296.0
276.00	5186.0
277.00	2961.0
277.80	440.0
283.00	230.0
283.90	242.0

Instrument :

BNA_G

ClientSampleId :

DFTPP57

285.00	453.0
292.10	274.0
293.00	643.0
295.10	190.0
296.00	9198.0
296.90	1281.0
301.00	159.0
303.00	950.0
303.90	295.0
308.00	249.0
313.90	478.0
314.90	1207.0
316.10	720.0
321.10	357.0
322.00	155.0
323.00	2903.0
323.90	469.0
327.00	535.0
327.90	283.0
333.00	189.0
334.00	1326.0
335.10	346.0
341.00	282.0
346.00	497.0
346.90	168.0
352.00	746.0
352.90	466.0
354.10	1152.0
355.00	174.0
356.10	168.0
365.00	3644.0
366.00	589.0
369.90	151.0
371.10	218.0
372.00	1046.0
373.00	473.0
383.00	286.0
402.00	370.0
403.00	714.0
404.00	266.0
421.00	663.0
422.10	648.0
423.00	2995.0
424.00	620.0
441.00	10276.0
442.00	64776.0
443.00	12537.0
444.10	1480.0

Instrument :
BNA_G
ClientSampleId :
DFTPP57