

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101617\
 Data File : BG029291.D
 Acq On : 17 Oct 2017 4:03
 Operator : SJ/JU
 Sample : I5906-12
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 WH-TWS-02

Quant Time: Oct 17 05:33:33 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 17:32:15 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	74362	20.00	ng	0.00
21) Naphthalene-d8	10.99	136	361241	20.00	ng	0.00
38) Acenaphthene-d10	14.79	164	229785	20.00	ng	0.00
63) Phenanthrene-d10	17.52	188	513798	20.00	ng	0.00
75) Chrysene-d12	21.79	240	556545	20.00	ng	0.00
86) Perylene-d12	25.10	264	572479	20.00	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.72	112	592991	133.22	ng	0.00
7) Phenol-d6	7.32	99	753792	120.64	ng	0.00
23) Nitrobenzene-d5	9.33	82	507601	89.29	ng	0.00
41) 2,4,6-Tribromophenol	16.27	330	347675	117.19	ng	0.00
44) 2-Fluorobiphenyl	13.42	172	1300430	83.00	ng	0.00
78) Terphenyl-d14	20.11	244	1882205	76.94	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
31) Naphthalene	11.04	128	934199	51.890	ng	98
37) 2-Methylnaphthalene	12.63	142	155384	11.842	ng	91
45) 1,1'-Biphenyl	13.62	154	53087	2.688	ng	97
49) Dimethylphthalate	14.23	163	36725	2.048	ng	# 98
51) Acenaphthene	14.85	154	241110	16.436	ng	99
54) Dibenzofuran	15.18	168	296173	14.142	ng	99
57) Fluorene	15.83	166	232441	13.436	ng	96
70) Phenanthrene	17.56	178	591176	22.272	ng	98
71) Anthracene	17.65	178	62532	2.346	ng	95
72) Carbazole	17.92	167	369600	14.436	ng	99
74) Fluoranthene	19.56	202	142896	4.603	ng	99
77) Pyrene	19.92	202	97793	2.897	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101617\
Data File : BG029291.D
Acq On : 17 Oct 2017 4:03
Operator : SJ/JU
Sample : I5906-12
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
WH-TWS-02

Quant Time: Oct 17 05:33:33 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 17:32:15 2017
Response via : Initial Calibration

