


4,6-Dinitro-2-methylphenol




Resp Ratio = $1.27 \times 10^{-1} \times \text{Amt} - 3.17 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_G\METHODS\8270-BG101617.M
Calibration Table Last Updated: Mon Oct 16 17:32:15 2017

2,4-Dinitrophenol

Response Ratio




Resp Ratio = $1.96 \times 10^{-1} \times \text{Amt} - 6.51 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_G\METHODS\8270-BG101617.M
Calibration Table Last Updated: Mon Oct 16 17:32:15 2017

Hexachlorocyclopentadiene

Response Ratio



Resp Ratio = $2.94 \times 10^{-1} \times \text{Amt} - 6.84 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_G\METHODS\8270-BG101617.M
Calibration Table Last Updated: Mon Oct 16 17:32:15 2017