


Data Path : Z:\HPCHEM1\BNA G\Data\BG101617\
 Data File : BG029267.D
 Acq On : 16 Oct 2017 10:41
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG101617.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 16 17:32:15 2017



AutoFind: Scans 2494, 2495, 2496; Background Corrected with Scan 2488

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.2	21422	PASS
68	69	0.00	2	1.1	217	PASS
69	198	0.00	100	33.2	19680	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	41.4	24549	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	59226	PASS
199	198	5	9	6.5	3869	PASS
275	198	10	60	27.2	16127	PASS
365	198	1	100	3.7	2199	PASS
441	442	0.01	24	16.9	9526	PASS
442	198	50	100	94.9	56213	PASS
443	442	15	24	18.8	10568	PASS