


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101620\
 Data File : BG046923.D
 Acq On : 16 Oct 2020 8:50
 Operator : CG/JU
 Sample : DFTPP39
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP39

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG101520MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Oct 19 11:15:40 2020



AutoFind: Scans 2473, 2474, 2475; Background Corrected with Scan 2468

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.7	39208	PASS
68	69	0.00	2	1.5	625	PASS
69	198	0.00	100	36.4	42376	PASS
70	69	0.00	2	0.2	98	PASS
127	198	10	80	40.4	46946	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	116328	PASS
199	198	5	9	7.2	8429	PASS
275	198	10	60	29.5	34296	PASS
365	198	1	100	3.8	4389	PASS
441	198	0.01	100	11.8	13701	PASS
442	198	50	100	73.0	84906	PASS
443	442	15	24	20.2	17157	PASS

Scan 2468 (17.841 min): BG046923.D
DFTPP39

m/z	Abundance
40.00	171.0
44.10	681.0
190.80	155.0
206.80	344.0
371.90	187.0

Instrument :
BNA_G
ClientSampleId :
DFTPP39

m/z	Abundance
36.10	177.0
38.20	377.0
39.10	1965.0
40.00	351.0
44.00	166.0
49.10	343.0
50.10	7752.0
51.10	33816.0
52.10	2101.0
55.00	153.0
55.90	1167.0
57.00	2032.0
57.90	379.0
61.00	699.0
62.00	625.0
62.90	1454.0
64.10	175.0
64.90	901.0
69.00	35800.0
72.80	571.0
74.10	3202.0
75.00	5659.0
76.00	2362.0
77.00	36536.0
78.10	2434.0
79.10	2615.0
80.10	1737.0
81.00	2965.0
82.00	622.0
82.90	557.0
85.10	293.0
86.00	623.0
87.00	531.0
88.00	223.0
90.90	826.0
91.90	704.0
92.90	4828.0
93.80	349.0
96.10	151.0
98.00	3349.0
99.00	2477.0
99.90	423.0
101.00	1835.0
102.00	159.0
102.90	617.0
103.90	963.0
105.00	992.0
105.70	361.0
107.00	11699.0
108.00	1872.0
109.00	447.0
110.00	20864.0
111.00	2997.0
112.10	480.0
112.80	177.0
116.00	775.0
117.00	11154.0
118.10	671.0
119.00	162.0
119.80	203.0
121.90	842.0
122.90	1389.0
124.00	609.0
124.90	622.0
127.00	39040.0
128.00	3186.0
129.00	15029.0
129.90	1132.0
131.00	417.0
132.00	159.0
134.00	784.0
135.00	1519.0
136.00	619.0
137.00	490.0
137.50	239.0
140.30	278.0
141.00	2248.0
141.90	609.0
143.00	447.0
144.60	274.0
147.00	1536.0
147.90	2996.0
148.70	364.0
150.00	248.0
150.80	222.0
151.60	211.0
153.00	771.0
154.00	633.0
155.00	1037.0
155.90	2302.0
156.90	371.0
157.90	332.0

Instrument :
BNA_G
ClientSampleId :
DFTPP39

159.00	519.0
159.90	647.0
161.00	969.0
161.90	520.0
163.00	152.0
164.90	1179.0
165.90	1142.0
167.00	3107.0
167.90	2655.0
168.90	508.0
169.90	244.0
170.90	337.0
171.80	466.0
172.90	667.0
173.80	879.0
174.90	1594.0
176.00	545.0
176.80	856.0
177.80	474.0
178.80	2600.0
179.90	2026.0
180.90	882.0
181.70	222.0
182.50	233.0
183.80	276.0
184.90	1566.0
185.90	11896.0
186.90	3725.0
188.00	639.0
188.90	763.0
189.60	172.0
190.80	341.0
191.90	1266.0
192.90	1393.0
193.80	252.0
194.70	310.0
196.00	2415.0
197.90	85392.0
198.90	7430.0
199.80	550.0
201.30	370.0
202.00	192.0
203.00	1110.0
203.90	3391.0
204.90	5867.0
205.90	23328.0
206.90	3526.0
207.90	1211.0
208.90	451.0
209.90	599.0
210.90	614.0
215.00	444.0
216.00	510.0
216.90	6848.0
217.70	515.0
219.00	252.0
220.80	3085.0
221.60	763.0
222.90	1501.0
223.90	13092.0
224.90	4283.0
225.90	509.0
226.90	6299.0
227.90	855.0
228.90	1037.0
230.80	323.0
232.80	201.0
233.80	305.0
234.70	425.0
235.90	605.0
236.90	408.0
238.90	273.0
239.90	255.0
240.90	442.0
241.80	633.0
242.80	585.0
243.90	8875.0
244.90	1569.0
245.90	2161.0
246.80	518.0
248.80	554.0
251.80	183.0
252.90	181.0
253.70	517.0
254.90	51440.0
255.90	8295.0
256.90	819.0
257.80	3299.0
258.90	836.0
264.90	1156.0
265.80	207.0
270.80	192.0
272.90	1774.0
273.90	4092.0
274.90	23624.0
275.90	3186.0

Instrument :
BNA_G
ClientSampleId :
DFTPP39

276.90	2120.0
277.90	441.0
278.90	191.0
281.00	372.0
283.80	172.0
285.00	191.0
292.50	195.0
293.00	301.0
294.00	366.0
295.80	5930.0
296.90	747.0
298.00	156.0
301.80	180.0
302.70	560.0
303.80	181.0
314.90	881.0
315.80	184.0
320.70	270.0
322.90	2091.0
323.80	497.0
326.70	359.0
328.00	157.0
332.60	153.0
333.90	1512.0
334.80	399.0
340.70	264.0
345.80	418.0
351.90	480.0
352.80	292.0
353.80	769.0
354.80	212.0
364.80	2993.0
365.90	376.0
370.70	152.0
372.00	1263.0
382.60	349.0
401.90	333.0
402.70	436.0
420.80	264.0
421.90	215.0
422.80	3012.0
423.80	644.0
440.80	7627.0
441.80	46760.0
442.80	9167.0
443.80	1067.0
461.00	170.0

Instrument :
BNA_G
ClientSampleId :
DFTPP39

m/z	Abundance
36.90	200.0
37.20	184.0
38.10	537.0
39.10	3457.0
40.00	585.0
44.00	709.0
50.00	12955.0
51.10	46480.0
52.10	2753.0
55.10	162.0
55.30	160.0
56.00	1543.0
57.00	2995.0
57.90	233.0
61.00	596.0
62.00	807.0
63.10	2526.0
64.10	237.0
65.00	907.0
68.00	1289.0
69.00	49744.0
69.90	237.0
73.00	674.0
74.00	4934.0
75.00	7388.0
76.10	3760.0
77.00	53768.0
78.00	3251.0
79.00	4326.0
80.00	2631.0
81.00	3903.0
82.00	1101.0
83.00	873.0
85.00	597.0
86.00	1258.0
87.00	674.0
89.00	192.0
90.90	978.0
92.10	813.0
93.00	7603.0
93.90	601.0
95.00	171.0
96.00	231.0
98.00	6595.0
99.00	4207.0
99.70	250.0
100.00	259.0
101.00	2278.0
102.90	906.0
104.00	1714.0
105.00	1226.0
107.00	18432.0
108.00	2329.0
110.00	28096.0
111.00	5201.0
112.00	532.0
117.00	15413.0
118.00	1082.0
120.10	212.0
122.00	1055.0
122.90	1805.0
123.90	676.0
124.90	1210.0
127.00	54696.0
128.00	3787.0
128.90	25144.0
129.90	2276.0
130.90	454.0
132.00	187.0
132.70	189.0
133.90	762.0
134.90	2103.0
136.00	774.0
137.00	1386.0
138.00	285.0
138.80	173.0
139.90	869.0
140.90	3202.0
142.00	710.0
142.70	908.0
143.80	190.0
145.00	230.0
145.90	732.0
147.00	1883.0
147.90	3854.0
148.90	1065.0
149.60	320.0
150.90	259.0
152.00	388.0
152.80	1442.0
154.00	1032.0
155.00	1304.0

Instrument :
BNA_G
ClientSampleId :
DFTPP39

155.90	3242.0
156.90	640.0
157.70	378.0
158.90	695.0
159.90	972.0
160.90	1589.0
162.00	412.0
162.90	154.0
163.80	230.0
165.00	1262.0
166.00	1276.0
166.90	6614.0
167.90	2847.0
169.00	549.0
170.70	373.0
171.90	752.0
173.00	1262.0
173.90	1129.0
175.00	2994.0
175.90	900.0
176.90	1462.0
177.90	447.0
178.90	5856.0
180.00	4018.0
180.90	1575.0
181.90	221.0
182.90	259.0
183.90	525.0
184.90	2319.0
186.00	19472.0
186.90	6116.0
187.90	762.0
188.80	721.0
189.90	342.0
191.00	744.0
191.90	1643.0
192.90	1919.0
193.90	359.0
194.60	236.0
195.90	4135.0
197.90	141248.0
198.90	9653.0
199.90	728.0
201.20	403.0
201.70	294.0
202.80	1018.0
203.90	5921.0
205.00	10379.0
205.90	42288.0
207.00	6468.0
207.90	1081.0
208.90	816.0
209.90	683.0
210.90	1535.0
212.00	171.0
214.60	393.0
215.90	1097.0
216.90	11137.0
217.90	1076.0
219.00	152.0
220.90	5159.0
221.70	1560.0
222.80	2109.0
223.90	23224.0
224.90	6804.0
225.90	766.0
226.90	11117.0
227.90	1716.0
228.80	2175.0
229.90	387.0
230.90	1337.0
233.80	664.0
234.80	759.0
235.90	453.0
236.90	1478.0
237.90	254.0
238.90	265.0
239.90	327.0
240.90	797.0
241.90	1550.0
242.90	1428.0
243.90	17528.0
244.90	1997.0
245.80	4612.0
247.00	695.0
248.90	479.0
249.70	323.0
251.80	240.0
252.90	486.0
254.90	93856.0
255.90	13204.0
256.90	1043.0
257.90	5057.0
258.90	706.0
260.90	389.0
264.00	190.0

Instrument :
BNA_G
ClientSampleId :
DFTPP39

264.80	1952.0
265.80	377.0
269.80	173.0
271.90	612.0
272.90	2859.0
273.90	7580.0
274.90	40472.0
275.80	4030.0
276.80	3438.0
277.90	490.0
278.90	410.0
281.80	258.0
282.80	422.0
284.00	375.0
284.80	987.0
290.80	196.0
291.90	314.0
292.80	1236.0
294.00	214.0
295.80	11112.0
296.80	1743.0
301.70	232.0
302.80	1109.0
303.90	273.0
304.70	153.0
310.00	159.0
312.90	266.0
313.80	805.0
314.80	1246.0
315.90	669.0
320.70	398.0
321.80	403.0
322.80	3748.0
323.90	361.0
324.60	197.0
326.90	659.0
327.90	479.0
331.80	285.0
332.80	505.0
333.90	2432.0
334.90	1122.0
340.80	478.0
342.80	228.0
345.90	758.0
346.80	156.0
351.90	938.0
352.90	755.0
353.80	877.0
354.90	233.0
362.90	170.0
364.80	5173.0
365.80	746.0
371.80	1431.0
372.90	336.0
382.80	585.0
383.80	283.0
389.90	541.0
390.60	175.0
401.80	503.0
402.80	1015.0
403.80	327.0
420.80	452.0
421.70	863.0
422.80	5556.0
423.80	1200.0
424.70	255.0
425.30	172.0
436.20	170.0
440.80	15413.0
441.80	97256.0
442.80	20440.0
443.80	1454.0

Instrument :
BNA_G
ClientSampleId :
DFTPP39

m/z	Abundance
37.10	264.0
38.10	349.0
39.10	3424.0
40.00	244.0
44.00	900.0
49.00	402.0
50.10	9935.0
51.00	37328.0
52.10	1194.0
55.00	221.0
56.00	1310.0
57.00	2328.0
59.80	219.0
61.00	656.0
62.00	488.0
63.00	2017.0
64.10	302.0
65.10	956.0
67.90	587.0
69.00	41584.0
70.20	294.0
73.00	465.0
74.00	4162.0
75.00	8894.0
76.10	2410.0
77.00	42952.0
78.00	3283.0
79.00	3617.0
80.00	2786.0
81.00	3222.0
82.10	1080.0
83.00	936.0
84.90	559.0
85.90	1241.0
86.90	717.0
91.00	624.0
92.00	853.0
93.00	5009.0
94.10	338.0
96.00	321.0
98.00	4031.0
99.00	3425.0
100.00	190.0
101.00	2175.0
102.10	313.0
102.90	575.0
104.00	1631.0
104.90	818.0
105.90	694.0
107.00	14307.0
107.90	2359.0
109.00	653.0
110.00	24064.0
111.00	4556.0
112.00	652.0
112.80	242.0
116.00	1040.0
117.00	15549.0
118.00	779.0
122.00	994.0
123.00	2209.0
123.90	803.0
125.00	390.0
127.00	47104.0
127.90	4161.0
129.00	21536.0
129.90	1878.0
131.00	420.0
131.70	196.0
133.90	860.0
134.90	2172.0
136.00	820.0
136.90	1204.0
138.30	193.0
139.80	273.0
140.90	3035.0
141.90	875.0
142.80	828.0
145.00	168.0
145.90	985.0
147.00	1821.0
148.00	2502.0
148.90	737.0
149.80	210.0
151.00	368.0
152.10	311.0
153.00	868.0
153.90	616.0
155.00	1892.0
155.90	2721.0
157.00	207.0
157.90	665.0

Instrument :
BNA_G
ClientSampleId :
DFTPP39

158.90	460.0
159.90	773.0
160.90	1671.0
162.00	441.0
163.20	174.0
163.80	408.0
164.90	1249.0
165.90	1202.0
166.90	5763.0
167.90	2920.0
169.10	447.0
169.60	153.0
170.90	368.0
171.90	765.0
172.80	832.0
173.90	1237.0
175.00	3065.0
175.90	831.0
176.90	953.0
177.90	706.0
178.90	5099.0
179.90	3056.0
180.90	1161.0
182.00	378.0
182.80	274.0
183.70	351.0
184.00	326.0
185.00	2430.0
185.90	17808.0
186.90	5053.0
187.90	489.0
188.90	1317.0
190.80	844.0
191.90	2139.0
192.90	1443.0
194.00	461.0
194.80	184.0
195.90	3630.0
197.90	122344.0
198.90	8205.0
199.90	744.0
201.30	604.0
202.90	980.0
203.90	5433.0
204.90	9144.0
205.90	38864.0
206.90	5711.0
207.90	1072.0
208.90	587.0
209.80	797.0
210.80	1025.0
211.50	480.0
215.00	616.0
215.90	1243.0
216.90	10477.0
217.80	1518.0
219.20	197.0
220.00	289.0
220.90	5286.0
221.70	1714.0
222.90	3085.0
223.90	21880.0
224.90	5209.0
225.80	491.0
226.90	10239.0
227.90	1469.0
228.90	1296.0
229.90	169.0
231.00	995.0
232.90	223.0
233.80	417.0
234.90	844.0
235.80	590.0
236.80	972.0
237.80	253.0
238.90	310.0
240.00	223.0
240.90	637.0
241.90	1234.0
242.90	1374.0
243.90	17112.0
244.90	2034.0
245.80	3664.0
246.90	781.0
248.90	351.0
249.90	155.0
250.90	233.0
251.90	254.0
252.80	470.0
254.90	89136.0
255.90	11823.0
257.00	1367.0
257.90	4858.0
258.90	837.0
259.70	188.0
260.80	243.0

Instrument :
BNA_G
ClientSampleId :
DFTPP39

264.90	2275.0
265.50	190.0
266.00	379.0
270.00	160.0
270.90	300.0
272.90	3003.0
273.90	8258.0
274.90	38792.0
275.90	4516.0
276.80	3716.0
277.90	578.0
282.90	348.0
283.70	329.0
284.90	575.0
288.80	230.0
289.90	184.0
290.80	266.0
291.80	313.0
292.90	1006.0
295.90	11198.0
296.80	2032.0
297.60	163.0
300.80	326.0
302.00	335.0
302.90	1421.0
303.90	396.0
308.80	277.0
313.90	523.0
314.90	1699.0
315.80	432.0
317.00	172.0
321.00	548.0
321.80	154.0
322.90	4004.0
323.90	882.0
326.70	728.0
327.90	353.0
329.00	172.0
331.90	303.0
333.00	587.0
333.80	2406.0
334.80	671.0
340.90	445.0
345.80	706.0
346.90	280.0
351.80	1360.0
352.90	305.0
353.90	924.0
364.80	5001.0
365.80	815.0
370.70	268.0
371.80	1382.0
373.00	361.0
382.80	530.0
384.90	195.0
389.90	388.0
401.90	795.0
402.70	839.0
403.80	257.0
420.70	950.0
421.70	782.0
422.80	6309.0
423.80	1371.0
424.80	158.0
440.80	18064.0
441.80	110704.0
442.80	21864.0
443.80	1977.0

Instrument :
BNA_G
ClientSampleId :
DFTPP39