

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101716\
 Data File : BG024415.D
 Acq On : 18 Oct 2016 11:40
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02033

Quant Time: Oct 18 12:26:30 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 18 03:43:57 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.456	0.535	-17.3	97	0.00
3 S	1,4-Dioxane-d8	0.421	0.404	4.0	85	0.00
4	Benzaldehyde	1.313	1.456	-10.9	81	0.00
5 S	Phenol-d5	1.866	1.954	-4.7	84	0.00
6	Phenol	1.894	2.063	-8.9	86	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.261	1.274	-1.0	82	0.00
8	Bis(2-Chloroethyl)ether	1.394	1.404	-0.7	79	0.00
9 S	2-Chlorophenol-d4	1.285	1.342	-4.4	82	0.00
10	2-Chlorophenol	1.302	1.330	-2.2	82	0.00
11	2-Methylphenol	1.403	1.457	-3.8	84	0.00
12	2,2'-oxybis(1-Chloropropane	2.784	2.636	5.3	74	0.00
13 S	4-Methylphenol-d8	1.509	1.595	-5.7	84	0.00
14	Acetophenone	2.240	2.454	-9.6	85	0.00
15 P	N-Nitroso-di-n-propylamine	1.351	1.411	-4.4	82	0.00
16	4-Methylphenol	1.488	1.639	-10.1	89	0.00
17	Hexachloroethane	0.602	0.603	-0.2	83	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
19 S	Nitrobenzene-d5	0.142	0.159	-12.0	96	0.00
20	Nitrobenzene	0.449	0.468	-4.2	89	0.00
21	Isophorone	0.816	0.823	-0.9	83	0.00
22 S	2-Nitrophenol-d4	0.159	0.191	-20.1#	106	0.00
23 C	2-Nitrophenol	0.173	0.189	-9.2	93	0.00
24	2,4-Dimethylphenol	0.426	0.428	-0.5	85	0.00
25	Bis(2-Chloroethoxy)methane	0.448	0.460	-2.7	85	0.00
26 S	2,4-Dichlorophenol-d3	0.346	0.347	-0.3	85	0.00
27 C	2,4-Dichlorophenol	0.332	0.346	-4.2	89	0.00
28	Naphthalene	0.963	0.968	-0.5	86	0.00
29 S	4-Chloroaniline-d4	0.363	0.416	-14.6	94	0.00
30	4-Chloroaniline	0.369	0.421	-14.1	94	0.00
31 C	Hexachlorobutadiene	0.266	0.287	-7.9	91	0.00
32	Caprolactam	0.126	0.122	3.2	83	0.00
33 C	4-Chloro-3-methylphenol	0.406	0.420	-3.4	89	0.00
34	2-Methylnaphthalene	0.766	0.787	-2.7	90	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
36	1,2,4,5-Tetrachlorobenzene	0.590	0.641	-8.6	95	0.00
37	Hexachlorocyclopentadiene	0.432	0.438	-1.4	89	0.00
38 C	2,4,6-Trichlorophenol	0.382	0.396	-3.7	92	0.00
39	2,4,5-Trichlorophenol	0.420	0.442	-5.2	93	0.00
40	1,1'-Biphenyl	1.295	1.323	-2.2	88	0.00
41	2-Chloronaphthalene	1.015	1.052	-3.6	91	0.00
42	2-Nitroaniline	0.386	0.424	-9.8	95	0.00
43 S	Dimethylphthalate-d6	1.396	1.399	-0.2	89	0.00
44	Dimethylphthalate	1.369	1.434	-4.7	91	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101716\
 Data File : BG024415.D
 Acq On : 18 Oct 2016 11:40
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02033

Quant Time: Oct 18 12:26:30 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 18 03:43:57 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.266	0.296	-11.3	95	0.00
46 S	Acenaphthylene-d8	1.600	1.650	-3.1	89	0.00
47	Acenaphthylene	1.539	1.583	-2.9	90	0.00
48	3-Nitroaniline	0.251	0.261	-4.0	91	0.00
49 C	Acenaphthene	1.100	1.116	-1.5	90	0.00
50	2,4-Dinitrophenol	0.178	0.183	-2.8	96	0.00
51 S	4-Nitrophenol-d4	0.248	0.238	4.0	86	0.00
52	4-Nitrophenol	0.300	0.303	-1.0	87	0.00
53	Dibenzofuran	1.619	1.643	-1.5	88	0.00
54	2,4-Dinitrotoluene	0.394	0.428	-8.6	94	0.00
55	2,3,4,6-Tetrachlorophenol	0.413	0.443	-7.3	94	0.00
56	Diethylphthalate	1.443	1.451	-0.6	86	0.00
57 S	Fluorene-d10	1.320	1.368	-3.6	89	0.00
58	Fluorene	1.339	1.352	-1.0	87	0.00
59	4-Chlorophenyl-phenylether	0.743	0.754	-1.5	89	0.00
60	4-Nitroaniline	0.289	0.280	3.1	86	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.120	0.126	-5.0	97	0.00
63	4,6-Dinitro-2-methylphenol	0.126	0.133	-5.6	92	0.00
64	N-Nitrosodiphenylamine	0.566	0.575	-1.6	86	0.00
65	4-Bromophenyl-phenylether	0.232	0.247	-6.5	94	0.00
66	Hexachlorobenzene	0.258	0.275	-6.6	93	0.00
67	Atrazine	0.231	0.247	-6.9	91	0.00
68 C	Pentachlorophenol	0.157	0.158	-0.6	88	0.00
69	Phenanthrene	1.010	1.028	-1.8	86	0.00
70 S	Anthracene-d10	0.903	0.917	-1.6	85	0.00
71	Anthracene	1.028	1.035	-0.7	85	0.00
72	Carbazole	0.807	0.865	-7.2	82	0.00
73	Di-n-butylphthalate	0.986	1.046	-6.1	88	0.00
74 C	Fluoranthene	1.020	1.195	-17.2	87	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
76 S	Pyrene-d10	0.933	0.974	-4.4	87	0.00
77	Pyrene	1.068	1.109	-3.8	87	0.00
78	Butylbenzylphthalate	0.429	0.432	-0.7	86	0.00
79	3,3'-Dichlorobenzidine	0.372	0.418	-12.4	94	0.00
80	Benzo(a)anthracene	1.110	1.171	-5.5	90	0.00
81	Bis(2-ethylhexyl)phthalate	0.607	0.611	-0.7	86	0.00
82	Chrysene	1.057	1.093	-3.4	87	0.00
83 I	Perylene-d12	1.000	1.000	0.0	96	0.00
84	Di-n-octyl phthalate	0.938	0.990	-5.5	87	0.00
85	Benzo(b)fluoranthene	1.110	1.112	-0.2	91	0.00
86	Benzo(k)fluoranthene	1.070	1.085	-1.4	91	0.00
87 S	Benzo(a)pyrene-d12	0.900	0.906	-0.7	91	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101716\
Data File : BG024415.D
Acq On : 18 Oct 2016 11:40
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02033

Quant Time: Oct 18 12:26:30 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 18 03:43:57 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.058	1.066	-0.8	91	0.00
89	Indeno(1,2,3-cd)pyrene	1.253	1.323	-5.6	95	0.00
90	Dibenzo(a,h)anthracene	1.059	1.106	-4.4	95	-0.01
91	Benzo(a,h,i)perylene	1.053	1.072	-1.8	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0