

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101716\
 Data File : BG024378.D
 Acq On : 17 Oct 2016 11:56
 Operator : UM/SJ
 Sample : H5239-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BC746

Quant Time: Oct 17 12:59:17 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 15 05:09:35 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.28	152	154095	20.00	ng/ul	0.00
18) Naphthalene-d8	11.11	136	718608	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.90	164	616447	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.64	188	1333946	20.00	ng/ul	0.00
75) Chrysene-d12	21.94	240	1522099	20.00	ng/ul	0.00
83) Perylene-d12	25.38	264	1531100	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.64	96	3990	1.23	ng/uL	0.00
5) Phenol-d5	7.41	99	73045	5.08	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.60	67	199343	20.51	ng/ul	0.00
9) 2-Chlorophenol-d4	7.81	132	170933	17.27	ng/ul	0.00
13) 4-Methylphenol-d8	8.97	113	137644	11.84	ng/ul	0.00
19) Nitrobenzene-d5	9.45	128	119346	23.40	ng/ul	0.00
22) 2-Nitrophenol-d4	10.18	143	130801	22.93	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.72	165	238254	19.17	ng/ul	0.00
29) 4-Chloroaniline-d4	11.25	131	1462	0.11	ng/ul	0.00
43) Dimethylphthalate-d6	14.29	166	967090	22.47	ng/ul	0.00
46) Acenaphthylene-d8	14.59	160	1103271	22.38	ng/ul	0.00
51) 4-Nitrophenol-d4	15.06	143	39514	5.18	ng/ul	0.00
57) Fluorene-d10	15.89	176	935856	23.00	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.99	200	200239	24.98	ng/ul	0.00
70) Anthracene-d10	17.74	188	1554808	25.80	ng/ul	0.00
76) Pyrene-d10	20.01	212	1822057	25.65	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.15	264	1893052	27.48	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101716\
Data File : BG024378.D
Acq On : 17 Oct 2016 11:56
Operator : UM/SJ
Sample : H5239-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BC746

Quant Time: Oct 17 12:59:17 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 15 05:09:35 2016
Response via : Initial Calibration

