

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101717\
 Data File : BG029316.D
 Acq On : 18 Oct 2017 00:49
 Operator : SJ/JU
 Sample : I5867-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Oct 18 04:07:29 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 17:32:15 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	71408	20.00	ng	0.00
21) Naphthalene-d8	10.98	136	345660	20.00	ng	0.00
38) Acenaphthene-d10	14.79	164	223142	20.00	ng	0.00
63) Phenanthrene-d10	17.52	188	507405	20.00	ng	0.00
75) Chrysene-d12	21.79	240	532462	20.00	ng	0.00
86) Perylene-d12	25.10	264	535761	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.72	112	199476	46.67	ng	0.00
7) Phenol-d6	7.32	99	174903	29.15	ng	0.00
23) Nitrobenzene-d5	9.33	82	517738	95.18	ng	0.00
41) 2,4,6-Tribromophenol	16.27	330	263730	91.54	ng	0.00
44) 2-Fluorobiphenyl	13.42	172	1349961	88.73	ng	0.00
78) Terphenyl-d14	20.11	244	1971704	84.24	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101717\
Data File : BG029316.D
Acq On : 18 Oct 2017 00:49
Operator : SJ/JU
Sample : I5867-02
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :

Quant Time: Oct 18 04:07:29 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 17:32:15 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 8.17 min Scan# 831
Delta R.T. -0.00 min
Lab File: BG029316.D
Acq: 18 Oct 2017 00:49

Instrument :
BNA_G
ClientSampleId :

Tot Ion:152 Resp: 71408
Ion Ratio Lower Upper
152 100
150 154.5 126.6 189.8
115 58.7 53.0 79.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 46.67 ng
RT: 5.72 min Scan# 414
Delta R.T. -0.00 min
Lab File: BG029316.D
Acq: 18 Oct 2017 00:49

Tgt Ion:112 Resp: 199476
Ion Ratio Lower Upper
112 100
64 59.5 56.3 84.5
63 31.4 30.1 45.1



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 63.00 (62.70 to 63.70): BGC





#7

Phenol-d6

Concen: 29.15 ng

RT: 7.32 min Scan# 686

Delta R.T. -0.01 min

Lab File: BG029316.D

Acq: 18 Oct 2017 00:49

Instrument :

BNA_G

ClientSampleId :


Tot Ion: 99 Resp: 174903

Ion Ratio Lower Upper




99 100

42 19.0 14.1 21.1

71 30.4 26.1 39.1



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 10.98 min Scan# 1310

Delta R.T. -0.01 min

Lab File: BG029316.D

Acq: 18 Oct 2017 00:49

Tgt Ion: 136 Resp: 345660


Ion Ratio Lower Upper

136 100



137 11.5 9.2 13.8


54 8.9 10.3 15.5#

68 6.4 4.5 6.7



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
Nitrobenzene-d5
Concen: 95.18 ng
RT: 9.33 min Scan# 1029
Delta R.T. -0.01 min
Lab File: BG029316.D
Acq: 18 Oct 2017 00:49




Instrument :
BNA_G
ClientSampled :


Tot Ion: 82 Resp: 517738
Ion Ratio Lower Upper
82 100
128 45.8 29.7 44.5#
54 52.4 43.1 64.7



#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.79 min Scan# 1957
Delta R.T. -0.01 min
Lab File: BG029316.D
Acq: 18 Oct 2017 00:49

Tgt Ion: 164 Resp: 223142
Ion Ratio Lower Upper
164 100
162 104.3 78.8 118.2
160 45.9 35.4 53.0









#41
 2,4,6-Tribromophenol
 Concen: 91.54 ng
 RT: 16.27 min Scan# 2209
 Delta R.T. -0.01 min
 Lab File: BG029316.D
 Acq: 18 Oct 2017 00:49

Instrument :
 BNA_G
 ClientSampleId :

Tot Ion:330 Resp: 263730
 Ion Ratio Lower Upper
 330 100
 332 96.1 77.0 115.6
 141 34.1 25.2 37.8




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 88.73 ng
 RT: 13.42 min Scan# 1724
 Delta R.T. -0.00 min
 Lab File: BG029316.D
 Acq: 18 Oct 2017 00:49

Tgt Ion:172 Resp: 1349961
 Ion Ratio Lower Upper
 172 100
 171 39.1 30.0 45.0
 170 25.7 19.5 29.3



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.52 min Scan# 2423
Delta R.T. -0.01 min
Lab File: BG029316.D
Acq: 18 Oct 2017 00:49

Instrument :
BNA_G
ClientSampleId :

Tot Ion:188 Resp: 507405
Ion Ratio Lower Upper
188 100
94 7.9 6.9 10.3
80 9.0 7.7 11.5




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BGC
Ion 80.00 (79.70 to 80.70): BGC






#75
Chrysene-d12
Concen: 20.00 ng
RT: 21.79 min Scan# 3150
Delta R.T. -0.01 min
Lab File: BG029316.D
Acq: 18 Oct 2017 00:49

Tgt Ion:240 Resp: 532462
Ion Ratio Lower Upper
240 100
120 9.5 6.8 10.2
236 26.9 20.9 31.3



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#78

Terphenyl-d14

Concen: 84.24 ng

RT: 20.11 min Scan# 2864

Delta R.T. -0.01 min

Lab File: BG029316.D

Acq: 18 Oct 2017 00:49

Instrument :

BNA_G

ClientSampleId :


Tot Ion:244 Resp: 1971704

Ion Ratio Lower Upper

244 100

212 8.3 5.8 8.8




122 10.3 7.0 10.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#86

Perylene-d12

Concen: 20.00 ng

RT: 25.10 min Scan# 3713

Delta R.T. -0.01 min

Lab File: BG029316.D

Acq: 18 Oct 2017 00:49


Tgt Ion:264 Resp: 535761

Ion Ratio Lower Upper

264 100

260 24.3 17.4 26.0

265 20.2 17.7 26.5



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

