

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101717\
 Data File : BG029313.D
 Acq On : 17 Oct 2017 22:50
 Operator : SJ/JU
 Sample : I5817-06
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 HW-3

Quant Time: Oct 18 00:56:02 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 17:32:15 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	66521	20.00	ng	0.00
21) Naphthalene-d8	10.99	136	322898	20.00	ng	0.00
38) Acenaphthene-d10	14.79	164	213706	20.00	ng	0.00
63) Phenanthrene-d10	17.52	188	480534	20.00	ng	0.00
75) Chrysene-d12	21.80	240	498750	20.00	ng	0.00
86) Perylene-d12	25.10	264	509451	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.72	112	174212	43.75	ng	0.00
7) Phenol-d6	7.32	99	152340	27.26	ng	0.00
23) Nitrobenzene-d5	9.33	82	499738	98.35	ng	0.00
41) 2,4,6-Tribromophenol	16.27	330	255625	92.65	ng	0.00
44) 2-Fluorobiphenyl	13.41	172	1318868	90.51	ng	0.00
78) Terphenyl-d14	20.12	244	1848228	84.30	ng	0.00
Target Compounds						
49) Dimethylphthalate	14.23	163	34862	2.091	ng	Qvalue # 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101717\
Data File : BG029313.D
Acq On : 17 Oct 2017 22:50
Operator : SJ/JU
Sample : I5817-06
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
HW-3

Quant Time: Oct 18 00:56:02 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 17:32:15 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.17 min Scan# 831
Delta R.T. 0.00 min
Lab File: BG029313.D
Acq: 17 Oct 2017 22:50

Instrument :
BNA_G
ClientSampleId :
HW-3

Tot Ion:152 Resp: 66521
Ion Ratio Lower Upper
152 100
150 157.9 126.6 189.8
115 56.8 53.0 79.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 43.750 ng
RT: 5.72 min Scan# 414
Delta R.T. 0.00 min
Lab File: BG029313.D
Acq: 17 Oct 2017 22:50

Tgt Ion:112 Resp: 174212
Ion Ratio Lower Upper
112 100
64 59.5 56.3 84.5
63 30.6 30.1 45.1



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 63.00 (62.70 to 63.70): BGC





#7

Phenol-d6

Concen: 27.256 ng

RT: 7.32 min Scan# 686

Delta R.T. -0.01 min

Lab File: BG029313.D

Acq: 17 Oct 2017 22:50

Instrument :

BNA_G

ClientSampled :

HW-3


Tot Ion: 99 Resp: 152340

Ion Ratio Lower Upper




99 100

42 18.3 14.1 21.1

71 32.8 26.1 39.1



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.99 min Scan# 1310

Delta R.T. -0.01 min

Lab File: BG029313.D

Acq: 17 Oct 2017 22:50

Tgt Ion: 136 Resp: 322898


Ion Ratio Lower Upper

136 100



137 12.1 9.2 13.8


54 8.6 10.3 15.5#

68 6.0 4.5 6.7



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
 Nitrobenzene-d5
 Concen: 98.350 ng
 RT: 9.33 min Scan# 1029
 Delta R.T. -0.01 min
 Lab File: BG029313.D
 Acq: 17 Oct 2017 22:50




Instrument :
 BNA_G
 ClientSampleId :
 HW-3


Tot Ion: 82 Resp: 499738
 Ion Ratio Lower Upper
 82 100
 128 44.6 29.7 44.5#
 54 52.0 43.1 64.7



#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.79 min Scan# 1957
 Delta R.T. -0.01 min
 Lab File: BG029313.D
 Acq: 17 Oct 2017 22:50

Tgt Ion:164 Resp: 213706
 Ion Ratio Lower Upper
 164 100
 162 104.6 78.8 118.2
 160 45.3 35.4 53.0









#41
 2,4,6-Tribromophenol
 Concen: 92.646 ng
 RT: 16.27 min Scan# 2209
 Delta R.T. -0.01 min
 Lab File: BG029313.D
 Acq: 17 Oct 2017 22:50

Instrument :
 BNA_G
 ClientSampleId :
 HW-3

Tot Ion	Ratio	Lower	Upper
330	100		
332	95.5	77.0	115.6
141	33.7	25.2	37.8




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 90.513 ng
 RT: 13.41 min Scan# 1723
 Delta R.T. -0.01 min
 Lab File: BG029313.D
 Acq: 17 Oct 2017 22:50

Tgt Ion	Ratio	Lower	Upper
172	100		
171	37.7	30.0	45.0
170	25.2	19.5	29.3



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#49

Dimethylphthalate

Concen: 2.091 ng

RT: 14.23 min Scan# 1862

Delta R.T. -0.01 min

Lab File: BG029313.D

Acq: 17 Oct 2017 22:50

Instrument :

BNA_G

ClientSampleId :

HW-3


Tot Ion:163 Resp: 34862

Ion Ratio Lower Upper

163 100

194 5.2 3.4 5.2#




164 10.7 8.2 12.4



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.52 min Scan# 2423

Delta R.T. -0.01 min

Lab File: BG029313.D

Acq: 17 Oct 2017 22:50


Tgt Ion:188 Resp: 480534

Ion Ratio Lower Upper

188 100

94 7.8 6.9 10.3



80 9.2 7.7 11.5




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC





#75

Chrysene-d12

Concen: 20.000 ng

RT: 21.80 min Scan# 3150

Delta R.T. -0.01 min

Lab File: BG029313.D

Acq: 17 Oct 2017 22:50

Instrument :

BNA_G

ClientSampleId :

HW-3


Tot Ion:240 Resp: 498750

Ion Ratio Lower Upper

240 100

120 8.8 6.8 10.2




236 27.0 20.9 31.3



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#78

Terphenyl-d14

Concen: 84.303 ng

RT: 20.12 min Scan# 2864

Delta R.T. -0.01 min

Lab File: BG029313.D

Acq: 17 Oct 2017 22:50


Tgt Ion:244 Resp: 1848228

Ion Ratio Lower Upper

244 100

212 8.6 5.8 8.8



122 10.3 7.0 10.4






Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#86
Pervlene-d12
Concen: 20.000 ng
RT: 25.10 min Scan# 3713
Delta R.T. -0.01 min
Lab File: BG029313.D
Acq: 17 Oct 2017 22:50

Instrument :
BNA_G
ClientSampleId :
HW-3

Tot Ion	Ratio	Lower	Upper
264	100		
260	24.7	17.4	26.0
265	22.5	17.7	26.5

