

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101717\
 Data File : BG029325.D
 Acq On : 18 Oct 2017 9:15
 Operator : SJ/JU
 Sample : I5868-04
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 101117-TIFR-E-D-M

Quant Time: Oct 18 15:38:15 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 17:32:15 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.16	152	66384	20.00	ng	0.00
21) Naphthalene-d8	10.99	136	326918	20.00	ng	0.00
38) Acenaphthene-d10	14.79	164	219673	20.00	ng	0.00
63) Phenanthrene-d10	17.52	188	504038	20.00	ng	0.00
75) Chrysene-d12	21.80	240	528383	20.00	ng	0.00
86) Perylene-d12	25.10	264	532556	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.72	112	210298	52.92	ng	0.00
7) Phenol-d6	7.32	99	186400	33.42	ng	0.00
23) Nitrobenzene-d5	9.33	82	515490	100.20	ng	0.00
41) 2,4,6-Tribromophenol	16.27	330	222916	78.60	ng	0.00
44) 2-Fluorobiphenyl	13.41	172	1394111	93.08	ng	0.00
78) Terphenyl-d14	20.12	244	2172548	93.54	ng	0.00


Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101717\
Data File : BG029325.D
Acq On : 18 Oct 2017 9:15
Operator : SJ/JU
Sample : I5868-04
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
101117-TIFR-E-D-M

Quant Time: Oct 18 15:38:15 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 17:32:15 2017
Response via : Initial Calibration






#1



1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.16 min Scan# 830
Delta R.T. -0.01 min
Lab File: BG029325.D
Acq: 18 Oct 2017 9:15

Instrument :
BNA_G
ClientSampleId :
101117-TIFR-E-D-M

Tot Ion:152 Resp: 66384
Ion Ratio Lower Upper
152 100
150 158.0 126.6 189.8
115 58.2 53.0 79.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 52.921 ng
RT: 5.72 min Scan# 414
Delta R.T. 0.00 min
Lab File: BG029325.D
Acq: 18 Oct 2017 9:15

Tgt Ion:112 Resp: 210298
Ion Ratio Lower Upper
112 100
64 58.2 56.3 84.5
63 30.3 30.1 45.1



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 63.00 (62.70 to 63.70): BGC





#7

Phenol-d6

Concen: 33.418 ng

RT: 7.32 min Scan# 686

Delta R.T. -0.01 min

Lab File: BG029325.D

Acq: 18 Oct 2017 9:15

Instrument :

BNA_G

ClientSampleId :

101117-TIFR-E-D-M


Tot Ion: 99 Resp: 186400

Ion Ratio Lower Upper




99 100

42 18.0 14.1 21.1

71 33.8 26.1 39.1



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.99 min Scan# 1310

Delta R.T. -0.01 min

Lab File: BG029325.D

Acq: 18 Oct 2017 9:15

Tgt Ion: 136 Resp: 326918


Ion Ratio Lower Upper

136 100



137 11.7 9.2 13.8

54 8.5 10.3 15.5#

68 6.0 4.5 6.7



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
Nitrobenzene-d5
Concen: 100.202 ng
RT: 9.33 min Scan# 1029
Delta R.T. -0.01 min
Lab File: BG029325.D
Acq: 18 Oct 2017 9:15




Instrument :
BNA_G
ClientSampleId :
101117-TIFR-E-D-M


Tot Ion	Ratio	Lower	Upper
82	100		
128	46.1	29.7	44.5#
54	51.5	43.1	64.7



#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.79 min Scan# 1957
Delta R.T. -0.01 min
Lab File: BG029325.D
Acq: 18 Oct 2017 9:15

Tgt Ion	Ratio	Lower	Upper
164	100		
162	100.2	78.8	118.2
160	43.7	35.4	53.0








#41
 2,4,6-Tribromophenol
 Concen: 78.597 ng
 RT: 16.27 min Scan# 2209
 Delta R.T. -0.01 min
 Lab File: BG029325.D
 Acq: 18 Oct 2017 9:15

Instrument :
 BNA_G
 ClientSampleId :
 101117-TIFR-E-D-M

Tot Ion	Ratio	Lower	Upper
330	100		
332	95.1	77.0	115.6
141	33.3	25.2	37.8




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 93.078 ng
 RT: 13.41 min Scan# 1723
 Delta R.T. -0.01 min
 Lab File: BG029325.D
 Acq: 18 Oct 2017 9:15

Tgt Ion	Ratio	Lower	Upper
172	100		
171	38.0	30.0	45.0
170	25.7	19.5	29.3



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.52 min Scan# 2423

Delta R.T. -0.01 min

Lab File: BG029325.D

Acq: 18 Oct 2017 9:15

Instrument :

BNA_G

ClientSampleId :

101117-TIFR-E-D-M


Tot Ion:188 Resp: 504038

Ion Ratio Lower Upper

188 100

94 8.6 6.9 10.3




80 9.5 7.7 11.5



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC



#75

Chrysene-d12

Concen: 20.000 ng

RT: 21.80 min Scan# 3150

Delta R.T. -0.01 min

Lab File: BG029325.D

Acq: 18 Oct 2017 9:15


Tgt Ion:240 Resp: 528383

Ion Ratio Lower Upper

240 100

120 8.9 6.8 10.2



236 25.2 20.9 31.3



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#78

Terphenyl-d14

Concen: 93.538 ng

RT: 20.12 min Scan# 2864

Delta R.T. -0.01 min

Lab File: BG029325.D

Acq: 18 Oct 2017 9:15

Instrument :

BNA_G

ClientSampleId :

101117-TIFR-E-D-M


Tot Ion:244 Resp: 2172548

Ion Ratio Lower Upper

244 100

212 8.6 5.8 8.8




122 10.2 7.0 10.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 25.10 min Scan# 3713

Delta R.T. -0.01 min

Lab File: BG029325.D

Acq: 18 Oct 2017 9:15


Tgt Ion:264 Resp: 532556

Ion Ratio Lower Upper

264 100

260 24.6 17.4 26.0

265 22.5 17.7 26.5



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

