

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG101720\
 Data File : BG046972.D
 Acq On : 18 Oct 2020 13:07
 Operator : CG/JU
 Sample : L4402-07
 Misc : GCMS CONFIRMATION
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0AB8

Quant Time: Oct 19 05:03:36 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG101520MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 19 04:55:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.06	152	68935	20.00	ng/ul	0.00
18) Naphthalene-d8	10.88	136	260187	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.70	164	175668	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.44	188	344805	20.00	ng/ul	0.00
78) Chrysene-d12	21.71	240	377454	20.00	ng/ul	0.00
86) Perylene-d12	24.95	264	378500	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL	
5) Phenol-d5	0.00	99	0	0.00	ng/ul	
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0	0.00	ng/ul	
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul	
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul	
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul	
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul	
29) 4-Chloroaniline-d4	10.89	131	262	0.05	ng/ul	-0.12
44) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul	
47) Acenaphthylene-d8	0.00	160	0	0.00	ng/ul	
52) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
58) Fluorene-d10	15.71	176	168	0.01	ng/ul	0.02
63) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul	
71) Anthracene-d10	17.44	188	345057	20.66	ng/ul	-0.10
79) Pyrene-d10	19.86	212	107	0.01	ng/ul	0.04
90) Benzo(a)pyrene-d12	24.77	264	65	0.00	ng/ul	0.04


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
70) Phenanthrene	17.48	178	21957	1.113	ng/ul#	95
77) Fluoranthene	19.49	202	27292	1.133	ng/ul#	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG101720\
Data File : BG046972.D
Acq On : 18 Oct 2020 13:07
Operator : CG/JU
Sample : L4402-07
Misc : GCMS CONFIRMATION
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0AB8

Quant Time: Oct 19 05:03:36 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG101520MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 19 04:55:35 2020
Response via : Initial Calibration





#70

Phenanthrene

Concen: 1.113 ng/ul

RT: 17.48 min Scan# 2407

Delta R.T. 0.00 min

Lab File: BG046972.D

Acq: 18 Oct 2020 13:07

Instrument :

BNA_G

ClientSampleId :

C0AB8


Tot Ion:178 Resp: 21957

Ion Ratio Lower Upper

178 100

179 11.5 12.5 18.7#




176 17.9 15.0 22.4



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#77

Fluoranthene

Concen: 1.133 ng/ul

RT: 19.49 min Scan# 2748

Delta R.T. -0.00 min

Lab File: BG046972.D

Acq: 18 Oct 2020 13:07


Tgt Ion:202 Resp: 27292

Ion Ratio Lower Upper

202 100

101 7.4 4.9 7.3#

100 6.0 4.6 6.8



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): E

