

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG101816\
 Data File : BG024431.D
 Acq On : 18 Oct 2016 22:34
 Operator : UM/SJ
 Sample : H5223-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 BDNH5

Quant Time: Oct 19 02:11:02 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG0100716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 19 02:09:29 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.28	152	174820	20.00	ng/ul	0.00
18) Naphthalene-d8	11.11	136	842880	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.90	164	709487	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.64	188	1444724	20.00	ng/ul	0.00
75) Chrysene-d12	21.94	240	1645525	20.00	ng/ul	0.00
83) Perylene-d12	25.38	264	1757978	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.64	96	10595	2.88	ng/uL	0.00
5) Phenol-d5	7.41	99	308623	18.92	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.60	67	189371	17.17	ng/ul	0.00
9) 2-Chlorophenol-d4	7.81	132	205814	18.32	ng/ul	0.00
13) 4-Methylphenol-d8	8.97	113	247702	18.79	ng/ul	0.00
19) Nitrobenzene-d5	9.45	128	110412	18.46	ng/ul	0.00
22) 2-Nitrophenol-d4	10.19	143	135545	20.25	ng/ul	0.01
26) 2,4-Dichlorophenol-d3	10.72	165	283304	19.44	ng/ul	0.00
29) 4-Chloroaniline-d4	11.24	131	270919	17.69	ng/ul	0.00
43) Dimethylphthalate-d6	14.29	166	960088	19.38	ng/ul	0.00
46) Acenaphthylene-d8	14.59	160	1127736	19.87	ng/ul	0.00
51) 4-Nitrophenol-d4	15.06	143	164482	18.73	ng/ul	0.00
57) Fluorene-d10	15.89	176	907913	19.38	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.99	200	177337	20.42	ng/ul	0.00
70) Anthracene-d10	17.74	188	1333671	20.44	ng/ul	0.00
76) Pyrene-d10	20.01	212	1584329	20.63	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.14	264	1672544	21.15	ng/ul	0.00

Target Compounds						Qvalue
6) Phenol	7.44	94	27815	1.68	ng/ul	95
44) Dimethylphthalate	14.34	163	614686	12.65	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG101816\
Data File : BG024431.D
Acq On : 18 Oct 2016 22:34
Operator : UM/SJ
Sample : H5223-01
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
BDNH5

Quant Time: Oct 19 02:11:02 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG0100716.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 19 02:09:29 2016
Response via : Initial Calibration

