


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG101816\
Data File : BG024416.D
Acq On : 18 Oct 2016 12:56
Operator : UM/SJ
Sample : DFTPP08
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP08

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG0100716.M
Title : SVOA CALIBRATION
Last Update : Wed Oct 19 15:18:40 2016



AutoFind: Scans 2586, 2587, 2588; Background Corrected with Scan 2578

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.3	280810	PASS
68	69	0.00	2	0.6	1813	PASS
69	198	0.00	100	40.8	284354	PASS
70	69	0.00	2	0.6	1677	PASS
127	198	10	80	44.3	308437	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	696853	PASS
199	198	5	9	7.6	53133	PASS
275	198	10	60	34.7	241920	PASS
365	198	1	100	5.0	34714	PASS
441	442	0.01	100	17.7	119533	PASS
442	198	50	100	96.7	673941	PASS
443	442	15	24	19.7	133032	PASS

m/z	Abundance
37.10	1888.0
38.10	2735.0
39.10	27320.0
40.10	1305.0
41.10	775.0
42.10	309.0
44.00	2876.0
45.00	465.0
46.20	226.0
47.10	643.0
49.10	4090.0
50.10	86768.0
51.10	344896.0
52.10	17800.0
53.00	777.0
55.10	1773.0
56.00	10529.0
57.00	24200.0
58.10	1058.0
61.00	5337.0
62.10	5973.0
63.00	16824.0
64.00	2280.0
65.00	7318.0
65.80	751.0
66.10	893.0
67.20	394.0
69.00	348992.0
70.10	2493.0
70.90	311.0
73.00	3854.0
74.10	37672.0
75.10	52560.0
76.10	16256.0
77.10	380736.0
78.10	30904.0
79.00	25464.0
80.00	22848.0
81.00	29128.0
82.10	5938.0
83.00	5210.0
84.10	1279.0
85.00	5062.0
86.00	8353.0
87.00	4521.0
88.10	1512.0
89.20	313.0
91.00	6786.0
92.00	5491.0
93.00	51520.0
94.00	4582.0
94.90	481.0
96.00	1401.0
96.90	714.0
98.00	35368.0
99.00	25832.0
100.00	2904.0
101.00	18248.0
102.00	1995.0
103.00	6183.0
104.00	8352.0
104.90	9509.0
107.00	128568.0
108.00	20712.0
110.00	192064.0
111.00	34056.0
112.10	3242.0
113.00	1162.0
114.20	364.0
116.00	7475.0
117.00	95136.0
118.00	4762.0
118.90	534.0
119.90	798.0
121.00	631.0
122.00	10403.0
123.00	10770.0
124.00	6588.0
125.00	4404.0
127.00	364032.0
128.00	28392.0
129.00	150592.0
130.00	14646.0
131.00	2402.0
132.10	944.0
132.90	167.0
134.00	5302.0
135.00	14724.0
135.90	5277.0
137.00	6329.0
138.10	1731.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

140.00	4076.0
141.00	21736.0
142.00	6539.0
143.00	5675.0
144.10	2201.0
145.00	1636.0
146.00	5268.0
147.00	12933.0
148.00	27864.0
149.00	5611.0
149.70	1091.0
150.20	715.0
151.00	2281.0
151.90	1269.0
153.00	9251.0
154.10	6409.0
155.00	14059.0
156.00	17824.0
157.00	3507.0
158.00	5397.0
159.00	2956.0
160.00	10323.0
161.00	10239.0
162.00	3483.0
162.70	462.0
163.20	841.0
164.00	1384.0
165.00	9231.0
166.00	7107.0
167.00	44592.0
168.00	18920.0
169.10	3166.0
170.10	2161.0
170.70	1394.0
171.90	4621.0
173.00	5134.0
174.00	8728.0
175.00	19768.0
176.10	4383.0
177.00	7143.0
177.80	1766.0
179.00	37248.0
180.00	22224.0
181.10	9999.0
182.00	1325.0
182.70	1057.0
184.00	2875.0
185.00	18040.0
186.00	134016.0
187.00	36104.0
188.00	4012.0
189.00	10608.0
190.00	1434.0
191.10	2874.0
192.00	12035.0
193.00	14099.0
194.00	2124.0
196.00	20912.0
198.00	762368.0
199.00	63368.0
200.00	5197.0
201.60	2904.0
203.00	6696.0
204.00	36600.0
205.00	60232.0
206.00	250496.0
207.00	35520.0
208.00	9670.0
208.90	3753.0
209.90	3264.0
211.00	8838.0
212.00	2170.0
213.10	1176.0
213.70	282.0
214.90	2925.0
216.10	6287.0
217.00	70784.0
217.90	9705.0
219.00	1141.0
221.00	28440.0
222.00	7369.0
223.00	17720.0
224.00	147904.0
225.00	34464.0
226.10	3280.0
227.00	60800.0
228.00	10903.0
229.00	12001.0
229.80	1598.0
230.10	1692.0
231.00	3778.0
232.00	653.0
233.10	1320.0
234.00	5192.0
235.00	4913.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

236.00	3996.0
237.00	5377.0
238.00	1077.0
239.10	2007.0
240.10	1812.0
241.00	3068.0
242.00	5166.0
243.00	7958.0
244.00	95456.0
245.00	13956.0
246.00	22656.0
247.00	4106.0
248.00	873.0
249.00	4125.0
250.00	1997.0
250.90	1460.0
251.70	1829.0
252.90	3994.0
255.00	563712.0
256.00	91776.0
256.90	6527.0
258.00	36344.0
258.90	6140.0
259.90	1491.0
261.00	1329.0
262.10	224.0
262.80	153.0
264.90	16720.0
266.00	2657.0
267.10	179.0
267.90	255.0
269.70	731.0
270.20	946.0
270.90	1830.0
272.10	3298.0
273.00	20048.0
274.00	40704.0
275.00	257728.0
276.00	37816.0
277.00	22376.0
277.90	4851.0
279.00	1068.0
281.00	619.0
282.00	985.0
283.00	2110.0
283.90	2212.0
285.00	4139.0
286.10	951.0
287.90	269.0
289.00	1418.0
290.10	866.0
291.00	577.0
292.10	2256.0
292.90	4667.0
293.80	1497.0
296.00	79496.0
297.00	12887.0
298.00	1663.0
299.00	471.0
301.00	1310.0
302.10	1978.0
303.00	8955.0
303.90	2016.0
305.80	288.0
308.00	2052.0
308.70	396.0
309.90	881.0
310.90	265.0
311.70	265.0
312.20	417.0
313.00	1302.0
314.00	2868.0
315.00	9536.0
316.00	4535.0
317.00	968.0
320.10	755.0
321.00	2928.0
322.00	1860.0
323.00	20248.0
324.00	5029.0
325.90	232.0
326.90	4122.0
327.90	1742.0
328.50	242.0
329.10	209.0
330.10	198.0
332.00	2570.0
333.00	2555.0
334.00	16512.0
335.00	4822.0
336.00	512.0
338.30	433.0
339.00	243.0
340.10	530.0
341.00	2271.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

342.20	1018.0
343.10	216.0
345.90	5292.0
347.10	1647.0
347.80	619.0
348.70	522.0
351.00	382.0
352.00	6636.0
353.00	4951.0
354.00	6847.0
355.00	897.0
356.10	279.0
359.00	348.0
361.50	336.0
362.50	179.0
364.90	33968.0
365.90	4625.0
367.10	633.0
370.00	936.0
372.00	12007.0
373.00	4138.0
373.90	205.0
377.20	268.0
381.70	359.0
382.90	3127.0
384.00	928.0
384.90	608.0
390.00	895.0
391.10	1886.0
391.80	575.0
397.10	201.0
400.70	616.0
402.00	6602.0
403.00	5688.0
403.80	1149.0
404.00	1132.0
405.10	592.0
410.30	233.0
412.50	750.0
413.40	174.0
414.40	326.0
414.70	246.0
415.00	258.0
418.20	527.0
421.00	4059.0
422.10	5723.0
423.00	42840.0
423.90	8643.0
425.00	888.0
428.40	384.0
428.60	384.0
431.30	170.0
432.70	627.0
433.90	155.0
434.40	318.0
436.20	394.0
437.10	566.0
438.80	475.0
439.10	331.0
439.30	351.0
439.90	831.0
441.00	96072.0
442.00	620928.0
443.00	130200.0
444.00	9314.0
444.90	1583.0
475.10	219.0
480.10	224.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
37.10	1572.0
38.10	3859.0
39.10	20560.0
40.10	1145.0
41.20	752.0
41.90	416.0
43.10	330.0
44.00	3441.0
45.10	937.0
47.00	546.0
49.10	2416.0
50.10	82704.0
51.10	324416.0
52.10	16424.0
53.20	478.0
55.00	2113.0
56.10	12291.0
57.10	21400.0
57.90	468.0
58.90	206.0
60.00	279.0
61.10	3571.0
62.00	4910.0
63.00	14483.0
64.10	1274.0
65.10	6601.0
66.00	706.0
67.10	923.0
68.00	5441.0
69.00	328896.0
70.00	1558.0
71.00	356.0
71.90	488.0
73.10	2413.0
74.10	30320.0
75.00	54568.0
76.10	18944.0
77.10	353792.0
78.10	26952.0
79.00	26672.0
80.00	22904.0
81.00	31584.0
82.00	7058.0
83.00	3884.0
84.00	930.0
85.10	3333.0
86.00	9337.0
87.10	2275.0
88.00	1418.0
89.10	412.0
91.00	6680.0
92.10	8704.0
93.00	42264.0
94.10	3989.0
94.90	172.0
96.00	2742.0
97.00	1527.0
98.00	35144.0
99.00	23120.0
100.10	2148.0
101.00	18624.0
101.90	820.0
103.00	5808.0
104.00	12314.0
105.00	10449.0
106.00	4105.0
107.00	121712.0
108.00	15694.0
110.00	189760.0
111.00	32104.0
112.10	3430.0
113.00	944.0
113.70	545.0
114.80	1205.0
116.10	7701.0
117.00	92632.0
118.10	5764.0
118.80	532.0
119.90	2340.0
121.00	566.0
122.00	8426.0
123.00	15343.0
123.90	5987.0
125.00	4427.0
127.00	359360.0
128.00	27856.0
129.00	161344.0
130.00	11842.0
131.00	2516.0
132.10	1692.0
133.00	652.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

134.00	5627.0
135.00	12893.0
136.00	8941.0
137.10	5531.0
138.10	1871.0
139.00	914.0
140.10	2476.0
141.00	26528.0
142.10	7861.0
143.00	3573.0
143.90	780.0
144.20	903.0
144.90	1020.0
145.20	927.0
145.90	3965.0
147.00	11174.0
148.00	25880.0
149.00	5477.0
150.00	1194.0
150.90	2820.0
152.00	898.0
153.00	8320.0
154.00	6056.0
155.00	12751.0
156.00	17096.0
156.90	3300.0
158.00	3476.0
159.00	4042.0
160.00	5267.0
161.00	10790.0
162.00	3489.0
163.20	458.0
163.90	643.0
165.00	9621.0
166.00	7217.0
167.00	34848.0
168.00	19776.0
169.00	3543.0
170.00	2287.0
171.00	1225.0
172.00	5375.0
173.00	6484.0
174.00	8627.0
175.00	20128.0
176.00	3785.0
177.00	9544.0
178.10	3399.0
179.00	41328.0
180.00	21768.0
181.00	12100.0
181.90	1386.0
183.00	1017.0
184.00	3991.0
185.10	18520.0
186.00	125224.0
187.00	40632.0
188.00	3795.0
189.00	9985.0
190.00	2418.0
191.00	4607.0
192.00	11625.0
193.00	13115.0
194.00	2904.0
195.00	2438.0
196.00	26080.0
198.00	812096.0
199.00	64704.0
200.00	4284.0
201.40	2504.0
203.00	8352.0
204.00	40952.0
205.00	64056.0
206.00	267200.0
207.00	38160.0
207.90	7578.0
209.00	4591.0
210.10	3719.0
211.00	10125.0
211.80	1936.0
213.00	1484.0
214.10	930.0
215.10	4951.0
216.00	7578.0
217.00	75040.0
218.00	10085.0
219.00	1161.0
219.90	796.0
221.00	30136.0
221.90	11341.0
223.00	20784.0
224.00	147328.0
225.00	45392.0
226.00	4634.0
227.00	76616.0
228.00	10861.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

229.00	14156.0
230.00	1629.0
231.00	5683.0
232.10	1221.0
232.90	1496.0
234.00	4348.0
235.00	5164.0
235.90	1868.0
237.00	5661.0
238.00	645.0
239.00	1560.0
240.00	2210.0
240.90	4873.0
241.90	10411.0
243.00	8713.0
244.00	112376.0
245.00	15954.0
246.00	25408.0
247.00	6355.0
247.90	1722.0
248.90	4040.0
249.90	1260.0
251.10	1483.0
252.00	1698.0
252.40	1121.0
253.10	3572.0
255.00	621696.0
256.00	103248.0
257.10	8955.0
258.00	37960.0
259.00	7248.0
260.00	874.0
260.80	937.0
261.90	579.0
263.00	1021.0
264.10	969.0
265.00	17544.0
265.90	2261.0
267.10	485.0
268.00	467.0
269.90	1071.0
270.90	1238.0
271.50	1568.0
272.10	1733.0
273.00	20560.0
274.00	51472.0
275.00	282752.0
276.00	31704.0
277.00	26168.0
277.90	4712.0
279.00	863.0
281.00	539.0
281.80	742.0
283.00	3051.0
284.00	4547.0
285.10	4544.0
286.00	913.0
288.80	795.0
289.70	404.0
290.00	542.0
290.80	505.0
292.00	2087.0
293.00	5714.0
293.80	1820.0
294.80	2590.0
296.00	90344.0
297.00	12966.0
297.90	930.0
299.10	267.0
301.00	1261.0
302.10	2074.0
303.00	10174.0
304.00	1892.0
305.10	442.0
308.10	2345.0
308.90	1164.0
309.90	894.0
310.90	222.0
312.10	712.0
312.90	625.0
314.00	5855.0
315.00	9620.0
316.10	5039.0
316.90	931.0
320.00	174.0
321.00	1985.0
322.10	617.0
323.00	27336.0
324.00	5176.0
325.00	900.0
326.00	691.0
326.90	5097.0
327.90	1971.0
331.90	2583.0
333.00	2593.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

334.00	14993.0
335.00	5607.0
335.90	632.0
339.00	413.0
340.20	269.0
341.00	2176.0
341.90	669.0
343.50	592.0
344.80	209.0
345.90	6331.0
346.90	1041.0
349.90	794.0
351.00	851.0
352.00	9040.0
352.90	6806.0
354.00	10296.0
354.90	1100.0
356.00	382.0
357.10	560.0
359.00	255.0
359.70	270.0
360.20	483.0
360.90	407.0
362.00	300.0
362.90	441.0
363.70	205.0
365.00	37472.0
365.90	4841.0
366.90	462.0
370.30	1131.0
370.90	2563.0
372.00	17224.0
372.90	2433.0
373.60	343.0
376.70	329.0
378.00	396.0
382.90	3873.0
384.00	835.0
384.90	453.0
388.80	260.0
390.00	1978.0
391.00	1541.0
391.90	1092.0
392.70	236.0
395.10	633.0
396.10	299.0
397.00	579.0
401.00	1265.0
402.00	4476.0
403.00	6992.0
404.10	3876.0
405.00	662.0
408.80	183.0
410.20	285.0
410.80	306.0
414.70	228.0
415.20	182.0
420.90	8335.0
422.00	6595.0
423.00	51872.0
424.00	8674.0
425.00	874.0
429.00	152.0
430.80	164.0
431.40	557.0
434.40	242.0
434.90	151.0
435.50	157.0
436.60	700.0
437.30	355.0
437.80	519.0
438.60	178.0
439.20	531.0
441.00	146496.0
442.00	774784.0
443.00	151808.0
444.00	14930.0
444.90	460.0
519.30	178.0
530.10	221.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
36.10	456.0
37.00	293.0
38.10	2044.0
39.10	11080.0
40.10	1352.0
41.10	919.0
44.10	3030.0
45.00	590.0
46.00	180.0
47.10	697.0
49.00	2487.0
50.10	40072.0
51.10	173120.0
52.10	7026.0
55.00	434.0
56.10	5179.0
57.10	12932.0
58.10	838.0
59.00	195.0
60.40	201.0
61.10	1426.0
62.10	3796.0
63.10	9319.0
64.10	1784.0
65.00	3945.0
66.10	688.0
67.00	429.0
69.00	176640.0
70.10	981.0
71.30	299.0
73.00	2045.0
74.10	17784.0
75.00	32688.0
76.10	9421.0
77.10	192768.0
78.00	11508.0
79.00	13120.0
80.00	10222.0
81.00	16177.0
82.00	3473.0
83.00	2982.0
83.90	863.0
85.00	2890.0
86.10	4533.0
87.00	3356.0
88.00	632.0
91.10	3846.0
92.00	2878.0
93.00	23152.0
94.00	1712.0
95.10	791.0
96.00	965.0
98.00	21032.0
99.00	13779.0
99.90	1016.0
101.00	8820.0
102.00	679.0
103.00	3631.0
104.10	5205.0
105.00	7928.0
107.00	66728.0
108.00	10152.0
110.00	103744.0
111.00	17528.0
112.00	1420.0
113.00	1890.0
114.70	510.0
116.00	4332.0
117.00	50072.0
117.90	4031.0
119.00	452.0
119.80	555.0
121.10	697.0
122.10	4807.0
123.00	7982.0
123.90	2957.0
125.00	2721.0
127.00	201920.0
128.00	13615.0
129.00	94128.0
130.00	5966.0
131.10	2501.0
131.70	1432.0
133.00	259.0
133.90	2295.0
135.10	6423.0
136.10	2452.0
137.10	1879.0
138.00	680.0
139.20	439.0
140.00	1592.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

141.00	11912.0
142.00	3524.0
144.10	999.0
145.20	316.0
146.00	3013.0
147.00	6016.0
148.00	18888.0
149.00	3753.0
149.80	1075.0
151.10	1824.0
151.90	1287.0
153.00	3352.0
154.00	3740.0
155.00	7780.0
156.00	13270.0
156.90	1001.0
157.80	1678.0
159.10	2034.0
160.10	4804.0
160.90	5186.0
161.90	1884.0
163.20	312.0
164.90	5934.0
166.10	2706.0
167.00	29064.0
168.00	10415.0
169.00	2630.0
170.00	1049.0
171.10	1567.0
172.00	2713.0
173.00	4580.0
174.00	4245.0
175.00	9415.0
175.90	1520.0
176.30	1043.0
177.00	5669.0
178.00	2411.0
178.20	2387.0
178.90	22224.0
180.00	15077.0
181.00	7798.0
181.90	153.0
183.10	513.0
184.10	1523.0
185.10	13427.0
186.00	76664.0
187.00	22888.0
187.90	1500.0
189.00	4379.0
190.20	644.0
191.10	2455.0
192.00	5721.0
193.00	8026.0
194.00	2286.0
196.00	14187.0
198.00	516096.0
199.00	31328.0
200.00	3880.0
201.10	1312.0
201.80	1642.0
203.00	4647.0
204.00	25168.0
205.00	40328.0
206.00	167808.0
207.00	20376.0
208.00	6442.0
209.00	2756.0
210.20	3374.0
210.90	7282.0
211.70	1597.0
211.90	1576.0
213.00	1191.0
215.00	3538.0
216.00	5074.0
217.00	51520.0
217.90	6836.0
219.10	986.0
219.90	472.0
221.00	22064.0
222.00	6740.0
223.00	12347.0
224.00	102080.0
225.00	26968.0
225.90	1828.0
227.00	51848.0
228.00	6435.0
228.90	9272.0
230.00	852.0
231.00	4460.0
232.20	605.0
232.80	934.0
233.90	2703.0
234.90	4098.0
235.70	857.0
236.10	940.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

237.00	3852.0
237.90	795.0
239.00	1494.0
240.00	1398.0
241.00	3876.0
242.00	5592.0
243.10	4569.0
244.00	79328.0
245.00	9622.0
246.00	17624.0
246.90	4349.0
248.00	1223.0
249.00	3143.0
250.10	774.0
250.90	392.0
252.10	863.0
253.00	2163.0
253.30	2016.0
255.00	404864.0
256.00	57280.0
257.00	3479.0
258.00	25856.0
259.00	4173.0
260.00	854.0
260.90	158.0
262.80	186.0
264.00	999.0
265.00	10036.0
265.90	1007.0
267.30	371.0
268.00	435.0
269.70	623.0
270.20	765.0
271.10	1257.0
272.00	1713.0
273.00	12897.0
274.00	34864.0
275.00	185280.0
276.00	25512.0
277.00	16864.0
278.00	4193.0
278.90	559.0
279.50	275.0
281.10	880.0
282.00	741.0
283.10	2790.0
284.00	1008.0
285.00	3504.0
286.00	432.0
288.90	1480.0
290.10	1088.0
290.70	474.0
291.90	668.0
292.90	3837.0
293.90	1415.0
295.00	1769.0
296.00	65096.0
297.00	13853.0
298.10	610.0
300.90	1237.0
301.80	1337.0
303.00	6665.0
304.00	2397.0
304.60	436.0
305.10	463.0
307.20	284.0
308.00	1047.0
308.90	313.0
309.90	779.0
311.00	152.0
312.00	342.0
312.90	423.0
314.00	3630.0
314.90	7158.0
316.00	4340.0
317.10	695.0
318.80	615.0
320.20	393.0
321.00	1759.0
322.00	634.0
323.00	19368.0
324.00	2950.0
324.90	542.0
327.00	4129.0
328.10	2066.0
329.00	672.0
331.90	1293.0
333.00	1552.0
334.00	11130.0
335.00	4252.0
336.00	738.0
336.90	201.0
338.80	932.0
340.00	541.0
341.00	1875.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

341.90	879.0
343.60	177.0
345.00	247.0
346.00	5321.0
346.90	1042.0
348.00	254.0
349.20	261.0
349.90	231.0
351.00	399.0
352.00	6479.0
352.90	3552.0
354.00	6252.0
354.80	1012.0
356.10	205.0
356.80	296.0
357.90	418.0
358.90	951.0
360.90	231.0
365.00	32704.0
366.00	2908.0
370.80	1470.0
372.00	9118.0
372.90	3991.0
374.20	169.0
376.60	450.0
383.00	3569.0
383.90	666.0
385.00	202.0
387.30	185.0
388.70	368.0
390.00	2358.0
390.90	963.0
392.00	1123.0
396.00	206.0
401.00	1014.0
402.00	4895.0
403.00	8080.0
404.00	1212.0
404.90	567.0
405.90	270.0
406.50	344.0
410.00	245.0
414.90	513.0
416.20	215.0
419.70	298.0
421.00	5615.0
422.00	7420.0
423.00	38072.0
424.00	6291.0
425.00	1507.0
425.80	592.0
427.70	188.0
429.50	361.0
430.50	192.0
432.70	166.0
433.50	315.0
434.40	187.0
435.10	335.0
437.40	218.0
438.70	650.0
439.20	475.0
439.80	1261.0
441.00	116032.0
442.00	626112.0
443.00	117088.0
444.00	10132.0
445.00	1319.0
466.00	187.0
473.80	785.0
475.30	207.0
500.50	193.0
515.80	201.0
529.40	216.0
542.60	251.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
35.10	257.0
36.90	315.0
40.10	189.0
44.10	3133.0
48.20	199.0
49.00	1768.0
53.90	183.0
55.30	403.0
55.90	196.0
57.00	335.0
58.80	238.0
61.20	403.0
64.20	163.0
65.70	516.0
67.10	330.0
69.20	488.0
70.90	207.0
73.00	294.0
76.80	336.0
81.10	162.0
84.00	861.0
84.80	335.0
85.80	260.0
91.50	245.0
98.10	918.0
102.50	154.0
102.90	251.0
106.10	153.0
109.10	313.0
115.00	198.0
120.20	346.0
121.60	158.0
123.10	236.0
125.10	213.0
131.50	226.0
133.20	339.0
133.80	178.0
136.80	188.0
137.30	223.0
145.80	264.0
146.30	252.0
148.80	436.0
152.10	313.0
162.00	159.0
177.00	199.0
189.00	217.0
190.80	237.0
206.90	1160.0
208.10	554.0
213.30	211.0
214.10	305.0
220.90	253.0
227.80	205.0
242.00	285.0
259.10	633.0
265.90	219.0
266.90	415.0
268.10	158.0
281.10	681.0
283.00	436.0
292.60	193.0
298.90	372.0
301.40	191.0
314.40	424.0
323.90	211.0
328.10	177.0
338.70	396.0
345.40	403.0
349.30	181.0
354.90	425.0
366.30	164.0
369.30	532.0
386.30	185.0
389.00	589.0
441.60	155.0
476.40	151.0
526.10	183.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08