

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG101818\
 Data File : BG037412.D
 Acq On : 18 Oct 2018 9:53
 Operator : JU/SJ
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Oct 18 12:19:05 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 12:15:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.11	152	53095	20.00	ng	0.00
21) Naphthalene-d8	10.93	136	247750	20.00	ng	0.00
39) Acenaphthene-d10	14.74	164	169944	20.00	ng	0.00
64) Phenanthrene-d10	17.49	188	406851	20.00	ng	0.00
76) Chrysene-d12	21.77	240	381994	20.00	ng	0.00
87) Perylene-d12	25.11	264	390137	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.64	112	63947	21.20	ng	0.00
7) Phenol-d6	7.24	99	97821	21.36	ng	0.00
23) Nitrobenzene-d5	9.28	82	88054	23.37	ng	0.00
42) 2,4,6-Tribromophenol	16.23	330	36286	21.77	ng	0.00
45) 2-Fluorobiphenyl	13.36	172	237781	21.01	ng	0.00
79) Terphenyl-d14	20.09	244	393946	21.65	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.53	88	16996	10.031	ng	92
3) Pyridine	3.94	79	39320	9.774	ng	93
4) n-Nitrosodimethylamine	3.86	42	14118	9.037	ng	# 82
6) Aniline	7.43	93	58651	9.784	ng	96
8) 2-Chlorophenol	7.67	128	38520	10.909	ng	98
9) Benzaldehyde	7.25	77	34290	10.876	ng	97
10) Phenol	7.27	94	52049	10.792	ng	93
11) bis(2-Chloroethyl)ether	7.52	93	39175	9.943	ng	91
12) 1,3-Dichlorobenzene	7.99	146	42580	10.258	ng	96
13) 1,4-Dichlorobenzene	8.15	146	43956	10.459	ng	97
14) 1,2-Dichlorobenzene	8.46	146	42393	10.410	ng	98
15) Benzyl Alcohol	8.34	79	34373	9.596	ng	97
16) 2,2'-oxybis(1-Chloropropan	8.63	45	70404	9.039	ng	97
17) 2-Methylphenol	8.53	107	32995	10.215	ng	97
18) Hexachloroethane	9.19	117	14337	9.995	ng	98
19) n-Nitroso-di-n-propylamine	8.90	70	31799	9.172	ng	94
20) 3+4-Methylphenols	8.87	107	49884	11.045	ng	91
22) Acetophenone	8.93	105	63791	10.214	ng	# 97
24) Nitrobenzene	9.33	77	46833	11.603	ng	99
25) Isophorone	9.84	82	78628	9.140	ng	97
26) 2-Nitrophenol	10.04	139	17820	12.408	ng	97
27) 2,4-Dimethylphenol	10.08	122	37786	10.509	ng	97
28) bis(2-Chloroethoxy)methane	10.33	93	53556	9.951	ng	95
29) 2,4-Dichlorophenol	10.57	162	41179	11.282	ng	98
30) 1,2,4-Trichlorobenzene	10.78	180	44538	10.027	ng	95
31) Naphthalene	10.98	128	125404	10.415	ng	98
32) Benzoic acid	10.14	122	17372	8.750	ng	95
33) 4-Chloroaniline	11.09	127	58959	10.439	ng	93
34) Hexachlorobutadiene	11.25	225	25815	9.359	ng	95
35) Caprolactam	11.85	113	16706	11.227	ng	88
36) 4-Chloro-3-methylphenol	12.18	107	47045	10.913	ng	97
37) 2-Methylnaphthalene	12.58	142	89855	10.226	ng	98
38) 1-Methylnaphthalene	12.79	142	89490	10.428	ng	98
40) 1,2,4,5-Tetrachlorobenzene	12.94	216	54371	9.879	ng	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG101818\
 Data File : BG037412.D
 Acq On : 18 Oct 2018 9:53
 Operator : JU/SJ
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampled :

Quant Time: Oct 18 12:19:05 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 12:15:58 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	12.91	237	21707	9.568	ng	98
43) 2,4,6-Trichlorophenol	13.17	196	36059	11.294	ng	94
44) 2,4,5-Trichlorophenol	13.24	196	39049	11.368	ng	95
46) 1,1'-Biphenyl	13.58	154	144077	10.306	ng	98
47) 2-Chloronaphthalene	13.62	162	107844	10.213	ng	99
48) 2-Nitroaniline	13.83	65	31083	12.181	ng	95
49) Acenaphthylene	14.47	152	166694	10.319	ng	98
50) Dimethylphthalate	14.19	163	133827	10.020	ng	100
51) 2,6-Dinitrotoluene	14.32	165	28951	12.011	ng	95
52) Acenaphthene	14.80	154	101224	10.145	ng	98
53) 3-Nitroaniline	14.65	138	30775	11.013	ng	# 97
54) 2,4-Dinitrophenol	14.85	184	2719	4.069	ng	# 81
55) Dibenzofuran	15.14	168	171710	10.606	ng	97
56) 4-Nitrophenol	14.93	139	18490	8.489	ng	98
57) 2,4-Dinitrotoluene	15.10	165	34837	11.504	ng	# 96
58) Fluorene	15.78	166	129641	10.587	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.36	232	33793	11.111	ng	# 98
60) Diethylphthalate	15.54	149	138157	9.982	ng	97
61) 4-Chlorophenyl-phenylether	15.77	204	73405	10.256	ng	95
62) 4-Nitroaniline	15.81	138	31208	10.641	ng	91
63) Azobenzene	16.07	77	136580	10.292	ng	97
65) 4,6-Dinitro-2-methylphenol	15.86	198	11005	9.707	ng	91
66) n-Nitrosodiphenylamine	15.99	169	125830	10.532	ng	98
67) 4-Bromophenyl-phenylether	16.67	248	44320	10.094	ng	94
68) Hexachlorobenzene	16.78	284	46713	10.261	ng	99
69) Atrazine	16.93	200	45711	10.294	ng	99
70) Pentachlorophenol	17.12	266	23758	8.089	ng	99
71) Phenanthrene	17.53	178	219739	10.610	ng	98
72) Anthracene	17.62	178	217656	10.737	ng	97
73) Carbazole	17.89	167	214850	10.287	ng	98
74) Di-n-butylphthalate	18.43	149	237699	10.618	ng	98
75) Fluoranthene	19.53	202	251177	10.377	ng	98
77) Benzidine	19.72	184	132118	9.041	ng	98
78) Pyrene	19.90	202	261276	10.495	ng	97
80) Butylbenzylphthalate	20.77	149	97982	10.116	ng	95
81) Benzo(a)anthracene	21.75	228	232486	10.162	ng	98
82) 3,3'-Dichlorobenzidine	21.67	252	87188	9.879	ng	97
83) Chrysene	21.82	228	223031	10.220	ng	96
84) Bis(2-ethylhexyl)phthalate	21.63	149	137184	9.863	ng	100
85) Di-n-octyl phthalate	22.88	149	222064	9.836	ng	97
86) Indeno(1,2,3-cd)pyrene	28.92	276	212244	8.774	ng	# 91
88) Benzo(b)fluoranthene	24.04	252	213172	10.050	ng	98
89) Benzo(k)fluoranthene	24.10	252	212000	10.129	ng	98
90) Benzo(a)pyrene	24.95	252	202538	9.958	ng	98
91) Dibenzo(a,h)anthracene	29.00	278	179357	9.216	ng	97
92) Benzo(g,h,i)perylene	30.13	276	182391	9.384	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG101818\
 Data File : BG037412.D
 Acq On : 18 Oct 2018 9:53
 Operator : JU/SJ
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 Client Sampled :

Quant Time: Oct 18 12:19:05 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 12:15:58 2018
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 8.11 min Scan# 821
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampleId :


Tgt Ion	Resp	Lower	Upper
152	53095		
152	100		
150	151.8	126.0	189.0
115	54.7	45.5	68.3



#2
 1,4-Dioxane
 Concen: 10.031 ng
 RT: 3.53 min Scan# 42
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
88	16996		
88	100		
58	85.7	74.4	111.6
43	28.2	25.8	38.8









#3
 Pyridine
 Concen: 9.774 ng
 RT: 3.94 min Scan# 111
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
79	100		
52	82.7	74.2	111.2
51	42.9	34.6	51.8



#4
 n-Nitrosodimethylamine
 Concen: 9.037 ng
 RT: 3.86 min Scan# 97
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
42	100		
74	149.2	102.0	153.0
44	15.6	9.6	14.4#









#5
 2-Fluorophenol
 Concen: 21.196 ng
 RT: 5.64 min Scan# 401
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
112	100		
64	63.1	53.1	79.7
63	34.4	27.3	40.9



#6
 Aniline
 Concen: 9.784 ng
 RT: 7.43 min Scan# 705
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
93	100		
66	37.8	32.8	49.2
65	19.8	16.4	24.6









#7
 Phenol-d6
 Concen: 21.362 ng
 RT: 7.24 min Scan# 673
 Delta R.T. -0.01 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Instrument :
 BNA_G
 ClientSampled :

Tgt Ion	Resp	Lower	Upper
99	100		
42	18.9	15.0	22.4
71	30.2	25.5	38.3




Abundance Ion 99.00 (98.70 to 99.70): BGC
 Ion 42.00 (41.70 to 42.70): BGC
 Ion 71.00 (70.70 to 71.70): BGC






#8
 2-Chlorophenol
 Concen: 10.909 ng
 RT: 7.67 min Scan# 745
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
128	100		
130	30.4	12.0	52.0
64	54.1	32.9	72.9



Abundance Ion 128.00 (127.70 to 128.70): E
 Ion 130.00 (129.70 to 130.70): E
 Ion 64.00 (63.70 to 64.70): BGC









#9
Benzaldehyde
Concen: 10.876 ng
RT: 7.25 min Scan# 674
Delta R.T. -0.00 min
Lab File: BG037412.D
Acq: 18 Oct 2018 9:53




Instrument :
BNA_G
ClientSampled :


Tgt Ion	Resp	Lower	Upper
77	100		
105	97.1	72.6	112.6
106	92.4	71.3	111.3



#10
Phenol
Concen: 10.792 ng
RT: 7.27 min Scan# 678
Delta R.T. -0.00 min
Lab File: BG037412.D
Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
94	100		
65	27.6	9.7	49.7
66	30.7	15.9	55.9








#11
 bis(2-Chloroethyl)ether
 Concen: 9.943 ng
 RT: 7.52 min Scan# 721
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53





Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
93	100		
63	79.3	69.5	109.5
95	30.0	12.6	52.6



#12
 1,3-Dichlorobenzene
 Concen: 10.258 ng
 RT: 7.99 min Scan# 801
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
146	100		
148	66.4	49.8	74.8
75	27.7	23.2	34.8








#13
 1,4-Dichlorobenzene
 Concen: 10.459 ng
 RT: 8.15 min Scan# 827
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53





Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
146	43956		
146	100		
148	61.2	51.0	76.6
111	38.7	32.4	48.6



#14
 1,2-Dichlorobenzene
 Concen: 10.410 ng
 RT: 8.46 min Scan# 881
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
146	42393		
146	100		
148	65.2	50.4	75.6
111	44.2	34.6	51.8








#15
 Benzyl Alcohol
 Concen: 9.596 ng
 RT: 8.34 min Scan# 860
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53





Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
79	100		
108	79.1	65.8	98.8
77	64.0	52.9	79.3



#16
 2,2'-oxybis(1-Chloropropane)
 Concen: 9.039 ng
 RT: 8.63 min Scan# 909
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
45	100		
77	10.5	0.0	30.0
79	7.4	0.0	29.0









#17
 2-Methylphenol
 Concen: 10.215 ng
 RT: 8.53 min Scan# 893
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Instrument :
 BNA_G
 ClientSampled :

Tgt Ion	Resp	Lower	Upper
107	100		
108	110.8	87.9	131.9
77	36.9	35.1	52.7
79	41.9	34.5	51.7




Abundance
 Ion 107.00 (106.70 to 107.70): E
 Ion 108.00 (107.70 to 108.70): E
 Ion 77.00 (76.70 to 77.70): BG
 Ion 79.00 (78.70 to 79.70): BG






#18
 Hexachloroethane
 Concen: 9.995 ng
 RT: 9.19 min Scan# 1005
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
117	100		
119	88.8	74.6	111.8
201	101.3	80.8	121.2



Abundance
 Ion 117.00 (116.70 to 117.70): E
 Ion 119.00 (118.70 to 119.70): E
 Ion 201.00 (200.70 to 201.70): E








#19
 n-Nitroso-di-n-propylamine
 Concen: 9.172 ng
 RT: 8.90 min Scan# 956
 Delta R.T. -0.01 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Instrument :
 BNA_G
 ClientSampled :

Tgt Ion	Resp	Lower	Upper
70	100		
42	60.1	53.9	80.9
101	9.1	8.2	12.2
130	23.0	18.2	27.4




Abundance Ion 70.00 (69.70 to 70.70): BGC
 Ion 42.00 (41.70 to 42.70): BGC
 Ion 101.00 (100.70 to 101.70): BGC
 Ion 130.00 (129.70 to 130.70): BGC







#20
 3+4-Methylphenols
 Concen: 11.045 ng
 RT: 8.87 min Scan# 950
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
107	100		
108	78.5	69.9	109.9
77	28.4	11.2	51.2
79	25.4	6.6	46.6



Abundance Ion 107.00 (106.70 to 107.70): BGC
 Ion 108.00 (107.70 to 108.70): BGC
 Ion 77.00 (76.70 to 77.70): BGC
 Ion 79.00 (78.70 to 79.70): BGC






#21
 Naphthalene-d8
 Concen: 20.000 ng
 RT: 10.93 min Scan# 1301
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :

Tgt Ion:136 Resp: 247750

Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.6	14.4
54	10.8	7.9	11.9
68	6.7	4.8	7.2




Abundance Ion 136.00 (135.70 to 136.70): E
 Ion 137.00 (136.70 to 137.70): E
 Ion 54.00 (53.70 to 54.70): BGC
 Ion 68.00 (67.70 to 68.70): BGC





#22
 Acetophenone
 Concen: 10.214 ng
 RT: 8.93 min Scan# 961
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53


Tgt Ion:105 Resp: 63791

Ion	Ratio	Lower	Upper
105	100		
71	1.1	1.2	1.8#
51	28.7	25.0	37.6
120	22.8	18.6	28.0



Abundance Ion 105.00 (104.70 to 105.70): E
 Ion 71.00 (70.70 to 71.70): BGC
 Ion 51.00 (50.70 to 51.70): BGC
 Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 23.367 ng
 RT: 9.28 min Scan# 1020
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :




Tgt Ion	Resp	Lower	Upper
82	100		
128	43.7	34.6	51.8
54	53.7	45.3	67.9



#24
 Nitrobenzene
 Concen: 11.603 ng
 RT: 9.33 min Scan# 1028
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
77	100		
123	42.1	33.6	50.4
65	12.1	11.3	16.9









#25
 Isophorone
 Concen: 9.140 ng
 RT: 9.84 min Scan# 1116
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53


Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
82	100		
95	7.6	5.3	7.9
138	17.7	13.3	19.9



#26
 2-Nitrophenol
 Concen: 12.408 ng
 RT: 10.04 min Scan# 1149
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
139	100		
109	29.4	23.3	34.9
65	53.0	39.8	59.8








#27
 2,4-Dimethylphenol
 Concen: 10.509 ng
 RT: 10.08 min Scan# 1156
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53





Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
122	37786		
107	121.5	94.2	141.2
121	60.6	47.3	70.9



#28
 bis(2-Chloroethoxy)methane
 Concen: 9.951 ng
 RT: 10.33 min Scan# 1198
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
93	53556		
95	34.2	24.6	37.0
123	11.5	9.1	13.7









#29
 2,4-Dichlorophenol
 Concen: 11.282 ng
 RT: 10.57 min Scan# 1239
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
162	100		
164	63.2	43.2	83.2
98	33.2	16.0	56.0



#30
 1,2,4-Trichlorobenzene
 Concen: 10.027 ng
 RT: 10.78 min Scan# 1276
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
180	100		
182	102.9	77.7	116.5
145	28.9	23.3	34.9








#31
 Naphthalene
 Concen: 10.415 ng
 RT: 10.98 min Scan# 1310
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53





Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
128	125404		
129	10.8	9.0	13.6
127	14.8	11.1	16.7



#32
 Benzoic acid
 Concen: 8.750 ng
 RT: 10.14 min Scan# 1166
 Delta R.T. -0.07 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
122	17372		
105	124.2	106.7	146.7
77	82.7	72.2	112.2









#33
 4-Chloroaniline
 Concen: 10.439 ng
 RT: 11.09 min Scan# 1328
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
127	100		
129	31.9	27.1	40.7
65	27.6	26.6	39.8
92	17.1	16.4	24.6



#34
 Hexachlorobutadiene
 Concen: 9.359 ng
 RT: 11.25 min Scan# 1355
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
225	100		
223	63.4	48.2	72.4
227	60.1	51.6	77.4









#35
 Caprolactam
 Concen: 11.227 ng
 RT: 11.85 min Scan# 1458
 Delta R.T. -0.03 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
113	16706		
55	181.9	176.6	216.6
56	129.9	128.3	168.3



#36
 4-Chloro-3-methylphenol
 Concen: 10.913 ng
 RT: 12.18 min Scan# 1514
 Delta R.T. -0.01 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
107	47045		
144	23.1	20.7	31.1
142	78.6	64.4	96.6









#37
 2-Methylnaphthalene
 Concen: 10.226 ng
 RT: 12.58 min Scan# 1581
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
142	100		
141	89.2	71.9	107.9
115	32.8	27.8	41.8



#38
 1-Methylnaphthalene
 Concen: 10.428 ng
 RT: 12.79 min Scan# 1618
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
142	100		
141	92.3	75.2	112.8
116	3.9	3.3	4.9









#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.74 min Scan# 1949
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
164	100		
162	102.4	81.3	121.9
160	44.7	36.3	54.5



#40
 1,2,4,5-Tetrachlorobenzene
 Concen: 9.879 ng
 RT: 12.94 min Scan# 1642
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
216	100		
214	80.5	63.8	95.6
179	21.2	17.4	26.0
108	15.6	13.9	20.9









#41
 Hexachlorocyclopentadiene
 Concen: 9.568 ng
 RT: 12.91 min Scan# 1637
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
237	100		
235	64.5	42.5	82.5
272	11.3	0.0	31.6



#42
 2,4,6-Tribromophenol
 Concen: 21.771 ng
 RT: 16.23 min Scan# 2202
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
330	100		
332	96.5	78.4	117.6
141	38.3	32.2	48.2








#43
 2,4,6-Trichlorophenol
 Concen: 11.294 ng
 RT: 13.17 min Scan# 1682
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53





Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
196	100		
198	88.8	76.4	114.6
200	28.6	24.9	37.3



#44
 2,4,5-Trichlorophenol
 Concen: 11.368 ng
 RT: 13.24 min Scan# 1694
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
196	100		
198	99.7	74.2	111.2
97	43.9	34.1	51.1
132	26.3	20.2	30.4









#45
 2-Fluorobiphenyl
 Concen: 21.012 ng
 RT: 13.36 min Scan# 1715
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
172	100		
171	35.1	31.0	46.4
170	24.0	20.2	30.2



#46
 1,1'-Biphenyl
 Concen: 10.306 ng
 RT: 13.58 min Scan# 1751
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
154	100		
153	43.0	22.1	62.1
76	15.4	0.0	33.2









#47
 2-Chloronaphthalene
 Concen: 10.213 ng
 RT: 13.62 min Scan# 1759
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
162	107844		
127	37.0	30.5	45.7
164	33.6	26.4	39.6



#48
 2-Nitroaniline
 Concen: 12.181 ng
 RT: 13.83 min Scan# 1794
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
65	31083		
92	61.7	53.2	79.8
138	95.1	79.3	118.9








#49
 Acenaphthylene
 Concen: 10.319 ng
 RT: 14.47 min Scan# 1903
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
152	166694		
151	20.2	17.1	25.7
153	13.3	11.1	16.7



#50
 Dimethylphthalate
 Concen: 10.020 ng
 RT: 14.19 min Scan# 1855
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
163	133827		
194	4.5	3.5	5.3
164	10.5	8.5	12.7









#51
 2,6-Dinitrotoluene
 Concen: 12.011 ng
 RT: 14.32 min Scan# 1877
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
165	100		
63	54.4	43.4	65.2
89	50.3	45.8	68.6



#52
 Acenaphthene
 Concen: 10.145 ng
 RT: 14.80 min Scan# 1960
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
154	100		
153	118.4	93.4	140.0
152	59.0	44.7	67.1









#53
 3-Nitroaniline
 Concen: 11.013 ng
 RT: 14.65 min Scan# 1934
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
138	30775		
108	9.5	11.0	16.6#
92	116.6	95.2	142.8



#54
 2,4-Dinitrophenol
 Concen: 4.069 ng
 RT: 14.85 min Scan# 1968
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
184	2719		
184	100		
63	65.0	42.6	63.8#
154	67.4	41.8	62.6#








#55
 Dibenzofuran
 Concen: 10.606 ng
 RT: 15.14 min Scan# 2017
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53





Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
168	100		
139	35.0	29.4	44.0
169	12.7	11.0	16.6



#56
 4-Nitrophenol
 Concen: 8.489 ng
 RT: 14.93 min Scan# 1982
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
139	100		
109	65.2	49.5	89.5
65	97.2	76.8	116.8









#57
 2,4-Dinitrotoluene
 Concen: 11.504 ng
 RT: 15.10 min Scan# 2010
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
165	100		
63	39.3	33.4	50.0
89	67.5	57.2	85.8
182	2.1	2.6	4.0#



#58
 Fluorene
 Concen: 10.587 ng
 RT: 15.78 min Scan# 2127
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
166	100		
165	99.5	79.0	118.6
167	13.6	11.3	16.9









#59
 2,3,4,6-Tetrachlorophenol
 Concen: 11.111 ng
 RT: 15.36 min Scan# 2054
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
232	100		
131	42.9	33.7	50.5
130	2.0	2.1	3.1#
166	29.5	24.6	36.8



#60
 Diethylphthalate
 Concen: 9.982 ng
 RT: 15.54 min Scan# 2085
 Delta R.T. -0.01 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
149	100		
177	21.1	18.3	27.5
150	12.3	10.0	15.0









#61
 4-Chlorophenyl-phenylether
 Concen: 10.256 ng
 RT: 15.77 min Scan# 2125
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :




Tgt Ion	Resp	Lower	Upper
204	100		
206	32.1	26.6	39.8
141	52.1	45.4	68.2



#62
 4-Nitroaniline
 Concen: 10.641 ng
 RT: 15.81 min Scan# 2131
 Delta R.T. -0.01 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
138	100		
92	48.9	36.4	76.4
108	72.6	60.1	100.1









#63
 Azobenzene
 Concen: 10.292 ng
 RT: 16.07 min Scan# 2175
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53


Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
77	136580		
182	28.6	8.0	48.0
105	21.3	1.3	41.3
51	27.3	10.7	50.7



#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 17.49 min Scan# 2417
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
188	406851		
94	9.5	7.2	10.8
80	10.6	7.7	11.5









#65
 4,6-Dinitro-2-methylphenol
 Concen: 9.707 ng
 RT: 15.86 min Scan# 2139
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
198	11005		
198	100		
51	38.0	22.7	62.7
105	32.8	19.7	59.7



#66
 n-Nitrosodiphenylamine
 Concen: 10.532 ng
 RT: 15.99 min Scan# 2162
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
169	125830		
169	100		
168	71.7	55.7	83.5
167	38.7	31.3	46.9









#67
 4-Bromophenyl-phenylether
 Concen: 10.094 ng
 RT: 16.67 min Scan# 2278
 Delta R.T. 0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
248	100		
250	90.4	77.0	115.6
141	63.8	55.5	83.3



#68
 Hexachlorobenzene
 Concen: 10.261 ng
 RT: 16.78 min Scan# 2296
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
284	100		
142	35.9	28.1	42.1
249	37.0	29.8	44.8









#69
 Atrazine
 Concen: 10.294 ng
 RT: 16.93 min Scan# 2322
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
200	45711		
173	24.4	6.1	46.1
215	50.5	30.8	70.8



#70
 Pentachlorophenol
 Concen: 8.089 ng
 RT: 17.12 min Scan# 2355
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
266	23758		
268	68.1	55.0	82.6
264	73.8	58.9	88.3









#71
 Phenanthrene
 Concen: 10.610 ng
 RT: 17.53 min Scan# 2424
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
178	100		
176	18.7	16.1	24.1
179	15.5	13.0	19.4



#72
 Anthracene
 Concen: 10.737 ng
 RT: 17.62 min Scan# 2439
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
178	100		
176	17.4	15.0	22.4
179	15.0	12.8	19.2









#73
 Carbazole
 Concen: 10.287 ng
 RT: 17.89 min Scan# 2485
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
167	100		
166	22.6	18.3	27.5
139	11.8	10.5	15.7



#74
 Di-n-butylphthalate
 Concen: 10.618 ng
 RT: 18.43 min Scan# 2577
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
149	100		
150	9.0	8.2	12.2
104	4.7	4.0	6.0









#75
 Fluoranthene
 Concen: 10.377 ng
 RT: 19.53 min Scan# 2765
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
202	100		
101	11.1	0.0	30.7
203	17.4	0.0	38.5



#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 21.77 min Scan# 3146
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
240	100		
120	11.0	8.1	12.1
236	26.7	21.4	32.0









#77
Benzidine
Concen: 9.041 ng
RT: 19.72 min Scan# 2796
Delta R.T. -0.00 min
Lab File: BG037412.D
Acq: 18 Oct 2018 9:53




Instrument :
BNA_G
ClientSampled :


Tgt Ion	Resp	Lower	Upper
184	132118		
185	14.1	12.6	18.8
183	12.7	10.2	15.2



#78
Pyrene
Concen: 10.495 ng
RT: 19.90 min Scan# 2827
Delta R.T. 0.00 min
Lab File: BG037412.D
Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
202	261276		
200	21.3	17.8	26.6
203	17.4	15.2	22.8









#79
 Terphenyl-d14
 Concen: 21.654 ng
 RT: 20.09 min Scan# 2859
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Ratio	Lower	Upper
244	100		
212	6.8	6.5	9.7
122	10.2	9.0	13.4



#80
 Butylbenzylphthalate
 Concen: 10.116 ng
 RT: 20.77 min Scan# 2975
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Ratio	Lower	Upper
149	100		
91	65.7	57.0	85.4
206	21.7	17.8	26.6









#81
 Benzo(a)anthracene
 Concen: 10.162 ng
 RT: 21.75 min Scan# 3142
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
228	100		
226	26.9	22.6	33.8
229	19.6	16.3	24.5



#82
 3,3'-Dichlorobenzidine
 Concen: 9.879 ng
 RT: 21.67 min Scan# 3128
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
252	100		
254	62.8	52.0	78.0
126	9.7	8.2	12.4









#83
 Chrysene
 Concen: 10.220 ng
 RT: 21.82 min Scan# 3154
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
228	100		
226	29.4	25.7	38.5
229	19.8	16.5	24.7



#84
 Bis(2-ethylhexyl)phthalate
 Concen: 9.863 ng
 RT: 21.63 min Scan# 3122
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
149	100		
167	28.7	23.0	34.4
279	4.0	3.6	5.4









#85
 Di-n-octyl phthalate
 Concen: 9.836 ng
 RT: 22.88 min Scan# 3335
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	9.5	8.7	13.1



#86
 Indeno(1,2,3-cd)pyrene
 Concen: 8.774 ng
 RT: 28.92 min Scan# 4363
 Delta R.T. -0.02 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
276	100		
138	11.1	12.0	18.0#
277	20.9	20.6	30.8









#87
 Perylene-d12
 Concen: 20.000 ng
 RT: 25.11 min Scan# 3714
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Instrument :
 BNA_G
 ClientSampled :

Tgt Ion: 264 Resp: 390137




Ion	Ratio	Lower	Upper
264	100		
260	22.2	18.2	27.4
265	20.6	17.9	26.9




#88
 Benzo(b)fluoranthene
 Concen: 10.050 ng
 RT: 24.04 min Scan# 3532
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion: 252 Resp: 213172

Ion	Ratio	Lower	Upper
252	100		
253	23.8	18.2	27.2
125	9.7	7.8	11.6









#89
 Benzo(k)fluoranthene
 Concen: 10.129 ng
 RT: 24.10 min Scan# 3543
 Delta R.T. -0.01 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53




Instrument :
 BNA_G
 ClientSampled :


Tgt Ion	Resp	Lower	Upper
252	100		
253	22.3	17.4	26.2
125	10.6	7.4	11.0



#90
 Benzo(a)pyrene
 Concen: 9.958 ng
 RT: 24.95 min Scan# 3687
 Delta R.T. -0.00 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
252	100		
253	21.3	17.4	26.2
125	10.0	9.0	13.6









#91
 Dibenzo(a,h)anthracene
 Concen: 9.216 ng
 RT: 29.00 min Scan# 4377
 Delta R.T. -0.01 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Instrument :
 BNA_G
 ClientSampled :

Tgt Ion	Resp	Lower	Upper
278	179357		
139	17.0	11.5	17.3
279	22.9	18.6	27.8



#92
 Benzo(g,h,i)perylene
 Concen: 9.384 ng
 RT: 30.13 min Scan# 4569
 Delta R.T. -0.02 min
 Lab File: BG037412.D
 Acq: 18 Oct 2018 9:53

Tgt Ion	Resp	Lower	Upper
276	182391		
277	20.8	18.9	28.3
138	22.1	17.2	25.8

