

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG101818\
 Data File : BG037418.D
 Acq On : 18 Oct 2018 14:03
 Operator : JU/SJ
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG101818

Quant Time: Oct 18 14:51:00 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 14:06:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.607	0.585	3.6	101	0.00
3	Pyridine	1.529	1.535	-0.4	103	0.00
4	n-Nitrosodimethylamine	0.545	0.550	-0.9	105	0.00
5 S	2-Fluorophenol	1.196	1.203	-0.6	103	0.00
6	Aniline	2.207	2.231	-1.1	103	0.00
7 S	Phenol-d6	1.809	1.794	0.8	100	0.00
8	2-Chlorophenol	1.399	1.391	0.6	101	0.00
9	Benzaldehyde	1.132	1.118	1.2	101	0.00
10 C	Phenol	1.909	1.912	-0.2	102	0.00
11	bis(2-Chloroethyl)ether	1.450	1.427	1.6	102	0.00
12	1,3-Dichlorobenzene	1.558	1.534	1.5	102	0.00
13 C	1,4-Dichlorobenzene	1.594	1.541	3.3	100	0.00
14	1,2-Dichlorobenzene	1.533	1.504	1.9	101	0.00
15	Benzyl Alcohol	1.337	1.375	-2.8	103	0.00
16	2,2'-oxybis(1-Chloropropane	2.594	2.581	0.5	102	0.00
17	2-Methylphenol	1.262	1.268	-0.5	101	0.00
18	Hexachloroethane	0.543	0.543	0.0	102	0.00
19 P	n-Nitroso-di-n-propylamine	1.246	1.258	-1.0	100	0.00
20	3+4-Methylphenols	1.804	1.797	0.4	101	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
22	Acetophenone	0.502	0.502	0.0	101	0.00
23 S	Nitrobenzene-d5	0.357	0.365	-2.2	103	0.00
24	Nitrobenzene	0.371	0.382	-3.0	103	0.00
25	Isophorone	0.652	0.676	-3.7	102	0.00
26 C	2-Nitrophenol	0.167	0.186	-11.4	108	0.00
27	2,4-Dimethylphenol	0.298	0.300	-0.7	100	0.00
28	bis(2-Chloroethoxy)methane	0.427	0.434	-1.6	101	0.00
29 C	2,4-Dichlorophenol	0.332	0.334	-0.6	101	0.00
30	1,2,4-Trichlorobenzene	0.361	0.363	-0.6	103	0.00
31	Naphthalene	0.982	0.978	0.4	102	0.00
32	Benzoic acid	0.194	0.230	-18.6	116	0.00
33	4-Chloroaniline	0.462	0.462	0.0	99	0.00
34 C	Hexachlorobutadiene	0.214	0.217	-1.4	101	0.00
35	Caprolactam	0.133	0.142	-6.8	104	0.00
36 C	4-Chloro-3-methylphenol	0.375	0.379	-1.1	101	0.00
37	2-Methylnaphthalene	0.716	0.716	0.0	100	0.00
38	1-Methylnaphthalene	0.695	0.686	1.3	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	0.645	0.662	-2.6	103	0.00
41 P	Hexachlorocyclopentadiene	0.308	0.317	-2.9	100	0.00
42 S	2,4,6-Tribromophenol	0.218	0.229	-5.0	100	0.00
43 C	2,4,6-Trichlorophenol	0.431	0.449	-4.2	101	0.00
44	2,4,5-Trichlorophenol	0.469	0.489	-4.3	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG101818\
 Data File : BG037418.D
 Acq On : 18 Oct 2018 14:03
 Operator : JU/SJ
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG101818

Quant Time: Oct 18 14:51:00 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 14:06:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.326	1.324	0.2	101	0.00
46	1,1'-Biphenyl	1.656	1.656	0.0	101	0.00
47	2-Chloronaphthalene	1.253	1.253	0.0	101	0.00
48	2-Nitroaniline	0.380	0.401	-5.5	104	0.00
49	Acenaphthylene	1.885	1.930	-2.4	102	0.00
50	Dimethylphthalate	1.547	1.563	-1.0	100	0.00
51	2,6-Dinitrotoluene	0.342	0.357	-4.4	101	0.00
52 C	Acenaphthene	1.161	1.169	-0.7	102	0.00
53	3-Nitroaniline	0.380	0.402	-5.8	101	0.00
54 P	2,4-Dinitrophenol	0.100	0.121	-21.0	123	0.00
55	Dibenzofuran	1.923	1.901	1.1	100	0.00
56 P	4-Nitrophenol	0.281	0.303	-7.8	104	0.00
57	2,4-Dinitrotoluene	0.449	0.492	-9.6	104	0.00
58	Fluorene	1.440	1.424	1.1	100	0.00
59	2,3,4,6-Tetrachlorophenol	0.402	0.418	-4.0	102	0.00
60	Diethylphthalate	1.588	1.611	-1.4	101	0.00
61	4-Chlorophenyl-phenylether	0.840	0.829	1.3	99	0.00
62	4-Nitroaniline	0.378	0.401	-6.1	102	0.00
63	Azobenzene	1.516	1.503	0.9	101	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	0.093	0.105	-12.9	111	0.00
66 c	n-Nitrosodiphenylamine	0.602	0.609	-1.2	100	0.00
67	4-Bromophenyl-phenylether	0.220	0.223	-1.4	101	0.00
68	Hexachlorobenzene	0.228	0.228	0.0	100	0.00
69	Atrazine	0.218	0.224	-2.8	100	0.00
70 C	Pentachlorophenol	0.152	0.163	-7.2	103	0.00
71	Phenanthrene	1.016	1.014	0.2	101	0.00
72	Anthracene	0.998	1.015	-1.7	102	0.00
73	Carbazole	0.998	1.010	-1.2	100	0.00
74	Di-n-butylphthalate	1.110	1.149	-3.5	101	0.00
75 C	Fluoranthene	1.153	1.159	-0.5	100	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzidine	0.680	0.720	-5.9	102	0.00
78	Pyrene	1.307	1.328	-1.6	101	0.00
79 S	Terphenyl-d14	0.936	0.938	-0.2	100	0.00
80	Butylbenzylphthalate	0.535	0.571	-6.7	102	0.00
81	Benzo(a)anthracene	1.183	1.201	-1.5	101	0.00
82	3,3'-Dichlorobenzidine	0.459	0.470	-2.4	99	0.00
83	Chrysene	1.138	1.143	-0.4	100	0.00
84	Bis(2-ethylhexyl)phthalate	0.750	0.801	-6.8	102	0.00
85 c	Di-n-octyl phthalate	1.223	1.313	-7.4	102	0.00
86	Indeno(1,2,3-cd)pyrene	1.220	1.251	-2.5	100	0.00
87 I	Perylene-d12	1.000	1.000	0.0	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG101818\
Data File : BG037418.D
Acq On : 18 Oct 2018 14:03
Operator : JU/SJ
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG101818

Quant Time: Oct 18 14:51:00 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG101818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 18 14:06:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.096	1.104	-0.7	99	0.00
89	Benzo(k)fluoranthene	1.074	1.109	-3.3	101	0.00
90 C	Benzo(a)pyrene	1.042	1.065	-2.2	100	0.00
91	Dibenzo(a,h)anthracene	0.961	0.977	-1.7	98	0.00
92	Benzo(q,h,i)perylene	0.972	0.971	0.1	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0