

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101915\
 Data File : BG019245.D
 Acq On : 19 Oct 2015 17:29
 Operator : UM/NP
 Sample : G4075-17
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 JG9M9

Quant Time: Oct 20 04:09:13 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG101515.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 20 04:07:05 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ng/ul	-8.00
18) Naphthalene-d8	0.00	136	0	0.00	ng/ul	-10.81
36) Acenaphthene-d10	0.00	164	0	0.00	ng/ul	-14.64
62) Phenanthrene-d10	0.00	188	0	0.00	ng/ul	-17.39
78) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.65
86) Perylene-d12	0.00	264	0	0.00	ng/ul	-24.86

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units
3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL
5) Phenol-d5	0.00	99	0	0.00	ng/ul
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0	0.00	ng/ul
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul
29) 4-Chloroaniline-d4	0.00	131	0	0.00	ng/ul
44) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul
47) Acenaphthylene-d8	0.00	160	0	0.00	ng/ul
52) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
58) Fluorene-d10	0.00	176	0	0.00	ng/ul
63) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul
71) Anthracene-d10	0.00	188	0	0.00	ng/ul
79) Pyrene-d10	0.00	212	0	0.00	ng/ul
90) Benzo(a)pyrene-d12	0.00	264	0	0.00	ng/ul

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101915\
Data File : BG019245.D
Acq On : 19 Oct 2015 17:29
Operator : UM/NP
Sample : G4075-17
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
JG9M9

Quant Time: Oct 20 04:09:13 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG101515.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 20 04:07:05 2015
Response via : Initial Calibration

