


Data Path : Z:\HPCHEM1\BNA G\Data\BG101915\
Data File : BG019240.D
Acq On : 19 Oct 2015 11:20
Operator : UM/NP
Sample : DFTPP04
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP04

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG101515.M
Title : SVOA CALIBRATION
Last Update : Tue Oct 20 06:03:40 2015



AutoFind: Scans 2548, 2549, 2550; Background Corrected with Scan 2543

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.7	8222	PASS
68	69	0.00	2	1.1	80	PASS
69	198	0.00	100	40.0	7350	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.3	8511	PASS
197	198	0.00	2	0.5	83	PASS
198	198	100	100	100.0	18390	PASS
199	198	5	9	7.3	1347	PASS
275	198	10	60	30.2	5562	PASS
365	198	1	100	5.0	917	PASS
441	442	0.01	24	16.0	2479	PASS
442	198	50	100	84.4	15528	PASS
443	442	15	24	20.5	3188	PASS

Scan 2543 (17.781 min): BG019240.D
DFTPP04

m/z	Abundance
40.10	245.0
43.10	185.0
44.00	486.0
45.00	158.0
73.10	218.0
281.00	155.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04

m/z	Abundance
39.10	468.0
40.00	207.0
41.10	222.0
43.10	253.0
44.00	778.0
45.10	189.0
50.10	1601.0
51.10	6772.0
52.10	276.0
55.10	243.0
56.00	244.0
57.10	483.0
63.00	320.0
68.00	242.0
69.00	6133.0
73.00	170.0
74.10	685.0
75.10	1136.0
76.10	390.0
77.10	7001.0
78.10	474.0
79.00	539.0
80.00	417.0
81.10	567.0
82.10	218.0
85.10	184.0
86.10	252.0
93.10	1042.0
98.00	809.0
98.90	386.0
101.00	207.0
105.00	270.0
107.00	2413.0
108.00	484.0
110.10	3858.0
111.00	635.0
117.10	1941.0
118.10	184.0
122.10	198.0
123.00	351.0
127.10	6779.0
128.00	538.0
129.00	2993.0
130.00	478.0
135.10	424.0
141.10	652.0
141.90	188.0
147.00	216.0
148.00	365.0
153.00	219.0
155.00	207.0
156.10	450.0
160.10	161.0
161.20	205.0
165.00	177.0
167.00	562.0
168.10	392.0
175.10	454.0
176.10	171.0
176.90	168.0
179.00	587.0
180.00	380.0
181.00	219.0
185.10	381.0
186.10	2243.0
187.10	641.0
189.20	214.0
192.00	173.0
193.00	384.0
196.00	358.0
198.00	13130.0
199.00	1210.0
203.00	250.0
204.00	780.0
205.10	1105.0
206.10	4012.0
207.10	626.0
207.90	360.0
217.00	1041.0
218.00	182.0
221.00	565.0
222.00	228.0
223.10	202.0
224.00	2644.0
225.00	535.0
227.00	1449.0
227.90	209.0
229.00	175.0
243.00	167.0
244.10	1554.0
245.10	261.0
246.00	424.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04

255.00	9153.0
256.00	1210.0
258.00	549.0
264.90	249.0
273.00	322.0
274.00	797.0
275.10	3827.0
276.10	479.0
277.00	376.0
296.00	1070.0
297.00	301.0
314.90	223.0
323.00	514.0
334.10	409.0
365.00	524.0
372.10	243.0
423.00	519.0
424.00	166.0
441.00	1464.0
442.00	8442.0
443.00	1820.0
444.00	174.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04

m/z	Abundance
39.10	700.0
40.00	192.0
43.10	245.0
44.00	532.0
50.10	2723.0
51.10	9811.0
52.10	441.0
56.10	336.0
57.00	693.0
60.00	152.0
61.00	175.0
63.10	378.0
65.10	203.0
69.00	8669.0
73.10	228.0
74.10	974.0
75.10	1556.0
76.10	701.0
77.10	9791.0
78.10	846.0
79.10	805.0
80.00	590.0
81.10	836.0
83.10	307.0
86.10	196.0
87.10	181.0
91.00	224.0
92.10	275.0
93.10	1275.0
98.10	870.0
99.10	716.0
101.00	570.0
103.90	346.0
105.10	328.0
107.00	3122.0
108.00	542.0
109.20	173.0
110.00	5390.0
111.10	1022.0
116.10	383.0
117.10	2676.0
118.10	186.0
122.10	249.0
123.00	387.0
127.10	10105.0
128.00	944.0
129.00	4314.0
130.00	560.0
134.00	184.0
135.00	369.0
136.00	176.0
137.10	168.0
139.90	151.0
141.10	689.0
143.00	165.0
147.00	406.0
148.00	651.0
153.00	319.0
155.00	245.0
156.10	474.0
159.00	183.0
160.00	266.0
160.90	269.0
165.00	406.0
165.90	222.0
167.00	1076.0
168.00	580.0
173.00	222.0
174.00	224.0
175.00	520.0
176.10	284.0
177.00	259.0
179.00	773.0
180.00	597.0
180.80	247.0
185.00	476.0
186.10	3192.0
187.10	1201.0
189.00	240.0
192.10	323.0
192.90	440.0
196.00	637.0
196.90	249.0
198.00	21128.0
199.00	1431.0
200.00	191.0
201.00	172.0
203.10	300.0
204.10	971.0
205.00	1661.0
206.10	6100.0
207.00	1140.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04

208.00	402.0
210.10	173.0
211.00	274.0
215.00	164.0
216.00	213.0
217.00	1907.0
218.10	236.0
221.10	828.0
222.00	269.0
223.00	479.0
224.10	3767.0
225.00	841.0
227.00	1902.0
228.00	216.0
228.90	275.0
242.00	316.0
242.90	172.0
244.10	2700.0
245.00	393.0
246.00	647.0
249.00	165.0
255.00	15289.0
256.00	2279.0
257.00	199.0
258.00	896.0
259.00	159.0
264.00	162.0
265.10	519.0
273.00	421.0
274.00	1252.0
275.00	5977.0
276.00	783.0
277.00	702.0
292.90	154.0
296.00	2647.0
297.00	327.0
303.00	290.0
315.00	274.0
323.10	603.0
324.00	184.0
327.00	152.0
334.10	534.0
354.00	206.0
365.00	915.0
372.00	298.0
421.00	209.0
422.00	221.0
423.00	1005.0
441.10	2722.0
442.10	17088.0
443.00	3582.0
444.00	472.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04

m/z	Abundance
39.10	721.0
40.00	292.0
43.20	205.0
44.10	584.0
45.10	187.0
50.10	2143.0
51.10	8084.0
52.10	308.0
55.10	187.0
56.10	303.0
57.10	494.0
61.00	180.0
63.10	204.0
65.10	207.0
69.00	7248.0
73.10	228.0
74.10	805.0
75.00	1051.0
76.00	368.0
77.10	8004.0
78.10	650.0
79.10	697.0
80.10	466.0
81.00	773.0
82.00	188.0
83.00	205.0
85.10	156.0
86.00	220.0
91.00	218.0
93.10	974.0
98.10	818.0
99.10	626.0
101.00	332.0
104.10	307.0
105.00	301.0
105.90	156.0
107.10	2918.0
108.00	409.0
110.00	4401.0
111.00	743.0
112.10	236.0
117.00	2146.0
118.10	197.0
122.10	353.0
123.10	272.0
124.00	153.0
127.10	8651.0
128.00	764.0
129.00	4232.0
130.00	363.0
131.00	181.0
135.10	433.0
137.20	178.0
141.00	474.0
142.00	200.0
146.10	187.0
147.00	398.0
148.00	662.0
149.10	164.0
153.00	246.0
155.00	222.0
156.10	645.0
161.00	352.0
165.00	436.0
167.00	1142.0
168.00	417.0
174.00	252.0
175.00	403.0
177.00	285.0
179.00	953.0
180.00	718.0
181.00	212.0
185.00	427.0
186.10	3354.0
187.00	948.0
189.00	218.0
192.00	224.0
193.00	430.0
196.00	565.0
198.00	20912.0
199.00	1402.0
203.00	244.0
204.00	890.0
205.00	1455.0
206.00	6331.0
207.00	992.0
208.00	335.0
211.00	282.0
215.10	160.0
216.00	226.0
217.00	2113.0
217.90	322.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04

221.00	870.0
222.10	349.0
223.00	494.0
224.10	3622.0
225.10	1029.0
226.00	172.0
227.10	2315.0
228.00	269.0
229.10	393.0
231.00	194.0
235.10	155.0
240.10	163.0
242.00	280.0
243.00	253.0
244.10	2474.0
245.00	376.0
246.00	759.0
247.00	190.0
252.90	171.0
255.00	15190.0
256.10	2329.0
257.00	176.0
258.00	1131.0
258.90	183.0
265.00	403.0
273.10	508.0
274.00	1325.0
275.00	6883.0
276.00	813.0
277.00	796.0
285.10	249.0
296.00	2108.0
297.00	325.0
303.00	364.0
315.00	255.0
323.10	655.0
334.00	590.0
346.00	255.0
352.10	258.0
353.00	196.0
365.00	1312.0
366.00	183.0
372.00	309.0
382.90	168.0
401.90	244.0
403.00	191.0
421.10	302.0
422.10	226.0
423.00	1249.0
424.00	214.0
441.00	3253.0
442.00	21056.0
443.00	4164.0
444.10	395.0

Instrument :
BNA_G
ClientSampleId :
DFTPP04