

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101916\
 Data File : BG024436.D
 Acq On : 19 Oct 2016 10:26
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02004

Quant Time: Oct 19 17:27:52 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 19 10:58:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.456	0.458	-0.4	92	0.00
3 S	1,4-Dioxane-d8	0.421	0.342	18.8	79	0.00
4	Benzaldehyde	1.313	1.488	-13.3	92	0.00
5 S	Phenol-d5	1.866	1.963	-5.2	93	0.00
6	Phenol	1.894	1.996	-5.4	92	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.261	1.199	4.9	86	0.00
8	Bis(2-Chloroethyl)ether	1.394	1.394	0.0	86	0.00
9 S	2-Chlorophenol-d4	1.285	1.406	-9.4	95	0.00
10	2-Chlorophenol	1.302	1.346	-3.4	92	0.00
11	2-Methylphenol	1.403	1.519	-8.3	97	0.00
12	2,2'-oxybis(1-Chloropropane	2.784	2.646	5.0	82	0.00
13 S	4-Methylphenol-d8	1.509	1.640	-8.7	96	0.00
14	Acetophenone	2.240	2.394	-6.9	92	0.00
15 P	N-Nitroso-di-n-propylamine	1.351	1.406	-4.1	90	0.00
16	4-Methylphenol	1.488	1.689	-13.5	102	0.00
17	Hexachloroethane	0.602	0.602	0.0	92	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
19 S	Nitrobenzene-d5	0.142	0.154	-8.5	99	0.00
20	Nitrobenzene	0.449	0.487	-8.5	99	0.00
21	Isophorone	0.816	0.879	-7.7	95	0.00
22 S	2-Nitrophenol-d4	0.159	0.179	-12.6	106	0.00
23 C	2-Nitrophenol	0.173	0.191	-10.4	100	0.00
24	2,4-Dimethylphenol	0.426	0.455	-6.8	96	0.00
25	Bis(2-Chloroethoxy)methane	0.448	0.474	-5.8	94	0.00
26 S	2,4-Dichlorophenol-d3	0.346	0.378	-9.2	99	0.00
27 C	2,4-Dichlorophenol	0.332	0.365	-9.9	100	0.00
28	Naphthalene	0.963	1.014	-5.3	96	0.00
29 S	4-Chloroaniline-d4	0.363	0.421	-16.0	102	0.00
30	4-Chloroaniline	0.369	0.415	-12.5	100	0.00
31 C	Hexachlorobutadiene	0.266	0.294	-10.5	100	0.00
32	Caprolactam	0.126	0.129	-2.4	94	0.00
33 C	4-Chloro-3-methylphenol	0.406	0.444	-9.4	101	0.00
34	2-Methylnaphthalene	0.766	0.819	-6.9	100	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
36	1,2,4,5-Tetrachlorobenzene	0.590	0.684	-15.9	105	0.00
37	Hexachlorocyclopentadiene	0.432	0.428	0.9	91	0.00
38 C	2,4,6-Trichlorophenol	0.382	0.434	-13.6	105	0.00
39	2,4,5-Trichlorophenol	0.420	0.464	-10.5	102	0.00
40	1,1'-Biphenyl	1.295	1.409	-8.8	98	0.00
41	2-Chloronaphthalene	1.015	1.110	-9.4	101	0.00
42	2-Nitroaniline	0.386	0.427	-10.6	100	0.00
43 S	Dimethylphthalate-d6	1.396	1.467	-5.1	97	0.00
44	Dimethylphthalate	1.369	1.427	-4.2	95	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101916\
 Data File : BG024436.D
 Acq On : 19 Oct 2016 10:26
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02004

Quant Time: Oct 19 17:27:52 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 19 10:58:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.266	0.305	-14.7	102	0.00
46 S	Acenaphthylene-d8	1.600	1.733	-8.3	98	0.00
47	Acenaphthylene	1.539	1.675	-8.8	99	0.00
48	3-Nitroaniline	0.251	0.280	-11.6	102	0.00
49 C	Acenaphthene	1.100	1.151	-4.6	97	0.00
50	2,4-Dinitrophenol	0.178	0.170	4.5	93	0.00
51 S	4-Nitrophenol-d4	0.248	0.244	1.6	92	0.00
52	4-Nitrophenol	0.300	0.309	-3.0	93	0.00
53	Dibenzofuran	1.619	1.733	-7.0	97	0.00
54	2,4-Dinitrotoluene	0.394	0.442	-12.2	101	0.00
55	2,3,4,6-Tetrachlorophenol	0.413	0.463	-12.1	102	0.00
56	Diethylphthalate	1.443	1.446	-0.2	89	0.00
57 S	Fluorene-d10	1.320	1.372	-3.9	93	0.00
58	Fluorene	1.339	1.413	-5.5	95	0.00
59	4-Chlorophenyl-phenylether	0.743	0.774	-4.2	95	0.00
60	4-Nitroaniline	0.289	0.298	-3.1	95	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.120	0.134	-11.7	102	0.00
63	4,6-Dinitro-2-methylphenol	0.126	0.140	-11.1	96	0.00
64	N-Nitrosodiphenylamine	0.566	0.618	-9.2	92	0.00
65	4-Bromophenyl-phenylether	0.232	0.273	-17.7	102	0.00
66	Hexachlorobenzene	0.258	0.309	-19.8	103	0.00
67	Atrazine	0.231	0.260	-12.6	95	0.00
68 C	Pentachlorophenol	0.157	0.176	-12.1	96	0.00
69	Phenanthrene	1.010	1.105	-9.4	91	0.00
70 S	Anthracene-d10	0.903	0.983	-8.9	90	0.00
71	Anthracene	1.028	1.138	-10.7	92	0.00
72	Carbazole	0.807	0.931	-15.4	88	0.00
73	Di-n-butylphthalate	0.986	1.113	-12.9	93	0.00
74 C	Fluoranthene	1.020	1.267	-24.2	91	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
76 S	Pyrene-d10	0.933	0.986	-5.7	90	0.00
77	Pyrene	1.068	1.131	-5.9	90	0.00
78	Butylbenzylphthalate	0.429	0.445	-3.7	90	0.00
79	3,3'-Dichlorobenzidine	0.372	0.425	-14.2	97	0.00
80	Benzo(a)anthracene	1.110	1.194	-7.6	93	0.00
81	Bis(2-ethylhexyl)phthalate	0.607	0.631	-4.0	90	0.00
82	Chrysene	1.057	1.131	-7.0	91	0.00
83 I	Perylene-d12	1.000	1.000	0.0	98	0.00
84	Di-n-octyl phthalate	0.938	1.015	-8.2	91	0.00
85	Benzo(b)fluoranthene	1.110	1.178	-6.1	98	0.00
86	Benzo(k)fluoranthene	1.070	1.073	-0.3	92	0.00
87 S	Benzo(a)pyrene-d12	0.900	0.941	-4.6	96	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG101916\
Data File : BG024436.D
Acq On : 19 Oct 2016 10:26
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02004

Quant Time: Oct 19 17:27:52 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG0100716.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 19 10:58:30 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.058	1.093	-3.3	94	0.00
89	Indeno(1,2,3-cd)pyrene	1.253	1.345	-7.3	99	-0.01
90	Dibenzo(a,h)anthracene	1.059	1.130	-6.7	99	0.00
91	Benzo(a,h,i)perylene	1.053	1.129	-7.2	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0