

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG101918\
 Data File : BG037464.D
 Acq On : 19 Oct 2018 23:46
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 20 00:23:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 14:06:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.607	0.586	3.5	113	0.00
3	Pyridine	1.529	1.546	-1.1	117	0.00
4	n-Nitrosodimethylamine	0.545	0.550	-0.9	118	0.00
5 S	2-Fluorophenol	1.196	1.181	1.3	113	0.00
6	Aniline	2.207	2.149	2.6	112	0.00
7 S	Phenol-d6	1.809	1.755	3.0	110	0.00
8	2-Chlorophenol	1.399	1.336	4.5	110	0.00
9	Benzaldehyde	1.132	1.064	6.0	108	0.00
10 C	Phenol	1.909	1.846	3.3	111	0.00
11	bis(2-Chloroethyl)ether	1.450	1.426	1.7	114	0.00
12	1,3-Dichlorobenzene	1.558	1.511	3.0	113	0.00
13 C	1,4-Dichlorobenzene	1.594	1.506	5.5	110	0.00
14	1,2-Dichlorobenzene	1.533	1.442	5.9	109	0.00
15	Benzyl Alcohol	1.337	1.303	2.5	109	0.00
16	2,2'-oxybis(1-Chloropropane	2.594	2.594	0.0	116	0.00
17	2-Methylphenol	1.262	1.231	2.5	110	0.00
18	Hexachloroethane	0.543	0.538	0.9	114	0.00
19 P	n-Nitroso-di-n-propylamine	1.246	1.212	2.7	109	0.00
20	3+4-Methylphenols	1.804	1.757	2.6	111	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
22	Acetophenone	0.502	0.487	3.0	109	0.00
23 S	Nitrobenzene-d5	0.357	0.361	-1.1	113	0.00
24	Nitrobenzene	0.371	0.376	-1.3	112	0.00
25	Isophorone	0.652	0.659	-1.1	110	0.00
26 C	2-Nitrophenol	0.167	0.177	-6.0	115	0.00
27	2,4-Dimethylphenol	0.298	0.289	3.0	107	0.00
28	bis(2-Chloroethoxy)methane	0.427	0.422	1.2	109	0.00
29 C	2,4-Dichlorophenol	0.332	0.322	3.0	108	0.00
30	1,2,4-Trichlorobenzene	0.361	0.348	3.6	109	0.00
31	Naphthalene	0.982	0.962	2.0	111	0.00
32	Benzoic acid	0.194	0.216	-11.3	121	0.00
33	4-Chloroaniline	0.462	0.451	2.4	108	0.00
34 C	Hexachlorobutadiene	0.214	0.208	2.8	108	0.00
35	Caprolactam	0.133	0.132	0.8	107	0.01
36 C	4-Chloro-3-methylphenol	0.375	0.362	3.5	107	0.00
37	2-Methylnaphthalene	0.716	0.699	2.4	108	0.00
38	1-Methylnaphthalene	0.695	0.674	3.0	110	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	0.645	0.648	-0.5	109	0.00
41 P	Hexachlorocyclopentadiene	0.308	0.327	-6.2	112	0.00
42 S	2,4,6-Tribromophenol	0.218	0.219	-0.5	104	0.00
43 C	2,4,6-Trichlorophenol	0.431	0.435	-0.9	106	0.00
44	2,4,5-Trichlorophenol	0.469	0.473	-0.9	106	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG101918\
 Data File : BG037464.D
 Acq On : 19 Oct 2018 23:46
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 20 00:23:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG101818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 18 14:06:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.326	1.291	2.6	107	0.00
46	1,1'-Biphenyl	1.656	1.636	1.2	108	0.00
47	2-Chloronaphthalene	1.253	1.237	1.3	108	0.00
48	2-Nitroaniline	0.380	0.402	-5.8	113	0.00
49	Acenaphthylene	1.885	1.887	-0.1	108	0.00
50	Dimethylphthalate	1.547	1.533	0.9	107	0.00
51	2,6-Dinitrotoluene	0.342	0.350	-2.3	108	0.00
52 C	Acenaphthene	1.161	1.146	1.3	109	0.00
53	3-Nitroaniline	0.380	0.394	-3.7	108	0.00
54 P	2,4-Dinitrophenol	0.100	0.123	-23.0	136	0.00
55	Dibenzofuran	1.923	1.864	3.1	107	0.00
56 P	4-Nitrophenol	0.281	0.290	-3.2	108	0.00
57	2,4-Dinitrotoluene	0.449	0.478	-6.5	110	0.00
58	Fluorene	1.440	1.405	2.4	107	0.00
59	2,3,4,6-Tetrachlorophenol	0.402	0.407	-1.2	108	0.00
60	Diethylphthalate	1.588	1.575	0.8	108	0.00
61	4-Chlorophenyl-phenylether	0.840	0.811	3.5	105	0.00
62	4-Nitroaniline	0.378	0.394	-4.2	109	0.00
63	Azobenzene	1.516	1.494	1.5	109	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	0.093	0.104	-11.8	120	0.00
66 c	n-Nitrosodiphenylamine	0.602	0.589	2.2	106	0.00
67	4-Bromophenyl-phenylether	0.220	0.216	1.8	107	0.00
68	Hexachlorobenzene	0.228	0.218	4.4	104	0.00
69	Atrazine	0.218	0.216	0.9	105	0.00
70 C	Pentachlorophenol	0.152	0.157	-3.3	108	0.00
71	Phenanthrene	1.016	0.984	3.1	106	0.00
72	Anthracene	0.998	0.981	1.7	107	0.00
73	Carbazole	0.998	0.994	0.4	107	0.00
74	Di-n-butylphthalate	1.110	1.134	-2.2	108	0.00
75 C	Fluoranthene	1.153	1.130	2.0	106	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
77	Benzidine	0.680	0.761	-11.9	118	0.00
78	Pyrene	1.307	1.290	1.3	107	0.00
79 S	Terphenyl-d14	0.936	0.905	3.3	105	0.00
80	Butylbenzylphthalate	0.535	0.567	-6.0	110	0.00
81	Benzo(a)anthracene	1.183	1.162	1.8	106	0.00
82	3,3'-Dichlorobenzidine	0.459	0.466	-1.5	106	0.00
83	Chrysene	1.138	1.124	1.2	107	0.00
84	Bis(2-ethylhexyl)phthalate	0.750	0.791	-5.5	109	0.00
85 c	Di-n-octyl phthalate	1.223	1.304	-6.6	111	0.00
86	Indeno(1,2,3-cd)pyrene	1.220	1.225	-0.4	106	0.00
87 I	Perylene-d12	1.000	1.000	0.0	110	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG101918\
Data File : BG037464.D
Acq On : 19 Oct 2018 23:46
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 20 00:23:26 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG101818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 18 14:06:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.096	1.060	3.3	105	0.00
89	Benzo(k)fluoranthene	1.074	1.076	-0.2	109	0.00
90 C	Benzo(a)pyrene	1.042	1.042	0.0	108	0.00
91	Dibenzo(a,h)anthracene	0.961	0.935	2.7	104	0.00
92	Benzo(g,h,i)perylene	0.972	0.947	2.6	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0