


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101920\
 Data File : BG046980.D
 Acq On : 19 Oct 2020 9:19
 Operator : CG/JU
 Sample : DFTPP41
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP41

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG101520MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Oct 20 10:59:53 2020



AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.7	43037	PASS
68	69	0.00	2	0.4	174	PASS
69	198	0.00	100	35.4	45154	PASS
70	69	0.00	2	0.4	171	PASS
127	198	10	80	42.0	53597	PASS
197	198	0.00	2	0.4	550	PASS
198	198	100	100	100.0	127554	PASS
199	198	5	9	7.1	9068	PASS
275	198	10	60	28.6	36506	PASS
365	198	1	100	3.2	4085	PASS
441	198	0.01	100	11.7	14882	PASS
442	198	50	100	72.4	92344	PASS
443	442	15	24	19.6	18085	PASS

Scan 2467 (17.833 min): BG046980.D
DFTPP41

m/z	Abundance
40.20	227.0
44.00	253.0
55.00	198.0
163.20	275.0
178.20	169.0
207.10	668.0
280.80	208.0
500.10	253.0

Instrument :
BNA_G
ClientSampleId :
DFTPP41

m/z	Abundance
38.10	317.0
39.10	2798.0
40.00	486.0
41.10	249.0
44.00	425.0
49.10	336.0
50.10	9136.0
51.10	37080.0
52.20	1752.0
56.10	1313.0
57.00	3182.0
61.10	400.0
62.10	887.0
63.10	1676.0
64.20	246.0
65.10	856.0
66.90	242.0
69.00	38880.0
70.00	231.0
74.10	3933.0
75.10	6344.0
76.10	2375.0
77.10	43440.0
78.10	2685.0
79.00	2869.0
80.00	2623.0
81.00	3970.0
82.10	719.0
83.10	550.0
84.90	562.0
86.10	1316.0
87.00	293.0
88.20	158.0
91.00	557.0
92.10	708.0
93.00	5326.0
94.00	318.0
96.00	229.0
97.00	288.0
98.00	3953.0
99.00	3404.0
100.10	369.0
101.10	1775.0
102.20	193.0
103.10	597.0
103.90	778.0
105.00	1314.0
106.00	398.0
107.10	13270.0
108.00	2650.0
110.00	23904.0
111.10	3560.0
112.00	636.0
113.00	294.0
117.00	10569.0
118.00	697.0
121.10	184.0
121.90	1123.0
123.00	1092.0
123.80	447.0
125.00	466.0
127.10	42088.0
128.00	4175.0
129.00	17912.0
130.00	1956.0
130.90	376.0
133.10	169.0
133.90	492.0
135.10	1456.0
136.00	407.0
137.00	1290.0
138.60	168.0
139.90	369.0
141.00	3022.0
142.00	689.0
142.90	562.0
146.10	594.0
147.00	1340.0
148.10	3241.0
149.00	581.0
150.90	196.0
151.50	152.0
153.00	915.0
154.10	584.0
155.00	1110.0
156.10	1980.0
157.20	545.0
158.10	479.0
159.00	468.0
160.00	786.0
161.10	1502.0
162.10	550.0

Instrument :
BNA_G
ClientSampleId :
DFTPP41

165.00	1369.0
166.10	824.0
167.00	4234.0
168.00	2585.0
169.00	325.0
170.10	437.0
170.90	319.0
172.00	349.0
172.90	509.0
174.00	952.0
175.00	1972.0
176.10	421.0
177.00	1485.0
178.00	698.0
179.00	4130.0
180.00	2944.0
181.00	1046.0
181.80	247.0
182.90	165.0
184.00	462.0
185.00	1958.0
186.00	13886.0
187.00	4265.0
187.90	413.0
189.00	1143.0
190.10	270.0
191.00	342.0
192.00	1617.0
193.00	1748.0
193.90	286.0
195.10	360.0
196.00	2824.0
198.00	92680.0
199.00	6985.0
200.10	599.0
201.50	628.0
202.10	259.0
203.00	703.0
204.00	3243.0
205.00	6984.0
206.10	27000.0
207.10	4112.0
208.00	903.0
209.10	245.0
210.00	469.0
210.80	678.0
211.10	722.0
214.90	441.0
216.00	512.0
217.00	8356.0
218.00	862.0
219.90	335.0
221.00	4283.0
222.10	1115.0
223.00	1760.0
224.00	14341.0
225.00	4419.0
226.00	704.0
227.00	7373.0
228.00	1151.0
229.00	1283.0
229.80	161.0
230.20	230.0
230.90	1089.0
232.10	167.0
232.90	168.0
234.00	453.0
235.10	460.0
236.10	340.0
236.90	606.0
239.00	283.0
240.90	414.0
242.00	730.0
243.00	645.0
244.10	11653.0
245.10	1581.0
245.90	2139.0
247.00	608.0
248.10	226.0
249.00	488.0
251.10	207.0
253.10	317.0
255.00	59520.0
256.00	8596.0
257.00	845.0
258.00	3534.0
259.10	501.0
261.00	159.0
265.00	1297.0
265.90	160.0
270.90	151.0
273.10	1761.0
274.00	4585.0
275.00	25920.0
276.00	3433.0
277.00	2379.0

Instrument :
BNA_G
ClientSampleId :
DFTPP41

277.80	429.0
281.30	425.0
281.90	293.0
283.00	305.0
285.10	307.0
291.10	159.0
293.00	556.0
295.10	410.0
296.00	6917.0
297.10	1052.0
302.90	840.0
304.00	469.0
307.00	187.0
308.00	185.0
309.10	354.0
313.90	593.0
315.10	1010.0
315.90	585.0
323.00	2248.0
324.10	863.0
325.90	163.0
327.00	318.0
328.00	183.0
332.10	234.0
333.10	317.0
334.10	1814.0
335.10	271.0
341.00	277.0
346.00	294.0
352.00	824.0
353.10	335.0
354.10	719.0
365.00	2164.0
365.90	280.0
371.10	258.0
372.10	1009.0
373.00	262.0
390.10	224.0
402.30	344.0
403.00	673.0
404.10	168.0
420.80	313.0
422.10	357.0
423.10	2430.0
423.90	401.0
441.10	8544.0
442.10	48248.0
443.10	11345.0
444.00	753.0

Instrument :
BNA_G
ClientSampleId :
DFTPP41

m/z	Abundance
37.10	251.0
38.20	609.0
39.10	3526.0
40.10	285.0
44.00	416.0
45.00	207.0
49.00	279.0
50.10	15079.0
51.10	54512.0
52.10	2161.0
55.10	195.0
56.10	1528.0
57.10	3572.0
57.90	390.0
59.00	162.0
61.10	670.0
62.10	1006.0
63.10	2619.0
64.10	327.0
65.10	685.0
69.00	54848.0
69.90	283.0
73.10	648.0
74.10	6599.0
75.00	8569.0
76.10	2863.0
77.10	60296.0
78.10	4151.0
79.10	4660.0
80.10	3421.0
81.00	4113.0
82.00	1066.0
83.00	1135.0
85.10	643.0
86.00	1823.0
87.10	615.0
87.90	223.0
89.20	173.0
91.10	1194.0
92.10	1115.0
93.10	7923.0
94.00	523.0
98.00	5728.0
99.00	4740.0
100.10	381.0
101.00	2170.0
103.00	947.0
104.00	1924.0
105.00	1603.0
105.90	431.0
107.00	18752.0
108.00	2764.0
110.00	30712.0
111.00	5564.0
112.10	863.0
113.00	238.0
116.10	1351.0
117.00	16848.0
118.00	1133.0
118.90	171.0
120.00	429.0
122.00	1372.0
123.00	2265.0
123.90	946.0
125.00	1089.0
127.00	67792.0
128.10	5585.0
129.00	25520.0
130.00	2772.0
131.00	448.0
132.00	225.0
134.00	622.0
135.10	2360.0
136.00	1274.0
137.10	1240.0
137.80	178.0
140.00	445.0
141.00	3106.0
142.00	971.0
143.10	913.0
144.10	185.0
145.00	364.0
146.00	829.0
147.00	1403.0
148.00	4267.0
149.00	759.0
150.00	186.0
151.40	629.0
153.00	1023.0
154.00	1581.0
155.00	1927.0
156.00	2933.0

Instrument :
BNA_G
ClientSampleId :
DFTPP41

157.10	296.0
158.00	740.0
158.90	264.0
160.00	1472.0
161.00	1886.0
161.90	538.0
164.10	193.0
165.00	1550.0
166.00	1125.0
167.00	5514.0
168.00	3679.0
169.10	540.0
170.10	311.0
171.00	447.0
172.00	943.0
173.00	1051.0
174.10	1964.0
175.10	2938.0
176.10	897.0
177.00	1485.0
178.00	707.0
179.00	6670.0
180.00	3533.0
181.10	2188.0
182.30	219.0
183.00	293.0
183.90	408.0
185.00	2639.0
186.10	21216.0
187.00	6170.0
188.00	461.0
189.00	1750.0
190.10	211.0
191.00	1260.0
192.10	1944.0
193.00	2091.0
194.00	410.0
196.00	4121.0
196.70	1650.0
198.00	152768.0
199.00	10666.0
200.00	923.0
200.70	159.0
201.50	689.0
202.20	205.0
203.00	1083.0
204.00	6995.0
205.00	10265.0
206.00	43712.0
207.00	7660.0
208.10	1635.0
209.00	516.0
210.20	844.0
210.90	1733.0
213.10	174.0
215.00	860.0
216.10	1109.0
217.00	12782.0
218.00	1658.0
218.90	232.0
219.60	336.0
220.90	5645.0
223.00	2921.0
224.00	25200.0
225.00	6262.0
226.10	953.0
227.00	12112.0
228.00	1899.0
229.00	2010.0
230.00	492.0
231.00	1096.0
232.20	226.0
234.00	620.0
235.00	628.0
235.90	780.0
237.00	827.0
238.00	183.0
239.10	445.0
240.10	359.0
241.00	783.0
242.10	1376.0
243.10	1234.0
244.00	18928.0
245.00	2712.0
246.00	3317.0
246.90	797.0
249.00	862.0
249.90	314.0
252.10	438.0
253.20	753.0
253.80	920.0
255.00	97312.0
256.00	14492.0
257.10	1265.0
258.00	5708.0
259.00	1406.0

Instrument :
BNA_G
ClientSampleId :
DFTPP41

261.00	360.0
263.90	178.0
265.00	2455.0
266.00	357.0
271.10	153.0
272.00	367.0
273.00	2786.0
274.00	7376.0
275.00	40976.0
276.00	4332.0
277.00	3581.0
277.90	896.0
279.00	179.0
282.90	620.0
284.00	266.0
284.40	291.0
285.20	483.0
289.90	200.0
292.00	390.0
293.00	867.0
294.00	243.0
296.00	11571.0
297.00	1703.0
298.00	272.0
300.90	154.0
302.00	304.0
303.10	1829.0
303.90	272.0
304.90	240.0
307.70	191.0
308.00	169.0
309.00	188.0
311.20	196.0
314.00	590.0
315.00	1216.0
316.10	545.0
316.90	157.0
321.00	297.0
321.80	209.0
323.10	3380.0
324.10	607.0
326.90	674.0
327.90	268.0
332.00	177.0
333.00	356.0
334.00	2275.0
335.00	539.0
340.90	272.0
341.10	252.0
342.00	204.0
346.00	816.0
352.00	1112.0
353.10	741.0
354.10	823.0
355.00	266.0
356.00	224.0
365.00	5115.0
365.90	955.0
370.00	215.0
372.00	1641.0
373.10	481.0
382.00	153.0
383.00	389.0
383.80	266.0
391.00	216.0
392.10	150.0
400.90	236.0
402.10	874.0
403.10	714.0
404.00	336.0
420.90	1040.0
422.10	747.0
423.10	5556.0
424.10	793.0
424.90	270.0
441.10	16544.0
442.10	103472.0
443.10	18888.0
444.10	2298.0
446.00	183.0

Instrument :
BNA_G
ClientSampleId :
DFTPP41

m/z	Abundance
38.20	530.0
39.10	3103.0
40.00	230.0
44.00	438.0
49.00	159.0
50.10	10775.0
51.10	37520.0
52.10	1943.0
55.10	181.0
56.00	1481.0
57.10	2419.0
61.10	542.0
62.10	316.0
63.10	1986.0
64.10	366.0
65.10	985.0
68.00	524.0
69.00	41736.0
73.00	570.0
74.10	4521.0
75.10	6606.0
76.10	2590.0
77.10	45848.0
78.10	3765.0
79.00	3786.0
80.00	2589.0
81.00	3655.0
82.00	570.0
83.10	1024.0
85.00	534.0
86.00	1243.0
87.00	899.0
88.10	291.0
91.10	530.0
92.10	1464.0
93.00	5038.0
94.10	303.0
96.00	281.0
98.00	3958.0
99.10	3402.0
100.00	300.0
101.00	2126.0
103.00	594.0
104.00	1340.0
105.10	1591.0
107.10	15190.0
108.00	2403.0
108.90	660.0
110.00	25928.0
111.10	4220.0
112.10	653.0
113.20	209.0
115.10	331.0
116.00	1105.0
117.00	12894.0
118.10	1105.0
118.90	153.0
122.00	1282.0
123.00	1803.0
123.90	1199.0
125.00	541.0
127.00	50912.0
128.00	4184.0
129.00	24264.0
130.00	2090.0
131.10	549.0
133.20	238.0
134.10	612.0
135.00	1983.0
136.00	771.0
137.00	906.0
138.10	230.0
139.80	258.0
141.00	3250.0
141.90	749.0
142.90	587.0
143.80	167.0
145.00	153.0
146.00	680.0
147.10	2406.0
148.00	3141.0
149.20	714.0
150.00	225.0
151.00	423.0
151.90	168.0
153.00	1106.0
153.90	809.0
155.10	1459.0
156.00	2504.0
157.00	508.0
158.00	639.0
159.00	388.0

Instrument :
BNA_G
ClientSampleId :
DFTPP41

160.00	1160.0
161.00	1429.0
162.10	746.0
162.90	227.0
163.90	214.0
165.00	1100.0
166.00	1139.0
167.00	5609.0
168.00	2368.0
168.90	540.0
170.30	234.0
171.00	721.0
172.00	1050.0
173.00	692.0
174.00	1076.0
175.00	2535.0
176.00	968.0
177.00	1670.0
178.00	617.0
179.00	5098.0
180.00	3751.0
180.90	1339.0
182.00	355.0
182.80	218.0
183.20	201.0
184.20	443.0
185.00	2674.0
186.00	19528.0
187.00	4940.0
188.00	746.0
189.00	1752.0
191.10	1149.0
192.00	1595.0
193.00	1768.0
194.20	529.0
196.00	4092.0
198.00	137216.0
199.00	9554.0
199.90	606.0
201.00	352.0
201.80	502.0
203.00	1059.0
204.10	5330.0
205.00	10703.0
206.00	37464.0
207.00	5425.0
208.00	923.0
209.00	317.0
210.10	685.0
211.10	1555.0
211.90	236.0
213.90	259.0
215.20	262.0
217.00	10438.0
218.00	1563.0
218.90	188.0
220.00	245.0
221.00	7284.0
222.00	1488.0
223.00	2663.0
224.00	22920.0
225.00	6603.0
226.10	470.0
227.00	11264.0
228.00	1838.0
229.00	2680.0
230.00	327.0
231.10	758.0
231.90	189.0
232.90	212.0
234.10	397.0
235.00	942.0
236.20	355.0
237.00	1197.0
237.90	150.0
238.90	493.0
240.00	516.0
241.00	559.0
242.00	1752.0
243.10	1131.0
244.10	16992.0
245.10	2446.0
246.00	3579.0
246.80	713.0
249.00	567.0
250.30	220.0
250.90	364.0
253.00	858.0
255.00	95960.0
256.00	13006.0
257.00	1129.0
258.00	5000.0
259.00	1088.0
263.90	262.0
265.00	1903.0
266.10	261.0

Instrument :
BNA_G
ClientSampleId :
DFTPP41

270.00	204.0
270.80	231.0
272.00	293.0
273.00	2371.0
274.00	7087.0
275.00	42624.0
276.00	6517.0
277.00	4309.0
278.10	408.0
278.80	204.0
279.30	183.0
282.20	278.0
283.10	329.0
284.00	315.0
285.10	622.0
289.10	224.0
292.00	156.0
293.10	472.0
294.10	366.0
294.80	405.0
296.00	12877.0
297.00	1940.0
302.00	227.0
303.00	1540.0
304.00	790.0
305.00	168.0
309.80	171.0
312.30	244.0
313.90	626.0
315.00	1689.0
316.00	996.0
317.10	474.0
319.90	173.0
321.10	437.0
322.10	316.0
323.10	3409.0
324.10	510.0
327.00	734.0
328.00	420.0
331.90	352.0
333.00	597.0
334.00	2705.0
335.00	768.0
340.90	445.0
342.10	246.0
345.90	604.0
347.10	298.0
352.00	1128.0
353.00	908.0
354.00	1081.0
355.10	427.0
365.00	4976.0
366.10	383.0
370.60	213.0
371.10	287.0
372.10	2017.0
372.90	843.0
383.00	317.0
385.00	166.0
389.90	308.0
391.10	298.0
402.00	827.0
403.00	616.0
404.10	223.0
421.10	700.0
422.10	858.0
423.10	6849.0
424.00	1606.0
425.10	198.0
438.90	179.0
441.00	19560.0
442.10	125312.0
443.10	24024.0
444.00	2439.0

Instrument :
BNA_G
ClientSampleId :
DFTPP41