

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101921\
 Data File : BG050628.D
 Acq On : 19 Oct 2021 9:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 19 11:42:39 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 10:56:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.693	0.636	8.2	79	0.00
3	Pyridine	1.892	1.798	5.0	83	0.00
4	n-Nitrosodimethylamine	0.892	0.875	1.9	88	0.00
5 S	2-Fluorophenol	1.337	1.298	2.9	86	0.00
6	Aniline	2.250	2.141	4.8	86	0.00
7 S	Phenol-d6	1.936	1.887	2.5	87	0.00
8	2-Chlorophenol	1.288	1.281	0.5	89	0.00
9	Benzaldehyde	1.429	1.253	12.3	84	0.00
10 C	Phenol	1.883	1.829	2.9	88	0.00
11	bis(2-Chloroethyl)ether	1.469	1.394	5.1	85	0.00
12	1,3-Dichlorobenzene	1.486	1.414	4.8	86	0.00
13 C	1,4-Dichlorobenzene	1.498	1.436	4.1	87	0.00
14	1,2-Dichlorobenzene	1.431	1.355	5.3	86	0.00
15	Benzyl Alcohol	1.530	1.489	2.7	88	0.00
16	2,2'-oxybis(1-Chloropropane	2.396	2.218	7.4	84	0.00
17	2-Methylphenol	1.239	1.210	2.3	90	0.00
18	Hexachloroethane	0.567	0.538	5.1	85	0.00
19 P	n-Nitroso-di-n-propylamine	1.411	1.315	6.8	84	0.00
20	3+4-Methylphenols	1.744	1.705	2.2	87	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22	Acetophenone	0.569	0.521	8.4	86	0.00
23 S	Nitrobenzene-d5	0.483	0.454	6.0	86	0.00
24	Nitrobenzene	0.480	0.447	6.9	86	0.00
25	Isophorone	0.857	0.783	8.6	85	0.00
26 C	2-Nitrophenol	0.176	0.177	-0.6	91	0.00
27	2,4-Dimethylphenol	0.305	0.286	6.2	86	0.00
28	bis(2-Chloroethoxy)methane	0.490	0.450	8.2	85	0.00
29 C	2,4-Dichlorophenol	0.311	0.297	4.5	88	0.00
30	1,2,4-Trichlorobenzene	0.353	0.333	5.7	88	0.00
31	Naphthalene	1.071	1.000	6.6	87	0.00
32	Benzoic acid	0.226	0.155	31.4#	60	0.00
33	4-Chloroaniline	0.449	0.429	4.5	89	0.00
34 C	Hexachlorobutadiene	0.222	0.205	7.7	86	0.00
35	Caprolactam	0.119	0.113	5.0	87	0.00
36 C	4-Chloro-3-methylphenol	0.388	0.365	5.9	88	0.00
37	2-Methylnaphthalene	0.739	0.685	7.3	87	0.00
38	1-Methylnaphthalene	0.716	0.654	8.7	86	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
40	1,2,4,5-Tetrachlorobenzene	0.591	0.556	5.9	87	0.00
41 P	Hexachlorocyclopentadiene	0.341	0.277	18.8	72	0.00
42 S	2,4,6-Tribromophenol	0.191	0.183	4.2	88	0.00
43 C	2,4,6-Trichlorophenol	0.416	0.397	4.6	86	0.00
44	2,4,5-Trichlorophenol	0.435	0.423	2.8	88	0.00
45 S	2-Fluorobiphenyl	1.329	1.243	6.5	86	0.00
46	1,1'-Biphenyl	1.459	1.361	6.7	86	0.00
47	2-Chloronaphthalene	1.120	1.076	3.9	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101921\
 Data File : BG050628.D
 Acq On : 19 Oct 2021 9:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 19 11:42:39 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 10:56:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.446	0.446	0.0	89	0.00
49	Acenaphthylene	1.835	1.725	6.0	87	0.00
50	Dimethylphthalate	1.512	1.439	4.8	88	0.00
51	2,6-Dinitrotoluene	0.303	0.292	3.6	87	0.00
52 C	Acenaphthene	1.179	1.082	8.2	83	0.00
53	3-Nitroaniline	0.358	0.351	2.0	89	0.00
54 P	2,4-Dinitrophenol	0.188	0.158	16.0	74	0.00
55	Dibenzofuran	1.824	1.704	6.6	87	0.00
56 P	4-Nitrophenol	0.288	0.273	5.2	86	0.00
57	2,4-Dinitrotoluene	0.427	0.424	0.7	90	0.00
58	Fluorene	1.401	1.304	6.9	86	0.00
59	2,3,4,6-Tetrachlorophenol	0.379	0.353	6.9	84	0.00
60	Diethylphthalate	1.588	1.504	5.3	88	0.00
61	4-Chlorophenyl-phenylether	0.767	0.718	6.4	86	0.00
62	4-Nitroaniline	0.373	0.369	1.1	91	0.00
63	Azobenzene	1.839	1.700	7.6	86	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
65	4,6-Dinitro-2-methylphenol	0.120	0.114	5.0	88	0.00
66 c	n-Nitrosodiphenylamine	0.568	0.519	8.6	86	0.00
67	4-Bromophenyl-phenylether	0.202	0.189	6.4	88	0.00
68	Hexachlorobenzene	0.200	0.188	6.0	88	0.00
69	Atrazine	0.224	0.209	6.7	87	0.00
70 C	Pentachlorophenol	0.136	0.109	19.9	73	0.00
71	Phenanthrene	1.054	0.967	8.3	87	0.00
72	Anthracene	1.047	0.974	7.0	88	0.00
73	Carbazole	1.044	0.992	5.0	90	0.00
74	Di-n-butylphthalate	1.225	1.198	2.2	93	0.00
75 C	Fluoranthene	1.296	1.208	6.8	89	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
77	Benzydine	0.648	0.585	9.7	89	0.00
78	Pyrene	1.419	1.316	7.3	90	0.00
79 S	Terphenyl-d14	1.026	0.978	4.7	93	0.00
80	Butylbenzylphthalate	0.600	0.601	-0.2	96	0.00
81	Benzo(a)anthracene	1.323	1.280	3.3	95	0.00
82	3,3'-Dichlorobenzidine	0.439	0.417	5.0	92	0.00
83	Chrysene	1.245	1.184	4.9	93	0.00
84	Bis(2-ethylhexyl)phthalate	0.853	0.848	0.6	96	0.00
85 c	Di-n-octyl phthalate	1.422	1.446	-1.7	97	0.00
86 I	Perylene-d12	1.000	1.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	1.393	1.345	3.4	95	-0.01
88	Benzo(b)fluoranthene	1.225	1.176	4.0	94	0.00
89	Benzo(k)fluoranthene	1.205	1.178	2.2	96	0.00
90 C	Benzo(a)pyrene	1.041	1.009	3.1	94	0.00
91	Dibenzo(a,h)anthracene	1.152	1.101	4.4	94	-0.01
92	Benzo(g,h,i)perylene	1.128	1.081	4.2	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG101921\
Data File : BG050628.D
Acq On : 19 Oct 2021 9:56
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 19 11:42:39 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 14 10:56:22 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0