

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG102016\
 Data File : BG024463.D
 Acq On : 20 Oct 2016 15:15
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02013

Quant Time: Oct 21 17:32:56 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG102016-MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 21 17:27:43 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2 S	1,4-Dioxane-d8	0.374	0.325	13.1	91	0.00
3 S	Phenol-d5	1.819	1.947	-7.0	108	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	1.124	1.205	-7.2	104	0.00
5 S	2-Chlorophenol-d4	1.280	1.317	-2.9	102	0.00
6 S	4-Methylphenol-d8	1.517	1.600	-5.5	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.150	0.160	-6.7	114	0.00
9 S	2-Nitrophenol-d4	0.174	0.181	-4.0	103	0.00
10 S	2,4-Dichlorophenol-d3	0.353	0.348	1.4	99	0.00
11 S	4-Chloroaniline-d4	0.367	0.413	-12.5	108	0.00
12	1-Methylnaphthalene	0.743	0.776	-4.4	106	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
14 S	Dimethylphthalate-d6	1.349	1.338	0.8	107	0.00
15 S	Acenaphthylene-d8	1.605	1.578	1.7	102	0.00
16 S	4-Nitrophenol-d4	0.231	0.216	6.5	105	0.00
17 S	Fluorene-d10	1.285	1.265	1.6	108	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
19 S	4,6-Dinitro-2-methylphenol-	0.134	0.125	6.7	108	0.00
20 S	Anthracene-d10	0.889	0.905	-1.8	107	0.00
21	1,2,3,4-Tetrachlorobenzene	0.305	0.302	1.0	102	0.00
22	Pentachlorobenzene	0.320	0.320	0.0	106	0.00
23 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
24 S	Pyrene-d10	0.869	0.914	-5.2	105	0.00
25 I	Perylene-d12	1.000	1.000	0.0	102	0.00
26 S	Benzo(a)pyrene-d12	0.875	0.898	-2.6	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0