

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102021\
 Data File : BG050642.D
 Acq On : 20 Oct 2021 12:37
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 20 13:54:19 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 10:56:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	35.210	12.0	85	0.00
3	Pyridine	40.000	36.298	9.3	88	0.00
4	n-Nitrosodimethylamine	40.000	37.463	6.3	94	-0.01
5 S	2-Fluorophenol	80.000	74.826	6.5	93	0.00
6	Aniline	40.000	37.752	5.6	95	0.00
7 S	Phenol-d6	80.000	76.797	4.0	96	0.00
8	2-Chlorophenol	40.000	38.740	3.1	97	0.00
9	Benzaldehyde	40.000	40.841	-2.1	93	0.00
10 C	Phenol	40.000	38.573	3.6	98	0.00
11	bis(2-Chloroethyl)ether	40.000	37.885	5.3	95	0.00
12	1,3-Dichlorobenzene	40.000	37.654	5.9	95	0.00
13 C	1,4-Dichlorobenzene	40.000	37.869	5.3	96	0.00
14	1,2-Dichlorobenzene	40.000	37.666	5.8	96	0.00
15	Benzyl Alcohol	40.000	39.092	2.3	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.268	4.3	97	0.00
17	2-Methylphenol	40.000	38.834	2.9	99	0.00
18	Hexachloroethane	40.000	38.672	3.3	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.288	6.8	94	0.00
20	3+4-Methylphenols	40.000	38.475	3.8	96	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	-0.01
22	Acetophenone	40.000	36.291	9.3	96	-0.01
23 S	Nitrobenzene-d5	80.000	74.086	7.4	96	0.00
24	Nitrobenzene	40.000	36.580	8.6	95	0.00
25	Isophorone	40.000	36.573	8.6	96	0.00
26 C	2-Nitrophenol	40.000	40.340	-0.9	103	0.00
27	2,4-Dimethylphenol	40.000	36.809	8.0	95	-0.01
28	bis(2-Chloroethoxy)methane	40.000	37.240	6.9	98	-0.01
29 C	2,4-Dichlorophenol	40.000	38.340	4.1	99	0.00
30	1,2,4-Trichlorobenzene	40.000	37.327	6.7	98	-0.01
31	Naphthalene	40.000	37.346	6.6	98	-0.01
32	Benzoic acid	40.000	35.475	11.3	87	0.01
33	4-Chloroaniline	40.000	37.524	6.2	98	0.00
34 C	Hexachlorobutadiene	40.000	37.357	6.6	98	0.00
35	Caprolactam	40.000	36.891	7.8	95	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.068	7.3	97	0.00
37	2-Methylnaphthalene	40.000	36.770	8.1	97	0.00
38	1-Methylnaphthalene	40.000	36.630	8.4	97	-0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.993	5.0	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.525	3.7	95	0.00
42 S	2,4,6-Tribromophenol	80.000	74.747	6.6	96	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.864	5.3	96	0.00
44	2,4,5-Trichlorophenol	40.000	39.622	0.9	101	0.00
45 S	2-Fluorobiphenyl	80.000	75.659	5.4	98	-0.01
46	1,1'-Biphenyl	40.000	37.810	5.5	98	0.00
47	2-Chloronaphthalene	40.000	38.087	4.8	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102021\
 Data File : BG050642.D
 Acq On : 20 Oct 2021 12:37
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 20 13:54:19 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 10:56:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.145	2.1	98	0.00
49	Acenaphthylene	40.000	37.259	6.9	96	0.00
50	Dimethylphthalate	40.000	37.400	6.5	97	0.00
51	2,6-Dinitrotoluene	40.000	38.790	3.0	98	0.00
52 C	Acenaphthene	40.000	37.357	6.6	95	-0.01
53	3-Nitroaniline	40.000	38.419	4.0	98	0.00
54 P	2,4-Dinitrophenol	40.000	33.438	16.4	83	0.00
55	Dibenzofuran	40.000	37.125	7.2	97	0.00
56 P	4-Nitrophenol	40.000	36.987	7.5	94	0.00
57	2,4-Dinitrotoluene	40.000	38.982	2.5	100	0.00
58	Fluorene	40.000	37.386	6.5	97	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.742	8.1	93	0.00
60	Diethylphthalate	40.000	37.271	6.8	97	0.00
61	4-Chlorophenyl-phenylether	40.000	37.439	6.4	96	-0.01
62	4-Nitroaniline	40.000	37.917	5.2	97	0.00
63	Azobenzene	40.000	37.053	7.4	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.060	4.8	97	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.716	8.2	95	0.00
67	4-Bromophenyl-phenylether	40.000	37.320	6.7	97	0.00
68	Hexachlorobenzene	40.000	37.167	7.1	97	0.00
69	Atrazine	40.000	36.937	7.7	95	0.00
70 C	Pentachlorophenol	40.000	35.811	10.5	90	0.00
71	Phenanthrene	40.000	36.318	9.2	95	0.00
72	Anthracene	40.000	36.536	8.7	95	0.00
73	Carbazole	40.000	36.606	8.5	96	0.00
74	Di-n-butylphthalate	40.000	37.562	6.1	98	0.00
75 C	Fluoranthene	40.000	35.737	10.7	94	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	95	-0.01
77	Benzydine	40.000	38.013	5.0	94	0.00
78	Pyrene	40.000	38.733	3.2	94	0.00
79 S	Terphenyl-d14	80.000	79.533	0.6	97	0.00
80	Butylbenzylphthalate	40.000	39.804	0.5	96	0.00
81	Benzo(a)anthracene	40.000	38.556	3.6	95	0.00
82	3,3'-Dichlorobenzidine	40.000	39.134	2.2	95	0.00
83	Chrysene	40.000	37.941	5.1	93	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.353	1.6	95	-0.01
85 c	Di-n-octyl phthalate	40.000	40.171	-0.4	96	0.00
86 I	Perylene-d12	20.000	20.000	0.0	95	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.499	3.8	94	-0.02
88	Benzo(b)fluoranthene	40.000	38.295	4.3	93	-0.01
89	Benzo(k)fluoranthene	40.000	39.207	2.0	96	0.00
90 C	Benzo(a)pyrene	40.000	38.858	2.9	94	-0.01
91	Dibenzo(a,h)anthracene	40.000	38.301	4.2	94	-0.02
92	Benzo(g,h,i)perylene	40.000	38.595	3.5	94	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102021\
Data File : BG050642.D
Acq On : 20 Oct 2021 12:37
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 20 13:54:19 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 14 10:56:22 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0