

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102122\
 Data File : BG055237.D
 Acq On : 21 Oct 2022 17:02
 Operator : CG/JU
 Sample : N5217-24
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SB-E3-COMP

Quant Time: Oct 22 01:29:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG101922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 20 09:21:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.288	152	46758	20.000	ng	0.00
21) Naphthalene-d8	11.114	136	196032	20.000	ng	0.00
39) Acenaphthene-d10	14.898	164	123947	20.000	ng	0.00
64) Phenanthrene-d10	17.630	188	258659	20.000	ng	0.00
76) Chrysene-d12	21.907	240	216108	20.000	ng	-0.01
86) Perylene-d12	25.321	264	225992	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.809	112	293865	113.072	ng	0.00
7) Phenol-d6	7.424	99	394035	99.500	ng	0.00
23) Nitrobenzene-d5	9.457	82	252001	62.061	ng	0.00
42) 2,4,6-Tribromophenol	16.373	330	131610	102.473	ng	0.00
45) 2-Fluorobiphenyl	13.523	172	510022	65.727	ng	0.00
79) Terphenyl-d14	20.204	244	695893	60.000	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102122\
Data File : BG055237.D
Acq On : 21 Oct 2022 17:02
Operator : CG/JU
Sample : N5217-24
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SB-E3-COMP

Quant Time: Oct 22 01:29:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG101922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 20 09:21:42 2022
Response via : Initial Calibration

