



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102122\
 Data File : BG055234.D
 Acq On : 21 Oct 2022 14:24
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG101922.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Oct 20 09:21:42 2022






AutoFind: Scans 2509, 2510, 2511; Background Corrected with Scan 2503

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.5	39413	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	41.6	36843	PASS
70	69	0.00	2	0.7	246	PASS
127	198	10	80	51.0	45152	PASS
197	198	0.00	2	0.3	245	PASS
198	198	100	100	100.0	88552	PASS
199	198	5	9	7.1	6256	PASS
275	198	10	60	26.8	23712	PASS
365	198	1	100	3.8	3367	PASS
441	198	0.01	100	13.4	11838	PASS
442	442	50	100	100.0	72120	PASS
443	442	15	24	19.6	14161	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102122\
Data File : BG055234.D
Acq On : 21 Oct 2022 14:24
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 27 13:21:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 20 17:56:56 2021
Response via : Initial Calibration



TIC: BG055234.D\data.ms

(77) Benzidine

19.833min (+ 0.133) 0.00 ng




response 660104

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	15.04
183.00	13.20	12.72
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102122\
Data File : BG055234.D
Acq On : 21 Oct 2022 14:24
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 27 13:21:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 20 17:56:56 2021
Response via : Initial Calibration



TIC: BG055234.D\data.ms

(70) Pentachlorophenol (C)

17.283min (+ 0.162) 0.00 ng

response 89021

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.60
264.00	61.60	63.14
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/21/2022	BNA_G	BG055234.D
Compound Name	Response	Retention Time
DDT	289385	21.067
DDD	8521	20.614
DDE	454	20.121
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8975	298360	3.01

Instrument :
BNA_G
ClientSampleId :
DFTPP