

Data Path : Z:\HPCHEM1\BNA G\DATA\BG102315\
 Data File : BG019326.D
 Acq On : 24 Oct 2015 8:33
 Operator : UM/NP
 Sample : SSTDC0020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02032

Quant Time: Oct 25 02:37:00 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102315.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Oct 25 01:11:12 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	130	0.00
2	1,4-Dioxane	0.404	0.459	-13.6	158#	0.00
3 S	1,4-Dioxane-d8	0.360	0.437	-21.4#	141	0.00
4	Benzaldehyde	1.092	1.386	-26.9#	144	0.00
5 S	Phenol-d5	1.760	1.943	-10.4	145	0.00
6	Phenol	1.850	2.040	-10.3	148	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.051	1.232	-17.2	145	0.00
8	Bis(2-Chloroethyl)ether	1.257	1.457	-15.9	148	0.00
9 S	2-Chlorophenol-d4	1.136	1.197	-5.4	137	0.00
10	2-Chlorophenol	1.155	1.242	-7.5	146	0.00
11	2-Methylphenol	1.432	1.459	-1.9	132	0.00
12	2,2'-oxybis(1-Chloropropane	0.926	1.089	-17.6	160#	0.00
13 S	4-Methylphenol-d8	1.372	1.461	-6.5	142	0.00
14	Acetophenone	2.305	2.497	-8.3	141	0.00
15 P	N-Nitroso-di-n-propylamine	1.417	1.558	-10.0	150	0.00
16	4-Methylphenol	1.519	1.622	-6.8	138	0.00
17	Hexachloroethane	0.592	0.599	-1.2	133	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	138	0.00
19 S	Nitrobenzene-d5	0.152	0.162	-6.6	144	0.00
20	Nitrobenzene	0.541	0.587	-8.5	142	0.00
21	Isophorone	0.945	1.030	-9.0	147	0.00
22 S	2-Nitrophenol-d4	0.173	0.183	-5.8	143	0.00
23 C	2-Nitrophenol	0.193	0.193	0.0	134	0.00
24	2,4-Dimethylphenol	0.489	0.522	-6.7	141	0.00
25	Bis(2-Chloroethoxy)methane	0.457	0.496	-8.5	143	0.00
26 S	2,4-Dichlorophenol-d3	0.378	0.403	-6.6	138	0.00
27 C	2,4-Dichlorophenol	0.369	0.384	-4.1	140	0.00
28	Naphthalene	0.996	1.048	-5.2	136	0.00
29 S	4-Chloroaniline-d4	0.376	0.445	-18.4	138	0.00
30	4-Chloroaniline	0.390	0.461	-18.2	138	0.00
31 C	Hexachlorobutadiene	0.382	0.391	-2.4	138	0.00
32	Caprolactam	0.118	0.130	-10.2	150#	0.00
33 C	4-Chloro-3-methylphenol	0.475	0.495	-4.2	139	0.00
34	2-Methylnaphthalene	0.829	0.841	-1.4	136	0.00
35	1-Methylnaphthalene	0.772	0.818	-6.0	136	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	132	0.00
37	1,2,4,5-Tetrachlorobenzene	0.817	0.865	-5.9	137	0.00
38	Hexachlorocyclopentadiene	0.653	0.673	-3.1	131	0.00
39 C	2,4,6-Trichlorophenol	0.496	0.513	-3.4	135	0.00
40	2,4,5-Trichlorophenol	0.555	0.554	0.2	132	0.00
41	1,1'-Biphenyl	1.443	1.474	-2.1	134	0.00
42	2-Chloronaphthalene	1.135	1.155	-1.8	132	0.00
43	2-Nitroaniline	0.448	0.450	-0.4	132	0.00
44 S	Dimethylphthalate-d6	1.658	1.676	-1.1	130	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG102315\
 Data File : BG019326.D
 Acq On : 24 Oct 2015 8:33
 Operator : UM/NP
 Sample : SSTDC0020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02032

Quant Time: Oct 25 02:37:00 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102315.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Oct 25 01:11:12 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.656	1.693	-2.2	134	0.00
46	2,6-Dinitrotoluene	0.348	0.345	0.9	132	0.00
47 S	Acenaphthylene-d8	1.835	1.891	-3.1	131	0.00
48	Acenaphthylene	1.861	1.907	-2.5	134	0.00
49	3-Nitroaniline	0.290	0.305	-5.2	127	0.00
50 C	Acenaphthene	1.244	1.294	-4.0	133	0.00
51	2,4-Dinitrophenol	0.277	0.273	1.4	129	0.00
52 S	4-Nitrophenol-d4	0.270	0.272	-0.7	130	0.00
53	4-Nitrophenol	0.445	0.447	-0.4	118	0.00
54	Dibenzofuran	1.888	1.908	-1.1	131	0.00
55	2,4-Dinitrotoluene	0.512	0.529	-3.3	136	0.00
56	2,3,4,6-Tetrachlorophenol	0.620	0.592	4.5	123	0.00
57	Diethylphthalate	1.650	1.655	-0.3	129	0.00
58 S	Fluorene-d10	1.556	1.557	-0.1	130	0.00
59	Fluorene	1.640	1.644	-0.2	131	0.00
60	4-Chlorophenyl-phenylether	1.007	0.986	2.1	126	0.00
61	4-Nitroaniline	0.342	0.344	-0.6	129	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	124	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.132	0.135	-2.3	122	0.00
64	4,6-Dinitro-2-methylphenol	0.138	0.146	-5.8	130	0.00
65	N-Nitrosodiphenylamine	0.509	0.521	-2.4	125	0.00
66	4-Bromophenyl-phenylether	0.248	0.250	-0.8	123	0.00
67	Hexachlorobenzene	0.257	0.267	-3.9	124	0.00
68	Atrazine	0.260	0.267	-2.7	121	0.00
69 C	Pentachlorophenol	0.183	0.177	3.3	117	0.00
70	Phenanthrene	1.011	1.057	-4.5	124	0.00
71 S	Anthracene-d10	0.916	0.953	-4.0	125	0.00
72	Anthracene	1.021	1.079	-5.7	126	0.00
73	1,2,3,4-Tetrachlorobenzene	0.277	0.299	-7.9	132	0.00
74	Pentachlorobenzene	0.285	0.301	-5.6	129	0.00
75	Carbazole	0.831	0.915	-10.1	126	0.00
76	Di-n-butylphthalate	0.980	1.059	-8.1	123	0.00
77 C	Fluoranthene	1.357	1.525	-12.4	123	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	126	0.00
79 S	Pyrene-d10	0.876	0.881	-0.6	124	0.00
80	Pyrene	1.121	1.127	-0.5	124	0.00
81	Butylbenzylphthalate	0.344	0.361	-4.9	130	0.00
82	3,3'-Dichlorobenzidine	0.396	0.424	-7.1	123	0.00
83	Benzo(a)anthracene	1.160	1.206	-4.0	128	0.00
84	Bis(2-ethylhexyl)phthalate	0.476	0.502	-5.5	128	0.00
85	Chrysene	1.066	1.127	-5.7	130	0.00
86 I	Perylene-d12	1.000	1.000	0.0	132	0.00
87	Di-n-octyl phthalate	0.843	0.898	-6.5	133	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG102315\
Data File : BG019326.D
Acq On : 24 Oct 2015 8:33
Operator : UM/NP
Sample : SSTDCCC020
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02032

Quant Time: Oct 25 02:37:00 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102315.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Oct 25 01:11:12 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.223	1.228	-0.4	128	0.00
89	Benzo(k)fluoranthene	1.167	1.176	-0.8	130	0.00
90 S	Benzo(a)pyrene-d12	1.020	1.067	-4.6	132	0.00
91 C	Benzo(a)pyrene	1.133	1.178	-4.0	132	-0.01
92	Indeno(1,2,3-cd)pyrene	1.279	1.322	-3.4	132	0.00
93	Dibenzo(a,h)anthracene	1.061	1.094	-3.1	133	0.00
94	Benzo(a,h,i)perylene	1.058	1.099	-3.9	133	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0