

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG102516\
 Data File : BG024502.D
 Acq On : 25 Oct 2016 18:14
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040

Quant Time: Oct 26 02:13:13 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 25 17:25:08 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.27	152	113711	20.00	ng	0.00
21) Naphthalene-d8	11.10	136	441084	20.00	ng	0.00
38) Acenaphthene-d10	14.89	164	340135	20.00	ng	0.00
63) Phenanthrene-d10	17.63	188	786590	20.00	ng	0.00
75) Chrysene-d12	21.93	240	1108292	20.00	ng	0.00
86) Perylene-d12	25.36	264	1196461	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.80	112	514582	74.17	ng	0.00
7) Phenol-d6	7.40	99	696644	73.20	ng	0.00
23) Nitrobenzene-d5	9.45	82	731854	75.54	ng	0.00
41) 2,4,6-Tribromophenol	16.37	330	403210	79.35	ng	0.00
44) 2-Fluorobiphenyl	13.51	172	1530144	74.77	ng	0.00
78) Terphenyl-d14	20.21	244	2727404	73.53	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.67	88	110282	38.91	ng	89
3) Pyridine	4.08	79	318898	37.58	ng	89
4) n-Nitrosodimethylamine	4.00	42	163836	37.29	ng	# 91
6) Aniline	7.59	93	433481	35.95	ng	98
8) 2-Chlorophenol	7.83	128	253634	35.96	ng	96
9) Benzaldehyde	7.41	77	215836	36.70	ng	95
10) Phenol	7.43	94	354303	36.12	ng	95
11) bis(2-Chloroethyl)ether	7.68	93	265298	36.00	ng	93
12) 1,3-Dichlorobenzene	8.16	146	321638	36.83	ng	99
13) 1,4-Dichlorobenzene	8.31	146	319781	37.07	ng	93
14) 1,2-Dichlorobenzene	8.63	146	308610	37.21	ng	96
15) Benzyl Alcohol	8.51	79	317899	35.91	ng	93
16) 2,2'-oxybis(1-Chloropropan	8.79	45	490416	35.86	ng	96
17) 2-Methylphenol	8.70	107	239330	36.82	ng	97
18) Hexachloroethane	9.37	117	125694	37.91	ng	94
19) n-Nitroso-di-n-propylamine	9.07	70	258479	36.85	ng	95
20) 3+4-Methylphenols	9.02	107	319737	36.63	ng	94
22) Acetophenone	9.10	105	428124	37.79	ng	# 93
24) Nitrobenzene	9.49	77	402946	37.39	ng	96
25) Isophorone	10.01	82	675233	37.47	ng	97
26) 2-Nitrophenol	10.21	139	153749	38.43	ng	94
27) 2,4-Dimethylphenol	10.24	122	260889	37.25	ng	94
28) bis(2-Chloroethoxy)methane	10.49	93	345052	37.01	ng	99
29) 2,4-Dichlorophenol	10.73	162	272852	37.12	ng	97
30) 1,2,4-Trichlorobenzene	10.95	180	320464	36.75	ng	99
31) Naphthalene	11.15	128	820062	37.27	ng	96
32) Benzoic acid	10.35	122	225183	38.61	ng	99
33) 4-Chloroaniline	11.25	127	359040	37.58	ng	# 95
34) Hexachlorobutadiene	11.42	225	257064	37.67	ng	94
35) Caprolactam	12.02	113	102464	38.08	ng	97
36) 4-Chloro-3-methylphenol	12.34	107	341194	38.45	ng	97
37) 2-Methylnaphthalene	12.73	142	601615	37.47	ng	95
39) 1,2,4,5-Tetrachlorobenzene	13.09	216	415723	36.24	ng	98
40) Hexachlorocyclopentadiene	13.07	237	247319	38.28	ng	91


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG102516\
 Data File : BG024502.D
 Acq On : 25 Oct 2016 18:14
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040

Quant Time: Oct 26 02:13:13 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 25 17:25:08 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.33	196	265469	38.50	ng	98
43) 2,4,5-Trichlorophenol	13.39	196	286092	38.37	ng	98
45) 1,1'-Biphenyl	13.73	154	883220	37.42	ng	98
46) 2-Chloronaphthalene	13.77	162	667682	37.15	ng	96
47) 2-Nitroaniline	13.98	65	292793	39.79	ng	95
48) Acenaphthylene	14.62	152	1029299	37.34	ng	99
49) Dimethylphthalate	14.33	163	907347	37.58	ng	98
50) 2,6-Dinitrotoluene	14.45	165	201002	38.99	ng	92
51) Acenaphthene	14.95	154	774412	37.69	ng	97
52) 3-Nitroaniline	14.79	138	196735	36.88	ng	91
53) 2,4-Dinitrophenol	14.99	184	142442	38.13	ng	90
54) Dibenzofuran	15.28	168	1061970	37.38	ng	99
55) 4-Nitrophenol	15.07	139	184191	39.63	ng	94
56) 2,4-Dinitrotoluene	15.24	165	293649	39.20	ng	# 98
57) Fluorene	15.93	166	891388	37.84	ng	97
58) 2,3,4,6-Tetrachlorophenol	15.50	232	311356	40.28	ng	93
59) Diethylphthalate	15.68	149	942062	37.21	ng	95
60) 4-Chlorophenyl-phenylether	15.92	204	497039	37.60	ng	92
61) 4-Nitroaniline	15.95	138	228380	39.19	ng	90
62) Azobenzene	16.20	77	954120	37.79	ng	98
64) 4,6-Dinitro-2-methylphenol	15.99	198	214063	39.50	ng	85
65) n-Nitrosodiphenylamine	16.13	169	792609	36.86	ng	99
66) 4-Bromophenyl-phenylether	16.81	248	355426	37.22	ng	95
67) Hexachlorobenzene	16.93	284	400736	37.98	ng	# 89
68) Atrazine	17.06	200	347453	37.67	ng	98
69) Pentachlorophenol	17.27	266	279979	40.21	ng	90
70) Phenanthrene	17.67	178	1510041	37.66	ng	98
71) Anthracene	17.76	178	1527319	37.76	ng	98
72) Carbazole	18.03	167	1394302	37.80	ng	98
73) Di-n-butylphthalate	18.56	149	1630938	38.34	ng	98
74) Fluoranthene	19.67	202	2006292	39.58	ng	96
76) Benzidine	19.84	184	994461	38.08	ng	98
77) Pyrene	20.03	202	2042522	37.70	ng	99
79) Butylbenzylphthalate	20.89	149	841178	37.24	ng	99
80) Benzo(a)anthracene	21.90	228	2266063	37.22	ng	99
81) 3,3'-Dichlorobenzidine	21.82	252	906287	36.90	ng	97
82) Chrysene	21.98	228	2176562	37.54	ng	97
83) Bis(2-ethylhexyl)phthalate	21.77	149	1203971	37.26	ng	98
84) Di-n-octyl phthalate	23.05	149	2056902	38.36	ng	95
85) Indeno(1,2,3-cd)pyrene	29.31	276	2912989	36.39	ng	# 97
87) Benzo(b)fluoranthene	24.26	252	2397855	37.22	ng	100
88) Benzo(k)fluoranthene	24.34	252	2353465	37.68	ng	96
89) Benzo(a)pyrene	25.20	252	2360402	37.35	ng	99
90) Dibenzo(a,h)anthracene	29.38	278	2508842	36.17	ng	99
91) Benzo(g,h,i)perylene	30.56	276	2503177	36.12	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 8.27 min Scan# 925

Delta R.T. 0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040


Tgt Ion:152 Resp: 113711

Ion Ratio Lower Upper

152 100

150 156.0 114.0 171.0




115 64.4 48.1 72.1



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 38.91 ng

RT: 3.67 min Scan# 141

Delta R.T. 0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14


Tgt Ion: 88 Resp: 110282

Ion Ratio Lower Upper

88 100

58 87.3 79.4 119.2



43 37.7 33.8 50.6




Abundance Ion 88.00 (87.70 to 88.70): BGC

Ion 58.00 (57.70 to 58.70): BGC

Ion 43.00 (42.70 to 43.70): BGC





#3

Pyridine

Concen: 37.58 ng

RT: 4.08 min Scan# 211

Delta R.T. -0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040





Tgt Ion: 79 Resp: 318898

Ion Ratio Lower Upper

79 100

52 85.7 77.8 116.6

51 47.4 43.2 64.8



#4

n-Nitrosodimethylamine

Concen: 37.29 ng

RT: 4.00 min Scan# 197

Delta R.T. 0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14




Tgt Ion: 42 Resp: 163836


Ion Ratio Lower Upper

42 100

74 104.9 76.7 115.1

44 4.0 6.1 9.1#





#5

2-Fluorophenol

Concen: 74.17 ng

RT: 5.80 min Scan# 504

Delta R.T. 0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14





Instrument :

BNA_G

ClientSampleId :

SSTDCCC040

Tgt Ion:	112	Resp:	514582
Ion	Ratio	Lower	Upper
112	100		
64	75.3	61.8	92.6
63	40.8	37.2	55.8



#6

Aniline

Concen: 35.95 ng




RT: 7.59 min Scan# 809


Delta R.T. 0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14

Tgt Ion:	93	Resp:	433481
Ion	Ratio	Lower	Upper
93	100		
66	44.1	35.4	53.2
65	24.5	21.2	31.8





#7

Phenol-d6

Concen: 73.20 ng

RT: 7.40 min Scan# 777

Delta R.T. 0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040


Tgt Ion: 99 Resp: 696644

Ion Ratio Lower Upper




99 100

42 22.9 20.5 30.7

71 46.9 37.6 56.4



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#8

2-Chlorophenol

Concen: 35.96 ng

RT: 7.83 min Scan# 849

Delta R.T. -0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14


Tgt Ion: 128 Resp: 253634

Ion Ratio Lower Upper



128 100


130 29.0 11.4 51.4

64 63.3 46.6 86.6



Abundance Ion 128.00 (127.70 to 128.70): BGC
Ion 130.00 (129.70 to 130.70): BGC
Ion 64.00 (63.70 to 64.70): BGC





#9

Benzaldehyde

Concen: 36.70 ng

RT: 7.41 min Scan# 778

Delta R.T. 0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040


Tgt Ion: 77 Resp: 215836

Ion Ratio Lower Upper

77 100

105 83.0 60.9 100.9




106 82.6 55.1 95.1



Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 106.00 (105.70 to 106.70): BGC



#10

Phenol

Concen: 36.12 ng

RT: 7.43 min Scan# 782

Delta R.T. 0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14


Tgt Ion: 94 Resp: 354303

Ion Ratio Lower Upper

94 100

65 34.8 20.6 60.6



66 56.4 35.5 75.5




Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC









#11
bis(2-Chloroethyl)ether
Concen: 36.00 ng
RT: 7.68 min Scan# 824
Delta R.T. 0.00 min
Lab File: BG024502.D
Acq: 25 Oct 2016 18:14

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 265298
Ion Ratio Lower Upper
93 100
63 90.4 78.5 118.5
95 30.8 9.8 49.8




Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC






#12
1,3-Dichlorobenzene
Concen: 36.83 ng
RT: 8.16 min Scan# 906
Delta R.T. -0.00 min
Lab File: BG024502.D
Acq: 25 Oct 2016 18:14

Tgt Ion: 146 Resp: 321638
Ion Ratio Lower Upper
146 100
148 62.8 49.2 73.8
75 42.6 33.4 50.2



Abundance Ion 146.00 (145.70 to 146.70): BGC
Ion 148.00 (147.70 to 148.70): BGC
Ion 75.00 (74.70 to 75.70): BGC









#13
 1,4-Dichlorobenzene
 Concen: 37.07 ng
 RT: 8.31 min Scan# 931
 Delta R.T. -0.00 min
 Lab File: BG024502.D
 Acq: 25 Oct 2016 18:14




Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
146	100		
148	60.0	52.2	78.2
111	40.5	37.0	55.6



#14
 1,2-Dichlorobenzene
 Concen: 37.21 ng
 RT: 8.63 min Scan# 986
 Delta R.T. 0.00 min
 Lab File: BG024502.D
 Acq: 25 Oct 2016 18:14

Tgt Ion	Ratio	Lower	Upper
146	100		
148	69.1	54.6	81.8
111	44.7	39.7	59.5





#15

Benzyl Alcohol

Concen: 35.91 ng

RT: 8.51 min Scan# 965

Delta R.T. 0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040





Tgt Ion: 79 Resp: 317899

Ion Ratio Lower Upper

79 100

108 62.7 45.5 68.3

77 63.5 54.3 81.5



#16

2,2'-oxybis(1-Chloropropane)

Concen: 35.86 ng

RT: 8.79 min Scan# 1013

Delta R.T. -0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14




Tgt Ion: 45 Resp: 490416


Ion Ratio Lower Upper

45 100

77 12.9 0.0 30.6

79 8.3 0.0 28.5









#17
2-Methylphenol
Concen: 36.82 ng
RT: 8.70 min Scan# 997
Delta R.T. -0.00 min
Lab File: BG024502.D
Acq: 25 Oct 2016 18:14

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:	107	Resp:	239330
Ion	Ratio	Lower	Upper
107	100		
108	104.7	85.6	128.4
77	61.8	51.4	77.2
79	59.7	49.2	73.8




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BGC
Ion 79.00 (78.70 to 79.70): BGC






#18
Hexachloroethane
Concen: 37.91 ng
RT: 9.37 min Scan# 1111
Delta R.T. 0.00 min
Lab File: BG024502.D
Acq: 25 Oct 2016 18:14

Tgt Ion:	117	Resp:	125694
Ion	Ratio	Lower	Upper
117	100		
119	98.6	69.8	104.6
201	118.5	95.4	143.0



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E









#19
n-Nitroso-di-n-propylamine
Concen: 36.85 ng
RT: 9.07 min Scan# 1061
Delta R.T. -0.00 min
Lab File: BG024502.D
Acq: 25 Oct 2016 18:14




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040


Tgt Ion:	70	Resp:	258479
Ion	Ratio	Lower	Upper
70	100		
42	64.6	56.1	84.1
101	10.3	7.5	11.3
130	17.6	14.6	21.8



#20
3+4-Methylphenols
Concen: 36.63 ng
RT: 9.02 min Scan# 1053
Delta R.T. -0.00 min
Lab File: BG024502.D
Acq: 25 Oct 2016 18:14

Tgt Ion:	107	Resp:	319737
Ion	Ratio	Lower	Upper
107	100		
108	99.2	70.8	110.8
77	49.0	25.8	65.8
79	39.6	20.1	60.1





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 11.10 min Scan# 1406

Delta R.T. -0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14


Instrument :

BNA_G

ClientSampleId :

SSTDCCC040

Tgt Ion:	136	Resp:	441084
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.9	13.3
54	9.9	9.3	13.9
68	5.5	5.1	7.7






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BGC

Ion 68.00 (67.70 to 68.70): BGC



#22

Acetophenone

Concen: 37.79 ng


RT: 9.10 min Scan# 1066

Delta R.T. 0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14

Tgt Ion:	105	Resp:	428124
Ion	Ratio	Lower	Upper
105	100		
71	0.5	0.9	1.3#
51	36.9	34.0	51.0
120	23.4	17.3	25.9






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BGC

Ion 51.00 (50.70 to 51.70): BGC

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 75.54 ng
 RT: 9.45 min Scan# 1125
 Delta R.T. -0.00 min
 Lab File: BG024502.D
 Acq: 25 Oct 2016 18:14




Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040


Tgt Ion: 82 Resp: 731854
 Ion Ratio Lower Upper
 82 100
 128 34.4 26.4 39.6
 54 59.1 49.4 74.2



#24
 Nitrobenzene
 Concen: 37.39 ng
 RT: 9.49 min Scan# 1132
 Delta R.T. -0.00 min
 Lab File: BG024502.D
 Acq: 25 Oct 2016 18:14

Tgt Ion: 77 Resp: 402946
 Ion Ratio Lower Upper
 77 100
 123 34.9 26.1 39.1
 65 16.5 12.4 18.6





#25

Isophorone

Concen: 37.47 ng

RT: 10.01 min Scan# 1220

Delta R.T. 0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040





Tgt Ion: 82 Resp: 675233

Ion Ratio Lower Upper

82 100

95 8.9 7.5 11.3

138 13.7 12.3 18.5



#26

2-Nitrophenol

Concen: 38.43 ng

RT: 10.21 min Scan# 1254

Delta R.T. 0.00 min

Lab File: BG024502.D

Acq: 25 Oct 2016 18:14

Tgt Ion: 139 Resp: 153749

Ion Ratio Lower Upper

139 100

109 39.0 38.6 58.0

65 71.2 57.1 85.7

