


Data Path : Z:\HPCHEM1\BNA G\DATA\BG102516\
 Data File : BG024501.D
 Acq On : 25 Oct 2016 17:36
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG102516.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Oct 26 11:14:13 2016



AutoFind: Scans 2584, 2585, 2586; Background Corrected with Scan 2576

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.3	16771	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	40.3	18632	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	39.4	18206	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	46261	PASS
199	198	5	9	6.2	2883	PASS
275	198	10	60	35.6	16490	PASS
365	198	1	100	3.8	1771	PASS
441	442	0.01	24	14.7	6731	PASS
442	198	50	100	99.2	45904	PASS
443	442	15	24	19.2	8824	PASS