


Data Path : Z:\HPCHEM1\BNA_G\Data\BG102517\
 Data File : BG029433.D
 Acq On : 25 Oct 2017 8:41
 Operator : SJ/JU
 Sample : DFTPP61
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP61

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG101417.M
 Title : SVOA CALIBRATION
 Last Update : Thu Oct 26 09:08:53 2017



AutoFind: Scans 2493, 2494, 2495; Background Corrected with Scan 2486

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.8	72133	PASS
68	69	0.00	2	0.5	352	PASS
69	198	0.00	100	32.3	66936	PASS
70	69	0.00	2	0.5	313	PASS
127	198	10	80	41.4	85854	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	207296	PASS
199	198	5	9	6.7	13892	PASS
275	198	10	60	28.3	58744	PASS
365	198	1	100	3.7	7636	PASS
441	442	0.01	100	15.0	26598	PASS
442	198	50	100	85.4	177080	PASS
443	442	15	24	19.9	35181	PASS

m/z	Abundance
37.00	423.0
38.10	621.0
39.00	3885.0
40.10	287.0
41.10	912.0
42.00	181.0
43.00	681.0
44.00	495.0
44.90	168.0
49.00	627.0
50.00	15046.0
51.00	56640.0
52.00	2859.0
53.00	436.0
55.00	1101.0
56.00	2127.0
57.00	4283.0
58.00	376.0
61.10	690.0
62.00	756.0
63.00	2099.0
64.00	275.0
65.00	1245.0
66.00	274.0
67.00	377.0
69.00	53048.0
69.90	303.0
70.90	548.0
73.00	511.0
74.00	5176.0
75.00	8822.0
76.00	3183.0
77.00	61192.0
78.00	4549.0
79.00	3611.0
80.00	3541.0
81.00	5043.0
82.00	1406.0
83.00	1402.0
84.10	242.0
85.00	1088.0
86.00	986.0
87.00	703.0
88.00	189.0
89.00	154.0
91.00	929.0
91.90	994.0
93.00	7603.0
94.00	429.0
95.00	794.0
95.90	683.0
97.10	557.0
98.00	5164.0
99.00	4693.0
100.00	389.0
101.00	2504.0
103.00	858.0
104.00	1582.0
105.00	1330.0
105.90	531.0
107.00	18288.0
108.00	3268.0
109.10	1040.0
110.00	32240.0
111.00	4824.0
111.90	600.0
113.00	266.0
115.00	302.0
115.90	1025.0
117.00	15015.0
118.00	1089.0
119.10	447.0
120.00	350.0
121.00	227.0
122.00	1039.0
122.90	2319.0
124.00	1032.0
125.00	939.0
127.00	64944.0
128.00	4580.0
129.00	24104.0
130.00	2164.0
131.00	589.0
132.10	256.0
133.10	175.0
133.90	823.0
135.00	1812.0
136.00	829.0
137.00	1353.0
137.80	265.0
139.10	180.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

140.00	425.0
141.00	3373.0
142.00	1202.0
143.00	1059.0
144.00	204.0
145.00	476.0
146.00	537.0
147.00	1709.0
148.00	3825.0
149.00	706.0
150.00	233.0
150.90	468.0
152.00	216.0
152.90	1495.0
154.00	1048.0
155.00	2068.0
156.00	2798.0
157.00	715.0
158.00	742.0
159.10	495.0
160.00	1296.0
160.90	1580.0
162.00	619.0
163.00	299.0
165.00	1254.0
166.00	1528.0
167.00	7024.0
168.00	4069.0
169.00	605.0
169.90	433.0
171.00	441.0
172.00	850.0
172.90	735.0
174.00	1594.0
175.00	2311.0
176.00	769.0
177.00	1506.0
177.90	815.0
179.00	5301.0
180.00	3382.0
181.00	2101.0
182.00	523.0
183.00	235.0
184.00	545.0
185.00	2905.0
186.00	20464.0
187.00	5461.0
188.00	950.0
189.00	1652.0
189.90	299.0
191.00	757.0
192.00	1812.0
193.00	1883.0
194.00	779.0
195.00	216.0
196.00	4534.0
197.90	145472.0
198.90	9827.0
199.90	705.0
201.40	720.0
201.90	632.0
203.00	1354.0
204.00	5415.0
205.00	8935.0
206.00	39640.0
207.00	5560.0
208.00	1795.0
209.00	613.0
210.30	690.0
211.00	1374.0
211.90	296.0
213.10	461.0
215.00	569.0
216.00	990.0
217.00	9106.0
218.00	1308.0
219.00	320.0
219.90	304.0
221.00	5805.0
222.00	1142.0
223.00	1860.0
224.00	21256.0
225.00	5355.0
226.10	685.0
227.00	9878.0
227.90	1494.0
229.00	1944.0
230.00	262.0
231.00	1164.0
232.00	224.0
233.00	209.0
233.90	551.0
235.10	667.0
235.90	418.0
237.00	758.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

239.00	420.0
240.10	152.0
240.90	403.0
242.00	1082.0
243.00	1245.0
244.00	15465.0
245.00	2212.0
246.00	3246.0
246.90	429.0
248.90	617.0
249.80	187.0
250.90	184.0
252.00	174.0
253.00	495.0
255.00	83832.0
256.00	12135.0
257.00	1119.0
258.00	4959.0
259.00	889.0
263.80	177.0
264.90	2371.0
266.00	300.0
271.10	204.0
272.00	474.0
273.00	2586.0
274.00	6434.0
275.00	36744.0
276.00	4820.0
277.00	3340.0
277.90	656.0
282.10	161.0
283.00	474.0
284.00	257.0
285.00	744.0
286.10	151.0
292.00	204.0
293.00	613.0
293.90	291.0
295.00	281.0
296.00	10384.0
297.00	1421.0
303.00	1254.0
304.00	499.0
307.90	177.0
309.00	206.0
314.00	508.0
315.00	986.0
315.90	760.0
316.90	262.0
321.00	223.0
321.90	190.0
323.00	3001.0
324.10	639.0
326.90	612.0
328.10	282.0
329.00	201.0
330.70	154.0
331.90	198.0
333.00	378.0
334.00	1743.0
335.00	460.0
341.10	324.0
341.90	226.0
346.00	730.0
347.10	166.0
352.00	772.0
353.00	572.0
354.10	825.0
355.10	159.0
365.00	4237.0
366.00	602.0
372.00	1227.0
373.00	387.0
382.90	504.0
390.00	228.0
402.00	608.0
403.00	878.0
404.00	376.0
421.10	578.0
422.00	697.0
423.00	4678.0
424.00	732.0
424.90	228.0
441.10	13722.0
442.00	92776.0
443.10	18808.0
444.10	1709.0
445.10	175.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

m/z	Abundance
36.90	292.0
38.00	1054.0
39.00	6537.0
40.00	417.0
41.00	1046.0
43.10	439.0
44.00	528.0
49.10	769.0
50.00	21760.0
51.00	85576.0
52.00	4931.0
53.10	348.0
55.00	1193.0
56.00	2308.0
57.00	6633.0
58.00	206.0
61.00	1026.0
62.00	982.0
63.10	3216.0
64.00	500.0
65.00	1703.0
66.00	267.0
67.00	374.0
69.00	78488.0
69.90	707.0
71.20	365.0
73.00	745.0
74.00	8436.0
75.00	13507.0
76.10	4865.0
77.00	93528.0
78.00	7174.0
79.00	6311.0
80.00	4991.0
81.00	6966.0
82.00	1609.0
83.00	1780.0
84.00	227.0
85.00	1315.0
86.00	1968.0
87.00	1043.0
88.00	511.0
88.90	202.0
91.00	2042.0
92.00	1945.0
93.00	11140.0
94.00	726.0
95.10	717.0
96.00	597.0
97.10	771.0
98.00	8552.0
99.00	6745.0
99.90	950.0
101.00	4133.0
101.90	186.0
103.00	1410.0
103.90	2842.0
105.00	2318.0
106.00	726.0
107.00	28736.0
108.00	5117.0
110.00	51848.0
111.00	8684.0
112.00	990.0
113.00	584.0
114.10	199.0
115.00	533.0
116.00	1600.0
117.00	23200.0
118.00	1722.0
119.20	407.0
120.00	524.0
121.10	273.0
122.00	1926.0
123.00	3432.0
124.00	1598.0
125.00	1692.0
127.00	100328.0
128.00	8064.0
129.00	40288.0
130.00	3769.0
131.00	891.0
132.00	219.0
134.00	1458.0
135.00	3041.0
136.00	1480.0
137.00	1909.0
138.10	359.0
138.90	232.0
140.00	790.0
141.00	5807.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

141.90	1786.0
143.00	1314.0
143.90	522.0
145.10	584.0
146.00	806.0
147.00	2836.0
148.00	6467.0
149.00	1290.0
149.90	411.0
151.00	860.0
151.90	486.0
153.00	1521.0
154.00	1537.0
155.00	3690.0
156.00	4667.0
156.90	1076.0
157.90	1152.0
159.00	1013.0
160.00	1563.0
161.00	2976.0
162.00	786.0
163.10	365.0
164.00	296.0
165.00	1787.0
166.00	1898.0
167.00	11396.0
168.00	5752.0
169.00	1190.0
169.90	474.0
170.90	611.0
171.90	1307.0
173.00	1372.0
174.00	2755.0
175.00	4572.0
176.00	1290.0
177.00	2243.0
177.90	1063.0
178.90	9157.0
180.00	7039.0
181.00	3078.0
182.00	407.0
183.00	177.0
184.00	540.0
185.00	4331.0
186.00	34104.0
187.00	9293.0
188.00	1175.0
189.00	2359.0
190.00	534.0
191.00	1138.0
192.00	2497.0
193.00	3577.0
194.00	911.0
195.10	765.0
196.00	6811.0
197.90	240128.0
198.90	17240.0
200.10	1551.0
201.40	1077.0
203.00	2081.0
204.00	9254.0
205.00	15933.0
206.00	66296.0
207.00	9262.0
208.00	2069.0
209.00	824.0
210.90	2689.0
212.10	361.0
213.00	434.0
215.00	678.0
215.90	1286.0
217.00	17192.0
217.90	2399.0
219.00	230.0
220.00	512.0
221.00	10351.0
221.90	2200.0
223.00	4342.0
224.00	37040.0
225.00	9422.0
226.00	1035.0
227.00	16352.0
228.00	2875.0
229.00	2920.0
229.80	354.0
231.00	1238.0
232.10	255.0
233.00	292.0
233.90	660.0
235.00	1328.0
236.00	589.0
237.00	1763.0
237.80	169.0
238.90	848.0
239.90	493.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

241.00	892.0
242.00	2116.0
243.00	2550.0
244.00	28912.0
245.00	3597.0
246.00	5971.0
247.00	1065.0
248.00	415.0
249.00	964.0
249.90	213.0
250.90	272.0
252.00	378.0
253.10	833.0
255.00	147456.0
256.00	21416.0
257.00	1920.0
258.00	10100.0
258.90	1241.0
259.90	348.0
261.00	288.0
264.00	221.0
265.00	3442.0
266.00	383.0
270.00	325.0
271.00	456.0
272.10	531.0
273.00	4931.0
274.00	12275.0
275.00	65648.0
276.00	9057.0
277.00	5528.0
278.00	923.0
281.90	198.0
283.00	715.0
284.00	454.0
285.00	826.0
289.00	181.0
290.00	192.0
292.00	252.0
293.00	1486.0
293.90	339.0
296.00	19064.0
297.00	3178.0
301.00	220.0
302.00	456.0
303.10	2286.0
304.00	917.0
308.00	427.0
308.80	276.0
309.90	390.0
313.00	183.0
314.00	775.0
315.00	1525.0
316.00	951.0
317.00	228.0
320.00	287.0
321.00	501.0
321.90	340.0
323.00	5384.0
324.00	1096.0
326.00	204.0
327.00	1254.0
328.00	465.0
332.00	489.0
333.00	619.0
334.10	3565.0
335.00	891.0
341.10	754.0
342.00	352.0
346.00	1014.0
346.90	335.0
351.00	207.0
352.00	1699.0
353.10	1226.0
354.00	2008.0
354.90	412.0
365.00	8556.0
366.00	1130.0
367.00	206.0
370.20	185.0
371.00	388.0
372.00	3206.0
373.00	923.0
382.90	850.0
390.00	459.0
392.10	204.0
402.00	994.0
403.00	1691.0
404.00	633.0
404.90	200.0
421.00	1334.0
422.00	1501.0
423.00	10050.0
424.10	2079.0
425.00	192.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

438.70	169.0
441.10	29488.0
442.00	193024.0
443.00	38696.0
444.10	3576.0
444.90	217.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

m/z	Abundance
37.00	284.0
38.00	1108.0
39.00	5618.0
40.00	288.0
41.00	698.0
43.00	465.0
44.00	715.0
44.90	182.0
49.00	620.0
50.00	19392.0
51.00	74184.0
52.00	3795.0
53.10	183.0
55.00	1122.0
56.00	2701.0
57.00	5824.0
58.00	311.0
60.90	1024.0
62.00	1392.0
63.00	3293.0
64.00	608.0
65.00	1446.0
65.90	259.0
67.10	467.0
68.00	1821.0
69.00	70992.0
70.00	578.0
71.10	248.0
73.00	576.0
74.00	7193.0
75.00	10983.0
76.10	4161.0
77.00	86008.0
78.00	4845.0
79.00	5313.0
80.00	4616.0
81.00	6181.0
82.00	1537.0
83.00	1850.0
83.90	209.0
85.00	1453.0
86.00	1512.0
87.00	717.0
87.90	353.0
89.00	263.0
91.00	1914.0
92.00	1500.0
93.00	9108.0
94.10	822.0
95.00	591.0
96.00	537.0
97.20	673.0
98.00	7364.0
99.00	5465.0
100.00	309.0
101.00	3202.0
101.90	297.0
103.00	1165.0
104.00	2138.0
105.00	2326.0
106.00	968.0
107.00	27160.0
108.00	4283.0
110.00	46112.0
111.00	7894.0
111.90	924.0
113.00	348.0
115.00	225.0
116.00	1363.0
117.00	20960.0
118.00	1452.0
119.00	466.0
120.00	182.0
121.00	302.0
122.00	2104.0
123.00	2898.0
124.00	1350.0
125.00	1197.0
127.00	92752.0
128.00	6894.0
129.00	36128.0
130.00	3206.0
131.00	873.0
132.20	324.0
132.70	153.0
133.20	387.0
133.90	1140.0
135.00	3585.0
136.00	1102.0
137.00	1798.0
137.90	465.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

139.10	263.0
140.10	856.0
141.00	5423.0
142.00	1629.0
143.00	1408.0
144.00	410.0
145.00	544.0
146.00	1295.0
147.00	3060.0
148.00	6052.0
149.00	1402.0
150.00	431.0
151.00	618.0
152.00	546.0
153.00	2205.0
153.90	1297.0
155.00	3386.0
156.00	4793.0
157.00	1122.0
158.00	1514.0
159.00	886.0
160.00	1997.0
160.90	2198.0
162.00	943.0
163.00	255.0
164.00	535.0
165.00	2271.0
166.00	2093.0
167.00	10706.0
168.00	5012.0
169.00	954.0
169.90	259.0
171.00	317.0
172.00	1221.0
173.00	1485.0
174.00	2469.0
175.00	4123.0
176.00	1012.0
177.00	2123.0
177.90	755.0
179.00	8953.0
180.00	6227.0
181.00	2480.0
182.00	417.0
182.90	289.0
184.00	1184.0
185.00	4792.0
186.00	32304.0
187.00	9961.0
188.00	939.0
189.00	2291.0
189.90	429.0
190.90	1353.0
192.00	2968.0
193.00	3090.0
194.10	832.0
195.20	626.0
196.00	6864.0
197.90	236288.0
198.90	15261.0
200.00	1361.0
201.40	983.0
203.00	1891.0
204.00	8776.0
205.00	16408.0
206.00	65720.0
207.00	8049.0
208.00	2286.0
209.00	637.0
211.00	2667.0
211.60	653.0
213.00	604.0
214.10	213.0
215.00	867.0
216.00	1320.0
217.00	17528.0
218.00	2436.0
219.10	333.0
221.00	10326.0
221.90	2569.0
223.00	4662.0
224.00	38944.0
225.00	9870.0
226.10	1475.0
227.00	15960.0
228.00	2491.0
229.00	3771.0
230.00	615.0
231.00	1504.0
233.00	489.0
234.00	1137.0
235.00	1109.0
236.10	560.0
237.00	1149.0
237.80	261.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

239.00	707.0
239.90	445.0
241.00	966.0
242.00	2523.0
243.00	2421.0
244.00	29584.0
245.00	4310.0
246.00	5662.0
246.90	1207.0
247.80	271.0
249.00	1293.0
249.90	247.0
251.00	342.0
251.90	396.0
253.00	1003.0
255.00	157312.0
256.00	24104.0
257.00	2181.0
258.00	9470.0
259.00	1314.0
259.90	278.0
261.00	203.0
262.00	155.0
263.10	184.0
263.80	257.0
265.00	3591.0
266.00	507.0
268.00	159.0
270.00	180.0
271.10	681.0
272.10	465.0
273.00	5610.0
274.00	13485.0
275.00	73840.0
276.00	10401.0
277.00	6228.0
278.00	878.0
278.90	274.0
281.10	202.0
282.00	228.0
283.00	734.0
283.90	364.0
285.00	1138.0
286.00	203.0
289.00	390.0
290.10	198.0
292.00	346.0
293.00	1477.0
294.00	583.0
294.90	704.0
296.00	20816.0
296.90	2816.0
298.00	217.0
301.00	178.0
301.90	288.0
303.00	2553.0
304.00	732.0
308.00	413.0
309.00	191.0
310.00	422.0
313.20	271.0
314.00	911.0
314.90	2618.0
316.00	1145.0
316.90	207.0
318.10	171.0
321.00	845.0
322.00	393.0
323.00	7399.0
324.10	1178.0
325.10	237.0
326.10	270.0
327.00	1051.0
328.00	777.0
332.00	671.0
333.00	710.0
334.00	4880.0
335.00	1053.0
335.90	223.0
341.00	503.0
342.20	198.0
346.00	1329.0
346.90	381.0
352.00	2387.0
353.00	1244.0
354.00	2364.0
355.10	320.0
359.00	233.0
365.00	10117.0
366.00	1283.0
370.10	248.0
371.00	506.0
372.00	3062.0
373.00	888.0
383.00	779.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

384.00	352.0
390.10	352.0
391.00	215.0
391.90	235.0
400.90	198.0
402.00	1291.0
403.00	1952.0
404.00	1091.0
421.00	1573.0
422.10	1945.0
423.00	11563.0
424.00	2926.0
425.10	313.0
441.00	36584.0
442.00	245440.0
443.00	48040.0
444.00	4695.0
445.00	361.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61

m/z	Abundance
39.00	226.0
41.00	599.0
42.00	182.0
43.00	523.0
44.00	539.0
49.00	220.0
54.10	152.0
55.00	701.0
56.10	311.0
57.00	578.0
67.00	240.0
68.10	255.0
69.00	573.0
70.10	216.0
71.10	325.0
76.90	175.0
79.00	163.0
81.00	540.0
82.00	152.0
83.10	625.0
84.10	269.0
85.10	200.0
91.00	203.0
93.00	161.0
95.00	536.0
96.10	197.0
97.00	402.0
105.00	227.0
109.10	391.0
111.00	446.0
123.10	215.0
127.00	154.0
128.00	160.0
133.00	285.0
134.00	154.0
135.10	170.0
137.10	246.0
143.00	189.0
149.10	210.0
151.10	151.0
159.10	249.0
167.10	168.0
197.10	154.0
199.00	217.0
207.00	418.0

Instrument :
BNA_G
ClientSampleId :
DFTPP61