

Data Path : Z:\HPCHEM1\BNA G\DATA\BG102615\
 Data File : BG019363.D
 Acq On : 26 Oct 2015 17:14
 Operator : UM/NP
 Sample : G4075-02MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 JG9H8MS

Quant Time: Oct 27 03:10:10 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102315.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 27 04:07:50 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.25	152	34070	20.00	ng/ul	0.00
18) Naphthalene-d8	11.08	136	143203	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.87	164	108077	20.00	ng/ul	0.01
62) Phenanthrene-d10	17.61	188	249335	20.00	ng/ul	0.01
78) Chrysene-d12	21.90	240	302292	20.00	ng/ul	0.01
86) Perylene-d12	25.31	264	303666	20.00	ng/ul	0.03

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.56	96	3106	5.06	ng/uL	0.00
5) Phenol-d5	7.40	99	104500	34.86	ng/ul	0.03
7) Bis-(2-Chloroethyl)ether-d	7.56	67	61792	34.52	ng/ul	0.00
9) 2-Chlorophenol-d4	7.78	132	67782	35.03	ng/ul	0.01
13) 4-Methylphenol-d8	8.96	113	86309	36.93	ng/ul	0.02
19) Nitrobenzene-d5	9.42	128	35840	32.96	ng/ul	0.01
22) 2-Nitrophenol-d4	10.14	143	44648	35.94	ng/ul	0.01
26) 2,4-Dichlorophenol-d3	10.70	165	95072	35.09	ng/ul	0.02
29) 4-Chloroaniline-d4	11.21	131	94326	35.05	ng/ul	0.01
44) Dimethylphthalate-d6	14.27	166	302045	33.71	ng/ul	0.01
47) Acenaphthylene-d8	14.57	160	346975	34.99	ng/ul	0.01
52) 4-Nitrophenol-d4	15.06	143	43406	29.79	ng/ul	0.03
58) Fluorene-d10	15.86	176	265469	31.58	ng/ul	0.01
63) 4,6-Dinitro-2-methylphenol	15.96	200	33003	20.11	ng/ul	0.00
71) Anthracene-d10	17.71	188	390060	34.16	ng/ul	0.01
79) Pyrene-d10	19.98	212	429837	32.45	ng/ul	0.01
90) Benzo(a)pyrene-d12	25.08	264	506350	32.69	ng/ul	0.04

Target Compounds


						Ovalue
6) Phenol	7.43	94	105329	33.43	ng/ul	95
10) 2-Chlorophenol	7.81	128	63671	32.36	ng/ul	99
15) N-Nitroso-di-n-propylamine	9.05	70	78226	32.41	ng/ul#	89
33) 4-Chloro-3-methylphenol	12.34	107	111940	32.94	ng/ul#	97
45) Dimethylphthalate	14.32	163	173100	19.34	ng/ul	99
50) Acenaphthene	14.94	153	181825	27.04	ng/ul	97
53) 4-Nitrophenol	15.08	109	61952	25.78	ng/ul#	80
55) 2,4-Dinitrotoluene	15.22	165	81602	29.49	ng/ul#	82
69) Pentachlorophenol	17.26	266	61470	26.93	ng/ul	98
80) Pyrene	20.01	202	393017	23.19	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG102615\
Data File : BG019363.D
Acq On : 26 Oct 2015 17:14
Operator : UM/NP
Sample : G4075-02MS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
JG9H8MS

Quant Time: Oct 27 03:10:10 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG102315.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 27 04:07:50 2015
Response via : Initial Calibration





#6

Phenol

Concen: 33.43 ng/ul

RT: 7.43 min Scan# 781

Delta R.T. 0.03 min

Lab File: BG019363.D

Acq: 26 Oct 2015 17:14

Instrument :

BNA_G

ClientSampleId :

JG9H8MS





Tot Ion: 94 Resp: 105329

Ion Ratio Lower Upper

94 100

65 33.6 29.3 43.9

66 52.3 45.0 67.6



#10

2-Chlorophenol

Concen: 32.36 ng/ul

RT: 7.81 min Scan# 846

Delta R.T. 0.02 min

Lab File: BG019363.D

Acq: 26 Oct 2015 17:14




Tgt Ion:128 Resp: 63671




Ion Ratio Lower Upper

128 100

64 66.1 53.2 79.8

130 34.3 25.9 38.9





#15

N-Nitroso-di-n-propylamine

Concen: 32.41 ng/ul

RT: 9.05 min Scan# 1057

Delta R.T. 0.01 min

Lab File: BG019363.D

Acq: 26 Oct 2015 17:14

Instrument :

BNA_G

ClientSampleId :

JG9H8MS

Tot Ion: 70 Resp: 78226





Ion Ratio Lower Upper

70 100

42 36.5 36.6 55.0#

101 7.0 5.8 8.8

130 15.7 14.6 22.0



#33

4-Chloro-3-methylphenol

Concen: 32.94 ng/ul

RT: 12.34 min Scan# 1617

Delta R.T. 0.02 min

Lab File: BG019363.D

Acq: 26 Oct 2015 17:14


Tgt Ion:107 Resp: 111940


Ion Ratio Lower Upper

107 100

144 21.5 13.0 19.6#

142 58.3 47.0 70.6





#45

Dimethylphthalate

Concen: 19.34 ng/ul

RT: 14.32 min Scan# 1953

Delta R.T. 0.01 min

Lab File: BG019363.D

Acq: 26 Oct 2015 17:14

Instrument :

BNA_G

ClientSampleId :

JG9H8MS


Tot Ion:163 Resp: 173100

Ion Ratio Lower Upper

163 100

194 4.8 4.3 6.5




164 9.8 7.5 11.3



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

Acenaphthene

Concen: 27.04 ng/ul

RT: 14.94 min Scan# 2059

Delta R.T. 0.01 min

Lab File: BG019363.D

Acq: 26 Oct 2015 17:14


Tgt Ion:153 Resp: 181825

Ion Ratio Lower Upper

153 100

152 46.3 38.8 58.2



154 86.6 71.5 107.3




Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E









#53
4-Nitrophenol
Concen: 25.78 ng/ul
RT: 15.08 min Scan# 2083
Delta R.T. 0.03 min
Lab File: BG019363.D
Acq: 26 Oct 2015 17:14




Instrument :
BNA_G
ClientSampleId :
JG9H8MS


Tot Ion	Ratio	Lower	Upper
109	100		
139	75.9	41.5	62.3#
65	104.0	73.7	110.5



#55
2,4-Dinitrotoluene
Concen: 29.49 ng/ul
RT: 15.22 min Scan# 2107
Delta R.T. 0.01 min
Lab File: BG019363.D
Acq: 26 Oct 2015 17:14

Tgt Ion	Ratio	Lower	Upper
165	100		
63	49.0	51.1	76.7#
182	2.4	3.3	4.9#









#69
 Pentachlorophenol
 Concen: 26.93 ng/ul
 RT: 17.26 min Scan# 2454
 Delta R.T. 0.01 min
 Lab File: BG019363.D
 Acq: 26 Oct 2015 17:14

Instrument :
 BNA_G
 ClientSampleId :
 JG9H8MS

Tot Ion	Ratio	Lower	Upper
266	100		
264	64.3	49.5	74.3
268	60.7	48.7	73.1



#80
 Pyrene
 Concen: 23.19 ng/ul
 RT: 20.01 min Scan# 2923
 Delta R.T. 0.01 min
 Lab File: BG019363.D
 Acq: 26 Oct 2015 17:14

Tgt Ion	Ratio	Lower	Upper
202	100		
101	5.8	4.3	6.5
100	5.2	3.6	5.4

