


Data Path : Z:\HPCHEM1\BNA G\DATA\BG102615\
 Data File : BG019355.D
 Acq On : 26 Oct 2015 10:02
 Operator : UM/NP
 Sample : DFTPP08
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP08

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG102315.M
 Title : SVOA CALIBRATION
 Last Update : Tue Oct 27 10:57:26 2015



AutoFind: Scans 2582, 2583, 2584; Background Corrected with Scan 2576

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.4	12877	PASS
68	69	0.00	2	1.8	326	PASS
69	198	0.00	100	34.6	18322	PASS
70	69	0.00	2	0.5	97	PASS
127	198	10	80	34.8	18394	PASS
197	198	0.00	2	0.2	128	PASS
198	198	100	100	100.0	52880	PASS
199	198	5	9	6.5	3435	PASS
275	198	10	60	29.4	15550	PASS
365	198	1	100	4.6	2448	PASS
441	442	0.01	24	16.0	6153	PASS
442	198	50	100	72.5	38352	PASS
443	442	15	24	18.9	7234	PASS

m/z	Abundance
41.10	179.0
42.90	266.0
44.00	284.0
45.00	213.0
55.10	237.0
56.00	254.0
56.90	288.0
66.90	255.0
69.00	416.0
71.00	196.0
73.00	620.0
76.90	171.0
80.70	283.0
83.90	179.0
87.00	185.0
90.90	189.0
93.00	186.0
95.00	219.0
95.80	177.0
96.90	204.0
103.00	168.0
104.90	235.0
109.00	221.0
120.00	279.0
129.00	258.0
142.10	207.0
147.00	202.0
162.90	191.0
175.10	204.0
191.10	204.0
192.90	177.0
206.90	642.0
207.90	232.0
209.00	277.0
220.90	153.0
222.00	192.0
238.90	157.0
267.90	196.0
281.10	254.0
541.90	438.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
38.00	262.0
39.00	1160.0
39.90	350.0
41.00	365.0
43.00	508.0
43.90	821.0
48.80	247.0
50.00	4083.0
51.00	17912.0
52.00	1022.0
53.10	183.0
54.90	594.0
56.00	643.0
56.90	1838.0
59.80	256.0
61.00	173.0
61.90	411.0
63.00	883.0
65.00	592.0
68.90	23600.0
69.90	238.0
70.80	188.0
73.00	1148.0
74.00	2672.0
75.00	3836.0
76.00	1392.0
77.00	25000.0
77.90	2129.0
78.90	1663.0
79.90	1131.0
80.90	2516.0
82.00	532.0
82.90	869.0
85.00	574.0
86.00	776.0
86.90	364.0
87.90	213.0
91.10	608.0
91.90	584.0
93.00	3720.0
93.90	256.0
96.10	224.0
97.10	306.0
97.90	2515.0
99.00	1721.0
99.90	233.0
100.90	1297.0
102.90	226.0
103.90	1070.0
104.90	754.0
105.90	219.0
107.00	8594.0
107.90	1515.0
108.90	579.0
109.90	11273.0
111.00	1775.0
113.10	204.0
115.90	629.0
116.90	5453.0
117.90	557.0
121.80	437.0
122.90	948.0
124.00	520.0
124.90	493.0
127.00	21912.0
128.00	1815.0
129.00	10169.0
130.00	864.0
134.00	412.0
135.00	676.0
137.00	312.0
137.60	312.0
138.80	234.0
140.00	191.0
140.90	926.0
141.90	690.0
144.90	158.0
146.20	547.0
146.90	1116.0
147.90	1966.0
148.80	443.0
151.10	226.0
152.90	362.0
154.10	314.0
154.90	600.0
155.90	807.0
158.00	260.0
159.00	405.0
159.90	446.0
160.90	580.0
161.90	209.0
163.80	237.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

165.00	788.0
165.70	313.0
166.10	418.0
167.00	3301.0
167.90	1275.0
169.00	385.0
170.90	231.0
171.90	594.0
173.10	276.0
174.10	858.0
175.00	1074.0
176.00	537.0
176.90	919.0
177.90	261.0
179.00	2118.0
179.90	2131.0
180.90	994.0
182.00	213.0
182.80	183.0
183.90	365.0
184.90	1533.0
186.00	10057.0
186.90	2529.0
187.90	445.0
188.90	803.0
191.00	289.0
192.00	882.0
192.90	1415.0
194.00	224.0
195.90	2246.0
196.70	385.0
197.90	58504.0
198.90	4324.0
199.80	312.0
203.00	457.0
204.00	3262.0
205.00	4415.0
206.00	18416.0
207.00	3381.0
207.80	690.0
208.90	356.0
209.90	628.0
210.90	640.0
211.70	179.0
214.90	288.0
216.00	272.0
216.90	5890.0
217.90	698.0
219.80	339.0
220.90	1981.0
221.80	668.0
222.90	1193.0
224.00	10988.0
225.00	3410.0
225.90	367.0
226.90	5594.0
227.90	607.0
228.90	953.0
230.00	409.0
231.00	316.0
232.90	229.0
234.80	324.0
235.30	186.0
235.80	315.0
236.80	411.0
238.80	431.0
239.80	206.0
240.90	595.0
242.10	623.0
243.00	473.0
243.90	7255.0
245.00	966.0
245.90	2075.0
246.90	553.0
247.90	197.0
252.40	173.0
252.80	251.0
253.90	625.0
254.90	44512.0
255.90	6235.0
256.90	747.0
257.90	2403.0
258.90	534.0
259.90	226.0
265.00	1381.0
265.80	330.0
267.10	171.0
270.00	173.0
272.90	1137.0
273.90	3297.0
275.00	16496.0
275.90	1996.0
276.90	1645.0
277.90	231.0
280.80	286.0
284.90	591.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

293.00	356.0
294.70	178.0
295.90	5119.0
296.90	1343.0
301.90	165.0
302.90	724.0
309.80	241.0
314.00	272.0
315.00	468.0
316.00	415.0
316.70	245.0
317.00	183.0
321.90	183.0
322.90	1757.0
324.00	359.0
326.90	203.0
331.90	242.0
334.00	1413.0
334.90	233.0
346.00	275.0
352.00	480.0
352.90	192.0
354.00	411.0
354.90	185.0
364.90	2191.0
365.90	301.0
371.10	209.0
372.10	697.0
382.90	404.0
402.90	411.0
403.90	311.0
420.90	249.0
421.90	339.0
423.00	1493.0
423.90	624.0
441.00	5809.0
442.00	35192.0
443.00	7366.0
443.90	525.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
38.00	299.0
39.10	858.0
39.90	219.0
41.00	168.0
43.00	367.0
44.00	497.0
45.00	244.0
48.90	175.0
50.00	4322.0
51.00	12913.0
52.00	939.0
53.00	175.0
55.00	250.0
55.90	808.0
57.00	1007.0
57.90	185.0
60.00	491.0
61.00	364.0
61.80	263.0
63.00	984.0
65.00	546.0
66.00	198.0
67.10	350.0
68.00	671.0
68.90	20832.0
70.10	291.0
71.00	317.0
73.00	968.0
74.00	2778.0
74.90	3871.0
76.00	1038.0
77.00	24456.0
78.00	1273.0
78.90	1884.0
79.90	1498.0
80.90	1578.0
82.00	503.0
83.00	416.0
84.30	232.0
85.00	359.0
85.80	962.0
87.00	405.0
90.90	675.0
91.90	663.0
93.00	3155.0
94.00	283.0
95.00	154.0
97.00	171.0
97.90	2785.0
99.00	1430.0
100.00	175.0
100.90	1000.0
101.90	320.0
102.90	477.0
103.90	950.0
104.90	723.0
107.00	8733.0
108.00	896.0
108.80	287.0
109.90	11957.0
111.00	2618.0
112.90	214.0
115.50	235.0
116.00	752.0
117.00	6123.0
117.90	826.0
118.90	219.0
119.80	265.0
121.90	447.0
122.90	932.0
124.10	527.0
124.40	433.0
125.00	289.0
126.90	21352.0
128.00	1612.0
128.90	10585.0
129.90	980.0
130.90	263.0
134.00	527.0
134.90	692.0
136.00	346.0
137.00	419.0
140.00	194.0
141.00	1610.0
141.80	356.0
142.80	261.0
146.10	215.0
146.90	1089.0
147.90	1598.0
148.90	587.0
150.90	204.0
151.20	208.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

151.80	304.0
152.90	192.0
154.00	361.0
155.00	1146.0
155.90	1254.0
156.90	351.0
157.60	280.0
158.10	433.0
158.90	457.0
159.80	522.0
161.00	822.0
161.90	170.0
163.80	196.0
164.90	636.0
165.80	609.0
166.90	2800.0
168.00	1338.0
169.00	497.0
169.90	152.0
170.90	180.0
172.00	558.0
173.00	578.0
174.00	1155.0
174.90	1098.0
175.90	236.0
176.80	744.0
178.10	319.0
178.90	2561.0
179.90	1512.0
180.90	904.0
182.20	204.0
182.90	280.0
183.80	272.0
185.00	1441.0
185.90	9096.0
186.90	3581.0
188.10	327.0
188.80	758.0
190.90	628.0
191.90	936.0
193.00	1008.0
194.00	401.0
195.10	236.0
196.00	1275.0
197.90	65872.0
198.90	3652.0
200.00	274.0
200.90	264.0
202.90	487.0
204.00	2357.0
205.00	5674.0
206.00	18080.0
206.90	3179.0
207.90	960.0
208.80	320.0
210.00	576.0
210.90	561.0
211.50	221.0
211.70	215.0
212.00	189.0
214.70	204.0
215.90	410.0
216.90	5828.0
217.90	736.0
219.20	159.0
220.90	2032.0
221.90	1137.0
222.90	1479.0
224.00	12843.0
225.00	3248.0
225.70	260.0
226.90	6001.0
227.90	672.0
228.90	1746.0
229.90	207.0
230.90	446.0
233.90	403.0
235.00	291.0
235.80	200.0
236.90	387.0
239.00	263.0
239.90	210.0
240.90	272.0
241.90	480.0
242.90	632.0
243.90	8163.0
244.90	937.0
245.90	1902.0
246.90	459.0
248.90	336.0
252.10	174.0
252.90	318.0
254.90	48472.0
255.90	6481.0
257.00	582.0
258.00	3891.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

258.90	439.0
260.90	242.0
265.00	1340.0
265.90	213.0
271.80	195.0
273.00	1132.0
273.90	3208.0
275.00	18640.0
275.90	2463.0
276.90	2004.0
277.90	430.0
281.10	335.0
282.00	224.0
282.90	488.0
283.90	158.0
284.90	433.0
286.00	191.0
290.90	174.0
292.80	457.0
295.00	199.0
296.00	6876.0
297.00	860.0
303.00	573.0
303.90	155.0
313.80	272.0
314.90	934.0
316.10	371.0
316.90	195.0
321.10	187.0
323.00	2065.0
323.80	233.0
326.90	571.0
328.70	159.0
331.90	157.0
333.20	203.0
334.00	1191.0
334.90	365.0
340.70	222.0
345.90	371.0
350.90	178.0
352.00	606.0
353.90	577.0
354.90	259.0
365.00	2951.0
365.80	279.0
368.10	205.0
372.00	1110.0
382.90	279.0
402.00	357.0
403.00	439.0
403.90	205.0
410.80	152.0
420.90	324.0
421.90	467.0
423.00	2361.0
423.90	533.0
438.70	185.0
441.00	6293.0
442.00	44288.0
443.00	8827.0
444.00	575.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
39.00	692.0
40.00	281.0
40.90	296.0
42.90	293.0
44.00	522.0
50.00	2376.0
51.00	7807.0
52.00	362.0
55.00	769.0
55.90	252.0
56.90	920.0
59.90	309.0
60.90	156.0
61.80	239.0
63.00	374.0
65.00	188.0
67.00	464.0
67.90	308.0
68.90	11783.0
73.10	760.0
74.00	1428.0
75.00	2378.0
76.00	460.0
77.00	12412.0
78.00	776.0
78.90	645.0
79.90	569.0
81.00	1201.0
81.90	301.0
82.90	383.0
85.10	266.0
85.80	276.0
86.70	179.0
87.10	190.0
91.10	377.0
91.90	353.0
92.90	1795.0
93.90	158.0
95.00	183.0
95.80	194.0
97.00	300.0
97.90	1083.0
99.00	1084.0
100.90	525.0
103.90	158.0
104.90	351.0
106.00	432.0
106.90	4097.0
107.90	715.0
108.90	402.0
109.90	6008.0
111.00	1233.0
116.00	540.0
116.90	3653.0
117.80	352.0
120.20	182.0
122.10	525.0
122.90	561.0
124.00	513.0
124.60	152.0
127.00	11918.0
128.10	668.0
129.00	5996.0
129.90	849.0
133.00	285.0
134.00	437.0
134.90	384.0
137.00	336.0
140.90	939.0
141.80	200.0
142.90	206.0
147.00	510.0
147.90	1178.0
148.80	163.0
152.90	493.0
153.90	289.0
154.90	707.0
156.00	1012.0
156.70	191.0
158.90	442.0
159.90	394.0
160.90	519.0
161.90	286.0
164.90	611.0
165.90	292.0
166.90	1476.0
168.00	952.0
168.70	245.0
173.80	225.0
174.90	766.0
176.00	225.0
176.70	193.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

178.90	1272.0
179.90	586.0
181.00	979.0
181.80	166.0
182.40	152.0
183.10	230.0
185.00	1021.0
186.00	5665.0
186.90	1666.0
188.90	391.0
190.10	157.0
190.90	548.0
191.90	381.0
193.00	591.0
194.90	199.0
195.90	869.0
196.90	356.0
197.90	34264.0
198.90	2331.0
200.00	162.0
200.90	192.0
201.40	336.0
203.10	190.0
203.90	1790.0
205.00	2969.0
206.00	12228.0
206.90	2621.0
208.00	733.0
209.00	390.0
210.30	207.0
210.90	390.0
214.80	204.0
216.00	153.0
216.90	3205.0
218.00	693.0
220.90	1324.0
221.90	459.0
222.90	832.0
224.00	7200.0
225.00	1886.0
226.00	394.0
226.90	3389.0
227.90	733.0
228.90	667.0
231.00	541.0
233.10	165.0
233.70	380.0
234.90	333.0
235.90	155.0
236.80	366.0
238.80	163.0
240.90	259.0
241.90	490.0
243.00	715.0
243.90	5411.0
244.90	865.0
245.90	1209.0
246.80	208.0
248.90	298.0
250.00	194.0
251.80	198.0
253.90	579.0
254.90	28984.0
255.90	4555.0
257.00	380.0
257.90	1489.0
258.90	414.0
265.00	940.0
265.90	251.0
271.00	241.0
272.90	1308.0
274.00	1670.0
274.90	11515.0
275.90	1363.0
277.00	1554.0
277.80	171.0
278.90	308.0
281.00	307.0
282.20	205.0
283.10	209.0
285.00	161.0
286.00	187.0
292.80	253.0
295.00	229.0
295.90	3833.0
296.90	698.0
302.90	298.0
304.10	223.0
313.90	362.0
314.90	479.0
315.90	327.0
320.80	275.0
323.00	1314.0
324.00	389.0
326.90	360.0
333.90	1078.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

335.00	296.0
341.90	154.0
345.80	211.0
351.90	352.0
353.00	179.0
354.00	516.0
364.90	2203.0
365.90	217.0
372.00	815.0
382.90	188.0
400.90	216.0
401.90	231.0
421.00	190.0
422.20	210.0
423.00	2203.0
423.90	264.0
426.50	272.0
441.00	6359.0
442.00	35576.0
442.90	5511.0
443.90	371.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08