

Data Path : Z:\HPCHEM1\BNA G\DATA\BG102616\
 Data File : BG024512.D
 Acq On : 26 Oct 2016 13:32
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 26 17:39:26 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 26 11:14:13 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	153#	0.00
2	1,4-Dioxane	0.499	0.508	-1.8	162#	0.00
3	Pyridine	1.493	1.460	2.2	155#	0.00
4	n-Nitrosodimethylamine	0.773	0.771	0.3	155#	0.00
5 S	2-Fluorophenol	1.220	1.162	4.8	153#	0.00
6	Aniline	2.121	2.037	4.0	150	0.00
7 S	Phenol-d6	1.674	1.629	2.7	153#	0.00
8	2-Chlorophenol	1.241	1.196	3.6	156#	0.00
9	Benzaldehyde	1.034	0.982	5.0	150	0.00
10 C	Phenol	1.725	1.681	2.6	154#	0.00
11	bis(2-Chloroethyl)ether	1.296	1.264	2.5	158#	0.00
12	1,3-Dichlorobenzene	1.536	1.452	5.5	154#	0.00
13 C	1,4-Dichlorobenzene	1.517	1.497	1.3	158#	0.00
14	1,2-Dichlorobenzene	1.459	1.418	2.8	156#	0.00
15	Benzyl Alcohol	1.557	1.522	2.2	154#	0.00
16	2,2'-oxybis(1-Chloropropane	2.405	2.299	4.4	152#	0.00
17	2-Methylphenol	1.143	1.089	4.7	152#	0.00
18	Hexachloroethane	0.583	0.570	2.2	162#	0.00
19 P	n-Nitroso-di-n-propylamine	1.234	1.195	3.2	149	0.00
20	3+4-Methylphenols	1.535	1.522	0.8	153#	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	139	0.00
22	Acetophenone	0.514	0.520	-1.2	148	0.00
23 S	Nitrobenzene-d5	0.439	0.456	-3.9	154#	0.00
24	Nitrobenzene	0.489	0.516	-5.5	154#	0.00
25	Isophorone	0.817	0.811	0.7	143	0.00
26 C	2-Nitrophenol	0.181	0.194	-7.2	157#	0.00
27	2,4-Dimethylphenol	0.318	0.346	-8.8	156#	0.00
28	bis(2-Chloroethoxy)methane	0.423	0.431	-1.9	153#	0.00
29 C	2,4-Dichlorophenol	0.333	0.344	-3.3	149	0.00
30	1,2,4-Trichlorobenzene	0.395	0.408	-3.3	153#	0.00
31	Naphthalene	0.998	1.008	-1.0	148	0.00
32	Benzoic acid	0.264	0.291	-10.2	161#	0.03
33	4-Chloroaniline	0.433	0.438	-1.2	141	0.00
34 C	Hexachlorobutadiene	0.309	0.319	-3.2	157#	0.00
35	Caprolactam	0.122	0.114	6.6	134	0.01
36 C	4-Chloro-3-methylphenol	0.402	0.421	-4.7	148	0.00
37	2-Methylnaphthalene	0.728	0.752	-3.3	152#	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	134	0.00
39	1,2,4,5-Tetrachlorobenzene	0.675	0.691	-2.4	151#	0.00
40 P	Hexachlorocyclopentadiene	0.380	0.417	-9.7	156#	0.00
41 S	2,4,6-Tribromophenol	0.299	0.312	-4.3	140	0.00
42 C	2,4,6-Trichlorophenol	0.405	0.433	-6.9	154#	0.00
43	2,4,5-Trichlorophenol	0.438	0.446	-1.8	144	0.00
44 S	2-Fluorobiphenyl	1.203	1.197	0.5	143	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG102616\
 Data File : BG024512.D
 Acq On : 26 Oct 2016 13:32
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 26 17:39:26 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 26 11:14:13 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.388	1.386	0.1	144	0.00
46	2-Chloronaphthalene	1.057	1.065	-0.8	147	0.00
47	2-Nitroaniline	0.433	0.453	-4.6	138	0.00
48	Acenaphthylene	1.621	1.614	0.4	139	0.00
49	Dimethylphthalate	1.420	1.422	-0.1	138	0.00
50	2,6-Dinitrotoluene	0.303	0.317	-4.6	139	0.00
51 C	Acenaphthene	1.208	1.135	6.0	133	0.00
52	3-Nitroaniline	0.314	0.322	-2.5	139	0.00
53 P	2,4-Dinitrophenol	0.220	0.231	-5.0	138	0.00
54	Dibenzofuran	1.670	1.656	0.8	139	0.00
55 P	4-Nitrophenol	0.273	0.276	-1.1	138	0.00
56	2,4-Dinitrotoluene	0.440	0.465	-5.7	137	0.00
57	Fluorene	1.385	1.379	0.4	137	0.00
58	2,3,4,6-Tetrachlorophenol	0.454	0.473	-4.2	141	0.00
59	Diethylphthalate	1.489	1.452	2.5	134	0.00
60	4-Chlorophenyl-phenylether	0.777	0.781	-0.5	142	0.00
61	4-Nitroaniline	0.343	0.357	-4.1	141	0.00
62	Azobenzene	1.485	1.474	0.7	137	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.00
64	4,6-Dinitro-2-methylphenol	0.138	0.156	-13.0	146	0.00
65 c	n-Nitrosodiphenylamine	0.547	0.560	-2.4	137	0.00
66	4-Bromophenyl-phenylether	0.243	0.256	-5.3	141	0.00
67	Hexachlorobenzene	0.268	0.280	-4.5	137	0.00
68	Atrazine	0.235	0.229	2.6	127	0.00
69 C	Pentachlorophenol	0.177	0.196	-10.7	140	0.00
70	Phenanthrene	1.019	1.011	0.8	133	0.00
71	Anthracene	1.028	1.031	-0.3	133	0.00
72	Carbazole	0.938	0.942	-0.4	135	0.00
73	Di-n-butylphthalate	1.081	1.073	0.7	131	0.00
74 C	Fluoranthene	1.289	1.292	-0.2	131	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	129	0.00
76	Benzidine	0.471	0.452	4.0	130	0.00
77	Pyrene	0.978	0.950	2.9	129	0.00
78 S	Terphenyl-d14	0.669	0.606	9.4	124	0.00
79	Butylbenzylphthalate	0.408	0.409	-0.2	136	0.00
80	Benzo(a)anthracene	1.099	1.055	4.0	134	0.00
81	3,3'-Dichlorobenzidine	0.443	0.435	1.8	134	0.00
82	Chrysene	1.046	1.022	2.3	134	0.00
83	Bis(2-ethylhexyl)phthalate	0.583	0.589	-1.0	138	0.00
84 c	Di-n-octyl phthalate	0.968	0.995	-2.8	139	0.00
85	Indeno(1,2,3-cd)pyrene	1.444	1.491	-3.3	147	0.00
86 I	Perylene-d12	1.000	1.000	0.0	134	0.00
87	Benzo(b)fluoranthene	1.081	1.051	2.8	141	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG102616\
Data File : BG024512.D
Acq On : 26 Oct 2016 13:32
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 26 17:39:26 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG102516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 26 11:14:13 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.044	1.025	1.8	141	0.00
89 C	Benzo(a)pyrene	1.056	1.053	0.3	143	0.00
90	Dibenzo(a,h)anthracene	1.159	1.137	1.9	147	0.00
91	Benzo(a,h,i)perylene	1.158	1.137	1.8	148	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0