


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG102620\
 Data File : BG047069.D
 Acq On : 26 Oct 2020 8:57
 Operator : CG/JU
 Sample : DFTPP46
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP46

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG102220MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Oct 26 10:36:50 2020



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.0	39010	PASS
68	69	0.00	2	0.6	257	PASS
69	198	0.00	100	35.9	39992	PASS
70	69	0.00	2	0.6	221	PASS
127	198	10	80	40.0	44570	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	111365	PASS
199	198	5	9	6.8	7556	PASS
275	198	10	60	28.1	31341	PASS
365	198	1	100	3.4	3826	PASS
441	198	0.01	100	11.3	12544	PASS
442	198	50	100	70.6	78602	PASS
443	442	15	24	19.1	15016	PASS

Scan 2466 (17.826 min): BG047069.D
DFTPP46

m/z	Abundance
40.00	282.0
42.90	163.0
44.00	580.0
193.00	250.0
207.00	513.0

Instrument :
BNA_G
ClientSampleId :
DFTPP46

m/z	Abundance
38.10	419.0
39.10	2811.0
39.90	195.0
43.10	166.0
44.10	680.0
49.00	240.0
50.10	9138.0
51.10	37160.0
52.10	1954.0
55.10	179.0
56.00	1106.0
57.10	2910.0
58.10	175.0
61.00	458.0
62.00	482.0
63.00	1385.0
64.10	274.0
65.10	517.0
68.00	773.0
69.00	35640.0
70.00	258.0
71.00	181.0
73.00	489.0
74.10	4418.0
75.10	5797.0
76.10	2454.0
77.10	34784.0
78.10	2547.0
79.10	2685.0
80.00	2403.0
81.10	3163.0
82.00	653.0
83.00	610.0
85.00	405.0
86.00	1359.0
87.10	565.0
91.10	717.0
92.00	883.0
93.00	5479.0
94.00	328.0
96.00	491.0
97.20	280.0
98.00	3240.0
99.00	2261.0
100.00	340.0
100.90	847.0
103.00	487.0
104.00	1216.0
105.00	884.0
105.90	430.0
107.00	12386.0
108.00	1832.0
109.10	585.0
110.00	18824.0
111.00	3495.0
111.90	155.0
116.10	818.0
117.00	10363.0
118.10	759.0
119.00	177.0
122.00	1012.0
123.00	1294.0
124.10	501.0
125.10	557.0
127.10	37944.0
128.00	3511.0
129.00	17104.0
130.00	1404.0
130.80	190.0
131.10	160.0
133.90	312.0
135.00	1372.0
136.10	459.0
137.10	665.0
138.20	426.0
138.90	165.0
140.00	189.0
141.00	2024.0
142.00	634.0
142.90	550.0
145.20	185.0
146.10	578.0
147.00	925.0
148.00	2337.0
149.00	462.0
151.10	469.0
153.00	872.0
154.00	609.0
155.00	1453.0
156.10	1570.0
157.00	318.0
158.10	518.0

Instrument :
BNA_G
ClientSampleId :
DFTPP46

159.10	327.0
160.00	545.0
161.10	995.0
161.90	310.0
163.20	198.0
164.00	213.0
165.00	535.0
166.20	747.0
167.00	4378.0
168.10	1746.0
169.00	292.0
170.10	303.0
171.20	213.0
171.90	540.0
173.10	906.0
174.00	913.0
175.00	1916.0
176.10	303.0
176.90	893.0
178.10	153.0
179.00	3544.0
180.00	2464.0
180.90	1088.0
182.00	207.0
182.80	435.0
183.90	334.0
185.00	1446.0
186.00	12814.0
187.00	3697.0
188.00	463.0
188.90	825.0
190.90	545.0
192.10	996.0
193.10	1308.0
193.90	297.0
196.10	1892.0
198.00	83024.0
199.00	6246.0
199.90	932.0
201.40	198.0
203.00	464.0
204.00	3261.0
205.10	5496.0
206.10	24960.0
207.00	4487.0
208.00	1290.0
209.00	395.0
211.00	925.0
216.10	673.0
217.00	7113.0
218.10	737.0
220.00	293.0
221.00	4228.0
222.00	889.0
222.90	1400.0
224.10	13168.0
225.00	4381.0
226.10	494.0
227.00	5016.0
228.00	701.0
229.10	1206.0
229.70	162.0
231.10	491.0
232.90	196.0
234.00	493.0
235.00	512.0
236.00	365.0
236.90	812.0
239.00	394.0
239.80	180.0
241.00	255.0
242.10	472.0
243.10	777.0
244.00	8729.0
245.10	1214.0
246.00	2323.0
246.90	183.0
247.90	574.0
248.90	326.0
250.80	220.0
253.10	370.0
255.00	49928.0
256.00	7643.0
257.00	646.0
258.00	2695.0
258.90	477.0
259.90	221.0
261.00	193.0
264.10	208.0
265.00	1491.0
265.80	154.0
272.10	218.0
273.10	1668.0
274.00	3640.0
275.00	21776.0
276.00	2785.0

Instrument :
BNA_G
ClientSampleId :
DFTPP46

277.10	1576.0
283.10	288.0
285.00	333.0
289.90	188.0
293.10	577.0
294.00	232.0
295.00	181.0
296.00	5486.0
297.00	1217.0
298.10	162.0
303.10	868.0
304.00	205.0
314.00	467.0
315.00	359.0
316.00	627.0
320.90	240.0
323.00	2166.0
323.90	421.0
326.90	497.0
332.00	221.0
332.80	204.0
334.00	1155.0
335.10	292.0
341.10	192.0
346.10	224.0
351.10	153.0
352.10	359.0
353.10	295.0
354.10	211.0
365.00	2394.0
366.10	473.0
372.20	627.0
383.10	258.0
390.00	231.0
401.90	262.0
403.10	714.0
403.80	188.0
422.10	461.0
423.00	3251.0
424.00	640.0
441.10	8606.0
442.10	43648.0
443.00	8922.0
444.10	837.0

Instrument :
BNA_G
ClientSampleId :
DFTPP46

m/z	Abundance
37.10	334.0
38.10	167.0
39.10	2954.0
40.20	275.0
44.10	824.0
49.10	452.0
50.10	11500.0
51.10	44992.0
52.10	2337.0
54.90	354.0
56.10	1661.0
57.00	2418.0
58.00	210.0
61.10	737.0
62.10	797.0
63.00	1622.0
64.00	382.0
65.20	1026.0
69.00	48328.0
70.00	231.0
73.00	483.0
74.10	5380.0
75.10	7781.0
76.10	2723.0
77.10	50880.0
78.10	3579.0
79.00	3346.0
80.10	2830.0
81.00	3963.0
82.00	986.0
83.00	1004.0
84.90	753.0
86.10	1041.0
87.10	740.0
91.00	842.0
92.10	937.0
93.10	7170.0
94.10	561.0
96.30	243.0
97.10	262.0
98.10	4747.0
99.10	3971.0
101.00	2483.0
103.10	895.0
104.00	1684.0
105.00	1482.0
107.00	16424.0
108.00	2858.0
109.10	947.0
110.00	27800.0
111.10	4529.0
112.10	746.0
112.90	219.0
116.10	1243.0
117.00	12957.0
117.90	1388.0
119.00	199.0
120.00	219.0
122.10	1073.0
123.10	1918.0
124.00	1029.0
125.10	633.0
127.00	53032.0
128.00	4903.0
129.00	22168.0
130.00	2473.0
130.90	586.0
131.70	177.0
133.10	172.0
134.00	630.0
135.00	1938.0
136.10	1124.0
137.00	1327.0
138.00	274.0
140.10	361.0
141.10	2767.0
142.10	1514.0
143.00	764.0
144.10	150.0
145.00	392.0
146.00	761.0
147.00	1511.0
148.00	3177.0
148.90	883.0
151.00	487.0
152.10	259.0
153.10	786.0
154.00	899.0
155.00	2014.0
156.10	3285.0
157.00	511.0
158.00	678.0

Instrument :
BNA_G
ClientSampleId :
DFTPP46

159.00	929.0
159.90	1339.0
161.00	1586.0
162.10	449.0
163.00	307.0
165.00	1584.0
166.00	1078.0
167.00	5885.0
168.00	2513.0
169.00	451.0
169.90	184.0
171.00	334.0
171.90	387.0
172.90	895.0
174.00	1078.0
175.10	2812.0
175.90	463.0
177.00	1417.0
177.80	192.0
179.00	5980.0
180.10	4029.0
181.00	1545.0
182.00	269.0
183.00	203.0
184.10	336.0
185.10	2511.0
186.10	17600.0
187.00	5255.0
188.00	759.0
189.00	1497.0
190.00	277.0
190.90	549.0
192.00	1630.0
193.10	1424.0
194.00	680.0
195.20	270.0
196.00	3076.0
198.00	129728.0
199.00	9095.0
200.10	825.0
201.60	765.0
202.90	1134.0
204.00	6352.0
205.00	9133.0
206.00	39520.0
207.00	5121.0
208.00	1265.0
209.00	710.0
209.90	628.0
211.00	1211.0
211.90	343.0
215.00	374.0
216.00	756.0
217.00	10325.0
218.00	1490.0
218.80	190.0
221.00	4545.0
221.90	1380.0
223.00	2432.0
224.00	22696.0
225.00	5324.0
226.00	994.0
227.00	8486.0
228.00	1978.0
229.10	1704.0
230.00	355.0
231.10	802.0
232.00	152.0
232.80	297.0
234.00	510.0
235.10	657.0
235.80	474.0
237.00	905.0
237.90	314.0
239.00	544.0
240.10	282.0
241.00	434.0
242.10	1251.0
242.90	601.0
244.10	14946.0
245.00	1815.0
246.00	2954.0
247.00	585.0
248.80	366.0
249.10	358.0
250.80	200.0
251.90	203.0
252.80	509.0
255.00	86968.0
256.00	12085.0
257.00	1379.0
258.00	5552.0
258.90	653.0
265.00	2365.0
266.00	171.0
272.00	331.0

Instrument :
BNA_G
ClientSampleId :
DFTPP46

273.00	2824.0
274.00	6913.0
275.00	35976.0
276.10	5741.0
277.00	2991.0
278.00	443.0
279.10	153.0
282.90	712.0
284.40	256.0
285.00	611.0
286.00	221.0
288.80	331.0
292.80	571.0
294.00	174.0
296.00	10088.0
297.00	1794.0
301.10	228.0
302.10	300.0
303.00	1267.0
304.00	392.0
309.20	182.0
312.10	210.0
313.10	235.0
313.90	425.0
315.00	1602.0
315.90	699.0
321.10	452.0
322.00	188.0
323.10	3961.0
323.90	472.0
327.00	621.0
332.90	536.0
334.00	1989.0
335.10	643.0
340.90	533.0
341.90	218.0
346.00	761.0
351.20	215.0
352.00	1030.0
353.10	565.0
354.10	762.0
359.10	250.0
365.00	4648.0
366.10	481.0
372.00	2016.0
372.90	803.0
378.30	167.0
382.90	303.0
383.90	191.0
390.00	151.0
390.80	170.0
402.00	583.0
403.00	751.0
404.10	425.0
421.00	585.0
421.90	569.0
423.10	4622.0
424.10	1041.0
441.10	13647.0
442.10	86696.0
443.10	17336.0
444.10	1682.0

Instrument :
BNA_G
ClientSampleId :
DFTPP46

m/z	Abundance
38.10	446.0
39.10	2478.0
40.10	251.0
44.00	565.0
45.00	151.0
49.10	495.0
50.10	10431.0
51.10	34880.0
52.10	1510.0
55.00	342.0
56.00	1318.0
57.00	2629.0
61.10	642.0
62.00	612.0
63.00	1376.0
64.10	218.0
65.10	592.0
69.00	36008.0
70.10	176.0
73.10	384.0
74.00	3837.0
75.00	6635.0
76.10	1965.0
77.10	40120.0
78.10	2626.0
79.10	2976.0
80.10	2922.0
81.00	3553.0
82.00	905.0
83.00	689.0
84.10	151.0
85.00	743.0
86.00	1110.0
87.10	586.0
88.00	155.0
91.10	425.0
92.00	970.0
93.00	5060.0
94.10	257.0
98.00	4578.0
99.00	3299.0
100.10	351.0
101.00	1789.0
103.20	640.0
104.00	1130.0
105.00	1097.0
107.10	13872.0
108.10	2303.0
108.80	400.0
110.00	20952.0
111.00	4152.0
112.10	312.0
116.10	670.0
117.00	11911.0
118.00	573.0
120.00	169.0
122.10	894.0
123.10	1535.0
124.10	811.0
125.00	673.0
127.00	42736.0
128.00	3562.0
129.00	20304.0
130.10	1582.0
131.00	266.0
132.10	207.0
134.10	259.0
135.00	1506.0
136.10	655.0
137.10	593.0
140.10	260.0
140.90	3205.0
142.00	1000.0
143.00	840.0
143.80	235.0
145.20	356.0
146.00	572.0
147.00	1450.0
148.00	2616.0
149.10	763.0
150.00	223.0
151.00	186.0
151.80	317.0
153.00	523.0
154.00	656.0
155.00	1344.0
156.10	2154.0
157.10	446.0
157.90	736.0
159.10	407.0
160.00	836.0
161.10	1611.0

Instrument :
BNA_G
ClientSampleId :
DFTPP46

162.10	493.0
165.10	1323.0
166.10	1006.0
167.10	6013.0
168.00	2119.0
169.00	295.0
169.20	294.0
170.10	242.0
170.90	290.0
171.90	453.0
172.30	370.0
173.00	612.0
174.10	1650.0
175.10	2152.0
175.90	697.0
177.00	831.0
178.10	666.0
179.00	4460.0
180.00	3838.0
181.00	1330.0
182.10	234.0
182.90	178.0
184.00	557.0
185.00	2228.0
186.10	15881.0
187.10	4815.0
188.10	549.0
189.00	941.0
190.20	182.0
191.00	780.0
192.00	839.0
193.00	1490.0
194.10	670.0
196.00	3280.0
198.00	121344.0
199.00	7329.0
200.00	894.0
201.40	390.0
201.90	282.0
202.10	276.0
203.00	1197.0
204.10	4312.0
205.00	8425.0
206.10	32848.0
207.00	5050.0
208.00	1707.0
209.10	377.0
211.00	995.0
213.10	170.0
215.10	442.0
216.20	1089.0
217.00	10241.0
218.00	1157.0
221.00	5069.0
221.90	1730.0
223.00	2070.0
224.10	19240.0
225.00	4452.0
226.10	561.0
227.00	8645.0
228.10	1095.0
229.00	1963.0
230.10	306.0
231.00	693.0
232.00	181.0
233.30	205.0
234.00	659.0
234.90	445.0
236.00	319.0
237.00	393.0
239.00	298.0
239.60	221.0
240.10	300.0
240.90	690.0
242.00	1463.0
243.00	1420.0
244.10	15526.0
245.00	2211.0
246.00	2789.0
247.00	780.0
247.90	155.0
249.10	593.0
249.90	185.0
253.00	291.0
253.40	252.0
255.00	77752.0
256.00	12006.0
257.00	844.0
258.00	5469.0
259.10	1019.0
260.10	228.0
260.90	173.0
264.10	204.0
265.00	2368.0
265.90	293.0
267.00	191.0

Instrument :
BNA_G
ClientSampleId :
DFTPP46

270.90	203.0
272.00	323.0
273.00	2640.0
274.00	6189.0
275.00	36272.0
276.00	4745.0
277.00	2608.0
277.90	437.0
281.90	216.0
282.90	395.0
283.90	158.0
285.10	523.0
292.00	186.0
293.00	546.0
294.00	272.0
294.80	264.0
296.00	11212.0
297.00	1498.0
300.80	151.0
303.00	1519.0
303.90	740.0
307.90	202.0
310.10	208.0
314.00	780.0
315.00	1008.0
316.00	608.0
321.00	445.0
322.00	386.0
323.00	3461.0
324.00	464.0
327.00	495.0
327.90	162.0
329.10	159.0
333.20	387.0
334.10	2408.0
335.10	467.0
339.20	210.0
340.90	222.0
346.00	842.0
347.00	204.0
352.00	1172.0
353.10	426.0
354.10	1436.0
365.00	4437.0
365.70	353.0
366.10	318.0
370.00	151.0
371.20	405.0
372.00	2159.0
373.10	291.0
383.10	302.0
402.00	1102.0
403.00	749.0
403.90	435.0
421.00	738.0
422.10	745.0
423.00	6276.0
424.10	770.0
425.10	179.0
441.10	15379.0
442.10	105464.0
443.10	18792.0
444.10	1428.0
444.70	181.0

Instrument :
BNA_G
ClientSampleId :
DFTPP46