

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102622\
 Data File : BG055310.D
 Acq On : 26 Oct 2022 22:43
 Operator : CG/JU
 Sample : N5218-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SP-1

Quant Time: Oct 27 10:43:45 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 24 16:54:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.271	152	37966	20.000	ng	0.00
21) Naphthalene-d8	11.102	136	159900	20.000	ng	0.00
39) Acenaphthene-d10	14.880	164	107352	20.000	ng	0.00
64) Phenanthrene-d10	17.618	188	247708	20.000	ng	0.00
76) Chrysene-d12	21.890	240	238506	20.000	ng	-0.01
86) Perylene-d12	25.286	264	249402	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.797	112	238467	109.980	ng	0.00
7) Phenol-d6	7.413	99	341221	108.623	ng	0.00
23) Nitrobenzene-d5	9.446	82	217113	69.334	ng	0.00
42) 2,4,6-Tribromophenol	16.361	330	122804	105.230	ng	0.00
45) 2-Fluorobiphenyl	13.511	172	476928	72.715	ng	0.00
79) Terphenyl-d14	20.192	244	817951	71.254	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102622\
Data File : BG055310.D
Acq On : 26 Oct 2022 22:43
Operator : CG/JU
Sample : N5218-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SP-1

Quant Time: Oct 27 10:43:45 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 27 16:54:05 2022
Response via : Initial Calibration

